


BGA244 Binary Gas Analyzer

User Manual



Certification

Stanford Research Systems certifies that this product met its published specification at the time of shipment.

Warranty

This Stanford Research Systems product is warranted against defects in materials and workmanship for a period of one (1) year from the date of shipment.

Service

For warranty service or repair, this product must be returned to a Stanford Research Systems authorized service facility. Contact Stanford Research Systems or an authorized representative for a RMA (Return Material Authorization) Number before returning this product for repair. These are available at www.thinksrs.com under Support, Repair/Calibration.

All users returning a BGA244 back to the factory for repair and/or service must submit a correctly completed "Declaration of Contamination of Equipment" form, available as part of the RMA process. The SRS personnel carrying out repair and service of the BGA244 must be informed of the condition of the components prior to any work being performed.

Warning



All returns to SRS must be free of harmful, corrosive, radioactive or toxic materials.

Dedication

In memory of Jim Williams, 1948 - 2011: Legendary Author and Analog Design Guru

Information in this document is subject to change without notice.

Copyright © Stanford Research Systems, Inc., 2016-2023. All rights reserved.

Stanford Research Systems, Inc.
1290-C Reamwood Avenue
Sunnyvale, California 94089
Phone: (408) 744-9040
Fax: (408) 744-9049
Email: info@thinksrs.com
www.thinksrs.com

Printed in the USA

Contents

Contents	i
Tables	vii
Figures	viii
Safety Procedures and Precautions	ix
Symbols You May Find on SRS Products	xi
Specifications	xii
Manual Convention	xvii
Unit Conventions and Abbreviations	xix
How this Manual is Organized	xx
Chapter 1: Getting Started	1
Binary Gas Analyzers	1
What is a Binary Gas Analyzer?	1
Uses for Binary Gas Analyzers	2
Features and Accessories	3
Display	3
Electrical Connections	3
Accessories	4
Unpacking	8
Quick Start	9
Quick Test	9
Installation and Configuration	10
Chapter 2: Installation Guide	13
Operating Environment	13
Access	14
Installation	15
Gas Fittings	16
BGA244	16
Gas Fitting Adapters	17
BGA244HP	18
Electrical Connections	19
Electrical Noise Precautions	19
Grounding	19
Electrical Connectors	20
Connector Pinouts	20
Terminal Strip Connectors	21
Power	22

Computer Interfaces	24
Analog I/O Connections	26
Event Relays	30
Pressure Transducers	31
Mounting Location	31
Types of Pressure Transducers	31
Chapter 3: Operation Guide	35
BGA244 User Interface	35
Navigation	35
Help Screens	35
Map of Interface Functions	36
Power On	37
Appearance	38
Binary Gas Analyzer	40
Details	42
Gas Purity Analyzer	44
Details	46
Physical Measurements	47
Details	49
Screen Messages	50
Analysis Messages	50
Other Messages	51
LED Blink Codes	53
Controls	54
Limits	54
Scale (Binary & Gas Purity)	55
Pressure (Physical Measurements)	55
Help	55
Event Relay 1 & 2	55
Setup	57
Setup	58
Run Stop	58
Instrument Mode	58
Selecting Gases	58
REL	61
Pressure	64
Temperature	67
Store Recall	67
Default Setup	69
Control Panel	73
Display	73
Units	74

Measurement	75
Status	77
Heater	78
Computer I/O	79
Analog I/O	81
Faults	85
Alerts	87
Message Log	89
Self Test	89
Password	91
About the BGA244	91
Hardware Reset	92
User Gases	92
Updating Firmware	92
Chapter 4: Application Guide	93
Gases	93
Gas Requirements	93
Gases that should be used with caution	94
Unusual Gas Properties	94
Condensation	95
Water Vapor in Air	96
Pressure	98
Operating Pressure	98
Pressure Effects in Gases	98
Minimum Operating Pressures	99
Ambient Pressure Variation	100
Operating Conditions	101
Interference	103
Measurements	104
Temperature Variations	104
Evacuating Cell	104
Using Averaging	104
REL to a Reference Gas	105
Binary Gas Analyzer vs Gas Purity Analyzer	108
When to use the Binary Gas Analyzer	108
When to use the Gas Purity Analyzer	108
Converting Molar Fraction vs Mass Fraction	109
Accuracy & Stability	110
Estimated Accuracy	111
Binary Gas Measurement Accuracy	111
Gas Purity Measurement Accuracy	113

Long Term Stability	113
High Purity Use	114
Fittings and tubing	114
Leak Testing	114
Dead Volumes	114
Outgassing	114
Degas	115
User Gases	117
User Gas Parameters	118
Adding a User Gas with BGAMon	119
Adding a User Gas over a Computer Interface	119
Adding Gases	120
Chapter 5: BGAMon	123
Installation	124
Drivers	124
Downloading BGAMon	124
BGAMon Quick Start	125
Using BGAMon	126
Menu Bar	126
System Area	127
Instrument Page	128
Status Bar	128
Instrument Pages	129
Snapshot	129
Analog I/O	130
Pressure	131
Limits	132
Events	132
Strip Chart	133
Gas Selection	138
Faults/Alerts	140
Computer I/O	140
Store / Recall	141
Loading and Saving Configurations	142
Updating Firmware	143
Troubleshooting BGAMon	144
Chapter 6: Remote Programming	145
Introduction	145
Interface Configuration	145
Front-Panel Indicators	145
Transmit and Receive Buffers	145

Break Signal	146
USB	146
RS-232	146
RS-422	147
Command Syntax	148
Parameter Conventions	148
Numeric Conventions	149
Measurement Errors	149
Missing Power Supplies	149
Abridged Index of Commands	150
Detailed Command List	154
Common IEEE-488.2 Commands	154
Instrument Status Commands	157
Interface Commands	164
Measurement and Related Commands	165
Configuration Commands	172
Analog I/O Commands	179
Miscellaneous Commands	185
Status Byte Definitions	189
Serial Poll Status Byte	190
Standard Event Status Register	190
Instrument Status Register Model	190
Event Status Register	193
Error Codes	194
Execution Errors	194
Query Errors	195
Parsing or Command Errors	195
Communication Errors	196
Other Errors	196
Using the USB Drivers	197
Chapter 7: Service	199
Troubleshooting	199
Power On	199
Self Tests	200
System Faults	201
Alerts	202
Screen Messages	203
Factory Safe Mode	205
Locked	206
Store and Recall Failure	206
Touchscreen Problems	207
Hardware Reset	207

Installing BGA-M	208
Maintenance	209
Replacing Gas Fittings	209
Replacing Gas Fittings for the BGA244E	210
Terminal Strips	211
Cleaning	211
Calibration	212
Performance Verification	212
Self Test	212
Analog I/O Accuracy	213
Measurement Accuracy	217
Appendix A: Gas Table	219
Factory Gas Table	219
Gas Table Properties	230
References	239
Appendix B: BGA244E	241
Operation	241
Environment	241
Access	242
Unpacking	243
Quick Start	243
Installation	244
Mounting	244
Gas Fittings	245
Conduit	246
Electrical	246
Appendix C: Mechanical Drawings	249
Appendix D: Declaration of Contamination	253
Appendix E: Installing USB Drivers	255
Appendix H: Revisions	257

Tables

Table 1: Gas Fitting Adapters 5

Table 2: Gas Fitting Adapters 17

Table 3: C3: RS-232..... 20

Table 4: C4: RS-422..... 20

Table 5: C5: +24V..... 21

Table 6: C6: Measure Out..... 21

Table 7: C7: Analog I/O..... 21

Table 8: C8: Relay Contacts 21

Table 9: Active Parameters..... 69

Table 10: Inactive Parameters..... 71

Table 11: External Power Supply Faults 85

Table 12: System Faults..... 86

Table 13: Analog I/O Alerts 87

Table 14: Temperature Alerts 88

Table 15: General Alerts..... 88

Table 16: Self Tests..... 90

Table 17: Recommended Minimum Operating Pressures for Pure Gases..... 100

Table 18: Heater Settling Time..... 101

Table 19: Response Time..... 101

Table 20: Zero Flow Transient Settling Time 102

Table 21: Transient Settling Time for Different Flow Rates 102

Table 22: Command Parameter Conventions 149

Table 23: Power Supply Fault Description..... 201

Table 24: System Fault Description 201

Table 25: Temperature Alert Description..... 202

Table 26: Power Supply Alert Description..... 202

Table 27: Analog I/O Alert Description..... 203

Table 28: Measure Out Test Limits..... 214

Table 29: Analog I/O Test Limits..... 216

Table 30: Pressure vs Altitude 218

Table 35: Factory Gas Table 219

Table 36: Gas Table Properties..... 231

Table 37: C9: RS-232 Connector for BGA244E 247

Figures

Figure 1: BGA244..... 6

Figure 2: BGA244 Display 6

Figure 3: BGA-M Metal Cover Accessory 7

Figure 4: RS-422 Signal Path..... 25

Figure 5: Voltage Output..... 26

Figure 6: Current Output..... 27

Figure 7: Voltage Input..... 27

Figure 8: Current Input..... 28

Figure 9: Current Input w/ Loop Power (suitable device)..... 28

Figure 10: Current Input w/ Loop Power (unsuitable device)..... 29

Figure 11: Event Relay Contacts..... 30

Figure 12: Map of User Interface Functions..... 36

Figure 13: BGA244 Home Page 38

Figure 14: Binary Gas Analyzer Home Page 41

Figure 15: Gas Purity Analyzer Home Page 45

Figure 16: Physical Measurements Home Page 48

Figure 17: Speed of Sound vs Pressure in SF6..... 98

Figure 18: Gas Concentration with and without Averaging for Argon and Air 105

Figure 19: BGAMon Snapshot Screen 125

Figure 20: BGAMon Layout 126

Figure 21: Snapshot Page..... 129

Figure 22: Analog I/O Page..... 130

Figure 23: Pressure Page..... 131

Figure 24: Limits Page 132

Figure 25: Events Page 132

Figure 26: Strip Chart Page..... 133

Figure 27: Print Preview Window 136

Figure 28: Export Window..... 137

Figure 29: Gas Selection Page 138

Figure 30: Fault/Alert Page 140

Figure 31: Computer I/O Page..... 140

Figure 32: Store/Recall Page 141

Figure 33: Setup for Measure Out Voltage Test..... 213

Figure 34: Setup for Measure Out Current Test 213

Figure 35: Setup for Analog I/O Voltage Test..... 215

Figure 36: Setup for Analog I/O Current Test 215

Figure 37: BGA244 Mechanical Drawing..... 250

Figure 38: BGA244E Mechanical Drawing (No Mounting Flanges) 251

Figure 39: BGA244E Mechanical Drawing (with Mounting Flanges) 252

Safety Procedures and Precautions

Observe the following general safety precautions during all phases of operation of this instrument. Failure to comply with these precautions or with other specific warnings elsewhere in this manual violates the safety standards of intended use of this instrument and may impair the protection provided by the equipment. Stanford Research Systems, Inc. assumes no liability for the customer's failure to comply with these requirements.

DO NOT SUBSTITUTE PARTS OR MODIFY THE INSTRUMENT

Do not install substitute parts or perform any unauthorized modification to the instrument. Return the instrument to Stanford Research Systems or an authorized representative for service and repair to ensure all safety features are maintained.

SERVICE BY QUALIFIED PERSONNEL ONLY

Operating personnel should not attempt any component replacement or internal adjustments. Any service should be performed by qualified service personnel only.

USE CAUTION WHEN OPERATING WITH HAZARDOUS MATERIALS

If hazardous materials are used, users must take responsibility to observe proper safety precautions, completely purge the instrument when necessary, and ensure the material used is compatible with materials in this product, including any sealing materials.

PURGE THE INSTRUMENT

After installing the unit or before removing it from a system, purge the unit completely with a clean dry gas to eliminate all traces of the previously used flow material.

USE PROPER PROCEDURES WHEN PURGING

Purge the instrument under a ventilation hood. Wear gloves for protection during this procedure.

EXPLOSIVE ENVIRONMENT WARNING

The BGA244 is not ATEX rated. Under normal operating conditions the BGA244 cannot ignite the gas being analyzed. However, if the instrument is used with flammable or explosive gas mixtures, SRS recommends the use of flame arrestors on both gas ports.

The BGA244 gas cavity's proof pressure (2,500 psia) is sufficient to contain the detonation of an explosive gas mixture of up to 30 psia. The instrument will likely be damaged by such an event.

USE PROPER FITTINGS AND TIGHTENING PROCEDURES

All instrument fittings must be consistent with instrument specifications and compatible with the intended use of the instrument. Assemble and tighten fittings according to manufacturer's directions.

CHECK FOR LEAK-TIGHT FITTINGS

Carefully check all connections to ensure leak tight installation.

OPERATE AT SAFE INLET PRESSURES

Never operate at pressures higher than the maximum operating pressure (refer to the product specifications for the maximum pressure).

INSTALL A SUITABLE BURST DISK

When operating from a pressurized gas source that may exceed the cavity proof pressure (2500 psia), install a suitable burst disk to prevent system explosion should the system pressure rise.

KEEP THE UNIT FREE OF CONTAMINATION

Do not allow contaminants to enter the unit before or during use. Contamination such as dust, dirt, lint, glass chips and metal chips may permanently damage the unit or contaminate the process.

ALLOW PROPER WARM UP TIME

The unit may not meet all specifications unless sufficient time is allowed for the unit to stabilize at the designed operating temperature. Do not REL or calibrate the unit until the warmup is complete.

GROUNDING

Proper operation of this instrument requires that it be connected to earth ground. If the power source does not provide the required grounding, you should add a protective ground to the device.

COVERS

Do not operate the unit with the instrument covers removed.

MOUNTING SCREWS




Use silver plated or iron oxide coated screws. The use of stainless steel mounting screws is not recommended as they may gall and seize, damaging the BGA244. If stainless steel screws are required, use a lubricant to prevent galling.

RETURNS

All returns to SRS must be free of harmful, corrosive, radioactive or toxic materials.

Users returning a BGA244 back to the factory for repair and/or service must submit a correctly completed "Declaration of Contamination of Equipment" form, available as part of the RMA process. The SRS personnel carrying out repair and service of the BGA244 must be informed of the condition of the components prior to any work being performed. See Appendix D: *Declaration of Contamination* for information required for the "Declaration of Contamination of Equipment" form.

Symbols You May Find on SRS Products

Symbol	Description
	Alternating Current
	Caution – risk of electrical shock
	Frame or Chassis terminal
	Caution – refer to accompanying document
	Earth (ground) terminal
	Battery
	Fuse
	Power On
	Power Off
	Power Standby

Specifications

Operational

Instrument Modes	Binary Gas Analyzer: Concentration in mole or mass fraction Gas Purity Analyzer: Deviation in speed of sound from ideal speed of sound / ideal speed of sound $\Delta W / W$ Physical Measurements: Measured Speed of Sound, Normalized Speed of Sound, temperature, and pressure
Operating Pressure	0 to 150 psia (1000 kPa)
Operating Temperature	-20 °C to +70 °C
Flow Rate	0 to 5000 sccm
Gas Species	~500 gases supported in Factory Gas Table, Users can add gases to the User Gas Table
Response Time for a step change (2000 sccm flow rate)	9 seconds to 90% 18 seconds to 99%, 27 seconds to 99.9%
Recommended Min Pressure	Gas Species Dependent (examples for pure gases) H ₂ 6 psia (41 kPa) He 6 psia (41 kPa) CH ₄ 3 psia (21 kPa) N ₂ 2 psia (14 kPa) Ar 2 psia (14 kPa) CO ₂ 12 psia (83 kPa) SF ₆ 1 psia (7 kPa)

Measurement

Measurement Technique	The Speed of Sound of the gas is measured using a cylindrical resonator using acoustic transducers. The gas temperature is measured using thermistors inside the resonant cavity. The Normalized Speed of Sound and Gas Concentration is calculated based on the cylindrical resonator characteristics and the thermodynamic properties of the gases.
Reading Rate	~4.4 Hz
Averaging	None, or 2 to 100 samples averaged
Temperature Measurement	Accuracy: ± 0.1 °C Resolution: 0.001 °C Aging: ± 0.001 °C/year
Speed of Sound Measurement	Range: 100 to 1500 m/s Resolution: 0.001 m/s Aging: $< \pm 10$ ppm/year Accuracy: $\pm 0.05\%$ (14.7 psia Argon at 200 sccm)

	Temp Co:	< ±50 ppm /°C (Argon)
Concentration Measurement ¹	Range:	0 – 100%, 0 – 1,000,000 ppm or 0 – 1.0 frac
	Resolution:	1 ppm
	Stability:	typically 10 ppm
	Accuracy:	typically 100 – 1000 ppm

¹ All concentration specifications depend on the gas species being measured. See *Accuracy* in the *Application Guide* for more information.

Heater (requires 24V)

Set Temperature	-30 °C to 70 °C
Current Limit	0.05 to 2.5 A
Settling Time to final temperature	16 minutes for 25 °C to 50 °C temperature step, with 2 A max current
Protection	Over temperature and Over Current Protected

Analog I/O (requires 24V)

Outputs

Function	Measure Out:	Gas Ratio, Gas Purity or Normalized Speed of Sound
	Output 1, 2:	Speed of Sound, Normalized Speed of Sound, Temperature, Pressure or User Value
Range	Voltage	0 to 5 V, 0 to 10 V
	Current	4 to 20 mA
Resolution	Voltage (5V range)	0.1 mV
	Voltage (10V range)	0.2 mV
	Current	0.4 µA
Accuracy ¹	Voltage ²	±0.1 % + 1 mV
	Current	±0.1 % + 10 µA

¹ Output accuracy specifications are for User values. For Measure Out or Linked output values the accuracy is determined by the measured parameter being output.

² Voltage accuracy specifications apply for V > 0.4% of the Range

Max Output Current (Voltage Out)	20 mA
Max Load Resistor (Current Out)	840 Ω

Inputs

Function	Input 1, 2	Pressure Sensor or User Value
Range	Voltage	0 to 10 V
	Current	4 to 20 mA, 4 to 20 mA with Loop Power
Resolution	Voltage	1 µV
	Current	1 nA

Accuracy	Voltage	±0.1 % + 1 mV
	Current	±0.1 % + 10 µA
Input Impedance	Voltage	10 MΩ
	Current	201 Ω
Absolute Maximum	Voltage	24 V
	Current	25 mA
Loop Power Voltage	Range	6 to 19 V
	Resolution	0.1 V
	Max Current	50 mA

Event Relay (requires 24V)

Function	2 independently configurable DPDT relays that can be set to switch on a combination of events
Events	Measurement limits, pressure limits, temperature limits, “No Signal” and System Fault
Relay Contact Rating	
Max Switching Power	30 W, 62.5 VA
Max Switching Voltage	220 V _{DC} , 250 V _{AC}
Max Switching Current	1 A
Max Carrying Current	2 A

Computer I/O

RS-232

Format	No parity, 8 bits, 1 stop bit, CTS/RTS flow control
Baud Rate	2400 – 115.2k
Max Cable Length	>100 meters for lower baud rates

USB

Connector	USB Type B
Format	WHQL high speed USB2.0
Drivers available	Virtual COM Port (VCP) and Direct Drivers (USB drivers +DLL)
OS Support	
Microsoft	Windows7, 8 (32, 64 bit), Windows Vista, XP (32, 64 bit), Windows XP Embedded, Windows CE 4.2, 5.0 & 6.0
Other	Mac OS-X, Linux 2.6 & greater, Android

RS-422 (requires 24V)

Format	4 wire, point-to-point, non-multidrop, no parity, 8 bits, 1 stop bit, no flow control
Baud Rate	2400 – 115.2k
Max Cable Length	>1000 meters for lower baud rates

Power**USB**

Connector	USB Type B
Voltage	+5 V _{DC} , +/- 0.25 V *
Current	0.35 A continuous, 0.45A max on startup

* The USB voltage must be >4.75V at the BGA244. Be sure to use a device that can supply enough current and a large enough wire gauge cable that can support this voltage. See *Power* in the *Installation Guide* for details.

Note that the USB current goes to 0 mA when +24V is connected.

+24 V

Connector	2 x 5.5 mm barrel jack, 2 wire terminal strip
Voltage	+24 V _{DC} , +/- 1 V
Ripple	<240 mV p-p
Current	
No Analog I/O, Heater	0.2 A
Max	2.7 A

Environment

Cavity Proof Pressure	2500 psi (17 MPa)
Operating Temperature	-20°C to 70 °C
Storage Temperature	80 °C max
Humidity	<90 % relative humidity, non-condensing
Altitude	≤ 4000 m (for applications above this altitude, contact SRS)
Pollution Degree	Category 2: (EN61010-1; only non-conductive pollution)
Inbound Helium Leak Rate:	1x10 ⁻⁸ sccs (BGA244HP only)

Physical

BGA244, BGA244 HP

Front Panel	Color TFT-LCD w/ touchscreen Power, Communication and Error LED Indicators
Dimensions	5.5" x 4.5" x 3.25" (WHD)
Weight	7 lbs (3.2 kg)
Swept Volume	130 cc (Acoustic Chamber Volume)
Gas Fittings	
BGA244	1/8"-27 female NPT
BGA244HP	1/4" male welded VCR
Wetted Materials	
BGA244, BGA244E	Electro polished 304 stainless steel, OFHC copper gaskets, nickel plated/immersion gold copper traces on 0.001" Kapton film, vented 316 stainless steel screws, 316 stainless steel NPT fittings, glass and nickel plated Dumet wire. Loctite 565 thread sealant is used to seal the NPT fittings.
BGA244HP	Electro polished 304 stainless steel, gold plated OFHC copper gaskets, nickel plated/immersion gold copper traces on 0.001" Kapton film, vented 316 stainless steel screws, 316 stainless steel male VCR fittings, glass, and nickel plated Dumet wire. The VCR fittings are welded in place.

BGA244E

When properly installed, the BGA244E (with Environmental Enclosure) will meet the following NEMA / UL-50 standards. See Appendix B (*Environmental Enclosure*) for installation instructions. All other specifications match the BGA244 unless otherwise noted.

Environmental Enclosure Ratings:	NEMA 6 (with catch latched) NEMA 6L (with screws installed) IP66, UL Type 4X, UL-50
Gas Fittings	1/8"-27 female NPT
Electrical	3/4" Flexible, Non-Metallic Conduit (FNMC)
Dimensions	10" x 11" x 5.5" (WHD)
Weight	11 lbs. (5 kg)

Manual Convention

The BGA244 is operated thru a multilevel graphical user interface. The following conventions are used in the manual to describe different functions of the interface.

[Key]

This key is either a navigation key that will take you to a different display or a direct action key that will perform the action listed on the key.



[On|Off]

This is an on/off or enable/disable key. Press the desired side of the key to activate that function.

This indicates On or Enabled



This indicates Off or Disabled.



<Entry>

This indicates that a value needs to be entered where “Entry” is the name of the parameter. Pressing the key will open up the alpha-numeric keypad for data entry.



If the entered value is out of range, it will be ignored and an “Invalid Entry” prompt will appear indicating the maximum allowable value.

[ESC]

This key on the alpha-numeric keypad allows you to exit without entering a new value.




<Entry ↓>

This indicates that a value needs to be selected from a list where “Entry ↓” is the name of the list. Pressing the key will open a drop list of possible selections. [Press] the desired selections to choose it. The current selection is normally highlighted in yellow.



○ Choice

This indicates a “radio button” that is used to select one of a number of possible choices. Press the circle to select that choice and de-select all other choices.



□ Choice

This indicates a check box that is used to enable a choice. Unlike Radio Buttons, you can select any or all of check boxes as needed.



(Home/Here/Over_There)

This denotes a location in the user interface hierarchy. To reach (Home/Here/Over_There) from the Home Page you would press [Here], followed by [Over_There].

[Home]

This returns you to the BGA244 Home Page. Note that the Home display has different appearances depending on the Analyzer Mode selected.

A rectangular button with a thin black border and the word "HOME" in black, uppercase letters centered on a light gray background.

[←] or [Back]

This returns up one level from the display you are currently on.

A rectangular button with a thin black border and a black left-pointing arrow centered on a light gray background.

[Help]

This takes you to the Help screen for that page. This page will have information about the settings and displayed parameters. [Page ↑] and [Page ↓] are active if the Help screen is more than one page long.

A rectangular button with a thin black border and the word "HELP" in black, uppercase letters centered on a light gray background.

Note: The touch screen on the BGA244 must be actually be pressed to activate functions. This is normally indicated by a key click or beep. Hovering over the button like you do on a cell phone or tablet won't activate the control.

Throughout the manual, text in *Italics* refers to another section of the manual.

Unit Conventions and Abbreviations

Normal Temperature and Pressure (NTP)

Several different conventions are used to normalize measurements to a fixed temperature and pressure. The BGA244 uses “Normal Temperature and Pressure” (NTP) for all of its normalized readings. This is defined as the following:

- Temperature: 20.00°C, 293.15°K or 68.00°F
- Pressure (absolute): 1 atm, 101.325 kPa or 14.696 psia

Flow

There are a few abbreviations used to describe flow rate.

SCCM: standard cc (ml) per minute

SCCS: standard cc (ml) per second

Speed of Sound

The Speed of Sound is occasionally abbreviated as “SOS” in the manual.

Normalized Speed of Sound

The speed of sound in a gas varies as a function of several environmental parameters. The dominant effects are caused by pressure and temperature. An additional effect is caused by the frequency at which the speed of sound is measured. The frequency dependent effect is due to vibrational population relaxation effects in the gas molecules. This effect is specific to the particular gas being measured.

The BGA244 normalizes the speed of sound to NTP at the measured frequency.

The Normalized Speed of Sound is occasionally abbreviated as “NSOS” in the manual.

Pressure

Pressure can be referred to in either absolute or gauge pressure units. Absolute pressure units are relative to vacuum, while gauge pressure units are relative to the ambient pressure on the outside of the gauge.

When using gauge units it is important to enter the ambient pressure, as the BGA244 uses absolute pressure in its calculations.

How this Manual is Organized

This manual provides instructions on how to install and operate a BGA244 Binary Gas Analyzer.

Before installing your BGA244 in a system and/or operating it, carefully read and familiarize yourself with all precautionary notes in the Safety and Installation sections at the beginning of this manual. In addition, observe and obey all WARNING and CAUTION notes provided throughout the manual.

Chapter 1: *Getting Started*

Describes the BGA244 including measurements and applications and a quick start guide

Chapter 2: *Installation Guide*

Explains the environmental requirements and describes how to install the BGA244 in your system

Chapter 3: *Operation Guide*

Describes how to configure and operate the instrument and explains all of its functionality in detail

Chapter 4: *Applications Guide*

Describes how to optimize the BGA244 for the best performance in your application

Chapter 5: *BGAMon*

Describes how to configure and control the BGA244 using the BGAMon Windows software

Chapter 6: *Remote Programming*

Describes how to control the BGA244 using the computer interfaces

Chapter 7: *Service*

Troubleshooting, Maintenance and Calibration of the BGA244

Appendix A: *Gas Table*

A list of all supported gases in the Factory Gas Table, plus a description of the data contained in the Gas Table

Appendix B: *BGA244E*

Description of the BGA244E, a NEMA enclosed version of the BGA244

Appendix C: *Mechanical Drawings*

Mechanical drawings and mounting locations for the BGA244 and its derivatives

Appendix D: *Declaration of Contamination*

Sample of the Declaration of Contamination form

Appendix E: *Installing the USB Drivers*

Instruction on manually installing the USB Drivers

Appendix H: *Revisions*

List of manual revisions

...

Trademarks

The trademarks of the products mentioned in this manual are held by the companies or organizations that produce them.

CAS Registry Number is a Registered Trademark of the American Chemical Society.

VRC, VCO and Swagelok are Registered Trademarks of Swagelok, Company.

Windows, Windows XP, Windows 7, Windows 8 and Windows 10 are Registered Trademarks of Microsoft Corporation.

Combicon and Phoenix Contact are trademarks of Phoenix Contact.

FTDI is a Trademark of Future Technology Devices International Ltd.

McMaster and McMaster-Carr are registered Trademarks of McMaster-Carr.

All other brand and product names are trademarks or registered trademarks of their respective companies.

Chapter 1: Getting Started

Binary Gas Analyzers

What is a Binary Gas Analyzer?

A Binary Gas Analyzer (or BGA) measures the ratio of two gases based on physical properties of the gases. Other instruments do this by measuring the thermal conductivity of the gas mixture or by measuring the speed of sound in the gas using a time-of-flight technique. Both the thermal conductivity and speed of sound of a gas vary inversely with their molar mass (along with several other gas properties). So the ratio of two gases can be calculated if you know the properties of the two gases and the thermal conductivity or speed of sound in the mixture.

A different method of measuring the speed of sound uses a resonant acoustic cell. Measuring the resonant frequencies within the cell and knowledge of the cell's geometry allows the speed of sound to be accurately calculated. This information combined with knowledge of the physical properties of the gases allows you to accurately determine the ratio of the two gases.

The BGA244 works by injecting a wideband acoustic signal into a gas cell using a speaker. The resonant modes of the cell are excited, producing large amplitudes at the resonant frequencies. This signal is measured using a sensitive, wideband microphone. Advanced signal processing extracts the frequencies of the resonant peaks and from them, the speed of sound.

Combining the speed of sound, temperature and pressure with a detailed list of about 500 different gases allow the BGA244 to make ratio measurements with a typical accuracy of better than 0.1%

This technique isn't just reserved for gases of a single species. Blended gases can also be measured, as long as physical data is available for each gas blend. The best known blended gas is air, made up of nitrogen, oxygen, argon and carbon dioxide. There are many other blended gases that can be used in the BGA244.

Most binary gas analyzers are factory configured to support a handful of gases, or only a single mixture over a limited range. In contrast, the BGA244 comes configured with data on nearly 500 gases that are characterized over a wide range of concentrations. This gas information is easily selected from the front panel or can be configured remotely, allowing over 50,000 different mixtures to be measured.

In addition to binary gas ratio measurements, the BGA244 can report purity of a gas expressed as the ratio of the measured speed of sound to the expected speed of sound. To support other research goals, measurements of the speed of sound, temperature and pressure can be reported directly.

Uses for Binary Gas Analyzers

Binary gas analyzers are used in a wide range of applications. They are often used in applications where no dedicated sensors exist for a gas or for indirectly controlled mixtures in a process.

A few typical applications are:

- Leak Detection
- Semiconductor Processing
- Food Processing
- Quality Control
- Environmental Monitoring
- Purge Gas Analysis
- Shielding or Blanketing Gas Monitoring
- Generator Cooling Gas Monitoring
- Heat Treating Gas Monitoring
- Gas purity monitoring
- Measuring Helium – Deuterium ratio
- Measuring He-3 / He-4 ratio
- Helium Recovery Systems

Features and Accessories

The BGA244 consists of a compact package which includes the resonant cell, acoustic transducers, sensitive electronics and advanced signal processing. It can be powered over USB or +24 V and can interface to external devices by computer interface or analog signals and features a touch screen LCD to configure and monitor measurements.

There are three models in the BGA244 series: the standard BGA244, the BGA244HP (high purity) and the BGA244E (enclosure). All models share the same acoustic resonant cell, transducers, signal processing and computer interfaces.

BGA244

The standard BGA244 is designed to operate in most common operating environments. It comes with $\frac{1}{8}$ "-27 female NPT fittings. There are several different accessories available. These include a number of different gas fitting adapters, a 24 V_{DC} power supply and protective covers for the display.

BGA244HP

The BGA244HP (high purity) is designed to operate in high purity or corrosive environments. It comes with welded-in-place $\frac{1}{4}$ " male VCR fittings in place of the female NPT fittings and is Helium leak checked. The BGA244HP operates the same as the standard BGA244. All accessories except the gas fitting adapters are available for the BGA244HP.

BGA244E

The BGA244E packages the standard BGA244 in a NEMA Type 6/6P enclosure for use in exposed locations. It comes with $\frac{1}{8}$ "-27 female NPT fittings. The BGA244E operates the same as the standard BGA244. All accessories are available for the BGA244E. See Appendix B for specific information on the BGA244E.

Display

The BGA244 includes a color TFT-LCD display with touchscreen. This can be used to configure the unit and display results. See Figure 2 (page 6)

Electrical Connections

The electrical connections and indicators are located on the front of the BGA244. See Figure 1 (page 6) for the connector locations.

All models of the BGA244 can be powered by either 24 V_{DC} or USB power. Several features require 24V_{DC} to operate including the analog inputs & outputs, heaters, event relays and the RS-422 interface. However the core functionality of the BGA244 will operate using USB power.

USB power can be provided by a computer, a powered USB hub or a USB charger. Be sure to use an appropriate USB cable when powering the BGA244 over USB. The cable that is included with the BGA244 is suitable. See *Power* (page 22) for information on suitable USB cables. 24 V_{DC} can be provided via a 3mm barrel plug or on a terminal strip.

All instrument functions can be controlled over the USB, RS-232 or RS-422 computer interfaces. There are 3 analog outputs, 2 analog inputs and 2 configurable event relays to interface to external devices.

There are three LEDs that indicate the status of the BGA244.

Name	Function
Power (green)	Shows that power is applied Flashes for power fault codes
Comm (green) (Communication)	Flashes when computer interface is active
Error (red)	Flashes for communication error Repetitive blink for various no signal errors. Constant on for "System Fault"

Accessories

There are several accessories available for the BGA244.

+24V Power Supply (BGA-24)

Accessory BGA-24 is a 50W universal input power supply that provides +24 V_{DC} at 2.5 A.

Shield (BGA-S)

Accessory BGA-S is an Acrylic shield used to protect the BGA244 Display. This cover is not needed for units using accessory BGA-M.

Note that the display touch screen will not operate through the Acrylic shield. It is necessary to first remove the shield before using the touch screen. Replace the Acrylic shield when you are done using the touch screen.

Metal Cover (BGA-M)

Accessory BGA-M is a rugged solid metal cover that can replace the color TFT display for applications where the display is not wanted (Figure 3, page 7). See *Installing BGA-M* (page 208) for information on installing the cover.

Units should either be configured prior to installing the metal cover or configured over one of the computer interfaces. The simplest way to configure the BGA244 over the computer interface is using the BGAMon program provided by SRS. See Chapter 5: *BGAMon* (page 123) for more information. The other configuration option is user written code. See Chapter 6: *Remote Programming* (page 145) for more information.

Gas Fitting Adapters

The following fitting adapters are available to interface the BGA244 to different gas fittings. These adapters connect to the 1/8"-27 female NPT fittings of the BGA244. See Chapter 2: *Installation Guide* (page 17) for more details on using the different gas fitting adapters.

TABLE 1: GAS FITTING ADAPTERS

Accessory	Gas Fitting Adapter
SS2A	1/8"-27 male NPT to 1/8"-27 female NPT
SS4RA2	1/8"-27 male NPT to 1/4" -18 female NPT
SS4VCR12	1/8"-27 male NPT to 1/4" male VCR
SS4VCO12	1/8"-27 male NPT to 1/4" male VCO Body
SS4HC12	1/8"-27 male NPT to 1/4" hose barb (1/4" ID hose)
SS40012	1/8"-27 male NPT to 1/4" compression fitting
SS6MO12	1/8"-27 male NPT to 6mm compression fitting
5482K119	1/8"-27 male NPT to 1/4", 37° flared AN tube, male, 7/16"- 20 TPI
4822T76	1/8"-27 male NPT to 1/8" BSPP female, 1/8"-28 TPI
4092K26	1/8"-27 male NPT to 1/8" BSPT female, 1/8"-28 TPI




FIGURE 1: BGA244




FIGURE 2: BGA244 DISPLAY



FIGURE 3: BGA-M METAL COVER ACCESSORY

Unpacking

Before You Open the Box

Do not remove the gas port caps until just prior to connecting the unit to your gas manifold to reduce the chance of contamination of the cell.

Read the Chapter 2: *Installation Guide* prior to installing the BGA244 into your system.

Read the Chapter 3: *Operations Guide* and Chapter 4: *Applications Guide* prior to operating the BGA244.

Inspect all components of the SRS BGA244 upon unpacking. Report any damage to Stanford Research Systems immediately. Compare the contents of the shipping container to the list below and report any discrepancies.

See www.thinksrs.com/downloads/PDFs/Manuals/BGA244m.pdf for the complete BGA244 User's Manual.

See Appendix B for information specific to the BGA244E.

What is included with the BGA244

1. One BGA244
2. One 6' (1.8 m) USB cable
3. One Quick Start Guide

Accessories (if ordered)

- | | |
|-------------|------------------------------------|
| 1. BGA-24 | +24 V _{DC} Power Supply |
| 2. BGA-S | Acrylic shield for the LCD display |
| 3. BGA-M | Metal cover to replace display |
| 4. Adapters | Gas fitting adapters |

What is included with the BGA244HP

1. BGA244HP
2. One 6' (1.8 m) USB cable
3. One Quick Start Guide

Accessories (if ordered)

- | | |
|-----------|------------------------------------|
| 1. BGA-24 | +24 V _{DC} Power Supply |
| 2. BGA-S | Acrylic shield for the LCD display |
| 3. BGA-M | Metal cover to replace display |


Quick Start

This guide is intended to help users get started making measurements with the BGA244 or BGA244HP. If the BGA244 has accessory BGA-M (metal cover) in place of the display it is recommended to use the BGAMon software to configure the unit and monitor the results. Refer to Chapter 5: *BGAMon* for details.


Quick Test

Connect power using either USB power or +24 V_{DC}. Make certain that the USB device and cable are sufficient to power the device if using USB power. See *Power* (page 22) for more information. If you are planning on operating with both USB and +24 V_{DC} connected, first connect +24 V_{DC}, wait until Self Test has completed and then connect the USB cable.

When power is first applied, a power-on splash screen will appear. After a few seconds the “About” information page will appear. This displays the unit serial number, firmware version, calibration date and hardware version.



After about 15 seconds the Gas Purity Analyzer Home page will appear. Assuming the unit has come from the factory and the gas port caps haven't been removed, it should display <0.1% deviation reading for Argon.



Disconnect power before continuing with the following steps.

Installation and Configuration

Plan your installation, including mounting and gas tubing connections. Depending on your application, this can be as simple as setting the BGA244 on a bench and connecting flexible tubing or may involve hard mounting the unit and forming metal tubing. Be sure to clean out gas lines prior to connecting them to the BGA244 to avoid contamination. Refer to Chapter 2: *Installation Guide* if you have any questions regarding these procedures as improper installation can damage the unit.

1. Mount the BGA244 as appropriate to your application.
2. Connect the gas tubing to the gas ports following the appropriate procedure for the specific gas fittings. Either gas port can be used for input or output.
3. Connect either USB power or +24 V_{DC}. If using USB power, be certain that the USB device and cable are sufficient to power the device. Refer to *Power* (page 22) for more information. The splash screen and “About” page should appear as described in Quick Test.
4. Configure the BGA244. Note that it may be easier to configure the unit prior to installing it into your system, especially if access is restricted. See Chapter 3: *Operation Guide* for more information on each of the following steps.
 - a. Set the Instrument Mode: [HOME] [SETUP] [INSTRUMENT MODE ↓]. There are 3 choices: Binary Gas Analyzer, Gas Purity Analyzer or Physical Measurements.
 - b. Select the Gases: [HOME] [SETUP] [SELECT GAS]. Set the Primary Gas and Secondary Gas (note that a Secondary Gas is not required for the Gas Purity Mode). Press [SELECT GAS] to open the gas selection window. Type the first few characters of its name, formula or CAS number, then press [ENTER] to open a list of gases that match that string. Select the desired gas from that list.
 - c. Configure the Pressure: [HOME] [SETUP] [PRESSURE]. Press [ENTER USER PRESSURE] to open the User pressure window. This is simplest way to enter the pressure. See Pressure (page Pressure64) for information on different methods of entering the operating pressure into the BGA244.
 - d. Set the Meter Scale and Limits: [HOME] [METER SCALE] and [HOME] [LIMITS] appropriately. (For the Binary Gas and Gas Purity Analyzers only).
5. Flow gas through the system. The BGA244 should display the selected gases and concentration, purity or physical measurements of the gas.

Windows Drivers for the BGA244

If the BGA244 is connected to the USB port of a computer running Microsoft Windows, you may be prompted with a “New Hardware Found” message and an invitation to search for the USB Driver. There are two USB drivers for the device (VCP and D2XX drivers). Depending on the version and configuration, Windows may either automatically install the drivers or prompt you to search for them. Allow it to install the drivers if you plan to configure or control the BGA244 using this computer. If there are difficulties installing the driver, see Appendix E for details on manually installing the drivers. If you have no plans of using this computer to configure or control the BGA244, cancel the driver installation.

Chapter 2: Installation Guide

The BGA244 can be installed in a variety of ways to interface with gas systems. For bench top experiments, it can operate freestanding on its non-slip feet. For more robust installations it can be bolted to a mechanical plate in any orientation. There is no preferred direction of gas flow. The LCD display can be rotated to any orientation for convenient viewing.

There is a wide range of different gas fitting adapters available to easily connect to different systems. Units can be power over the USB interface or by +24 V_{DC}. The BGA244 can be easily connected to an automated system over any of its three computer interfaces. In addition, there are configurable event relays and connections for analog input and output signals.

The standard BGA244 is designed to be operated in a clean and dry environment. The BGA244HP is designed for high purity applications or for use with corrosive gases. The BGA244E is recommended for locations that may be exposed to wet or dirty conditions. The BGA244E packages the standard BGA244 in a NEMA Type 6/6P enclosure for use in exposed locations. See Appendix B for installation information specific to the BGA244E.

Operating Environment

Temperature

The operating temperature range of the BGA244 is from -20° to 70 °C. Do not expose the BGA244 to bake out temperatures above 80 °C.

If the BGA244 is being operated at an ambient temperature below -20 °C use the heaters to raise the operating temperature. It may be necessary to insulate the unit to bring the temperature within the operating range.

Pressure

The BGA244 can make reliable measurements with pressures ranging from around 2 psia (14 kPa) to as high as 150 psi (1000 kPa) depending on the gas species. The design proof pressure is 2500 psi (17.2 MPa) making secondary containment chambers unnecessary.

Flow

The BGA244 is specified for flow rates from 0 to 5000 sccm. Customers have successfully operated at flow rates as high as 20,000 sccm.

Gas

Gases must be clean and dry (free of any solid or liquid particulates). In addition, they must be non-condensing at the operating temperature and pressure. See *Gases* (page 93) for more details.

Electrical and Magnetic Fields

Strong electrical or magnetic fields can interfere with the BGA244. These can make measurements noisy or even impossible to make. The best solution is to make sure that the BGA244 is separated from interfering sources. If this isn't possible, see *Interference* (page 103) in Chapter 4 for suggestions on shielding techniques.

Avoid passing currents through the BGA244 from the gas lines. This can generate interfering signals. Make sure that all pipes are properly grounded.

Access

Front Panel

As the BGA244 can be operated or viewed from the front panel, it is necessary to have a clear view and easy access to the LCD and touch screen. This is less important if operating over a computer interface, although it is useful to have access to the front panel for debugging purposes.

Visibility of the status LEDs is also helpful, especially when operating over the computer interfaces.

Wrench Clearance

Most of the gas fittings that connect to the BGA244 require one or more wrenches to fully tighten. Make sure there is sufficient clearance, both in length and rotation. Follow the recommended installation guidelines for details on the different gas fittings.

Cable Clearance

Typical USB and RS-232 and RS-244 cables require about 2.5" (65 mm) clearance from the BGA244. The +24 V barrel connector (with accessory BGA-24) and terminal strips (analog I/O and relay connections) need about 2.0" (50 mm) of clearance.

Ventilation Clearance

When the heaters are not operating, the BGA244 dissipates between 2 to 5 watts, depending on the operating condition. Therefore, ventilation clearance isn't required. When operating, the heaters adjust their power to maintain a constant temperature. This eliminates the need for any additional ventilation.

Service Access

Besides the terminal block connectors and the gas fittings there are no user serviceable parts in the BGA244. See Chapter 7: *Service* for details on servicing those parts.

Installation

Plan the location of the BGA244, gas pipes, fittings and electrical connectors prior to mounting the unit and forming tubing. Take into account cable routing to minimize electrical interference. Refer to *Electrical Connections* (page 19) for more details. The BGA244 can be mounted in any orientation. There is no preferred gas flow direction; either gas port can be input or output. See Appendix C for the location of the gas ports, electrical connectors and mounting points.

Clean out the gas lines before connecting the BGA244 to remove any particulates or oils. These can contaminate or damage the acoustic cell.

Strong mechanical vibrations may interfere with obtaining accurate measurements. Whenever possible mount the BGA244 in a location that is free of large vibrations or impacts.

Do not rely on the inlet and outlet tubing to support the weight of the BGA244 to avoid damaging either the tubing or the unit. Either mount the BGA244 to a rigid plate or rest it on its non-stick feet.

The distance between the two gas fittings is different for the BGA244 and the BGA244HP. Remember to take into account any gaskets (VCR), ferrules, adapters and thread insertion when determining pipe length. Refer to *Gas Fittings* (page 16) for more details.

The BGA244 is mounted using four 10-32 screws. The maximum penetration depth of the screws is 0.38" (9.6 mm). Either remove the rubber feet or relieve the area beneath them for hard mounting to a rigid plate. This helps to avoid vibration problems.

Note: Use silver plated or iron oxide coated screws. The use of stainless steel mounting screws is not recommended as they may gall and seize, damaging the BGA244. If stainless steel screws are required, use a lubricant to prevent galling.

Procedure

- Fasten the BGA244 to the mounting surface using the previously mentioned guidelines. If rigid preformed tubing is used, it may be necessary to remove the caps from the gas fittings before mounting.
- Connect the gas lines as described in *Gas Fittings* (page 16). Be sure to account for any washers or O-rings that may be required.
- Connect the electrical cabling for power, analog and computer I/O as described in *Electrical Connections* (page 19).

Gas Fittings

BGA244

The body of the BGA244 is machined with two gas port, each with 1/8"-27 female NPT threads. Male to female NPT adapters are installed in these ports to act as thread savers. These are installed using Loctite 565 thread sealant.

The insertion length of the BGA244 is shown in Figure 37 (Appendix C).

Dimension 'X' refers to a reference dimension on the gas fittings. Refer to the Swagelok part drawings listed for additional details.

BGA244 Gas Fitting	Part Number	Dimension X ¹
1/8" -27 Female NPT	SS-2-A	4.360"

¹Dimension "X" is measured from the outside dimension of the fittings. This is the outer most part of the fitting.

If damaged, the adapters can be replaced by the customer. See the Chapter 7: *Service* for details on the proper procedure to replace the gas fittings. It is not recommended that NPT pipes be directly connected to the BGA244. Instead use the installed NPT thread adapters to avoid damaging the port threads.

Use a 9/16th inch wrench and either a pipe wrench or pliers to tighten the tubing so as to not damage the unit. Use thread sealant when connecting tubing to the BGA244. SRS recommends Loctite 565 thread sealant to seal leaks and prevent galling/seizing of the threads. A 0.2-oz tube of the Loctite is available from SRS (accessory 45855K12). Be sure to clean out gas lines prior to connecting them to the BGA244 to avoid contamination.

Installation Guidelines

1. Apply a strip of thread sealant around the male pipe threads, leaving the first 2 turns uncovered.
2. Finger tighten the pipe into the fitting.
3. Hold the fitting with the 9/16th inch wrench. Wrench tighten the pipe 1.5 – 3.0 turns past finger tight using a pipe wrench or pliers.
4. Allow the Loctite 565 at least 24 hours to cure before applying pressure or vacuum to the system.

Gas Fitting Adapters

SRS offers a number of different adapters to enable the BGA244 to connect to a wide variety of different systems. Each adapter converts from 1/8"-27 male NPT to one of several different standards. Refer to the specific Swagelok or McMaster-Carr part drawings for specific details and dimensions for each fitting.

TABLE 2: GAS FITTING ADAPTERS

Accessory	Fitting Type	Part Number
SS2A	1/8"-27 female NPT	Swagelok SS-2-A
SS4RA2	1/4" -18 female NPT	Swagelok SS-4-RA-2
SS4VCR12	1/4" male VCR	Swagelok SS-4-VCR-1-2
SS4VCO12	1/4" male VCO Body	Swagelok SS-4-VCO-1-2
SS4HC12	1/4" hose barb (1/4" ID hose)	Swagelok SS-4-HC-1-2
SS40012	1/4" compression fitting	Swagelok SS-400-1-2
SS6M012	6mm compression fitting	Swagelok SS-6M0-1-2
5482K119	1/4", 37° flared AN tube, male, 7/16"- 20 TPI	McMaster 5482K119
4822T76	1/8" BSPP female, 1/8"-28 TPI	McMaster 4822T76
4092K26	1/8" BSPT female, 1/8"-28 TPI	McMaster 4092K26

Always use 2 wrenches to tighten the adapters so as to not damage the unit. Use thread sealant when connecting the adapters to the BGA244. SRS recommends Loctite 565 thread sealant to seal leaks and prevent galling/seizing of the threads. A 0.2-oz tube of the Loctite is available from SRS (accessory 45855K12).

Installation Guidelines

1. Apply a strip of thread sealant around the male pipe threads, leaving the first 2 turns uncovered.
2. Finger tighten the adapter into the fitting.
3. Wrench tighten the adapter 2 1/4 turns past finger tight using 2 wrenches.
4. Allow the Loctite 565 at least 24 hours to cure before applying pressure or vacuum to the system.

Ordinarily the adapters should be installed into the male-female thread savers that come with the BGA244. If minimum insertion length is critical, the male-female thread savers can be removed and the adapters can be directly fastened to the BGA244. Extreme caution should be exercised while changing the fittings to avoid damaging the unit. See the Chapter 7: *Service* for details on the proper procedure to replace the gas fittings.

Follow the manufacturer's guidelines when connecting external tubing the adapters. Swagelok and others have detailed instruction for connecting different types of tubing.

BGA244HP

The BGA244HP come with two welded, non-replaceable ¼" Male VCR fittings, intended for high purity systems. These fittings are assembled without any pipe thread sealant or tape. If these fittings are ever damaged, the unit needs to be returned to SRS for repair.

The insertion length of the BGA244HP is shown in Figure 37 (Appendix C).

Dimension 'X' refers to a reference dimension on the gas fittings. Refer to the Swagelok part drawings listed for additional details.

BGA244HP Gas Fitting	Part Number	Dimension X ¹
¼" Welded Male VCR	SS-4-VCR3-4MTW + SS-4-VCR-4	4.770"
VCR Gasket	SS-4-VCR-2	0.028"

¹Dimension "X" is from mating surface to mating surface and does not include the thickness of the VCR gasket. Be sure to account for this when planning installation.

Follow the Swagelok VCR Fitting Installation Instructions when connecting to the BGA244. Remember to install a new VCR gasket between male and female fittings. Use 2 wrenches to tighten the fittings so as to not damage the unit. Swagelok recommends tightening VCR fittings ½ turn past finger tight for stainless steel or nickel gaskets.

Electrical Connections



Be sure that your installation confirms to all safety and electrical code requirements.



For CE compliance it is recommended that all power and I/O cables are shielded and grounded.



The BGA244 has no line voltages connections. Applying line voltage to any pin of any connector on the BGA244 will cause severe damage to the instrument and is a fire and smoke hazard.

The BGA244 supports a wide variety of electrical connections for power, control and monitoring. Power is provided over USB (+5V_{DC}) or a separate +24 V_{DC} power supply. Computer interfaces include USB, RS-232 and RS-422. There are multiple user configurable analog inputs, analog outputs and two configurable event relays.

Electrical Noise Precautions

Electrical Noise can cause interference between different devices. AC line wiring, motors, pumps, relays and their control wiring are common noise sources. Wherever possible, route the BGA244 wiring separate from noise sources. This is a particular concern for the Analog I/O signals.

There are two main groups of connections used by the BGA244. Wires within each group can normally be bundled together without adverse effects. Try to avoid combining wires from different groups, especially for long cable runs.

Group 1: Power, ground, computer interfaces and relay signals

Group 2: Analog Inputs, Analog Outputs

Avoid passing current through the BGA244 and its gas tubing. This can be accomplished by providing a current return path for high current wiring and making sure that all metal surfaces are properly grounded.

Grounding



Make sure that the BGA244 is properly grounded. Depending on the installation, power supply wiring and gas tubing may not be properly connected to earth ground. If necessary, connect the BGA244 chassis ground lug (C1) to a suitable earth ground.

Electrical Connectors

The BGA244 features different options for power, computer control and analog input and output. The unit can be powered over either the USB port or by 24 V_{DC}. Computer control is available over USB, a RS-232 DCE port or a RS422 port. There are three separate analog output ports, two analog input ports and two event relays. See Figure 1 (page 6) for the connector locations.

- Ground Lug (C1)
- USB for Power and Computer Control (C2)
- RS-232 Interface for Computer Control (C3)
- RS-422 Interface for Computer Control (C4)
- +24V Power Connector (C5)
- Analog Measure Out and +24 V_{DC} Power (C6)
- 2 Analog Outputs (0-5 V, 0-10 V, 4-20 mA) (C7)
- 2 Analog Inputs (0-10 V, 4-20 mA, 4-20 mA w/ loop power) (C7)
- 2 Event Relay Contacts (C8)

Connector Pinouts

TABLE 3: C3: RS-232

Pin	Signal
1	CD
2	Rx
3	Tx
4	DTR
5	GND
6	DSR
7	RTS
8	CTS
9	RI

TABLE 4: C4: RS-422

Pin	Signal
1	Chassis GND
2	-RxD
3	+RxD
4	-TxD
5	+TxD

TABLE 5: C5: +24V

Pin	Signal
1 (Center)	+24 V _{DC}
2 (Outside)	GND (24V Return)

TABLE 6: C6: MEASURE OUT

Pin	Signal
1	Chassis GND
2	GND (Measure Return)
3	Measure Out
4	GND (24 V Return)
5	+24 V _{DC}

TABLE 7: C7: ANALOG I/O

Pin	Signal
1	Analog In 1 +
2	Analog In 1 -
3	Analog In 2 +
4	Analog In 2 -
5	Analog Out 1 +
6	GND (Out 1 Return)
7	Analog Out 2 +
8	GND (Out 2 Return)

TABLE 8: C8: RELAY CONTACTS

Pin	Signal
1	Relay 1 Normally Open
2	Relay 1 Common
3	Relay 1 Normally Closed
4	Relay 2 Normally Open
5	Relay 2 Common
6	Relay 2 Normally Closed

Terminal Strip Connectors

The BGA244 uses Phoenix Contact Combicon MC Series connectors for its terminal strips. Each terminal strip is made up of two pieces: a PCB soldered base strip and a detachable terminal block. Each terminal strip connector in the BGA244 comes with its terminal block inserted into the base strip.

Tip: It's usually easier to remove the terminal strip from the BGA244 before connecting wires. Unplug the terminal strip by pulling straight back from the unit. Loosen the screws before inserting the wires. Make sure to observe the correct pinouts!

The terminal blocks can accept wire sizes from 16 – 28 AWG (1.29 – 0.32 mm dia). Wires are secured by screw connection. Wires may be attached to the terminal block when it is disconnected from the base strip to simplify assembly. Make sure the wire installation is stripped back far enough to ensure good electrical contact.



Be sure to use the correct pinout for each connector. Failure to do so can result in damage to the BGA244. Pay special attention to C4 and C6, as they are adjacent and have the same number of pins. See *Terminal Strips* (page 211) in the Service section for information on replacement terminal blocks.

Power

USB Power

The BGA244 can be powered by its USB Type B connector (C2). Remember that some features require +24 V_{DC} to operate, including the heaters, analog inputs and outputs, relays and RS422.

It's frequently convenient to configure units at a desktop computer using USB power. All parameters can be configured under USB power, but features that depend on +24 V_{DC} won't operate until +24 V_{DC} is supplied.

When running, the BGA244 draws about 0.35 amps from the USB interface. It requires the USB voltage to be within 4.75 to 5.25 V_{DC}. If the voltage is outside this range, an Alert is displayed. If the voltage drops below 4.6 V_{DC} a Fault is generated and the unit will not operate. Refer to *Faults* (page 85) for more details.

Most USB chargers, desktop and laptop computers can supply the proper voltage and current. A dedicated USB charging port or charging downstream port is specified to supply enough current to operate the BGA244. A plain downstream port can probably supply enough current.

The USB voltage will droop below the acceptable range if a device cannot supply enough current. Add a powered hub to increase the voltage and current. Make sure that the powered hub is capable of providing around 0.5 amps over a single port.

USB Power Cables

All USB cables are not created equal!

Some USB cables have power wires as thin as 24 AWG. These produce large voltage drops that cause the BGA244 voltage to drop below its operating range. Thin USB cables typically have small gauge power wires.

Use USB cables that have 20 AWG power wires. The power wire gauge is frequently printed on the cable or packaging. SRS recommends Belkin Gold Series Hi-Speed USB 2.0 cables. These are available from Amazon, CDW and other distributors.

Length	Belkin p/n
6' (1.8M)	F3U133-06-GLD
10' (3.0M)	F3U133-10-GLD

+24V Power

The BGA244 can also be powered by 24 V_{DC} through either C5 or C6. If the USB port is connected to an external device, its current goes to 0 A when the BGA244 is powered by +24 V_{DC}.


The acceptable voltage range is 24, ±1 V_{DC}. The maximum ripple voltage is < 240 mV_{PP}. The maximum +24 V_{DC} power supply current is 2.7 A. However the BGA244 can be operated at much lower currents. With the heaters turned off, the +24 current draw is between 0.2 – 0.35 A, depending on the analog I/O configuration. The maximum heater current can be set to anywhere from 0.01 to 2.2 A.

Conditions	Current
No Heaters, No Analog I/O	0.2 A
Using Analog I/O	Add 0.15 A
Using Heaters	Add Max Heater Current setting (0.01 – 2.2 A)

Make sure that the wire gauge used to connect the +24 V power supply can support the maximum current required without excessive voltage drops. If operating at the maximum heater current, 100 feet (30 meters) of 18 AWG wire will have a voltage drop of ~3.5 V_{DC}. Lower currents and shorter wires will minimize this drop.

There are two different connectors that can be used to provide +24 V power. Power can be provided using connector C5, a 2x5.5 mm barrel jack (see figure) or on pins 4 and 5 of terminal strip C6. Make sure to connect the +24 V power supply with the correct polarity, to the correct pins. Failure to do so may cause serious damage to the BGA244.

C6	
C6 pin 5	+24 V _{DC}
C6 pin 4	Return



BGA-24 USB Power Supply

Accessory BGA-24 is a 100 – 240 V_{AC} input, 24 V_{DC} 2.5 A output power supply suitable to power the BGA244. It connects using the 2x5.5 mm barrel jack (C5).

Computer Interfaces

The BGA244 can be remotely operated over the USB interface, the RS-232 serial interface, or the RS-422 serial interface. Any host computer interfaced to the instrument can control and monitor all of its functions. For details on configuring and monitoring the interfaces, see *Computer I/O* (page 79). All interfaces communicate with the BGA244 using the commands listed in Chapter 6: *Remote Programming*.


USB

The USB type B connector (C2) is located on the front of the BGA244. The interface is USB 2.0, Full Speed compatible. USB 2.0 is specified for a maximum cable length of 5 meters. If powering the BGA244 over USB, be sure cable is able to support the operating current. See *USB Power Cables* in the previous section for details. Before a computer can control the BGA244 over USB, it needs to have the appropriate USB driver installed. If you are not using Microsoft Windows, see *Using the USB Drivers* (page 197) for details on installing and using the USB drivers.

The first time the BGA244 is connected to the USB port of a computer running Microsoft Windows, you will likely be prompted with a “New Hardware Found” message and an invitation to search for the USB Driver. There are two USB drivers for the device (VCP and D2XX drivers). Depending on the version and configuration, Windows may either automatically install the drivers or prompt you to search for them. Allow Windows to install the drivers. Occasionally only a single driver will install. In some cases disconnecting and reconnecting the BGA244 will cause the second driver to load correctly. If there are difficulties installing the driver, refer to Appendix E for details on manually installing the drivers.

RS-232

RS-232 connector C3 is located on the front of the BGA244. The connector is a standard 9 pin, type D, female connector configured as a DCE (transmit on pin 2, receive on pin 3). CTS and RTS are supported. See figure below for the connector pin numbering.



RS-232 Pinout

In order to communicate properly over RS-232, both the BGA244 and the host computer must be set to the same configuration. The RS-232 interface supports baud rates from 2400 to 115.2 k baud. In general, the highest baud rates will operate successfully for shorter cable lengths. At lower baud rates, cable lengths over 100m (300 ft) should be possible. Communication errors can be caused by excessive cable length, overly high baud rates or electrical noise. If errors occur, operating at a lower baud rate will usually help.

RS-422

The RS-422 connector (C4) can only be used if an external +24 VDC power supply is connected to the BGA244. Connections to the RS-422 interface are made using terminal block C4 located on the front of the BGA244.

The RS-422 interface is implemented as a 4 wire, point-to-point, non-multidrop connection. It supports a single transmitter and single receiver pair. Connections are made using a 5 pin terminal strip. Tx and Rx connections between the BGA244 and host should be made as follows. Note that the transmit pins on the BGA244 connect to the receive pins on the host and the receive pins of the BGA244 connect to the transmit pins of the host.

BGA244 pin number	BGA244 Connection	Host Connection
C4-5	+TxD	+RxD
C4-4	-TxD	-RxD
C4-3	+RxD	+TxD
C4-2	-RxD	-TxD
C4-1	GND	GND

RS-422 Pinout




FIGURE 4: RS-422 SIGNAL PATH

Use twisted pair cabling with an impedance of $\sim 100 \Omega$, especially for longer cable runs and higher data rates. Shielded Cat5 or Cat6 cable is a good choice. A 100Ω terminating resistor (R) can be added to each receive end for long cable runs or high data rates to improve signal quality. The resistor should be rated for at least $\frac{1}{4}$ watts. The resistor can be connected along with the RxD lines at the BGA244 terminal strip.

In order to communicate properly over RS-422, both the BGA244 and the host computer must be set to the same configuration. The RS-422 interface supports baud rates from 2400 to 115.2 k baud. RS-422 can operate at cable lengths over 1000 m (3250 ft). In general, the highest baud rates will operate successfully for shorter cable lengths. Communication errors can be caused by excessive cable length, missing terminators, overly high baud rates or electrical noise. If errors occur, adding a terminator and/or operating at a lower baud rate will usually help.

Analog I/O Connections

The BGA244 has three analog outputs and two analog inputs. These features can only be used if an external +24 V_{DC} power supply is connected to the BGA244. The analog I/O signals are located on C6 and C7. There are a number of different parameters that can be set for the analog inputs and outputs.

Pin	Signal
C7-1	Analog In 1 +
C7-2	Analog In 1 -
C7-3	Analog In 2 +
C7-4	Analog In 2 -
C7-5	Analog Out 1 +
C7-6	Analog Out 1 -
C7-7	Analog Out 2 +
C7-8	Analog Out 2 -
C6-3	Measure Out +
C6-2	Measure Out -

Analog I/O pinout

Analog Output

There are three separate Analog Outputs: Measure Out, Output 1 and Output 2. These can be independently set as voltage or current outputs. Measure Out is always linked to the Instrument Mode measured value. Outputs 1 and 2 can be linked to one of several different measured parameters or set explicitly by the user. The output full scale ranges can be scaled to match external devices. See *Analog I/O* (page 81) for information on configuring the outputs.

Voltage Outputs

The voltage outputs are unipolar and are ground referenced at the BGA244. The minus (-) outputs are connected to the BGA244s chassis ground. Avoid connecting the minus outputs to ground at the destination to avoid ground loops.

The voltage outputs have a maximum drive current of 20 mA and can drive capacitive loads of up to 1 μ F without oscillation. The output voltage may be reduced by resistive losses for long cable lengths and high currents. Make sure that the wire size, length and load current do not create excessive errors.




FIGURE 5: VOLTAGE OUTPUT

Current Outputs

The current outputs are unipolar and return to the BGA244s ground. The minus (-) outputs are connected to the BGA244s chassis ground. Avoid connecting the minus outputs to ground at the destination to avoid ground loops.

The current outputs have a compliance voltage of 16.5 V and can drive inductive loads up to 50 mH without oscillation. The maximum load resistance, including cable resistance, is 825 Ω . Cable resistance can be large depending on the length and wire size. Make sure that the cable resistance plus the load resistor is less than 825 Ω .




FIGURE 6: CURRENT OUTPUT

Analog Input

There are two separate Analog Inputs: Input 1 and Input 2. These can be independently set to measure voltage or current inputs. An internal loop power voltage source can be enabled for current input. Inputs 1 and 2 can be read on the front panel or over the computer interfaces. They can also be linked to a pressure transducer to monitor gas pressure. See *Analog I/O* (page 81) for details on configuring the inputs. See *Pressure Transducers* (page 31) for details on connecting pressure transducers to the analog inputs.

Voltage Input

The voltage input measures the differential voltage between the plus (+) and minus (-) lines. The input voltage range of either of the inputs is -0.1 to +20 V relative to ground. The differential voltage range from (+) to (-) is -2 to +10.2 V. This allows measurement of signals that are biased above ground.




FIGURE 7: VOLTAGE INPUT

Current Input

The current input measures the absolute value of the current from the plus (+) input to the minus (-) input. The signal can be either polarity, with a maximum of 24 mA. The voltage range at each input must be between -0.5 and + 20 V; the burden voltage is 5.5 V for 20 mA. The wide voltage range and low burden voltage allows for multiple sense resistors or an external power supply in series with the current loop.




FIGURE 8: CURRENT INPUT

Current with Loop Power Voltage Source

Current Input with Loop Power is a special configuration of the current input mode that allow both signal and power to be transmitted over a pair of wires.

The Loop Power voltage source is ground referenced at the BGA244 and requires that the measured current returns to the minus (-) input. Devices that ground either current loop lead or require a floating loop power supply must use an external supply.

Hint: In general, devices using this feature should float with respect to the BGA244's ground. This feature may not operate properly if either current loop lead is connected to ground.

The maximum current amplitude is 24 mA. The Loop Power supply can be set between 6 and 19 V with a maximum output current of 50 mA.




FIGURE 9: CURRENT INPUT W/ LOOP POWER (SUITABLE DEVICE)




FIGURE 10: CURRENT INPUT W/ LOOP POWER (UNSUITABLE DEVICE)

Event Relays

There are two separate SPDT relays that are linked to the BGA244's Event 1 and Event 2 conditions. Events can be configured to switch for things like exceeded limits, loss of the measurement signal and system faults. They can also be directly set and cleared on the front panel or over the computer interfaces. See *Events* (page 55) for detail on configuring and controlling the Event Relays.

Pin	Signal
C8-1	Relay 1 Normally Open
C8-2	Relay 1 Common
C8-3	Relay 1 Normally Closed
C8-4	Relay 2 Normally Open
C8-5	Relay 2 Common
C8-6	Relay 2 Normally Closed

Event Relay pinout

These relays will operate only if an external +24 V_{DC} power supply is connected to the BGA244. The relays are floating with respect to each other and chassis ground. Each relay consists of a common, a normally open and a normally closed connection. Normally open and normally closed refer to the relay in the de-energized state.




FIGURE 11: EVENT RELAY CONTACTS

The Event relay contacts are rated for the following conditions.

Max Switching Power	30 W, 62.5 VA
Max Switching Voltage	60 V _{DC} , 42.4 V _{PK} , 30 V _{AC}
Maximum Switching Current	1 A
Maximum Carrying Current	2 A
Lifetime (42 V _{DC} , 0.1 A resistive load)	10 ⁶ operations

Switching high power loads can dramatically reduce the relays lifetime. If driving an inductive load, make sure to use catch diode to minimize inductive fly-back.

Common uses of the relays include:

- Turning on an alarm when a limit is exceeded
- Opening a valve to control a process
- Computer control of devices

Pressure Transducers

In many cases real time monitoring of the gas pressure can improve measurement accuracy. The simplest way to do this is by connecting a pressure transducer to one of the BGA244's analog inputs. This requires that the BGA244 is powered over 24 V_{DC} to operate the analog inputs. Using the voltage or current analog inputs, the BGA244 can interface with a variety of different transducers. Scale factors for minimum and maximum pressures can be set, in either absolute or gauge pressure.

There is a wide variety of pressure transducers that can interface with the BGA244. Differences among them include gas fittings, pressure ranges and electrical connections. When selecting a pressure transducer, select one that matches the system operating pressure. Transducers typically have the best linearity and accuracy when not operated at their extreme limits.

Refer to *Analog Inputs* (page 81) for details on configuring the Analog Inputs.

Mounting Location

Locate the pressure transducer on the BGA244 side of any restriction to minimize offsets in the measured pressure. Make sure to account for any drop in pressure between the measured or regulated pressure and the BGA244 for large flow rates. Install the pressure transducers following their manufacturer's instructions.

Types of Pressure Transducers

The choice of which pressure transducer to use largely depends on the application it will be used in. Transducers come in many different ranges, fittings and accuracies. Choose a transducer with good sensitivity and accuracy and that matches the operating pressure of your system. If possible, use a transducer that reads in absolute pressure to eliminate ambient pressure variation.

The BGA244 can interface to most pressure transducers that output a voltage or current within the nominal analog input range. The BGA244 includes a selectable loop power voltage source that can be used to power many current output transducers. Transducers of this type are the simplest to integrate with the BGA244.

Input	Minimum	Maximum
Voltage	0 V	10 V
Current	4 mA	20 mA

Voltage Output Transducers

Voltage output pressure transducers have separate signal and power connections. In addition, they often have a drain wire that connects to the cable shield. They require an external power supply.

The BGA244 supports transducers with a full scale output range that lies between 0 and 10 V. Voltages must always be ≥ 0 V (no bipolar outputs). The output can be scaled to either absolute pressure or gauge pressure.

Connect an appropriate power supply per the manufacturer's instructions. If needed the Drain wire can be connected to ground at the BGA244 ground lug (C1). Connect the Signal and Signal Return (or +/- Signal) to +/- In 1 or In 2 on the BGA244. Refer to *Analog Input* (page 27) for details on connecting the transducer to the BGA244. Configure the BGA244 as described below.

- Set the analog input to: Enabled, Voltage.
- Select Absolute or Gauge units, depending on the transducers specifications. If Gauge units are selected, make sure to enter the ambient pressure.
- Set the Min and Max to the transducers minimum and maximum pressure values.
- Check "Use as Pressure Gauge".
- Set the Analysis Pressure to the appropriate input (Analog Input 1 or 2).

Current Output Transducers

Current output pressure transducers come in several different configurations. They can have separate power and signal lines or the power can be provided in series with the signal lines (loop power). In addition, they may have a drain wire that connects to the cable shield.

The BGA244 supports current output transducers with a full scale range of 4 – 20 mA. Currents can be either positive or negative. The output can be scaled for absolute pressure or gauge pressure.

The BGA244 can provide a loop power voltage source (6 – 19 V) if the transducer can accept a ground referenced voltage. See the pressure transducers manufacturers' information for specific details. Refer to *Analog Input* (page 27) for details on devices that are loop power compatible with the BGA244.

Hint: If either the + or – signal lines are connected to the pressure transducers body or drain wire, it probably won't work with the BGA244 loop power voltage source.

Transducers with External Power Supplies

Connect the appropriate power supply as described by the pressure transducer manufacturer. If needed, the Drain wire can be connected to ground at the BGA244 ground lug (C1). Connect the Signal and Signal Return (or +/- Signal) to +/- In 1 or In 2 on C7 on the BGA244. Refer to *Analog Input* (page 27) for details on connecting the transducer to the BGA244. Configure the BGA244 as follows.

- Set the analog input to: Enabled, Current
- Select Absolute or Gauge units, depending on the transducers specifications. If Gauge units are selected, make sure to enter the ambient pressure.
- Set the Min and Max to the transducers minimum and maximum values.
- Check: Use as Pressure Gauge.
- Set the Analysis Pressure to the appropriate input (Analog Input 1 or 2).

Transducers with Loop Power

Connect the transducer Signal (or +) wire to +In 1 or 2 on the BGA244. Connect the Return (or -) wire to –In 1 or 2 on the BGA244. If needed, the Drain wire can be connected to ground at the BGA244 ground lug (C1). Refer to *Analog Input* (page 27) for details on connecting the transducer to the BGA244. Configure the BGA244 as follows.

- Set the analog input to: Enabled, Current w/ Loop Power
- Set the Loop Power Voltage to the voltage specified by the pressure transducer manufacturer, typically 12-15 V.
- Select Absolute or Gauge units, depending on the transducers specifications. If Gauge units are selected, make sure to enter the ambient pressure.
- Set the Min and Max to the transducers minimum and maximum values.
- Check: Use as Pressure Gauge.
- Set the Analysis Pressure to the appropriate input (Analog Input 1 or 2).

Chapter 3: Operation Guide

BGA244 User Interface

The BGA244 features a touch screen display that allows the user to set and display all settings and measurements. Units can also be fully configured and monitored over one of the three computer interfaces.

The description of the BGA244 features in this chapter is directed toward operating the unit from the display. Chapter 5: *BGAMon* contains information on operating the BGA244 using the BGAMon software. Chapter 6: *Remote Programming* contains information on the remote commands that can also be used to configure the BGA244.

Navigation

There are several navigation keys that are present on many of the pages to make it easier to navigate the different instrument functions of the BGA244. See *Manual Conventions* (page xvii) at the beginning of this manual for a list of conventions used to describe the different functions of the interface. The following rules will help you navigate the menus.

- [HOME] returns you to the Home Page.
- [SETUP] take you to the configuration pages that aren't displayed on the Home page.
- [←] or back key returns up one level from the display you are currently on.
- [PAGE ↑] and [PAGE ↓] are used to scroll through a page that can't all be displayed on the screen at the same time.
- Different pages are referenced by their location in the interface hierarchy. (Home/Setup/Control Panel/Units) would be reached from the Home page by pressing [Setup] [Control Panel] [Units].

Help Screens

Most screens have a Help page associated with them accessed by a Help key. This page will have information about the settings and displayed parameters on the screen. The [Page ↑] and [Page ↓] are active if the Help screen is more than one page long.

Map of Interface Functions





FIGURE 12: MAP OF USER INTERFACE FUNCTIONS

Power On

When power is first applied, a power-on splash screen will appear.

After a few seconds the “About” information page will be displayed. This shows the unit serial number, firmware version, hardware information and calibration date.



After about 15 seconds the currently selected Home page will appear.

The three LEDs also indicate power on behavior. This is especially useful for units with the metal protective cover installed (Accessory BGA-M). When power is first applied, the Power LED will light for ~ 1 second; followed by all 3 LEDs lighting for 5 seconds. At this time the Power LED should blink once indicating the code is loading properly.

You can enter the screen calibration routine by pressing a finger to the screen when power is applied. Hold your finger down until the message “Release Screen to start Screen Calibration” appears, then follow the instructions to complete the routine. See Display later in this chapter for more information.


If the firmware detects a fatal error the unit will restart in the Factory Safe Mode. This will be indicated on the bottom of the About screen. Refer to *Factory Safe Mode* (page 205) for information.

Appearance

The BGA244 uses a touch screen display to configure measurements and display measurement results. The Home page displays the primary measurements and has controls to navigate the instrument functions and secondary measurements. You can always return to the Home page directly from any page by pressing [Home].

Each Instrument Mode has a unique Home page. Home pages consist of a Measurement Section that is specific to the Instrument Mode and a Control Section that is common to all Instrument Modes.

There are two possible appearances of the display depending on the orientation: vertical, with the control section below the measurement section and horizontal, with the control section to the right of the display section.



Home Page, Vertical View

Home Page, Horizontal View

FIGURE 13: BGA244 HOME PAGE

Measurement Section

The Measurement Section displays the operating mode, measurements and status information. The meter scale, limits and other functions are set using keys in the Control section. See the sections on *Binary Gas Analyzer* (page 40), *Gas Purity Analyzer* (page 44) and *Physical Measurements* (page 47) later in this chapter for details on each Instrument Mode.

Screen messages can appear in the Measurement section. These indicate operating states, errors or problems with the measurements. See *Screen Messages* (page 50) for details on these functions.

Control Section

The Control Section is common to all Home pages. The Control buttons are used to set parameters and access menus within the user interface. Press the appropriate button to access that control. See *Controls* later in this chapter for details on setting each control.

Note: The touch screen on the BGA244 must be actually be pressed to activate functions. This is normally indicated by a key click or beep. Hovering over the button like you would on a cell phone or tablet won't activate the control.



Instrument Modes

The BGA244 can operate in one of three Instrument Modes: Binary Gas Analyzer Mode, Gas Purity Analyzer Mode and Physical Measurement Mode. The Instrument Mode determines the measurements made by the BGA244. Each Instrument Mode has a unique home page and measurement specific parameters (like gas selection, limits, and meter scale). See *Instrument Mode* (page 58) for more information.

Select Instrument Mode

Press [SETUP] to access the Setup page. This page has controls to configure the all of the instrument function besides those in the Control section.

Press [Instrument Mode ↓] to open the Instrument Mode list. Press the desired mode to select it. The currently selected mode is highlighted in yellow.

Binary Gas Analyzer

The Binary Gas Analyzer reports the fraction of one gas in a two gas mixture. Both gases must be known.

The Binary Gas Ratio is reported as the ratio of the primary gas in the mixture. Results are reported in percent (%), parts per million (ppm) or fraction (0 – 1.0). Go to (Setup/Control Panel/Units) to change the units. See the *Units* (page 74) for details on selecting units.

Principle of Operation

The speed of sound in an ideal gas can be approximated by:

$$W = \sqrt{\frac{\gamma RT}{M}}$$

Where W is the speed of sound, γ is the ratio of specific heat capacities ($\gamma = C_p/C_v$), M the molar mass, T the absolute temperature and R the ideal gas constant. For a gas mixture, γ and M are determined by the properties of each gas and their ratio within the mixture. By measuring the speed of sound and temperature of the gas mixture, and knowing the properties of each gas, the mole fraction of each gas can be precisely determined.

The speed of sound in real gases is somewhat more complicated. There are a number of gas specific effects that complicate the ideal case. These include thermo-viscous frequency shifts, temperature dependencies of heat capacity, dispersive effects and intermolecular (virial) effects that depend on both pressure and temperature. The BGA244 Factory Gas table contains data for these effects for nearly 500 different gases. See Appendix A: *Gas Table* for a detailed description of the properties used to characterize gases for the BGA244.

Binary Gas Analyzer Home Page




FIGURE 14: BINARY GAS ANALYZER HOME PAGE

- Page Identification: Binary Gas Analyzer.
- Selected Gases: Primary Gas in Secondary Gas. Refer to *Selecting Gases* (page 58) for details.
- Measured Value: Concentration of Primary Gas in the mixture.
- Units: %, ppm or fraction, and Mole/Mass fraction. See (Setup/Control Panel/Units) for units and (Setup/Control Panel/Masurement) for Mole/Mass fraction.
- REL: Visible when the REL function is active.
- Pressure: Measured or entered pressure used for analysis. See the *Pressure* (page 64) for more information.
- Temperature: Measured gas temperature.
- Accuracy: Estimated accuracy.
- Upper & Lower Limits: Set by [LIMIT] in the Control section. See *Limits* (page 54) for more information.
- Upper and Lower Scale: Set by [METER SCALE] in the Control section. See *Scale* (page 55) for more information.
- Meter: Graphical display of the Measured Value on a graph scaled by the upper and lower scale values. The red areas are set by the upper and lower limit values.

Details

This is a description of some of the items specific to the Binary Gas Analyzer Instrument Mode.

Range

The maximum range for the Binary Gas Analyzer is -2% to +102% or its equivalent in fraction or ppm. The 2% over range is to allow for measurement inaccuracies. Outside of this range the display will limit at > (max) or < (min) to indicate out of range.

Selecting gases

Both the primary and secondary gases must be selected before measurements can be made. There are nearly 500 different gases contained in the BGA244 Factory Gas Table. Go to (Setup/Select Gas) to view the Gas Selection page. Here you can select the Primary and Secondary Gases by name, chemical formula or CAS#. See *Selecting Gases* (page 58) for more information.

Note that the BGA244 reports the ratio of Primary to the total mixture of the Primary and Secondary Gases. To report the concentration of the Secondary Gas, press [SWAP GASES] on the Gas Selection page.

Molar Fraction vs Mass Fraction

The BGA244 can report the gas ratio in either mole or mass fraction. These differ by the molar masses of the gases. Certain applications commonly use one or the other.

Go to (Setup/Control Panel/Measurement) to change between mole and mass fractions. See *Binary Gas Concentration* (page 75) for more details.

Accuracy Estimator

The BGA244 calculates a real time accuracy estimation of the gas ratio measurement. The accuracy estimation is based on temperature uncertainty of 0.1 °C and a pressure uncertainty of 1 psi.

REL

The REL (relative) function can be used to zero the ratio measurement to a reference gas. This can be either a pure gas, or a gas blend with a known ratio. See the *REL* (page 61) for more information. The REL indicator only appears when the REL function is active.

Dual Concentrations

There are a few gas combinations where there are two valid molar ratios for a given speed of sound. The ratio of these gases can be accurately determined for certain concentrations, but there can be two ratio values for a portion of the range. Both ratios are equally valid as far as the BGA244 can determine.



If a dual concentration is detected, the BGA244 reports both ratios. Concentration 1 is the smaller value and is displayed on top. Concentration 2 is the larger value and is displayed on the bottom.

Gas Purity Analyzer

This analyzer reports the purity of a gas as measured by the ratio of its measured speed of sound to its ideal speed of sound. The gas species must be known. This instrument mode is useful when measuring relatively pure gases. Results are reported in percent (%), parts per million (ppm) or fraction (0 – ±1.0). Go to (Setup/Control Panel/Units) to change the units. Refer to *Units* (page 74) for details on selecting units.

Principle of Operation

The speed of sound of a pure gas at a known temperature and pressure is well known. Impurities will change this speed as a function of their mole fraction, molar mass and γ per the equation described in the Binary Gas Analyzer mode.

$$W = \sqrt{\frac{\gamma RT}{M}}$$

The definition of the Gas Purity in the BGA244 is:

$$\frac{\Delta W}{W} = \frac{\text{measured speed of sound} - \text{expected speed of sound}}{\text{expected speed of sound}}$$

Both the expected speed of sound and the measured speed of sound are normalized to NTP (20°C, 1 atm) before calculating the ratio to eliminate temperature and pressure dependencies. Since temperature and pressure affect each gas differently, it's important to specify the gas being measured for best accuracy. See Appendix A: *Gas Table* for a detailed description of the properties used to characterize gases for the BGA244.

The gas purity is most sensitive when contaminants significantly heavier or lighter than the selected gas. If the purity measurement ($\Delta W/W$) is positive, the contaminating gas is ordinarily lighter than the selected gas. Similarly if $\Delta W/W$ is negative, the contaminating gas is ordinarily heavier than the selected gas.

Gas Purity Analyzer Home Page




FIGURE 15: GAS PURITY ANALYZER HOME PAGE

- Page Identification: Gas Purity Analyzer.
- Selected Gas: Gas name or reference speed of sound. See the *Selecting Gases* (page 58) for details.
- Measured Value: $\Delta W/W$
- Units: %, ppm or fraction. See (Setup/Control Panel/Units) for unit selection.
- REL: Visible when the REL function is active. (not shown)
- Pressure: Measured or entered pressure used for analysis. See the *Pressure* section later in this chapter for more information.
- Temperature: Measured gas temperature.
- Accuracy: Estimated accuracy
- Upper & Lower Limits: Set by [LIMIT] in the Control section. See the *Limits* section later in this chapter for more information.
- Upper and Lower Scale: Set by [METER SCALE] in the Control section. See the *Scale* section later in this chapter for more information.
- Meter: Graphical display of the Measured Value on a graph scaled by the upper and lower scale values. The red areas are set by the upper and lower limit values.

Details

This is a description of some of the items specific to the Gas Purity Analyzer Instrument Mode.

Select gas

The Primary (or Reference) Gas must be selected before accurate measurements can be made (the Secondary Gas is not used for Gas Purity Measurements). There are nearly 500 different gases contained in the BGA244 Factory Gas Table. Go to (Setup/Select Gas) to view the Gas Selection page. Here you can select the Primary and Secondary Gases by name, chemical formula or CAS#. See *Selecting Gases* (page 58) for more information.

If you are using a blended gas that is not in the Gas Table, select the dominant gas species as the Primary Gas. Then use the REL function to set the current speed of sound as the reference. If your mixture doesn't have a dominant gas species, select a Primary Gas with a similar speed of sound to the mixture and use the REL function.

Accuracy Estimator

The BGA244 calculates a real time accuracy estimation of the gas purity measurement. The accuracy estimation is based on temperature uncertainty of 0.1 °C and a pressure uncertainty of 1 psi.

REL

The REL (relative) function can be used to zero the measurement to a pure reference gas. See the *REL* section later in this chapter for more information on using this feature. The REL indicator only appears when the REL function is active.

Physical Measurements

The Physical Measurements Analyzer reports the physical measurements of a gas made by the BGA244. These include the

- Measured Speed of Sound
- Speed of Sound normalized to NTP (20°C, 1 atm)
- Temperature
- Pressure

The speeds of sound are reported in m/s, kph or mph. The temperature is reported in °C, °K or °F. The pressure is reported in psi, atm, bar, Pa, mmHg or torr. Go to (Setup/Control Panel/Units) to change units. See *Units* (page 74) for more details.

Principle of Operation

The measured speed of sound is based on the resonant frequency measured in the cell, properties of the selected gases, measured temperature, analysis pressure and some calibration factors. Since temperature and pressure affect each gas differently, it's important to correctly specify the gases being measured.

The normalized speed of sound is the measured speed of sound normalized to NTP (20°C, 1 atm). The normalized speed of sound is typically the most useful measurement, since it can be compared to data taken at other temperatures and pressures.

Note that both measured and NTP speeds of sound are reported at their measured frequency. This is distinct from the gases speed of sound in free space which may be slightly different. The BGA244 only uses the speeds of sound at the measured frequency.

The temperature is the measured gas temperature. The pressure is the analysis pressure.

Physical Measurements Analyzer Home Page




FIGURE 16: PHYSICAL MEASUREMENTS HOME PAGE

- Page Identification: Physical Measurements Analyzer.
- Speed of Sound Normalized to NTP (20°C, 1 atm).
- Measured Speed of Sound.
- Temperature: Measured gas temperature.
- Pressure: Measured or entered pressure used for analysis. See the *Pressure* section later in this chapter for more information.
- Units: See (Setup/Control Panel/Units).

The Upper and Lower Limits are not displayed in Measurement section since there is no graph, but they are active. See *Limits* (page 54) for more information.

Details

This is a description of some of the items specific to the Physical Measurements Instrument Mode.

Select gas

Both the Primary and Secondary gases must be selected before accurate measurements can be made. There are nearly 500 different gases contained in the BGA244 Factory Gas Table. Go to (Setup/Select Gas) to view the Gas Selection page. Here you can select the Primary and Secondary Gases by name, chemical formula or CAS#. See *Selecting Gases* (page 58) for more information.

The Primary and Secondary gas species do have an effect on the speed of sound measurements due to gas specific interactions with the BGA244. The Primary Gas is usually set to the dominant gas species. For pure gases, the Secondary Gas can be set to either a much lighter or much heavier gas (Hydrogen or SF₆ are good candidates). For mixed gases, the Secondary Gas should be set to the gas with the second highest concentration.

Screen Messages


There are a variety of operating conditions that can be reported by the BGA244 (display versions only). There are several groups of messages including Analysis, Entry, Fault and Alert messages.

There are two types of screen messages. Normal messages appear as yellow or red text on the screen overwriting part of the Home page. These are notifications only and are referred to as “Messages”. Other messages are dual purpose, being both a notification as well as a navigation button. Pressing them will take you to the relevant page where more details are available. These are referred to as “Message Buttons”.

Analysis Messages

Messages appear in the measurement section to indicate something is affecting analysis within the BGA244. These items include user settings that may halt analysis, loss of signal and other analysis errors. All of these messages appear as text in the numeric display portion of the display.

Note that these messages may briefly appear at power on. This does not indicate any problems with the unit or measurement.



Degas On

This indicates that the Degas heaters are on and measurements can't be made. To return to the normal operating state, turn off the Degas heaters. See *Heater* (page 78) for more details.

>102%, <-2%

>102% or <-2% (or their equivalent in ppm or fraction) indicate that the binary gas measurement is out of range for the selected gas combination. This can be caused by measurement inaccuracies in gases with similar speeds of sound, additional contaminating gas(es) in the cell or if the wrong gas(es) have been selected.

No Signal

This indicates that the BGA244 cannot recover the acoustic signal or identify a valid speed of sound for the gas. This most often occurs if the gas pressure in the cell is too low. See *Troubleshooting* (page 199) if this message is displayed with adequate gas pressure.

Invalid

This indicates that the BGA244 cannot report a valid result based on the measured speed of sound. Usually this only appears for a brief instant after a major disruption of the signal occurs. See *Troubleshooting* (page 199) if this message remains on for more than a few seconds.

Bad Pressure

This indicates that the measured pressure is invalid. It's usually caused by a problem with a pressure transducer or wiring connected to an Analog Input. See *Pressure* (page 65) for more information.

Condensation

This indicates that one or both of the gases is nearing its condensation point. When condensation occurs the BGA244 cannot correctly calculate the gas concentration, gas purity or normalized speed of sound. Refer to *Condensation* (page 95) for information on dealing with this problem.

Invalid Gas Message

This indicates that a recalled setting or configuration is referencing a nonexistent User gas. See *Gas Selection* (page 138) in Chapter 5 for more details.

Other Messages

Two Sol

Two Sol (Two Solutions) indicates there are two valid molar ratios for a given speed of sound. This can only occur for binary gas measurements. Concentration 1 is the smaller value and is displayed on top.

Concentration 2 is the larger value and is displayed on the bottom. This only occurs for a few gas pairs with similar masses, but different γ 's. Refer to *Gases* (page 93) for more information.



System Fault

This Message Button indicated that a serious problem has occurred with the BGA244. Pressing [SYSTEM FAULT] takes you to the Fault page. See *Faults* (page 85) for more details.



Alert

This Message Button indicates that non-critical problem has occurred. Pressing [ALERT] takes you to the Alert page. See *Alerts* (page 87) for more details.



Stopped

This indicates that the BGA244 has been placed in the STOP mode and no measurements are being made.

Pressing the [STOP] key will take you to the Setup page where the Run mode can be selected. Refer to RUN|STOP (page 58) for more details.



Invalid Entry

This message appears if an entered number is outside the allowable range. The minimum or maximum value for that entry is displayed.



LED Blink Codes

The Power and Error status LEDs are used to indicate the operating state of the BGA244. They can indicate power faults, system faults, analysis errors and communication errors using blink codes. These features are active on all units and are especially useful for units with the metal protective cover installed (Accessory BGA-M).. See Figure 1 (page 6) for the LED locations.

Normal Behavior

When power is first applied, the Power LED will light for ~ 1 second; followed by all 3 LEDs lighting for 5 seconds. At this time the Power LED should blink once indicating the code is loading properly then operational.

After this the Power LED remains on. The Comm (Communications) LED will flash during communication over any of the computer interfaces. The Error LED will remain off unless an error occurs.

ERROR LED Codes

These codes use the red ERROR LED to indicate various errors.

- System Fault: The ERROR LED is continuously on as long as the condition persists.
- Stopped, Degas, Invalid, Bad Pressure, Condensation or No Signal: The ERROR LED flashes at about 4 Hz as long as the condition persists.
- Communication Error: ERROR LED flashes once for each error.

POWER LED Codes

The BGA244 can indicate if the USB or +24 V power supplies are out of range using blink codes on the Power LED. Refer to *External Power Supply Faults* (page 85) for details. Note that the BGA244 must have a minimum power supply voltage to generate the blink codes.

- Active Power Supply Fault: The POWER LED will flash at about 4 Hz.

Minimum Operational Voltage: USB \geq 3.5V, +24 \geq 18 V

USB Fault: Voltage $<$ 4.5V or $>$ 5.3V

+24 V Fault: Voltage $<$ 20V or $>$ 26V

Controls

Each Home page has the same Control section. Control buttons are used to access menus within the user interface. Several functions can be directly accessed from the Home page while others are accessed through the Setup Menu.

Note that the Physical Measurements Home page omits the Meter Scale control button since there is no meter. It replaces it with a Pressure Meter button.



Limits

Process limits are used to indicate that the measurement is above or below a set value. Limits serve two purposes in the BGA244. They are displayed on the meter to provide visual feedback (Binary Gas and Purity Analyzers) and are used in Events to control external devices.

The [LIMITS] button on the Home page is a dual purpose control. At any time, pressing [LIMITS] takes you to the Process Limits Page where the limit values can be entered. Additionally, the [LIMIT] button will turn red if either the Upper or Lower Limit is exceeded.

Process Limits Page

The upper and lower limits are set in units of the measured value. Each limit can be separately enabled. The Hysteresis value provides a guard band around the limit thresholds to avoid oscillating between the limit & non-limit states. This is particularly helpful when using the Event Relays to stop relay chatter.

The Upper Limit must be greater than the Lower Limit and within the operating range of the measured parameter. If not, an “Invalid Entry” message will appear with guidelines for a legal value.

If the Upper Limit has been exceeded, the measurement must drop below {Upper Limit – Hysteresis} to deactivate the Upper Limit. Similarly if the Lower Limit has been exceeded, it must rise above {Lower Limit + Hysteresis} to deactivate the Lower Limit.


Choose a hysteresis value that is reasonable for your limits. A hysteresis value of 1% may never reset if the limits are set to 0.01%.

Example:

For an upper limit set at 90% and the hysteresis set at 5%. If the measurement increased above 90%, the Upper Limit would be active. When decreasing, the Upper Limit wouldn't deactivate until the measurement dropped below 85%.

Scale (Binary & Gas Purity)

The [SCALE] button on the Home page opens the Meter Scale page. The Upper and Lower scale values are set in units of the measured value for the Binary Gas Analysis and Gas Purity modes. This control doesn't appear for the Physical Measurements mode. The Upper Scale value must be greater than the Lower Scale value and within the operating range of the measured parameter. If not, an *Invalid Entry* message will appear with guidelines for a legal value.



Pressure (Physical Measurements)

The [PRESSURE METER] button on the Home page opens the Analysis Pressure Meter page. See *Pressure* later in this chapter for more details.

Help

[HELP] takes you to a Help screen for the particular page you are on. Each Help page describes the parameters and settings for that page. There may also be suggestions to resolve problems or errors found on that page.

Event Relay 1 & 2


The Event Relays can be used to control external devices that depend on different conditions in the BGA244. These conditions can be individually enabled and include things like exceeded limits, loss of the measurement signal and system faults. There are two independent events relays: Event Relay 1 and Event Relay 2. If any enabled conditions are true, the Event Relay becomes active. If +24 V_{DC} is not present, the Event Relays are not active.

The [EVENT RELAY] buttons are dual purpose controls. Pressing them takes you to the Event 1 or Event 2 configuration page where the event conditions are set. If an event is currently active, that event button will turn red as an indicator. If the Event Relay is in a Force On or Force Off state, that button will turn yellow to indicate that condition.

Event Relay Configuration Pages

Event 1 and Event 2 are set independently and have separate configuration pages. Any of the enabled conditions can activate an event. The following conditions can be selected. Use the [Page ↑] and [Page ↓] keys to navigate the Event list.

Any active condition in the Event list is displayed in red, whether it is enabled or not.



- Above Limit: Activates if the measured value is above the process upper limit. Setting the upper limit is described in *Limits* (page 54).

- Below Limit: Activates if the measured value is below the process lower limit. Setting the lower limit is described in *Limits* (page 54).
- Pressure Meter 1 Above Limit: Activates if Pressure Gauge 1 is above the upper pressure limit. This selection is only available if Pressure Gauge 1 has been configured. Configuring the pressure gauge and limits is described in *Pressure* (page 64).
- Pressure Meter 1 Below Limit: Activates if Pressure Gauge 1 is below the lower pressure limit. This selection is only available if Pressure Gauge 1 has been configured. Configuring the pressure gauge and limits is described in *Pressure* (page 64).
- Pressure Meter 2 Above Limit: Activates if Pressure Gauge 2 is above the upper pressure limit. This selection is only available if Pressure Gauge 2 has been configured. Configuring the pressure gauge and limits is described in *Pressure* (page 64).
- Pressure Meter 2 Below Limit: Activates if Pressure Gauge 2 is below the lower pressure limit. This selection is only available if Pressure Gauge 1 has been configured. Configuring the pressure gauge and limits is described in *Pressure* (page 64).
- Temperature Above Limit: Activates if the Temperature Gauge is above the upper temperature limit. Configuring the Temperature gauge and limits is described in *Temperature* (page 67).
- Temperature Below Limit: Activates if the Temperature Gauge is below the lower temperature limit. Configuring the Temperature gauge and limits is described in *Temperature* (page 67).
- No Measurement: Activates if the BGA244 cannot report a valid measurement. Most of the causes of this are indicated as an Analysis Screen Message. See *Troubleshooting* (page 199) for possible causes.
- System Fault: Activates if there is a serious fault with the BGA244. See *Faults* (page 85) for a list of possible System Faults.

Force On, Force Off Buttons

There are two additional buttons that manually control the Event Relays regardless of any event conditions. [FORCE ON] sets the relay to the ON state. [FORCE OFF] sets the relay to the OFF state. The selected Force button will turn yellow, as will the Event button on the Home page to indicate the force condition. Press the button a second time to release the force condition.

Event Relay Connections

Each Event Relay has three pins: normally open (NO), normally closed (NC) and common (COM). In the OFF state, NC is connected to COM and NO is open. In the ON state, NO is connected to COM and NC is open. The relays are in the OFF state when the BGA244 is powered off.

See *Event Relays* (Page 30) for details on wiring the relays to outside circuits. See the *Specifications* section for details on relay ratings and life time.

Using Event Relays

Event Relays are normally used to signal an out of range or error condition. They could be connected to a bell or horn to notify an operator of an out of range condition, or connected to a valve to increase the flow of a gas.

A common set up for Event 1 and Event 2 would be as follows.

- Event 1: Above Upper Limit, Below Lower Limit
- Event 2: System Fault, No Signal

The Event 1 Relay indicates if the measured value was above or below the set limits and Event Relay 2 would indicate that there was a problem with the measurement.


Setup

[SETUP] takes you to the Instrument Setup page which is described in the following section. From here you can access the remaining controls and settings of the BGA244.

Setup

The Setup menu accesses all the BGA244 controls that aren't directly available on the Home page. Press [SETUP] on the Home page to reach the Instrument Setup page.

The most important functions in the BGA244 can be directly accessed from the Setup page. Less commonly used functions are reached via the Control Panel. See the *Map of Interface Functions* (page 36) for the locations of different function controls.



Run|Stop

It's occasionally necessary to stop analysis and output updates. For example, you may not want alarm relays to activate or analog outputs to pin while changing gas cylinders or performing other system maintenance. Press [RUN] or [STOP] to switch between the run and stop states. When the BGA244 is stopped, a [STOPPED] message button will appear on the Home Page. Pressing this button takes you to the Setup Page where the [RUN|STOP] control is located.

When the BGA244 is stopped the following functions do not update:


- Gas Ratio, Gas Purity, Speeds of Sound
- Event Relays
- Analog Outputs
- Temperature Readings

Instrument Mode

The BGA244 can operate in one of three Instrument Modes: Binary Gas Analyzer, Gas Purity Analyzer and Physical Measurement. Each Instrument Mode has a unique Home page and measurement specific parameters (gas selection, limits, and meter scale).

Select Instrument Mode

Press [INSTRUMENT MODE ↓] to open the Instrument Mode list. Press the desired mode to select it. The currently selected mode is highlighted in yellow.



Selecting Gases

Measurements made with the BGA244 require selecting the gases to be measured. Two gases must always be entered: however the Gas Purity Analyzer only uses a single gas for its measurements.

The BGA244 Factory Gas Table contains nearly 500 different gases. Each entry includes the formula, common name, CAS # and up to two alternate names. All of these are scanned when selecting a gas. Besides naming information, an array of physical properties is stored with each gas for calculating gas ratios and speeds of sound. Appendix A contains a list of Factory Gas Table gases and a description of all of the gas properties.

Additional gases can be added into the User Gas Table. These can include mixtures or blended gases, made up of 2 or more single species gases. See *User Gases* (page 117) for more information.


To view the currently selected gases or to change gases go to (Setup/Select Gases). This page displays the information about the currently selected gases and has controls to change the gases.

The following information is displayed for each gas in the Factory Gas Table:

- Gas ID (Primary/Secondary/Reference)
- Chemical formula: The simple (non-structural) formula of the gas
- Common Name of the gas
- CAS # The CAS Number is a unique identifier for every chemical substance described in open scientific literature.
- MW (Molar Weight) of the gas molecule in AMUs (Binary Gas Analyzer and Physical Measurements)
- γ (gamma): The ratio of specific heat capacities (Binary Gas Analyzer and Physical Measurements)
- Speed of Sound at NTP (Gas Purity Analyzer only)

How to select gases

Gases are selected at (Setup/Select Gases). Press [SELECT PRIMARY] or [SELECT SECONDARY] to open the alpha-numeric keypad.



Begin typing the name, formula or CAS # of the desired gas. As you enter characters, the BGA scans the entire gas Table for matches to the entered text and displays that number in the “Matches” field. If no gases match the entered text it displays “Matches: 0”. If you press an incorrect key, press [←] to erase the last character or [CLR] to erase all characters.



When the number of matches is down to a manageable number (< 20 or so), press [ENT] or [ENTER]. This will open up a list of all of the matches from the gas table. Note that matches can occur at any position in the gas name, formula or CAS #. If there are more than 7 matches, use [Page ↑] and [Page ↓] to view the entire list. If you don't see the expected gas, press [←] to return to the alphanumeric entry window.



Select the desired gas by pressing its entry. This will enter the selected gas and return to the Select Gases page. If you wish to change the gas, press the desired gas selection key and re-enter the gas.


When selecting gases, User Gas Table gases will appear as “User/gas_name” to differentiate from a gas in the Factory Gas Table.

Swap Gases

For the Binary Gas Analyzer Mode, results are displayed for the concentration of the Primary Gas relative to the total mixture. If you want to view the concentration of Secondary Gas, press [SWAP GASES] which will exchange the two entries. This key will also swap the Reference Gas for the Secondary Gas in the Gas Purity Analyzer. It will also swap gases for the Physical Measurements mode, but this ordinarily has no effect on measurements.

Reference Gas (Gas Purity Analyzer)

For the Gas Purity Analyzer, the Primary Gas is referred to as the Reference Gas and the Secondary Gas is grayed out to indicate that only the Reference Gas is being used for this measurement.



REL

The BGA244 provides excellent accuracy using its internal calibration and its stored thermodynamic data. But there are circumstances where measurement accuracy can be enhanced using reference gases. Typical uses for the REL function include:

- Setting a carrier gas to 0 or 100% when working with small concentrations of a dopant gas.
- Zeroing readings to a “reference mixture” and measuring the deviation from that value.
- Using a REL value to compensate for an unknown pressure.
- Using REL to measure deviations from a “reference point” when complete thermodynamic data isn’t available for a User gas.
- Compensating for static impurities within gases.
- Creating compensation curves for complicated mixtures.

For Binary mode, there are three different REL functions: 1 Point REL, 2 Point REL and Multi Point REL. For the Purity Mode, only 1 Point REL is active. When active, “REL”, “2REL” or “MREL” indicators are displayed below the measurement units as shown on the Home pages.

Note: The REL function is applied to the value of the Primary Gas in the Binary Gas Analyzer Mode.

The 1 Point REL function performs the following operation:

$$\text{displayed value} = \text{measured value} - \text{Rel Value}$$

The 2 Point REL function uses a two point slope/offset equation to adjust the displayed value.

Multipoint REL uses multiple calibration points (up to 20). Measured values will be adjusted using linear interpolation between calibration points. Values above or below the highest and lowest calibration points will be calculated based on the slopes of the adjacent points.


In general, REL works best when operating near the conditions that the REL was performed at. Large deviations in temperature, pressure, flow rate or concentration may make the technique less effective. Situations and techniques where the REL function can improve measurement accuracy are described in the Chapter 4: *Application Guide*.

Gas Purity Mode

Pressing [REL] will take you to the Purity Mode Measurement REL page. Press Use REL or No REL to select or deselect the REL function.

Press [REL TO 0%] to set the measured value to 0%. (or 0 ppm or 0.0 fraction). The value required to force the displayed value to 0% is displayed in the [Rel Value] button in global ratio units.


You can also directly enter the Rel Value between -110% and +210% (or the equivalent in ppm or fraction). This value will be subtracted from the measured value to produce the displayed value.



Binary Gas Mode

At most, only one of the REL methods can be active at a time. However, all parameters can to be set at any time.


Pressing the REL button on the Setup page will take you to the Select REL page, which has 4 radio buttons and 3 navigation keys. The radio buttons select which REL mode is active, or "None" if none are active. Pressing the navigation keys takes you to the entry pages where the different parameters are set.



1 Point REL

From the select Rel box, press [1 Point REL] to enter the single point REL entry box.

Press [REL TO 100%] to set the measured value to 100% (or 1000000 ppm or 1.0 fraction). The value required to force the displayed value to 100% is displayed in the [Rel Value] button in global ratio units.



Press [REL TO 0%] to set the measured value to 0% (or 0 ppm or 0.0 fraction). The value required to force the displayed value to 0% is displayed in the [Rel Value] button in global ratio units.

You can also directly enter the Rel Value between -110% and +210% (or the equivalent in ppm or fraction). This value will be subtracted from the measured value to produce the displayed value.

2 Point REL

From the select Rel box, press [2 Point REL] to enter the 2 Point REL entry box.

The 2 Point REL page consists of 2 pairs of Raw/REL values. Raw values are the measurements from the BGA244 with no REL active. REL values are the expected value for a given Raw value.

Two Point Rel	
Point 1	
Raw Value	REL Value
98.000 %	100.000 %
Point 2	
Raw Value	REL Value
2.000 %	0.000 %

Clicking on either pair of values will open the EDIT Raw/REL value page for that point, where the values can be changed.

Click on either the Raw or Rel values to open the entry screen and then enter the appropriate values. Click [DONE] to load the values & return to the previous page or ← to return to the previous screen.

Note: Any measurements used to calculate the Raw/REL values must be made with the REL Mode set to None.

Edit Raw/REL Values

Raw Value

98.000 %

Rel Value

100.000 %

DONE

Multipoint REL

From the select Rel box, press [Multi Point REL] to open a list of Raw/REL values.

The Multi Point REL page consists of a list of Raw/REL value pairs. There can be up to 20 value pairs. Use PAGE up and PAGE down to navigate the entire list. While there can be duplicate REL values, each Raw value must be unique.

Multi Point Rel - Page 1	
Raw Value	REL Value
0.000 %	0.000 %
75.000 %	0.000 %
76.000 %	1.000 %
77.000 %	2.000 %
80.000 %	5.000 %
83.000 %	8.000 %
85.000 %	10.000 %

←
PAGE ↑
PAGE ↓
ADD
EDIT
DELETE

Existing entries are selected by clicking on one of the Raw/REL value pair, which will highlight that pair. Clicking on the value a second time, deselects the pair.

To add a new value to the table, press the ADD key, which will open the EDIT Raw/REL value box (described previously). When DONE is pressed (assuming a valid value), the new value is added to the list and the list is sorted in order of ascending Raw values.

To edit a value, select the desired Raw/REL value, then press EDIT. This will open the same Edit Raw/REL value box used for 2 Point REL for that selection. When DONE is pressed (assuming a valid value), the new value is entered to the list and the list is re-sorted in order of ascending Raw values.

To delete a value, select the desired Raw/REL value, then press DELETE. A warning “Delete Multi Point Entry Are you sure?” with Yes/Cancel buttons will appear. Clicking Yes will delete that entry and re-sort the list in order of ascending Raw values. Clicking No will revert to the list window without deleting the entry.

Pressure

The speed of sound in an ideal gas is independent of pressure. But the speed of sound in real gases does depend on pressure. For many gas(es) satisfactory measurements can be made without knowing the exact pressure. However to meet its specified accuracy for all gases, the BGA244 must know the gas pressure to within ± 1 psi (6.9 kPa). See the *Applications Guide* for information on the effects of pressure on measurement accuracy.

Pressure is the only external parameter that needs to be input to the BGA244 besides the gas species. It can be directly entered using the keypad or computer interface. Or a pressure transducer can be integrated with the BGA244 using one of the analog inputs.

Direct entry of the pressure is well suited for processes that operate at a known, fairly constant pressure. Typically, the pressure is known by the system design or by monitoring an external gauge. This is the only method available if the unit is powered over USB.

Other processes may experience pressure variations over time or operating conditions. These applications will benefit from integrating a pressure transducer using one of the BGA244's analog inputs. This provides a simple, integrated solution to provide the best accuracy across a range of operating conditions. Using an external pressure transducer requires +24 V_{DC}. Refer to *Pressure Transducers* (page 31) for information on connecting a pressure transducer to the BGA244.


Gauge Pressure vs Absolute Pressure

The BGA244 uses the absolute pressure of the gas as part of its calculations. However pressure can be input in either absolute pressure units or in gauge pressure units combined with the ambient pressure. This is true for both direct entry of User pressure or when integrating a pressure transducer.

Absolute pressure is relative to vacuum while gauge pressure is relative to the ambient pressure outside the gauge. The correct pressure units (gauge or absolute) must be selected in the BGA244 to avoid errors. If gauge pressure units are selected, the ambient pressure must also be entered. Refer to *Pressure* (page 98) for details.


Hint: If a pressure gauge or transducer can report negative pressures, it is reporting the gauge pressure. If it reports only positive pressures it can be either absolute or gauge. This information should be listed on the pressure gauge specifications.

Press [PRESSURE] to access the pressure gauge setup menu. Here you can view the current pressure or select the pressure entry method for analysis and configure the external pressure gauge if needed. Pressure units are set on the Units page (Setup/Control Panel/Units section). Allowable units include psi, atm, bar, Pa, mmHg and torr.



Selecting Analysis Pressure


Press <Use for Analysis ↓> to select the pressure entry method. There are 3 choices: User Entered, Analog Input 1 and Analog Input 2. The currently select method is highlighted in yellow. The Analog Inputs will be grayed out unless they have previously been configured as pressure gauges.



User Pressure


Press [ENTER USER PRESSURE] to manually enter the Pressure. Pressure can be entered in absolute pressure units (relative to vacuum) or in gauge pressure units (relative to ambient pressure). If you select gauge pressure units you must enter the ambient pressure.

Example: The display to the right shows a gauge pressure of 0 psi, with an ambient pressure of 14.7 psi.



Configure Pressure Gauge

Press [CONFIGURE INPUT 1|2] to open one of the Configure Pressure Gauge pages. Typically only one of the two inputs is configured as a pressure gauge, but both selected can be if desired. Check the “Use as Pressure Gauge” box and select absolute or gauge pressure units. Be sure to enter the ambient pressure if using Gauge Pressure. Verify that the analog input is enabled and check that the Min and Max input format matches your transducer (V vs mA). If not, configure the input as described below.



Assuming the formats match, enter the pressure values for the minimum and maximum analog input values. If the analog output of your pressure gauge doesn't match the span of the analog input, calculate what the pressure would be at those points.

Example:

Your pressure transducer outputs 0 - 5 V_{DC} for 0 - 50 psi


Set the following

Min (0V) = 0.00 psi


Max (10V) = 100.00 psi

Configure Input

Enable the Analog Input. Press [CONFIG INPUT] to open the Configure Analog Input page. Press <Type ↓> to select the appropriate input type: Voltage, Current and Current w/ Loop Power. The currently select method is highlighted in yellow.




Certain pressure transducers can operate using loop power provided by the BGA244. This is the simplest type of transducer to integrate since it only requires 2 wires. See *Analog Inputs* (page 81) for details on different input types. If Current with Loop Power is selected, enter the appropriate loop power supply voltage.



An improperly configured analog input, broken pressure transducer or wiring can lead to out of range pressure readings. If this occurs, a “BAD PRESSURE” screen message will appear in place of the measurement and no results will be reported. There are also several alerts that can occur for the analog inputs. These normally indicate a problem with the device connected to the Analog Input or wiring, rather than the input itself. See *Analog Inputs* (page 81) for details.

Pressure Meter

There are pressure meter pages associated with each analog input when it's configured as a pressure gauge. Press [METER] on either the Pressure Gauge Setup or Configure Pressure Meter pages to view the pressure meter associated with that input. If an analog input is not configured as a pressure gauge that selection will be grayed out.




The Pressure Gauges each have a meter scale and limits similar to the Home page displays. Press [METER SCALE] to set the graph scale and [LIMITS] to set the Pressure Limit values. These are the over and under pressure limit values used on the Event Relay page.

Temperature

The cell temperature can also be displayed on a meter with limits and a scale. Press [TEMP] from the Setup page to view the temperature meter.

The Temperature Gauge has a meter scale and limits similar to the Home page displays. Press [METER SCALE] to set the gauge scale and [LIMITS] to set the Temperature Limit values. These are the over and under values limit used on the Event Relay page.



Store|Recall

The BGA244 can save and recall up to 20 different setups. Each setup contains all parameters that can be set on the BGA244; including the instrument mode, gas selection, display setup and I/O parameters. Stored setups can be given an alpha numeric name, making it easy to identify details about each setup. Press [STORE-RECALL] to access the Store Recall menu.



If a stored setup is over written or erased, it cannot be retrieved.

There are two setups that cannot be changed or deleted. The *Default Setup* configures the BGA244 as described in the Default Setup Table (page 69). The *Factory Setup* places the BGA244 into the same configuration as when it left the factory. See *Factory Setup* (page 72) for details.

Note: Setups are stored into Flash memory with a life time of about 100,000 erase cycles. Don't continuously store settings especially over the computer interfaces to avoid wearing out the memory.

If storing, recalling or erasing a setup fails once, try a second time. If it fails repeatedly, try a different location. This failure indicates a problem with the storage memory. See *Troubleshooting* (page 199) for more information.

Setup List

Pressing [STORE], [RECALL] or [ERASE] will open the Stored Setup list. This list contains the Default Setup and 20 available setup locations and their names. Use [Page ↑] and [Page ↓] to navigate the entire list. Filled setup locations are named; if no alpha numeric name was entered, the name "Setup x" is used ("x" is the setup location). Empty setups are denoted by "**Empty**".



Store

Press [STORE] to open the Setup list.

Press the desired setup. If the selected location is currently filled, a prompt will appear asking if you want to continue. You cannot store a setup to Location 0 (Default Setup)

Enter the setup name using the alpha-numeric keypad. If you selected a currently filled location a prompt will appear asking if you want to change the setup name.

A confirmation prompt will appear containing the setup name and if the setup was stored successfully.

When entering a name for a setup, try to be descriptive. For example, "ARGON IN OXYGEN, 1%" could indicate Primary Gas = Argon, Secondary Gas = Oxygen, full scale output = 1%. This makes it easier to identify each setup.

Recall

Press [RECALL] to open the Setup list. Note that Setup 0 is the Default setup.

Press the desired setup.

A confirmation prompt will appear containing the setup name and if the setup was recalled successfully.



Erase

Press [ERASE] to open the Setup list.

Press the desired setup. A prompt will appear asking if you want to erase the selected setup.

A confirmation prompt will appear if the setup was successfully erased.



Default Setup

TABLE 9: ACTIVE PARAMETERS

Parameter	Value
Instrument Mode	Gas Purity Analyzer
Run Mode	Run
Gas Settings	
Primary/Reference Gas	Argon
Secondary Gas	Hydrogen
Pressure Settings	
Analysis Pressure	User
User Pressure	0 psig
User Pressure Units	Gauge
Ambient Pressure	14.7 psi
Meter Settings	
Purity Upper Scale	5%
Purity Lower Scale	-5%
Purity Upper Limit	2%
Purity Lower Limit	-2%
Purity Upper Limit Active	Active
Purity Lower Limit Active	Active
Purity Hysteresis	0.1 %
Purity REL Mode	No REL
Purity REL Value	0 %
Event Settings *	
Event 1 Configuration	Nothing Selected
Event 2 Configuration	Nothing Selected
Event 1 Force	No Force
Event 2 Force	No Force
Temperature Meter Settings	
Temperature Upper Scale	50 °C
Temperature Lower Scale	10 °C
Temperature Upper Limit	40 °C
Temperature Lower Limit	20 °C
Temperature Upper Limit Active	Active
Temperature Lower Limit Active	Active
Temperature Hysteresis	1 °C
Display Settings	
Display Orientation	0 degrees
Display Backlight	10
Display Key Click	On
Units Settings	
Ratio Units	%
Speed Units	m/s
Temperature Units	°C
Pressure Units	Psi
Measurement Settings	
Average Mode	On
# Averages	10
Amplitude	Auto

Parameter	Value
BGA Concentration	Mole Fraction
Relaxation Correction	On
Heater Settings *	
Block Heater Mode	Off
Heater Temperature	40 °C
Max Current	2.0 A
Degas Heater	Off
Computer I/O Settings	
RS-232	Enabled
Baud Rate	9600
RS-422 *	Enabled
Baud Rate *	9600
Analog Inputs Settings *	
Analog Input 1	Enabled
Analog Input 1 Type	Voltage
Analog Input 1 Loop Power Voltage	12 V
Analog Input 2	Enabled
Analog Input 2 Type	Voltage
Analog Input 2 Loop Power Voltage	12 V
Analog Output 1 Settings *	
Analog Output 1	Enabled
Analog Output 1 Type	0 – 10 V
Analog Input 1 Linked to	USER
Analog Output 1 User Value	0 V
Analog Output 2 Settings *	
Analog Output 2	Enabled
Analog Output 2 Type	0 – 10 V
Analog Output 2 Linked to	USER
Analog Output 2 User Value	0 V
Measure Output Settings *	
Measure Output	Enabled
Measure Output Type	0 – 10 V
Measure Output Linked to	Purity Deviation
Measure Out Purity Scale Min	-10 %
Measure Out Purity Scale Max	+10 %
Misc. Settings	
Alerts	All Enabled
Use Password	Off
Password	0000

* Settings marked with * are only operational when 24 V_{DC} is connected to the unit. They can be configured even if 24 V_{DC} is not present.

TABLE 10: INACTIVE PARAMETERS

Parameter	Value
Binary Gas Analyzer Settings	
BGA Upper Scale	105%
BGA Lower Scale	95%
BGA Upper Limit	102%
BGA Lower Limit	98%
BGA Upper Limit Active	Active
BGA Lower Limit Active	Active
BGA Hysteresis	1 %
BGA REL Mode	No REL
BGA REL Value	0 %
Physical Measurement Settings	
Physical Upper Limit	1000 m/s
Physical Lower Limit	0 m/s
Purity Upper Limit Active	Inactive
Purity Lower Limit Active	Inactive
Purity Hysteresis	1 m/s
Pressure Meter 1 Settings *	
Use Analog In 1 as Pressure Gauge 1	Inactive
Gauge 1 Units	Gauge
Gauge 1 Min	0 psi
Gauge 1 Max	150 psi
Gauge 1 Upper Scale	150 psi
Gauge 1 Lower Scale	0 psi
Gauge 1 Upper Limit	100 psi
Gauge 1 Lower Limit	10 psi
Gauge 1 Upper Limit Active	Inactive
Gauge 1 Lower Limit Active	Inactive
Gauge 1 Hysteresis	1 psi
Pressure Meter 2 Settings *	
Use Analog In 2 as Pressure Gauge 2	Inactive
Gauge 2 Units	Gauge
Gauge 2 Min	0 psi
Gauge 2 Max	150 psi
Gauge 2 Upper Scale	150 psi
Gauge 2 Lower Scale	0 psi
Gauge 2 Upper Limit	100 psi
Gauge 2 Lower Limit	10 psi
Gauge 2 Upper Limit Active	Inactive
Gauge 2 Lower Limit Active	Inactive
Gauge 2 Hysteresis	1 psi
Analog Output 1 Settings *	
Analog Output 1 Speed Scale Min	0 m/s
Analog Output 1 Speed Scale Max	2000 m/s
Analog Output 1 Speed NTP Scale Min	0 m/s
Analog Output 1 Speed NTP Scale Max	2000 m/s
Analog Output 1 Temp Scale Min	0 °C
Analog Output 1 Temp Scale Max	50 °C
Analog Output 1 Press 1 Scale Min	0 psi
Analog Output 1 Press 1 Scale Max	150 psi

Parameter	Value
Analog Output 1 Press 2 Scale Min	0 psi
Analog Output 1 Press 2 Scale Max	150 psi
Analog Output 2 Settings *	
Analog Output 2 Speed Scale Min	0 m/s
Analog Output 2 Speed Scale Max	2000 m/s
Analog Output 2 Speed NTP Scale Min	0 m/s
Analog Output 2 Speed NTP Scale Max	2000 m/s
Analog Output 2 Temp Scale Min	0 °C
Analog Output 2 Temp Scale Max	50 °C
Analog Output 2 Press 1 Scale Min	0 psi
Analog Output 2 Press 1 Scale Max	150 psi
Analog Output 2 Press 2 Scale Min	0 psi
Analog Output 2 Press 2 Scale Max	150 psi
Measure Out Settings *	
Measure Output Linked to	Gas Concentration 1
Measure Out BGA Scale Min	0 %
Measure Out BGA Scale Max	+100 %
Measure Out Physical Scale Min	200 m/s
Measure Out Physical Scale Max	2000 m/s

* Settings marked with * are only operational when 24 V_{DC} is connected to the unit. They can be configured even if 24 V_{DC} is not present.

Factory Setup



Caution: Loading the Factory Settings will erase all stored setups as well as the User Gas Table. This procedure cannot be “Un-done” so be certain that you want to do it before starting.

The Factory Setup is used to set the BGA244 to the condition it left the factory. This involves:

- Erase Stored Settings 1-20
- Delete the User Gas Table
- Recall Setup 0 (Default Setup)

Press [FACTORY] to load the factory setup. A prompt will appear asking if you want to recall the Factory Setup. Press [YES] to continue or [CANCEL] if you don’t want to recall the Factory Setup.



A second prompt will appear to confirm that you really want to restore the Factory Setup. Press [YES] to continue or [CANCEL] if you don’t want to recall the Factory Setup.



A confirmation prompt will appear to confirm that the factory setup was recalled successfully. Press [OK] to continue.



Control Panel

The remainder of the controls and measurements in the BGA244 are accessed through the Control Panel. Press [CONTROL PANEL] to access this menu. Use [PAGE ↑] or [PAGE ↓] to navigate the different functions. Press the Function name to enter that page. The Control Panel menu includes:

Display
Units
Measurement
Status
Heater
Computer I/O
Analog I/O
Faults
Alerts
Message Log
Self Test
Password
About

Display

Since the BGA244 can be mounted in any orientation, the display needs to be rotated for proper viewing. The back light intensity and key clicks can be set. In addition, the screen coordinates can be calibrated. Press [DISPLAY] to access the Display controls.

Display Orientation


The display can be viewed from 4 different orientations, indicated by the four arrow buttons. Press whichever arrow points up to rotate the display where that direction is up.

Backlight Intensity

The display backlight intensity can be varied from full on (10) to nearly off (1). In high brightness conditions the backlight should always be set to full intensity.

Reducing the backlight intensity minimize the power supply current. Setting it to the minimum reduces the current by nearly 100 mA. This is really only useful when operating the BGA244 off of poor quality USB power and isn't commonly done.

Press [↑] to increase the back light intensity. Press [↓] to decrease the back light intensity.



Key Click

The key click can be turned on and off. They should normally be left on to provide feedback that a key was pressed but can be turned off if necessary.

Press Key Click [ON] or [OFF] to turn the Key Click sound on and off. When turning the key click off you will hear the final click as the button is pressed. When turning it on there is a longer beep.

Screen Calibration


The resistive touch screen used in the BGA244 can drift over time. This can cause the active location of buttons and controls to not align correctly with the graphics. This drift is normally quite slow, but large temperature changes can accelerate it. If the key push location isn't aligned well with the screen graphics, use the following procedure to calibrate the screen.

- Press [CAL SCREEN] to begin the screen calibration process.
- Touch the center of each crosshair as prompted.
- When the calibration is complete the BGA244 will return to the Home page.

You can also enter the screen calibration routine by pressing a finger to the screen when power is applied. Be sure to continue holding your finger down until the message "Release Screen to start Screen Calibration" appears, then follow the instructions to complete the routine.

Units


The BGA244 displays several different types of measurements including ratios, speeds of sound, temperatures and pressures. Each measurement can have one of several different units.



Ratio

Ratio units are used in measurements that are ratios of one quantity to another. They are used for gas concentration (Binary Gas Analyzer) and ratio of speeds of sound (Gas Purity Analyzer).


Allowable units are: fraction (0 to ± 1.0), % (percent) and ppm (part-per-million). Press <Ratio ↓> to open the selection list. The currently selected unit will be highlighted in yellow. Press the desired value to select it.



Speed

Speed units are used for speeds of sound measurements. They are primarily used on the Physical Measurements page.


Allowable units are: m/s, kph (kilometers/hour) and mph (miles/hour). Press <Speed ↓> to open the selection list. The currently selected unit will be highlighted in yellow. Press the desired value to select it.



Temperature

Temperature units are used for gas temperature readings and heater settings.

Allowable units are: °C, °K and °F. Press <Temperature ↓> to open the selection list. The currently selected unit will be highlighted in yellow. Press the desired value to select it.



Pressure

Pressure units are used for gas pressure readings.


Allowable units are: psi, atm, bar, Pa, mm/Hg and torr. Press <Pressure ↓> to open the selection list. The currently selected unit will be highlighted in yellow. Press the desired value to select it.

Note that for pressure setting and entry the pressure units are defined as either absolute units (relative to vacuum) or gauge units (relative to atmospheric pressure). This selection is made on the Configure Pressure Gauge or Configure User Pressure pages. See the *Pressure* section earlier in the chapter for more information.



Measurement

The default configuration for the BGA244 is optimized for a wide range of gases, pressures, temperatures, flow rates and other environmental conditions. However there are circumstances when measurement and analysis can be tuned for better performance. Refer to Chapter 4: *Application Guide* for conditions when optimizing these parameters may be useful.



Averaging

Averaging successive measurements can improve accuracy and repeatability of the measurements by reducing interfering signals and noise. See *Using Averaging* (page 104) for more information. The BGA244 takes a measurement every 228 ms. This is quite a bit faster than most processes change, so a moderate number of averages doesn't create a noticeable delay.

The averaged output is continuously calculated, producing a new averaged measurement for every input measurement following the formula:

$$\text{Average}_N = [1/N * \text{Measure}_N] + [(1 - 1/N) * \text{Average}_{N-1}]$$

For signals without large transients, the averaged value settles to within 1% of its final value in about $5 * N$ measurements or $N * 1.14$ seconds. For the default value of 10 averages the averaged output settles in about 11.4 seconds.

Large transients may persist in averaged measurements for a long time. They may take several times as long to decay as the normal settling time. Similarly, any changes in a measurement will only be visible if they last for a sufficient number of samples.

Press [OFF] or [ON] to turn averaging on and off. Enter the number of averages between 2 and 100.

Amplitude

The amplitude of the speaker signal is optimized for the best signal to noise ratio of measurements within the acoustic cell. In rare cases this amplitude may cause the acoustic signal to over or under load the preamplifier. The BGA244 detects this and adjusts the amplitude accordingly. When this occurs there is a brief period when measurements are invalid until the acoustic signal settles to the new value, reported by an Invalid screen message.


Ordinarily the amplitude should be left in the default auto setting. If numerous Invalid screen messages due to amplitude changes appear, it may be helpful to set the amplitude to a lower fixed value. View the message log to see if an Invalid message is due to an amplitude change.

Press <Amplitude ↓> to open the selection list. The allowed selections are: auto, full, ½ and ¼ full scale. The currently selected method will be highlighted in yellow. Press the desired value to select it.



Binary Gas Concentration

Binary gas measurements can be reported in one of two formats: mole fraction and mass fraction. See *Converting Mole Fraction* (page 109) for a detailed description of the relationship between them.



Press <Binary Gas Measurement ↓> to open the selection list.

The currently selected method will be highlighted in yellow. Press the desired value to select it.

Relaxation Correction

Gases take different amount of times to reach thermodynamic equilibrium between their translational, rotational and vibrational degrees of freedom. If that amount of time is comparable with an acoustic cycle of the resonant mode, a small correction factor is applied to the applied to the heat capacity used to compute the speed of sound. Most gases don't exhibit this effect. For the few gases that do, the effect is small

and corrected. However in some cases, carbon dioxide (CO₂) can exhibit varying amounts of this effect depending on other gas species present.

Relaxation Correction should be left in the default setting (on) for nearly all cases. See *Relaxation Effects* (page 94) for details on when this should be turned off. Press [ON] or [OFF] to turn relaxation correction on and off.

Status

This page lists the present value of most of the parameters measured by the BGA244. If a parameter isn't currently configured in the BGA244 it will be dashed out (-----). Use [PAGE ↑] and [PAGE ↓] to navigate the list of measured parameters.

Status - Page 2	
Speed of Sound	325.132 m/s
Speed of Sound (NTP) (BGA)	319.550 m/s
Relay 1	OPEN
Relay 2	OPEN
Pressure (Ana. In 1)	21.99 psi (a)
Pressure (Ana. In 2)	-----
Pressure (User Ent.)	0.00 psi (g)
Analog Out 1 Status	OK

List of Monitored Parameters

- Speed of Sound (Measured)
- Speed of Sound (NTP)
- Gas Temperature
- Analysis Pressure
- Block Temperature
- PCB Temperature
- Heater Power
- USB Voltage
- External Voltage
- Analog In 1 Value
- Analog In 2 Value
- Measure Out Value
- Analog Out 1 Value
- Analog Out 2 Value

Heater

The BGA244 heaters serve several purposes. They are used to stabilize the measured gas temperature, to prevent condensation and to help with high purity bake out. There are two separate heaters: the Block Heater and the Degas Heater. Information on settling time and specific uses of the heaters is available in the Chapter 4: *Applications Guide*.

Note: The Gas temperature will usually not match the block temperature due to heat flow in the gas and acoustic cell. In addition, the Block and Set temperatures may differ by a few degrees. This is normal for the BGA244.

The Heaters can only be used if an external +24 VDC power supply is connected to the BGA244. The maximum Block Heater current drawn from the external supply can be set between 50 mA and 2.2 A. The current draw of the Degas Heater is fixed at about 140 mA.

The Block Heater can be used during either analysis or bake out. The Degas Heater cannot be used during analysis. If the Degas Heater is active, a “DEGAS ON” message is displayed on the Home page. See *Screen Messages* (page 50) for more information.

Block Heater

The Block Heater can apply over 50W of power to heat the cell temperature to as high as 70 °C. The Block Heater is over-temperature and over-current protected.

The Block Heater uses a temperature servo to maintain the temperature of the cell to the set point. It will hold the temperature stable assuming the following conditions are true.

- The Block Heater temperature set point is at least 3 degrees above the highest ambient temperature.
- The Maximum Heater Current is large enough so the heater can raise the cell temperature up to the Block Heater set temperature for the lowest ambient temperature.

Press Heater [ON] or [OFF] to turn the Block Heater on and off. Enter the Heater Temperature between -30 and 70 °C. Enter the Maximum Current between 0.05 and 2.2A.

The endplate temperature and power being applied to the BGA244 is reported. The Regulating Indicator shows when the temperature servo is actively regulating the temperature. A Heater Regulation alert will appear if the heater is not regulating. See *General Alerts* (page 88) for details.

Note: It is normal for the [Heater Regulation] alert to appear when the block heater is first turned on or for large change in the ambient or set temperatures.

Heater	
Heater	OFF ON
Heater Temp.	28 C
Max Current	2.00 A
Power	9.41 W
Block Temp.	28.185 °C
Regulation	REGULATING
Degas	OFF ON

This can last for a considerable period of time, especially when the set temperature is reduced by more than a few degrees.

If the heater cannot regulate after several minutes, try the following:

- If the cell temperature is above the set temperature, increase the set temperature. If the maximum set temperature is exceeded (70 °C), it may be necessary to either allow the BGA244 to operate without temperature regulation or to reduce the temperature of the gas or the environment. Remember to not exceed the maximum operating temperature (70 °C).
- If the cell temperature goes below the set temperature, increase the maximum heater current. If the maximum allowable heater current cannot raise the temperature to the required temperature, the operating environment temperature needs to be increased. A simple way to do this is to add insulation around the BGA244.

Degas Heater

The Degas Heater can only be used during system bake out. It is normally used in conjunction with the Block Heater to bake out the BGA244 cell for use high purity systems. The Degas heaters operate by heating the speaker and microphone Kapton membranes directly. They each draw about 70 mA and provide about 0.25 W of heat. The BGA244 cannot make gas measurements while the Degas Heater is active. This is indicated on the Home Page by the “DEGAS ON” screen message.

Note: The error LED will blink and a membrane resistance fault (*TST, bit 7) may occur when the Degas Heater is on.

Press Degas [ON] or [OFF] to turn the Degas Heater on and off.

Computer I/O



The BGA244 has three separate computer interfaces that can control and monitor all functions. Details on controlling the BGA244 over the computer interfaces are described in Chapter 6: *Remote Programming*.

The USB and RS-232 interfaces will operate for either USB or 24 V_{DC} power. The RS422 requires that an external +24 V_{DC} power supply is connected to operate.

The RS-232 and RS-422 interfaces can be enabled and disabled and have settable baud rates from 2400 to 115.2k baud. The USB has no enable or configuration functions. Press [USB], [RS-232] or [RS-422] to access that interfaces page.

Receive and Transmit Buffers


Each interface has its own set of receive and transmit buffers that show the most recent commands and responses over



that interface. They can be helpful when debugging communication problems. Press the [BUFFER] key on the interface page of interest to view.

USB


Nothing can be set for the USB interface. The only control on the USB page is the [BUFFER] key. Press it to view the transmit and receive buffers.




RS-232

Press [DISABLE|ENABLE] to enable or disable the RS-232 interface. Press [BUFFER] to view the transmit and receive buffers.

Press <Baud Rate ↓> to open the selection list. The currently selected rate will be highlighted in yellow. Press the desired value to select it. Most of the remaining parameters are read only.



Parity: None
 # of Bits: 8
 Stop Bits: 1
 Flow Control: CTS/RTS



The following RS-232 communication problems are reported in the *Message Log* (page 89).

- Parity Error
- Break
- Frame Error


RS-422

The RS-422 Interface page is similar to the RS-232 page. [ENABLE|DISABLE], <Baud Rate ↓> and [BUFFER] all operate in the same manner. There is no Flow Control for the RS-422.

Note that 24 V_{DC} must be present for the RS-422 interface to be operational.


The following RS-422 communication problems are reported in the *Message Log* (page 89).

- Parity Error
- Break
- Frame Error




Analog I/O

The BGA244 has two analog inputs and three analog outputs that can be used to interface to external devices, sensors or meters. Refer to *Analog I/O* (page 26) for details on connecting the analog inputs and outputs to external circuits. See the *Specifications* for details on the analog inputs and outputs performance. The Analog Inputs and Outputs will only operate if an external +24 V_{DC} power supply is connected.



Analog Inputs


The two Analog Inputs (INPUT 1 and INPUT 2) can each be set to measure either voltage, current or current with a loop power voltage source. The measured voltage or current can be viewed on the BGA244 or read over one of the computer interfaces. Press [INPUT 1] or [INPUT 2] to go to their configuration page.



Analog inputs can be used as either general purpose inputs, or together with a transducer as a pressure meter. If an analog input is configured as a pressure meter it may not be used for general purpose measurements. In either case the input must be configured properly for the desired measurement. The Input reading displayed on this page is in V or mA regardless if the analog input is used as a pressure meter.

[CONFIG PRESSURE] takes you to the Configure Pressure Gauge page as described in *Pressure* (page 64) earlier in this chapter. To set an input to general purpose, uncheck "Use as Pressure Gauge" on the Configure Pressure Gauge page.

Press [ENABLE | DISABLE] to enable or disable the analog input. Press <Type ↓> to open the selection list. The currently selected type will be highlighted in yellow. Press the desired value to select it.



The Loop Power Voltage can be entered if Current w/Loop Power is selected. This can vary from 6 and 19 V. See *Analog Input* (page 27) for details on devices that can use loop power.

There are several alerts that can occur for the Analog Inputs. They appear as red text at the bottom of the Analog Input page as well as on the Alert page. These normally indicate a problem with the device connected to the Analog Input or wiring, rather than the input itself.

Name	Valid for	Indicates
Undervoltage	Voltage	Input < -0.2 V
Overvoltage	Voltage	Input > 10.2 V
Undercurrent	Current	Input < 4 mA
Overcurrent ¹	Current	Input > 20 mA
Loop Power ²	Current w/ Loop Power	Loop Power current is > 50 mA

¹ The Input 1/2 Overcurrent condition switches out the current sense resistor when the input current exceeds 26 mA. It will automatically switch the current sense resistor back in when the input **voltage** drops below 5.8 V. In some cases it may be necessary to manually reset an overcurrent fault due to the behavior of the external circuitry.

² Loop Power 1/2 Alert switches out the Loop Power voltage source. It does not self clear and must be cleared manually.

Most of these Alerts except Loop Power self clear if the condition that caused them goes away. See the Message Log for a history of transient events. See *Alert* (page 87) for details on alert behavior.

Analog Outputs

There are three separate Analog Outputs that can be configured as 0 – 5 V, 0 – 10 V or 4 – 20 mA outputs. Measure Out is always linked to the measured value of the selected instrument mode. Outputs 1 and 2 can be linked to one of several different measured quantities or set explicitly by the user. Outputs linked to a measured quantity can be scaled to match a specific range using global units. For example, Measure Out can be scaled so a 0 to 1% gas concentration corresponds to a 0 to 10 V output to interface to an external device.

Use the Scale Min and Scale Max values to scale the analog output relative to the linked parameter. Scale Min is the value for the minimum analog output (0 V or 4 mA). Scale Max is the value for the maximum analog output (5 V, 10 V or 20 mA). Scale Min and Max are always set in the selected Global Unit for that parameter.

Example: In Binary Gas Analyzer mode, the Measurement Output is always linked to the Gas Concentration ratio (% , ppm or fraction). So Scale Min and Scale Max are entered as ratios.

Example: Analog Output 1 can be linked to the gas temperature (°C, °K or °F). In this case, Scale Min and Scale Max are entered as temperatures.

If the values of the linked parameters exceed the Scale Min or Scale Max values, the Analog Outputs will pin to their respective minimum or maximum values.

If the full scale ranges of the external device and the analog output don't match, you can still calculate what the measured output would be for the minimum and maximum analog output values.

Example: You need a 1 to 8 V output for a Gas Concentration of 10% to 20%.
 Use the 0 – 10 V setting for the Measure Output.
 7V corresponds to 10%, so 1V ~ 1.43%
 Scale Min (0V) = 10% - 1.43% = 8.57%
 Scale Max (10) = 20% + (2 * 1.43%) = 22.86%

There are two alerts that can occur for the Analog Outputs. They appear as red text at the bottom of each Analog Output page and on the Alert page. These normally indicate a problem with whatever is connected to the Analog Output rather than a problem with the output itself.


- IOut Alert indicates that the current output can't output the required current. This usually indicates a break in the 4-20 mA current loop or an overly large series resistor.
- Temperature Alert indicates the output driver has overheated and is in thermal limit. This usually indicates that the output is shorted.

These alerts self clear if the condition that caused them goes away. See the Message Log for a history of transient events. See *Alerts* (page 87) later in this chapter for details on alert behavior.

Measure Output


Measure Out is always linked to the measured parameter set by the Instrument Mode as follows. See the specific Instrument Modes for details on the measurements.

- Binary Gas Analyzer: Gas Concentration Ratio
- Gas Purity Analyzer: Speeds of Sound Ratio
- Physical Measurements: Normalized speed of sound




The present Output Value (voltage or current) is displayed just below Scale Max.

Press [ENABLE | DISABLE] to enable or disable the output. Press <Type ↓> to open the selection list. The currently selected type will be highlighted in yellow. Press the desired value to select it. Enter the Scale Min and Scale Max values.



Select Gas Concentration 1 or 2 using <Linked to ↓>. Normally these values are identical. However there are rare cases where there are two possible solutions for a given speed of sound. In this case Gas Concentration 1 is always the lower concentration and Gas Concentration 2 the higher one. Refer to *Gases* (page 93) for details.




Enter Scale Min and Scale Max to scale the Measure Output relative to the value of the measured parameter. They are in ratio units for the Binary Gas and Gas Purity, or in speed units for Physical Measurements.

Output 1 and Output 2

Analog Output 1 and 2 can be linked to several different parameters or be explicitly set from the GUI or computer interface as a User value. Analog Output 1 and Output 2 scale in the same manner as the Measure Output if linked to a measured parameter. Possible linked parameters include:


- Measured Speed of Sound
- Normalized Speed of Sound (to NTP)
- Gas Temperature
- Pressure Meter 1 or 2 (if configured)
- User: Set thru the GUI or over the computer interfaces.



See *Physical Measurements* (page 47) for information on the measured speed of sound, normalized speed of sound and the temperature. See *Pressure* (page 64) for details on pressure measurements.

Press [ENABLE|DISABLE] to enable or disable the output. Press <Type↓> to open the selection list. The currently selected type will be highlighted in yellow. Press the desired value to select it.

Press <Linked to↓> to open the selection list. The currently selected type will be highlighted in yellow. Press the desired value to select it.



If User Value is selected, a User Setting entry box will appear below “Linked to”. Enter a value within the full scale range for that output type. Otherwise enter the Scale Min and Scale Max values for the linked parameter.



The present Output Value (voltage or current) is displayed just below Scale Max or User Setting.

Faults

The BGA244 monitors several critical parameters. An Instrument Fault is generated if any of these goes out of range. These faults indicate that something serious is wrong with either the BGA244 or the power supplies connected to it. These faults are listed on the Faults page and can be read over the computer interface (see Remote Programming, page 160). See *Troubleshooting* (page 199) for further details.

Some Faults must remain active for ~60s before they are reported, to avoid spurious activity. Most Faults will self-clear if the problem causing them goes away. To view transient events, go to the Message Log. There are three main types of faults: External Power Supply Faults, General Faults and Self Test Faults.

Note: The behavior of the Fault Status registers is slightly different in terms of response time and self clearing. See *Status Byte Definitions* (page 189) for further details.

If any fault occurs, the red [System Fault] button appears on the Home page and the [Home] button turns red on all pages. Pressing [System Fault] takes you to the Faults page.



If an External Power Supply occurs, the POWER LED will flash at ~4 Hz. If a General or Self Test Fault occurs, the ERROR LED will stay on continuously.

External Power Supply Faults

If the BGA244's power supply input voltages fall outside the specified range, an external power supply fault will be generated. The +5V USB power supply is monitored only when the unit is powered over USB (no +24 V_{DC}). Otherwise the External +24 V_{DC} power supply is monitored. The following faults are displayed on the Faults page and are indicated by the Power LED blinking at about 4 Hz. See *Power* (page 22) for details on power supplies and cabling.

Note: There must be a minimum voltage (~ 3.3 V for USB, 18 V for the External +24) for the BGA244 to control the display or lite any LEDs.

TABLE 11: EXTERNAL POWER SUPPLY FAULTS

Fault Name	Meaning
USB Undervoltage	External USB Supply < 4.45V
USB Overvoltage	External USB Supply > 5.50V
24V Overvoltage	External 24 Supply > 28V

USB Undervoltage Fault: This is usually caused by a USB port that cannot provide sufficient current or a high resistance USB cable.

USB Overvoltage Fault: This fault may be caused by a faulty USB supply or poor quality cable.

24V Overvoltage Fault: This fault indicates a faulty or improper power supply. Confirm that the 24 V_{DC} Power Supply is appropriately rated.

General System Faults

General System Faults are critical errors that occur while the BGA244 is operating. These faults indicate a problem with the internal circuitry or a serious temperature problem. The following faults are displayed on the Faults page and are indicated by the ERROR LED staying on continuously.

TABLE 12: SYSTEM FAULTS


Fault Name	Meaning
Under Temperature	Cell Temperature reads < -20 °C
Over Temperature	Cell Temperature reads > 80 °C
Flash Fault	Fatal Error in Flash Memory
Firmware	Error in Firmware Update
Heater Fault	Block Temperature reads > 90 °C
1.4V Under Voltage	Internal 1.4V Supply too low
1.4V Over Voltage	Internal 1.4V Supply too high
3.3V Under Voltage	Internal 3.3V Supply too low
3.3V Over Voltage	Internal 3.3V Supply too high
5V_I Under Voltage	Internal 5V Supply too low
5V_I Over Voltage	Internal 5V Supply too high
Temperature Mismatch	>5°C Mismatch on Temperature Measurements

Self Test Fault

The BGA244 runs a series of Self Tests at power on, or when requested from the Self Test page or the *TST command over the computer interfaces. A Self Test fault is generated if any of the Self Tests fail. See *Self Test* (page 89) for more details.

Faults Page

The Faults page provides a list of all currently active Faults. Use [PAGE ↑] and [PAGE ↓] to navigate the list of Faults. If the faulting condition goes away, the Fault will self-clear and be removed from the list. To view transient events go to the Message Log.



Alerts

Alerts are warnings of non-critical conditions. These may be problems with the BGA244 or with external devices or wiring.

If an alert occurs, the yellow [Alert] button appears on the Home page and the [Home] button turns yellow on all pages. Pressing [Alert] takes you to the Alerts page. Each Alert can be cleared or disabled using its Alert Action Window.



Most Alerts will self-clear if the problem causing them goes away. Transient events can be viewed on the Message Log. There are three main types of alerts: Analog I/O Alerts, Temperature Alerts and General Alerts.

Analog I/O Alerts

There are a number of alerts that can occur for the Analog Inputs and Outputs. These normally indicate a problem with whatever is connected to the Analog Input or Output and not a problem with the BGA244. All Alerts except the Input 1/2 Loop Overcurrent will self clear when the condition causing them is eliminated. See *Analog I/O* (page 81) for details on the specific alerts.

TABLE 13: ANALOG I/O ALERTS

Alert Name	Meaning
Input 1 Undervoltage	Input 1 < -0.1 V (V In only)
Input 1 Overvoltage	Input 1 > 10.2 V (V In only)
Input 1 Undercurrent	Input 1 < 3.5 mA (I in only)
Input 1 Overcurrent	Input 1 > 26 mA (I in only) *
Input 2 Undervoltage	Input 2 < -0.1 V (V In only)
Input 2 Overvoltage	Input 2 > 10.2 V (V In only)
Input 2 Undercurrent	Input 2 < 3.5 mA (I in only)
Input 2 Overcurrent	Input 2 > 26 mA (I in only) *
Input 1 Loop Overcurrent	Input 1 Loop Power current is > 50 mA
Input 2 Loop Overcurrent	Input 2 Loop Power current is > 50 mA
Measure Output Current	Measure Out cannot output the required current
Measure Out Temp	Measure Out Over Temperature
Output 1 Current	Output 1 cannot output the required current
Output 1 Temp	Output 1 Over Temperature
Output 1 Current	Output 1 cannot output the required current
Output 1 Temp	Output 1 Over Temperature

* In some cases it may be necessary to manually reset an overcurrent fault due to the behavior of the external circuitry. See *Analog Inputs* (page 81) for details

Temperature Alerts

There are several alerts that indicate that the cell temperature is out of range or there are discrepancies between the two temperature sensors.

TABLE 14: TEMPERATURE ALERTS

Alert Name	Meaning
Under Temperature	Cell Temperature < 0 °C
Over Temperature	Cell Temperature > 70 °C
Temperature Mismatch	Mismatch between temperature sensors > 2 °C

General Alerts

These Alerts are for conditions not included in other alert tables.

TABLE 15: GENERAL ALERTS

Alert Name	Meaning
USB Undervoltage	External USB Supply < 4.6 V
USB Overvoltage	External USB Supply > 5.3 V
24V Undervoltage	External 24 Supply < 22 V
24V Overvoltage	External 24 Supply > 26 V
Heater Not Regulating	The Cell Heater is not regulating
Factory Safe Mode	A fatal error caused the unit to re-boot to Safe Mode.

The Under and Over Voltage Alerts indicate that the power supply input voltages are marginal. The +5V USB power supply is monitored only when the unit is powered over USB (no +24 V_{DC}). Otherwise the External +24 V_{DC} power supply is monitored. These alerts may indicate a problem with the power supply or cabling. See *Power* (page 22) for details on power supplies and cabling.


Heater Not Regulating indicates that the heater servo is set to its minimum or maximum values and is not correctly regulating the cell temperature. It is only active when the heaters are on. Note that this Alert normally appears when the heater is first turned on or if the ambient or set temperatures are changed. See *Heater* (page 78) for more details.

Factory Safe Mode indicates that the BGA244 software encountered a fatal error and restarted to the Factory Safe Mode. Refer to Factory Safe Mode, page 205 for details.

Alerts page

Each Alert appears as a separate multipurpose button on the Alerts page. Use [PAGE ↑] and [PAGE ↓] to navigate the list of Alerts.

Most Alerts self clear and vanish when the problem causing them goes away. Pressing a particular [ALERT] button will take you to its Alert Action Window that allows you to control the appearance and behavior of the alerts. Note that all alert actions are lost if the power is cycled.



Alert Actions

Press [CLEAR ALERT] to clear a currently active alert. This is only useful for alerts that are not self clearing. Other Alerts will immediately re-assert themselves after being cleared, since they are still active.



Press [DISABLE ONCE] to ignore a currently active alert. This will cause the alert to be ignored and clear the alert indicators (assuming this is the only active alert). The Alert button will turn lite grey to indicate its “disabled once” state. If this alert is cleared and later re-asserted, it will revert to the normal “un-ignored” state, with the alert indicators active.

Press [DISABLE ALWAYS] to permanently ignore an active alert. This will cause the alert to be ignored and clear the alert indicators (assuming this is the only active alert). The Alert button will turn dark grey to indicate its “disabled always” state. If this alert is cleared and later re-asserted, it will remain inactive with the Alert added to the list in dark grey until power is cycled.

Press [ENABLE ALL ALERTS] to return all alerts to the active state. This will clear any “Disable Once” or “Disable Always” behavior. Note that power cycling the BGA244 will always enable all alerts.

Message Log

The Message Log is a list of the most recent events that occur in the BGA244. This includes stored and recalled settings, alerts, faults and other messages. The Message Log can be particularly useful in diagnosing transient events that may disappear before they can be properly diagnosed.




Self Test

The Self Test page provides access to the Self Test functions and their results.

Running Self Test

Running Self Test will stop all measurements and can temporarily modify inputs, outputs and relay settings for several seconds. Self Test is the same procedure that is run upon power on or by the computer interface *TST command.



To begin Self Test, press [START SELF TEST]. The tests and their results will be written to the screen. Upon completion, the BGA244 will display “SELF TEST PASS” or “SELF TEST FAIL”. Use [PAGE↑] and [PAGE↓] to navigate the list of tests. If Self Test fails it will be listed on the Faults page.

See *Troubleshooting* (page 199) for information about Self Test failures.

TABLE 16: SELF TESTS

Test Name	Description
3.3 V Supply	Internal 3.3 V Power Supply is within range
1.4 V Supply	Internal 1.4 V Power Supply is within range
4.5 V Supply	Internal 4.5 V Power Supply is within range Not performed on Rev C PCB
USB Supply	External USB Power Supply is within range Not performed if +24V is present
24 V Supply	External 24 V Power Supply is within range Performed if +24V is present
5 V Industrial Supply	Internal 5 V Power Supply is within range Performed if +24V is present
CODEC	CODEC is detected
ADC 1	ADC 1 is detected
ADC 2	ADC 1 is detected
USB	FTDI USB chip is detected Performed only if USB power or USB host is connected Includes part number & serial number
Thermistors	Thermistor 1 & 2 resistance is within range
Thermistors Single Ended	Low side of thermistor terminals is not shorted to the case
Membrane 1	Membrane 1 resistance is within range
Membrane 2	Membrane 2 resistance is within range
Magnet 1	Magnet 1 field is within range
Magnet 2	Magnet 2 field is within range
Analog Output	Output DACs are detected Performed if +24V is present
Analog Input	ADC is detected Performed if +24V is present


Password

The BGA244 keypad can be locked out with a password enable to keep unauthorized personnel from changing settings on the BGA244. When locked out only the home page can be viewed.

If keypad is locked, a message appears at the top left of the screen. Any key press will open the numeric keypad to allow the user to enter the password. An “Invalid Password” prompt will appear if an incorrect password is entered. After the correct numeric password has been entered, all functions can be accessed. If locking is enabled, the keypad will be re-locked after 30 seconds of no keypad activity.



Press [ON] or [OFF] to turn locking on and off. A new password can be entered if locking is on. Enter a 4 digit numeric value to change the password.




If you forget the password, the unit can be reset using Hard Reset function described in the *Hardware Reset* section. This will disable password locking and reset the password to the default value of 0000.

About the BGA244

The About the BGA244 page displays the following information. This is the same information that is displayed on the display during power on.

- Serial Number
- Firmware Version
- Calibration Date
- PCB Version
- Cell Version
- (if active) Factory Safe Mode



“Factory Safe Mode” will only be displayed if the BGA244 encountered a fatal software error that caused the unit to re-boot. Refer to Factory Safe Mode, page 205 for details.

Hardware Reset

The BGA244 has a hardware reset button that can be used if the BGA244 is frozen in an inoperable state. There are two different types of resets with slightly different behaviors.

Use an unbent paper clip or a small screw driver to press the reset button.

Normal Reset

Press the reset button once. Releasing the button forces a Normal Reset. This acts the same as power cycling the BGA244. This is ordinarily used if the BGA244 is in an unknown state and it is difficult to remove power to restart the unit.

Hard Reset

Press the reset button once. Release it, then immediately press and hold the button down for at least 10 seconds to execute a Hard Reset. This acts the same as power cycling the BGA244 and Recalling Setup 0 (Default Setup). This is ordinarily done if something has occurred to put the BGA244 in an inoperable state where loading the Default Setup is impossible, or to reset a forgotten password.

Note that this will replace the current settings of all parameters with the Default Settings. See *Default Setup* (page 69) for details.

Factory Safe Mode

If the firmware has detected a fatal error the unit will restart in the Factory Safe Mode. This will be indicated on the bottom of the About screen. Refer to *Factory Safe Mode*, page 205 for information.

User Gases

Gases can be added to the User Gas Table using the BGAMon software or over one of the computer interfaces. See *User Gases* (page 117) in Chapter 4 for details on this procedure.

Updating Firmware

The BGA244 Firmware can be updated by the user using the BGAMon software. See *Updating Firmware* (page 143) in Chapter 5 for details on this procedure.

Chapter 4: Application Guide

This section provides information on using the BGA244. It includes details on gases, pressure, operating conditions, measurements, high purity use and adding User gases.

Gases

The BGA244 determines the concentration and purity for a wide variety of gases. It does this by measuring the speed of sound and temperature and then determines the mixture ratio or gas purity using the physical properties of the gases.

The fastest speed of sound that can be measured is ~1400 m/s, the speed of sound for hydrogen (2 amu). The slowest is less than 130 m/s, the speed of sound for SF₆ (146 amu). This range covers nearly all compounds that are gaseous within the operating temperature and pressure range of the BGA244.

Gas Requirements

Gases must be clean and dry. This means they cannot contain any solid or liquid particulates. Gases that contain solid particulates must be filtered before entering the BGA244. It may be necessary to operate the BGA244 at an elevated temperature to ensure there are no liquid particulates in the gas stream. Gases must not contain oils or waxes can coat the inside of the acoustic cell or Kapton acoustic transducers.

Gases must be non-condensing and contain no liquid particulates at the temperature and pressure the BGA244 is operating. If a gas is nearing its condensation point the BGA244 will indicate a condensation warning message. See *Condensation* (page 95) for more information.

Supported Gases

The BGA244 Factory Gas Table contains nearly 500 different gases. These include common industrial gases and a wide range of specialty gases. See Appendix A: *Gas Table* for a list of supported gases.

Gases not included in the Factory Gas Table can be added using the User Gas Table. This supports a simple model of a gas as compared to the Factory Gas Table. See *User Gases* (page 117) for more details.

Gases aren't constrained to single species. Gases mixtures or blended gases can also be used. Atmospheric air (N₂, O₂, Ar, CO₂) is included in the Factory Gas Table. Other mixed gases can be added to the User Gas Table using the procedure described later in this chapter.

Gases that should be used with caution

Certain gases or mixtures can be problematic for the BGA244. Caution should be used with gases that react with the wetted materials. See *Specifications* for a complete list of wetted materials. Contact SRS if you have questions about compatibility of a particular gas or mixture with the BGA244.

Explosive or Flammable Mixtures

The BGA244 is not ATEX rated. Under normal operating conditions the BGA244 cannot ignite the gas being analyzed. However, if the instrument is used with flammable or explosive gas mixtures SRS recommends the use of flame arrestors on both gas ports.

The BGA244 gas cavity's proof pressure (2,500 psia) is sufficient to contain the denotation of an explosive gas mixture of up to 30 psia (207 kPa). The instrument would not operate after such an event.

Corrosive Mixtures

Some gases can react with certain of the wetted materials. Copper is used in two places in the BGA244: the OHFC end gasket and the acoustic membrane.

SRS recommends using the BGA244HP for corrosive mixtures as it replaces the standard OFHC copper gasket with a gold flashed version.

The instrument's acoustic transducer consists of an ENIG (Electroless nickel, covered by a thin gold layer) plated copper spiral on a 40µm thick Kapton polyimide film. It can be susceptible to:

- Gases that react strongly with copper
- Strong bases (NaOH, KOH) that may react with Kapton film

We recommend corrosion testing of this component on a case-by-case basis. Contact SRS for additional information.

Unusual Gas Properties

Certain gases or gas combinations may have properties that may affect the accuracy of their measurements.

Relaxation Effect

Relaxation (or dispersion) is a non-linear molecular effect that can be modeled as a frequency dependence of the heat capacity of the gas. Most gases fully equilibrate (or relax) in less than one cycle of their resonant mode within the acoustic cavity. Of the gases that don't, most take far longer than a single resonant cycle. A very few gases do neither. These gases exhibit a pressure dependent shift in the speed of sound. In most cases this shift is small, only a few 100 ppm. Depending on the gas species involved, a second gas may cause the mixture to equilibrate more rapidly.

Carbon dioxide (CO₂) is unique in that it has a very large relaxation effect at the frequencies of its resonant modes for lower gas pressures. This distorts the acoustic signal, leading to larger errors or unrecoverable signals. To avoid this effect, the minimum recommended operating pressure for CO₂ is around 12 psia (83 kPa), although the detection threshold is less than 8 psia (55 kPa).

Most gases added to CO₂ will not affect its relaxation time, with the notable exception of water vapor. A mole fraction of 0.5% water vapor reduces the relaxation time of CO₂ by about 15 times which dramatically reduces its effect. Hydrogen will also reduce the relaxation time, but only by about 1.5 times. Alcohols may also have an effect.

The BGA244 includes data to correct relaxation effects for its measurements. This correction should be left in the default setting (on) for almost all cases. See *Relaxation Correction* on page 76 for details.

Low pressure measurements of CO₂ that contain water vapor may benefit from turning off relaxation correction. The water vapor can be either the second gas of a ratio measurement or a contaminant in the CO₂.

Reactive or Dimerizing gases

A reactive gas mixture can produce new gas species. A dimer is a molecule made up of two identical simpler molecules. The analysis techniques used in the BGA244 assumes that the molecules are non-reactive and non-dimerizing.

Condensation

The BGA244 cannot measure gases that are at or near their condensation point. There are two main reasons for this. Actual condensation will interfere with the acoustic cavity making it difficult or impossible to accurately determine the speed of sound. Plus the thermodynamic properties of a gas change radically as it transitions from its gaseous to liquid state. This change makes it impossible to calculate the gas composition and normalized speed of sound. In extreme cases, condensation can actually fill the cell with liquid.

Before actual condensation occurs, pre-condensation can interfere with measurements. During each cycle of the acoustic signal, the pressure within the acoustic chamber increases and decreases slightly. For a gas near condensation, this pressure increase can take the gas into the non-linear region causing large measurement inaccuracies.

Condensation is a function of a particular gas species, its current vapor pressure and temperature. The relative concentration of the gas in a mixture has no effect. Condensation will occur if the temperature is low enough that the vapor pressure of the gas exceeds its saturation value.

Ordinarily gases flowing through the BGA244 aren't near their condensation point. But there are several things that could cause them to condense.

- A reaction may occur at elevated temperatures relative to the BGA244. Condensation may occur when the gas passes into the cooler BGA244.
- The BGA244 can be used to monitor and control the output of a bubbler in a dilution flow. The temperature of the gas bath is adjusted to control the gas concentration. If the BGA244 is operated at a lower temperature than the bath, condensation can occur.
- The BGA244 may be operated at an ambient temperature that is below that of the gas, leading to condensation.

Condensation Warning

The BGA244 contains condensation data on the gases in the Factory Gas Table. By knowing the gas species, concentration, temperature and pressure the BGA244 can detect if the gas is near its condensation point and display a “Condensation Warning” screen message. The BGAMon software also provides information if the conditions and selected gas species are near condensation.

Note that the condensation warning depends on the BGA244 being filled with only to the two gases selected. It is possible for condensation to occur without warning if a non-selected gas gets into the chamber.

Avoiding Condensation

To avoid condensation, ensure that the saturation point (pressure at which condensation occurs) of the gas species exceeds the partial pressure of the gas within the BGA244. The following techniques can be used to eliminate condensation.

The Block Heaters can be used to raise the cell temperature. The temperature should be high enough so the saturation value of the gas exceeds the partial pressure of the gas. If this temperature is unknown, set the Block Heater to operate at least 3-5 °C above the highest expected gas manifold temperature up to the maximum of 70 °C. Make sure the power supply and Maximum Heater Current are large enough to raise the cell temperature to the set point (typically 0.5 to 2 amps). Refer to *Heater* (page 78) for information on setting the heater.

Water Vapor in Air

Atmospheric air usually contains some water vapor in addition to the usual nitrogen, oxygen, argon and carbon dioxide. This is referred to as humidity. It will affect measurements since water vapor is an additional gas species. Ordinarily mixtures containing water vapor as a third gas should be dried prior to measurement. However in some cases, drying can be avoided by using the REL function to compensate for the third gas.

For this technique to work the amount of water vapor in the “wet” air must be fixed. If it changes, a new REL must be performed. The amount of water vapor must also be low enough to avoid condensation. Note that the condensation point of water (or any liquid) is independent of any other gas species. But it does depend on both temperature and pressure. BGAMon can be used to see the saturation point (maximum amount of water vapor) for your operating temperature and pressure. It may be

necessary to operate at an elevated temperature if the water vapor is approaching condensation.

As with any REL, this technique works best near the conditions the REL was performed at. This means operating at a similar temperature, pressure and (most importantly) amount of water vapor in the wet air. Errors will increase the as the operating conditions move further away from the REL conditions. Refer to *REL to a Reference Gas* (page 105) for more details about this procedure.

Procedure

It is usually a good idea to know the approximate amount of water in the air before performing a REL. This serves two purposes. First it determines if the cell temperature should be increased using the heaters to avoid condensation. Second it provides a check to see if the amount of water vapor has changed since the unit was last REL'd.

Select dry air and water in the gas selection menu. Flow 100% wet air through the cell. If a condensation warning occurs, increase the cell temperature using the block heaters. Record the amount of water vapor in the wet air to compare with future measurements. A new REL should be performed whenever the amount of water vapor changes. It may be helpful to track the atmospheric temperature and relative humidity.

Next select dry air and the second gas to be measured in the gas selection menu. Be sure that wet air is still flowing through the cell. Go to the REL menu, select "Use REL" and press either [REL to 0%] or [REL to 100%], depending on your measurement. (Remember that the BGA244 graph reports the fraction of the primary gas in the mixture.) Return to the Home page. The reading should be very nearly the exact value selected on the REL page, either 0 or 100 %.

Water Vapor in Other Gases

This technique is generally applicable to any gas that contains water vapor.

Pressure

Operating Pressure

The BGA244 can operate over a wide range of pressures, starting from a few psi (~10 kPa) up to 150 psi (1000 kPa). Operating near atmospheric pressure, ~15 psi (~100 kPa), provides good results if you have the flexibility to select the operating pressure.

Pressure Effects in Gases

Several thermodynamic properties of gases have pressure dependent terms that cause their behavior to diverge from the ideal gas law. These effects are significant for some gases. Sudden changes in pressure may also cause transients in the measured readings. See *Transients* (page 102) in this chapter for more details.

The largest departure from the ideal gas law is caused by intermolecular effects (effects between gas molecules). This causes a change in the speed of sound as a function of pressure. These are commonly modeled as virial coefficients. The graph below shows the shift in the speed of sound for pure SF₆, a gas with a fairly large frequency shift vs pressure.




FIGURE 17: SPEED OF SOUND VS PRESSURE IN SF₆

These frequency shifts directly affect the Gas Purity and Physical Measurements modes. Their effect on the binary gas ratio is more complicated, since this also depends on the relative molecular masses of the two gases.

Is Pressure Information Necessary?

There are cases where the BGA244 can make satisfactory measurements without precise pressure information.

- Best accuracy is not required.
- Using gases that have small variations due to pressure.
- Using the REL function to improve accuracy at a fixed pressure.

Unless one or more of these is true, it's necessary to know the gas pressure. The *BGAMon* software can indicate measurement sensitivity to pressure variations.

To meet the specified accuracy for all gases, the pressure must be known to within ± 1 psi (6.9 kPa). This information can be directly entered into the BGA244 or measured by a pressure transducer interfaced to one of the analog I/O ports as described in the *Pressure Transducers* (page 31).

Estimating pressure

Tip: If your system vents to the atmosphere, the operating pressure is probably around 1 atm (14.7 psia or 101 kPa).

Compensating using REL

The REL function can be used to compensate for a lack of pressure information in some cases. This technique works best when deviations from a reference gas or reference mixture are more important than the absolute accuracy. It is most accurate for small deviations from the reference point; larger deviations are less well compensated. This REL will only be valid for the gas or mixture it was performed on. A new REL should be performed whenever anything that can affect the gas pressure is changed. This may include different flow rates, new gas cylinders, regulators or MFC's.

Minimum Operating Pressures

The minimum operating pressure that the BGA244 can make accurate measurements depends on the mass of the gas mixture. Below this minimum the accuracy of the measurements may be degraded. Table 17 lists the recommended minimum operating pressures for pure gases. All values are listed in absolute pressure units (relative to vacuum). Use Averaging at low pressures to improve the detection limit and stability of measurements. The default value of 10 averages is a good place to start, but values of 20 or more may be necessary some cases.

At very low pressures the BGA244 cannot reliably recover the acoustic signal. A "No Signal" message will be displayed in place of the measurement if this occurs. (See *Screen Messages*, page 50 for details).

TABLE 17: RECOMMENDED MINIMUM OPERATING PRESSURES FOR PURE GASES

Gas	Minimum Pressure
H ₂	6 psia (41 kPa)
He	6 psia (41 kPa)
CH ₄	3 psia (21 kPa)
N ₂	2 psia (14 kPa)
Ar	2 psia (14 kPa)
CO ₂	12 psia (83 kPa)
SF ₆	1 psia (7 kPa)

Pure CO₂ (carbon dioxide) has an unusually high minimum operating pressure due to relaxation effects. For best accuracy the operating pressure should be at least 12 psia (83 kPa). Refer to *Relaxation Effects* (page 94) for more details.

Since the average mass of the gas is what sets the minimum operating pressure, mixtures of He or H₂ and any other gas will be able to be measured at lower pressures than pure Helium or Hydrogen.

Ambient Pressure Variation

Ambient pressure varies only slightly due to weather. The default ambient pressure is 14.7 psia (101 kPa) which corresponds to the average atmospheric pressure at sea level. A typical low ambient pressure at sea level is about 14.5 psia (100 kPa); a typical high pressure is about 14.9 psia (103 kPa).

Elevation has a much larger effect on ambient pressure. The typical atmospheric pressure in Denver, Colorado at 5280' (1610 m) is 12.2 psia (84 kPa). Setting the ambient pressure based on the elevation and ignoring any weather effects should only introduce a small error for devices using gauge pressure units. However for best accuracy the ambient pressure should be measured and updated whenever it changes by more than 0.5 psi (3.4 kPa).

Operating Conditions

Warm Up Time

The BGA244 can make measurements immediately after power-on. The accuracy may improve slightly after warmup due to thermal gradients stabilizing. Depending on the operating conditions it can take up to 60 minutes for the internal temperature to completely stabilize in a closed cell. If gas is flowing through the cell this time will be reduced to less than 10 minutes. There may be a small change in reading during this time, depending on the specific measurement.

The temperature settling when the heaters are used depends on a number of factors including the ambient and set temperatures, the maximum heater current, gas temperature and flow. Examples for several different temperatures and heater current are listed below. See *Heater* (page 78) for information on setting the heaters.

TABLE 18: HEATER SETTling TIME

Initial Temp	Final Temp	Max Current	Regulating	Settled
25 °C	40 °C	0.5 A	22 min	32 min
25 °C	50 °C	1.0 A	24 min	29 min
25 °C	70 °C	2.5 A	19 min	27 min

Response Time

The response time of the BGA244 is primarily a function of the gas flow rate required to turn over the volume of the acoustic cell (~ 130 cc). The following table shows the step response time for a 100% step in gas species assuming a constant temperature, pressure and flow rate. At low flow rates, changes in temperature or pressure may affect the settling time as described in *Transients* (page 102).

TABLE 19: RESPONSE TIME

Flow (sccm)	Settling to 10%	Settling to 1%	Settling to 0.1%
100	180 s	360 s	540 s
200	90 s	180 s	270 s
500	36 s	72 s	108 s
1000	18 s	36 s	54 s
2000	9 s	18 s	27 s
5000	4 s	7 s	11 s

High Flow Rate

The BGA244 is specified for flow rates up to 5000 sccm (5 liters/min). Customers have successfully operated at rates of up to 20,000 sccm (20 liters/min). At very high flow rates there may be some noise in the readings that averaging may help stabilize. See *Using Averaging* (page 104) for details.

Transients

Changes in flow, pressure or gas temperature may cause small transients in the measured readings even if the gas does not change. This is caused by a disruption in the thermal profile in the BGA244 cell. The transient settles as the new thermal profile is established. The magnitude of the transient is a function of the measurement being made, the gases and the magnitude of change. The change in the speed of sound is typically less than 0.1%. The change for ratio measurements is typically less than the accuracy estimation. For the best accuracy allow the readings to settle after large changes in flow, pressure or gas temperature.

Changes in the thermal profile can be described as follows. A cell in equilibrium has a temperature profile that is a function of the gas mixture, flow rate, pressure, temperature of the gas and the temperature of the BGA244 gas cell. The cell thermistors measure the cell temperature at 2 specific points within that profile. If any of the conditions change, a new temperature profile will gradually be established. The thermistor temperature will lag the overall gas temperature until the new equilibrium is reached. How long it takes to reach this new equilibrium is a function of the thermal conductivity and flow rate of the gas.

For low flow rates (<100 sccm) settling occurs primarily through the thermal conductivity of the gas in the cell. This typically takes around 15 seconds and depends heavily on the gas species. The following table shows the transient settling time for a few different gases at zero flow.

TABLE 20: ZERO FLOW TRANSIENT SETTLING TIME

Gas	Transient Settling Time
H ₂	3 s
He	3 s
N ₂	10 s
Ar	15 s
SF ₆	21 s
Kr	28 s

At higher flow rates (> 100 sccm) settling becomes more of a property of the gas flow rate. The gas at the thermistors is replaced by outside gas rather than relying on the thermal conductivity to reach equilibrium. The following table shows the transient settling time to 1% for Nitrogen (N₂) at several flow rates.

TABLE 21: TRANSIENT SETTLING TIME FOR DIFFERENT FLOW RATES

Flow (sccm)	Transient Settling Time
100	10 s
200	10 s
500	10 s
1000	10 s
2000	9 s
5000	7 s

Interference

Vibration

Mechanical vibrations can be picked up by the BGA244. This causes an interfering signal that can make measurements noisy or even impossible to make. The best solution is to make sure that the BGA244 is mounted in a vibration and impact free location. Follow the mounting guidelines in Chapter 2: *Installation Guide*.

Averaging may help stabilize the readings. It may require a large number of averages to suppress a large interfering signal. You may need to experiment to find the proper balance between stable answers and response time. See *Using Averaging* (page 104) for more details.

Electrical and Magnetic Interference

Even though the BGA244 is shielded for EMI, motors or transformers can generate large enough fields to overwhelm the acoustic transducer signals. It may be necessary to add shielding if the BGA244 and the interfering source cannot be separated far enough apart. For low frequencies (<1 kHz) use cold rolled steel. At higher frequencies (> 1 kHz) conductive materials (copper or aluminum) can be used. Avoid passing currents through the BGA244 from the gas lines. This can generate interfering signals. Make sure that all pipes are properly grounded.

Averaging may help stabilize the readings. It may require a large number of averages to suppress a large interfering signal. You may need to experiment to find the proper balance between stable answers and response time. See *Using Averaging* (page 104) for more details.

Measurements

In most cases the BGA244 can make accurate measurements right out of the box without any adjustments or fine tuning. However certain measurements may benefit from the following techniques.

Temperature Variations

The BGA244 measures the gas temperature as part of its measurements. However large ambient temperature changes or extremely precise measurements may benefit from operating the Block Heater to stabilize the cell temperature.

The Block Heater is typically operated at least 3-5 °C above the highest expected ambient temperature. The Maximum Heater Current should be large enough to raise the cell temperature at least 5-10 °C above the set temperature. This is normally between 0.5 and 2 A.

Evacuating Cell

At very low flow rates it can take a long time for changes in gas concentrations to completely settle to their final value (see *Response Time*, page 101).

Rather than wait until the total volume of the cell turns over several times, it may be much faster to first evacuate the cell and then refill it. This speed is at the expense of additional complexity in the gas manifold. A vacuum pump and inlet and outlet valves are required. But in some cases, most of this may already be part of the system. For extremely precise measurements it may be necessary to allow the cell to reach thermal equilibrium prior to getting completely settled measurements.

Using Averaging

Averaging successive measurements can improve accuracy and repeatability of the measurements by reducing interfering signals and noise. However this occurs at the expense of having a slightly slower response to changes in the measurement. The BGA244 takes a measurement every 228 ms which is quite a bit faster than most processes change. A moderate number of averages doesn't affect the response time and can minimize variations in a measurement.

The averaging method used in the BGA244 cancels out any signals that are not synchronous to its measurement signal. This suppresses interfering signals and noise as a function of the number of averages.

Averaging can dramatically reduce the noise on a signal. The graph shown is for Argon in Air. Because the speeds of sound of the two gases are similar, the signal is fairly noisy. Using 10 averages decreases the peak variation by about 4 times.




FIGURE 18: GAS CONCENTRATION WITH AND WITHOUT AVERAGING FOR ARGON AND AIR

Larger numbers of averages may be required to suppress interfering signals or recover low pressure signals in the presence of a lot of noise. This will slow the response time. You may need to experiment to find the best balance between stable answers and response time. See *Averaging* (page 75) for details.

REL to a Reference Gas

Using REL can improve measurements by removing uncertainties in the temperature, pressure and other parameters. The procedure involves either flowing a reference gas through the cell and then nulling the measured value to the expected value or by using knowledge from previous measurements. REL works best near the conditions where the REL was performed. Errors will increase as the operating conditions move the further away from the REL conditions.

The BGA244 is relatively insensitive to variations in temperature, pressure, or flow. However large changes in the operating conditions will produce measurable changes in the measurements. It is best to perform a REL as close to the operating condition as possible. This means operating at the same pressure, flow rate and temperature.

Consider repeating REL whenever a significant change in the operating conditions occurs. This could include changes in temperature, pressure and flow or changing the gas used as the reference (new gas bottle).



Caution: The accuracy of readings after setting a REL depends on how well the REL was performed. Be sure that the measurements have completely settled before performing the REL.

Flowing a Reference Gas

A Reference Gas is usually:

- One of the two gases in a binary mixture
- An extremely well measured blend of the gases
- The pure Gas for the Gas Purity Mode

The usual magnitude of the correction in a REL for a Reference Gas is a few percent or less. A larger than expected REL value may indicate that the gas you are using for a reference isn't as pure as expected.

It is necessary to ensure that the BGA244 has only the reference gas in it prior to performing a REL. There are two techniques to ensure the acoustic cell is free of all residual non-Reference gas.

- Flow the Reference Gas until the measurements have completely settled. This may take many minutes to remove any residual gas, especially at low flow rates.
- Evacuate the acoustic cell to vacuum prior to flowing the Reference Gas. This takes only a few seconds, but requires additional valves and pumps.

Be sure to flow enough reference gas thru the unit until readings have completely stabilized. Unless the cell is first evacuated, this will require flowing at least 1 liter of gas over 5-10 minutes to completely turn over the original gas inside the BGA244.

Set the number of averages (N) between 10 and 30. Allow the reading to completely stabilize. Remember that it takes about $N * 1.14$ seconds for averaged measurements to settle. Refer to *Averaging* (page 75) for details.

Purity Mode and 1 Point REL

See *1 Point REL* (page 62) or *Gas Purity Mode* (page 62) for information on the REL controls.

Flow the reference gas thru the unit until readings have completely stabilized.

Go to the REL page and select (Use REL) or (Use 1 Point REL), then press the (1 Point REL) control button.

Press [REL to 100%], [REL to 0%] or enter the <REL Value> directly as described in *REL* (page 61).

Return to the Home page. The reading should be very nearly the exact value that was selected on the REL page.

2 Point REL

See *2 Point REL* (page 63) for information on the REL controls.

The values for 2 Point REL can be determined by either measuring a reference gas or by knowledge from previous measurements.

Note: Any measurements used to calculate the Raw/REL values must be made with the REL Mode set to None.

Flow the initial reference gas thru the unit until readings have completely stabilized. Enter the Raw value (the measurement with no REL active) and the expected (REL) value for Point 1. It may be useful to select "Use 2 Point REL", return to the Home screen and verify that the value is as expected. Be sure to select "Use None" before making any measurements for Point 2.

Repeat this procedure for the second reference gas. If there is no second reference gas, set the Raw and REL values to the same value (usually either 0 or 100%). Enter the Raw value (the measurement with no REL active) and the expected (REL) value for Point 2. Select "Use 2 Point REL", return to the Home screen and verify that the value is as expected.

Multi Point REL

See *Multi Point REL* (page 63) for information for information on the REL controls.

Multipoint REL is normally used for blended gas mixtures, gases not contained in the Factory Gas Table, or where the behavior of the gas mixture depends on something besides just the molar ratio. Creating Raw/REL values for multipoint REL typically requires multiple reference gases or the ability to create mixtures of different concentrations.

Note: Any measurements used to calculate the Raw/REL values must be made with the REL Mode set to None.

As with 2 Point REL, values can be determined by either measuring a reference gas or by knowledge from previous measurements. Use the procedures described for 2 Point REL to determine the Raw & REL values. After you have obtained the measured (Raw) values and the expected (REL) values, edit the table entries. Select Use Multi Point REL, return to the Home screen and verify the values are as expected.

Binary Gas Analyzer vs Gas Purity Analyzer

Both the Binary Gas Analyzer and the Gas Purity Analyzer can be used to measure the purity of a gas. There are benefits for each analyzer depending on the gases being measured.

When to use the Binary Gas Analyzer

If you know the contaminating gas species you ordinarily want to use the Binary Gas Analyzer. It will accurately report the concentration of the contaminating gas species.

However if the contaminating gas is misidentified it will report the wrong concentration.

Example: The Binary Gas Analyzer is configured to measure Helium in Methane. It reports 95% Helium.

If the contaminating gas was actually air instead of Methane, the reported value would be incorrect. In this case the true concentration of Helium would be 97.3%

When to use the Gas Purity Analyzer

The Gas Purity Analyzer is most useful when you don't know the contaminating gas species. It reports the gas purity based only on the speed of sound of the gas. So you can determine that the gas isn't pure, but you do not know the concentration of the contaminating gas.

Example: The Gas Purity Analyzer is set to measure Nitrogen and reports $\Delta W/W = -1\%$.

This answer can be obtained from a variety of different contaminating gases, each with its own concentration. All you really know is that the contaminating gas is heavier than Nitrogen. The following table lists a few of the possible contaminating gases and their concentration required to make $\Delta W/W = -1\%$.

Contaminating Gas	Concentration
Oxygen	13.9%
Argon	6.6%
Carbon Dioxide	2.9%
Butane	1%

Converting Molar Fraction vs Mass Fraction

The ratio of the two gases can be reported as either a mole fraction or a mass fraction. These differ only by the atomic masses of the gases. Mole fraction is the ratio of the number of molecules of one gas relative to the total number of molecules in a volume. Mass fraction is the ratio of the mass of one gas relative to the total mass of the gas in the volume.

Mole fraction is commonly used in gas blending, while Mass fraction is used in monitoring combustion products. The BGA244 can report results using either method.

The formula for converting from mole fraction to mass fraction is:

$$w1 = \frac{(x1 * M1)}{(x1 * M1) + (x2 * M2)}$$

For converting mass fraction to mole fraction:

$$x1 = \frac{(w1/M1)}{(w1/M1) + (w2/M2)}$$

Where $w1$, $x1$ and $M1$ are the mass fraction, mole fraction and Mass of the primary gas and $w2$, $x2$ and $M2$ are the mass fraction, mole fraction and Mass of the secondary gas.

Example: Determine the mass fraction of Nitrogen in air

For the simple 2 gas case (ignoring other gases), air made up of 78% Nitrogen molecules (N_2) and 22% Oxygen molecules (O_2). In this case, the mole fraction of Nitrogen in air is 78% or 0.78.

$$\text{Mass Fraction } N2 = \frac{0.78 * 28.01348}{(0.78 * 28.01348) + (0.22 * 31.9988)} = 0.756$$

Example: Determine the mole fraction of Hydrogen (H_2) for a mixture of 1 kg Hydrogen and 100 kg Carbon Dioxide (CO_2).

$$\text{Mole fraction of } H2 = \frac{\left(\frac{1}{101}/2.01588\right)}{\left(\frac{1}{101}/2.01588\right) + \left(\frac{100}{101}/44.0098\right)} = 0.179$$

Accuracy & Stability

The BGA244 makes two primary measurements: the speed of sound in the gas and the gas temperature. It combines those measurements with the gas pressure provided by the user and information about the gas species to report the ratio or purity of the gas. These measurements and data all have uncertainties that can lead to measurement errors.

The speed of sound for an ideal gas is given by the equation:

$$W = \sqrt{\frac{\gamma RT}{M}}$$

From this you can see that the speed of sound strongly depends on the gas temperature. Real gases have additional corrections that depend on the temperature, pressure and gas species.

Speed of Sound Measurement

Speed of sound of the gas can be resolved to <1 mm/s and is extremely stable over time and temperature. The stability and temperature coefficient are largely due to the crystal time base aging. These contribute an insignificant amount to the total measurement accuracy.

The speed of sound accuracy term ($\pm 0.05\%$) does contribute to the total accuracy specification of the BGA244.

Temperature Measurement

The temperature of the gas can be resolved to 0.001 °C and is extremely stable over time. The stability is largely a function of the aging of the glass bead thermistor used to measure the gas temperature. This contributes an insignificant amount to the total measurement accuracy.

The temperature accuracy term (± 0.1 °C) does contribute to the total accuracy specification of the BGA244.

Pressure

The gas pressure is the only measurement provided by the user. While pressure does not affect the accuracy to the same extent as speed of sound or temperature, it is necessary to know the correct pressure to within 1 psi (6.9 kPa) to meet the specified accuracy.

Failure to enter the operating pressure or improperly configuring a pressure gauge will lead to measurement errors. Be certain you are operating with the appropriate pressure units (either absolute or gauge units). Refer to Pressure (page 98) for more details.

Gas Table Data

The BGA244 Factory Gas Table contains detailed information on nearly 500 gases gathered from a number of different sources as described in Appendix A: *Factory Gas Table*. As a general rule, commonly used gases have more accurate data available for than some of the more obscure gases. In rare cases complete data may not exist for a gas. This can degrade accuracy when operating across a range of temperatures and pressures. Contact SRS if you have a question about a particular gas or gas mixture.

Using REL to Improve Accuracy

In some cases, the REL feature can be used to compensate for a number of different error sources, including:

- Uncertainties in the BGA gas model speed of sound (SOS) and calibration. Our gas model & calibration aren't perfect.
- Gas temperature errors. The BGA has a small uncompensated temperature coefficient of < +/-50ppm/degC.
- Pressure errors, either measured or in the BGA calibration. The BGA calibration doesn't perfectly compensate for SOS vs pressure variations. And of course pressure measurement errors will cause it to compensate for the wrong value.
- Thermal gradients: for slow or no flow cases, temperature gradients can form leading to small variations.

Refer to *REL to a Reference Gas* (page 105) for more details about this procedure.

Estimated Accuracy

The Binary Gas Analyzer and Gas Purity Analyzer display the estimated accuracy for the measurement based on a temperature uncertainty of 0.1 °C and a pressure uncertainty of 1 psi (6.9 kPa). The estimated accuracy does not use the REL value in its calculations.

Binary Gas Measurement Accuracy

In the Binary Gas Analysis mode, the gas concentration is determined by measuring the speed of sound of the gas in the cell and its temperature. The amount of each gas is calculated based on these measurements and their physical properties. At a given temperature T, the speed of sound is proportional to

$$\sqrt{\left(\frac{\gamma}{M}\right)}$$

Since γ (the ratio of specific heats) only varies between 1 and 1.67, changes in the speed of sound are largely due to the change of mass, which can vary from 2 to around 200. Because of this, the BGA244 has the best resolution and accuracy when the difference in molar mass of the two gases is largest. After the molar mass, the next most influential parameter is γ (gamma). After these there are several non-ideal gas properties that cause additional shifts. The BGA244 accounts for all these factors.

The best measurement sensitivity occurs when measuring a small amount of a heavy gas in a lighter gas. Absolute accuracies at the 10 ppm level are possible for a small amount of a heavier gas with a carrier of Helium or Hydrogen.

Example: Measuring the ratio of 1% SF₆ (mass 146) in Helium (mass 4) gives an estimated accuracy of ~0.002% (20 ppm).

The REL mode can be used to cancel out the some of the measurement uncertainties. With this relative accuracies of <100 ppm level are achievable for many gas mixtures.

Poor Sensitivity

The measurement sensitivity suffers for mixtures where both gases have similar speeds of sound. This leads to somewhat poorer estimated accuracy.

Example: Nitrogen (mass 28) and oxygen (mass 32) have similar speeds of sound (around 349 and 326 m/s). At an 80%/20% ratio, the estimated accuracy is 0.24%.

Problem Mixtures

Other gas mixtures can be even more problematic. For certain gas combinations there are two different gas ratios with the same speed of sound over a portion of their range. Both of these ratios are equally valid as far as the BGA244 can determine. The BGA244 detects if this occurs and reports both solutions (Two Solutions) as described in *Screen Messages* (page 51).

These mixtures can be measured with some restrictions. For a portion of their range there is only the single molar ratio value. Or if the approximate ratio is known, the larger or smaller solution can be selected.

Problem mixtures can occur for gases with similar molar masses and different molecular types (monatomic, diatomic or polyatomic). Use the Accuracy Calculator in BGAMon to determine if this problem exists for your gas combination, pressure and temperature. In some cases this effect can be minimized by operating at a different pressure or temperature.

Certain problem mixtures exist for all binary gas analyzers. They may be more visible with the BGA244 because it supports many thousands of different gas combinations and concentrations.

Example: A common mixture that exhibits this behavior is Argon (mass 36, $\gamma = 1.67$) in Oxygen (mass 32, $\gamma = 1.45$). At NTP there is a single solution reported for 0 – 43.5%, but two solutions above that.

These problem mixtures can also affect Physical Measurements accuracy, albeit to a far lesser extent. Calibration factors in the BGA244 depend on properties of gases in the resonant cell. These will be incorrect if the “wrong” concentration is used. However this affect is extremely small. A typical worst case error is approximately 30 ppm

Gas Purity Measurement Accuracy

In the Gas Purity Analysis mode the measured speed of sound of the gas in the cell is compared with expected speed of sound at that temperature and pressure. This measurement is sensitive to small amounts of an unknown contaminating gas. However the magnitude of the reported value depends on the species of the contaminating gas. Gas Purity measurements are most sensitive when the contaminating gas(es) have a much higher or lower speed of sound. Gases with similar speeds of sound are problematic. Higher speeds of sound are normally associated with lighter gases, while slower speeds of sound are associated with heavier gases.

Example:

1% of He (a lighter gas) in N₂ give a purity measurement of +0.491%

1% of SF₆ (a heavier gas) in N₂ gives a purity measurement of -2.482%

1% of O₂ (similar mass) in N₂ gives a purity measurement of -0.073%

This problem becomes more difficult when dealing with multiple contaminating gases. As a rule, if the reading is positive the combined contaminants are lighter gases; if negative, the contaminants are usually heavier. However it is possible for the contaminant to be a mixture of both heavier and lighter gases. In this case measurements can be unpredictable.

Long Term Stability

Measurements made by the BGA244 are extremely stable, assuming no dramatic changes in the operating conditions. Long term aging is on the order of 10 ppm/year for SOS measurements. This is primarily due to aging in the crystal time base and the glass bead thermistors.

Short term drift is usually caused by a change in the gases being measured brought about by leakage or outgassing of the Kapton transducers.

High Purity Use

Certain applications are extremely sensitive to any contamination. These may involve low flow rates, sensitivity to contamination, operating below atmospheric pressure or a combination of these. The following section provides some guidelines to help minimize contamination in sensitive systems. For the most stringent applications SRS recommends using the BGA244HP.

Fittings and tubing

Metal tubing and all metal seals should be used. Avoid elastomeric seals as they are porous and prone to outgassing. VCR fittings are recommended for applications operating below atmospheric pressure.

Leak Testing

The BGA244HP is Helium leak tested at the factory, while the standard BGA244 is not. Leak tests are commonly performed on systems after any significant changes are made to the gas manifold or fittings. Depending on the application either outbound or inbound leak testing may be appropriate.

Dead Volumes

There are several small volumes in the BGA244 that are poorly vented to the cell. These total about 1 cc and vent to the cell by diffusion through fairly low conductance paths.

For normal operation these volumes contain the same gas as the rest of the cell and have no effect on measurements. However if the gas species has changed significantly, the gas in these volumes will gradually diffuse into the cell over a few minutes. The diffusion occurs faster for lighter gases than for heavy gases. For low flow rates this may introduce a small error for the first few minutes after the gas species has changed.

If this presents a problem, the cell can be briefly evacuated under vacuum to vent these volumes. Otherwise flow gas for a few minutes to allow these volumes to diffuse into the cell.

Outgassing

Outgassing is only a concern when a cell has been exposed to moist air, solvents, or if the gas species in the cell have changed. Normally it can only be detectable at very low flow rates and will diminish over time. To accelerate outgassing, follow the Degas procedure later in this section.

The primary source of outgassing in the BGA244 is the Kapton acoustic transducers. While Kapton has excellent resistance to nearly all chemicals, it is hydrophilic and porous to some gases. Substances that are readily absorbed by Kapton include water, alcohols and many solvents (generally polar molecules).

There is about 2.9 in² (18.5 cm²) of 1 mil (0.025 mm) thick Kapton in the two acoustic transducers. If fully saturated, they can absorb around 2.7 g of water. At NTP this becomes about 3.3 cc of water vapor. Similar amounts of other polar molecules will also be absorbed.

Kapton is also somewhat porous to many gases. Gases present in the cell will diffuse into the transducers. If the gas is changed or the cell is put under a vacuum, this gas will gradually diffuse out. This will have no effect if the gas species hasn't changed, but it can be detected if the cell is filled with a different gas or under vacuum.

Does outgassing matter? In most cases it will be undetectable. For a fully saturated membrane at room temperature, the outgassing rate of Argon, Nitrogen or water vapor starts at about 1×10^{-4} sccs and decreases over time. For a 500 sccm flow this corresponds to a contamination of about 12 ppm. However for very slow (or zero) flow rates it can become a significant source of contamination.

The gases absorbed by the Kapton can be largely removed using the following Degas procedure. This will reduce the effect of outgassing from the Kapton to below 0.1 ppm for the same 500 sccm flow.

Always minimize exposure to the atmosphere or other contaminants to reduce the amount absorbed. Fill the cell with a dry gas and cap it off if not in use. Follow the Degas procedure if exposure has occurred or different gasses are being used.

Degas

The Degas and Block Heaters can be used to dramatically reduce the outgassing rate of the Kapton transducers. This procedure is optimally done under vacuum where the outgas rate can be monitored. If necessary it can also be done while flowing a dry gas through the cell for an extended period of time.

Degas requires an external +24 V power supply rated for at least 1A. If degas is being done under vacuum, a residual gas analyzer, ion gauge or other suitable vacuum gauge can be used to track the outgassing rate until it reaches an acceptable level. The SRS RGA100/200/300 and IGC100 are well suited for this. Refer to *Heaters* (page 78) for details on operating the heaters.

Under Vacuum

Set the block heaters to 70 °C and turn on Degas. The initial outgassing rate of the Kapton transducers will be around 1×10^{-4} sccs for water vapor, Argon or Nitrogen. After the heaters are first turned on, the rate will initially increase. Typically the rate should be around 5×10^{-5} sccs after a few minutes and then decrease exponentially over time.

Run degas until the outgas rate has dropped to around 1×10^{-6} sccs or whatever level is acceptable if the outgas rate is being monitored. Otherwise run for at least 3 hours.

When complete, turn off Degas and the Block Heater. Remember that the BGA244 cannot make gas measurements under vacuum. Fill the cell with a dry gas and cap it off if not in use.

While Flowing Gas

This procedure can also be performed while flowing a dry gas through the BGA244. It may take longer to reach the same outgas rate since the Kapton transducers are cooled by the flowing gas. Ideally the dry gas should be one of the gases being used in operation so a different gas species isn't absorbed into the membrane.

Set the block heaters to 70 °C and turn on Degas. Flow at least 10 sccm of the dry gas through the BGA244. Run Degas for at least 3 hours. This should reduce the outgas rate to around 1×10^{-6} sccs.

When Degassing is completed, turn off Degas and the Block Heater. Remember that the BGA244 cannot make gas measurements when Degas is on. Fill the cell with a dry gas and cap it off if not in use.

User Gases

The BGA244 Factory Gas Table contains nearly 500 different gases. However there are many more gases that can be measured, including mixtures or blended gases. Blended gases are made up of two or more single species gases. The BGA244 supports a User Gas Table that allows a nearly unlimited number of new gases to be added by the end user.

Gases in the BGA244s Factory Gas Table have detailed information about how a gas behaves over a wide temperature and pressure range, as well as how it interacts with any other gas. However gases don't always need to be specified this completely to make accurate measurements, especially at fairly constant temperatures and pressures.

User Gases are modeled using the ideal gas law. This is a reasonable approximation of the behavior of many gases, especially at constant temperature and pressure. However it can lead to errors when operating over wide ranges of temperature, pressure or ratios. But remember that repeatability is often more important than accuracy. As long as you operate at a similar temperature and pressure, the measured values won't change much.

Contact SRS if you are unable to obtain adequate results with a User Gas. In some cases we are able to create a complete model of a gas and add it to the Factory Gas Table.

Factory Gas Table

The Factory Gas Table contains a comprehensive list of data for each gas, including names, formula, molar mass and various thermodynamic properties. This information allows accurate measurements to be made on a wide variety of gases over the entire operating range of the BGA244.

See Appendix A (*Gas Table*) for a detailed explanation of all of the terms contained in the Factory Gas Table and their effects on measurements.

User Gas Table

The User Gas Table is used to store data on gases that aren't available in the Factory Gas Table. This information is stored using the BGAMon Software. The following parameters can be entered for User gases: Name, molar mass and γ (ratio of specific heats).

Selecting a User Gas

User gases show up in gas selection searches as "User/Gas Name". See *Selecting Gases* in the *Operations Guide* for more information.

User Gas Parameters

User gases must have a name, molar mass and a single thermodynamic parameter γ (gamma). The name is used when selecting the gas and to display on the Home page. The molar mass is the atomic weight of the molecules multiplied by their concentration. γ is ratio of specific heats.

The following section describes the technique for adding a User Gas to the BGA244. Most User Gases added will be blends of other gases since the BGA244 Factory Gas Table contain nearly every non-blended gas.

Mass

Finding the mass of the gas is fairly simple. For a single species gas, it's just the molecular mass in AMUs (atomic mass units).

To calculate the molar mass of a blended gas, take the molar (mole) fraction of each gas times its mass. Assuming MF1 and MF2 are the mole fractions of each gas.

$$mass = (mass1 \times MF1) + (mass2 \times MF2)$$

This equation can be extended to all the gases that make up the blend.

γ

γ (gamma) is the ratio of specific heats or heat capacity ratio in a gas. γ is a unitless quantity that can be expressed in terms of C_p (heat capacity at constant pressure) and C_v (heat capacity at constant volume) or C_p and R (the ideal gas constant).

$$\gamma = \frac{C_p}{C_v} = \frac{C_p}{C_p - R}$$

γ (gamma) and C_p values for many gases can be found in literature or on line. Both C_p and R can be expressed in different units so make sure that you use the same units for both when calculating γ .

The following rules are generally true for single species gases at room temperature and near atmospheric pressure:

- Monoatomic gases $\gamma \approx 1.66$
- Diatomic gases $\gamma \approx 1.4$
- Polyatomic gases $\gamma \approx 1.05 - 1.25$

Try to obtain values for γ at or near the temperature and pressure the BGA244 will be operating for best accuracy.

Since the Factory Gas Table contains nearly all of the monatomic and diatomic gases most added gases are either polyatomic or blends. γ varies from 1.1 – 1.2 for most polyatomic gases.

Hint: If you can't find a value for γ for the gas you are adding, use a value from a similar gas.

Determining γ is a bit more complicated for blended gases. Assuming that

- γ_1 is for the gas 1, γ_2 is for gas 2 and γ_b is for the blend
- MF1 and MF2 are the mole fractions of each gas.

Then

$$\frac{\gamma_b}{\gamma_b - 1} = \left(\frac{\gamma_1}{\gamma_1 - 1} \times MF1 \right) + \left(\frac{\gamma_2}{\gamma_2 - 1} \times MF2 \right)$$

This equation can be extended to all the gases that make up the blend.

Adding a User Gas with BGAMon

User gases can be added using BGAMon See Chapter 5: *BGAMon* for information on installing and using the BGAMon software.

Click [Page] [Gas Table] to access the Gas Selection page. Press [Add Gas] in the User Gas Table area to open the Add User Gas dialog box. Enter the name, molar mass and γ (ratio of specific heats). Click [OK] to save the new gas to the BGA244 User Gas Table.

User Gases are selected the same way as gases from the Factory Gas Table. They are identified by "USER/*gas name*".

Adding a User Gas over a Computer Interface

User Gases can also be added over the Remote Interface using the AUSR command. The arguments for the AUSR command are the molar mass, gamma and name.

Example

An approximation for air is 79% N₂ and 21% O₂. This has a molar mass of 28.86 and a gamma of 1.4. So the command to enter this over the remote interface is:

```
AUSR 28.86,1.4,Air
```

User Gases are selected using the same commands as Factory Gases. Instead of using the CAS #, use the string: "User/*gas name*".

Adding Gases

Example: Adding a Single Species Gas (Argon)

To add a gas, you need the three previously mentioned items: a name, molar mass and γ . For this example we will add a simple model for Argon to the User Gas Table.

- Name = Argon
- Molar Mass = 39.95 (same as the atomic mass)
- $\gamma = 1.66$ (Argon is monatomic)

Example: Adding a Blended Gas (Air)

Blended gases are made up from two or more single species gases. The molar mass and γ must be calculated based on the ratio of the gases in the blend.

The most familiar blended gas is air. As a commercial product, air is usually considered to be 79% Nitrogen and 21% Oxygen.

To find the molar mass of the gas, take the mole fraction of each gas times its mass.

Gas	Mol Fraction	Mol Mass	Fractional Mass
N ₂	0.79	28.02	22.14
O ₂	0.21	32.00	6.72
Sum			28.86

Note that the molar mass is slightly different than the value from the Factory Gas Table. This is because Air in the Factory Gas Table is atmospheric air (78.09% N₂, 20.95% O₂, 0.93% Ar). The extra mass of Argon (39.95) increases the total mass slightly. However since it is such a small fraction of the total it has no real effect on the thermodynamic data (γ).

- Name = Air
- Molar Mass = 28.86
- $\gamma = 1.4$ (both gases are diatomic)

Example: Adding a Blended Gas (Heliox)

Another common blended gas is Heliox 80:20 (He:O₂). The calculations are a little more complicated than air since Helium is monatomic and Oxygen is diatomic.

To find the molar mass of the gas, take the mole fraction of each gas times its mass.

Gas	Mol Fraction	Mol Mass	Fractional Mass
He	0.80	4.03	3.22
O ₂	0.20	32.00	6.40
Sum			9.62

γ mixes as the mole fraction of the two gases.

To find γ for the blend, apply γ and the mole fraction of each gas to the formula above.

Gas	Mol Fraction	γ	Blend γ
He	0.80	1.66	
O ₂	0.20	1.40	
			1.58

- Name = Heliox 80-20
- Molar Mass = 9.62
- $\gamma = 1.58$ (calculated above)

Operating with User Gases

Typically there will be larger pressure and temperature dependent errors associated with User entered gases, since they are not specified as thoroughly as factory gases. Try to operate at a fairly constant temperature and pressure to avoid these errors.

Changes in gas parameters with temperature and pressure may create significant errors in some cases. But there are plenty of situations where they will have an insignificant effect.

- If the User gas is used as a dopant to a much lighter factory gas, the differences in mass may overwhelm any small errors in the gas model.
- The thermodynamic properties of the user gas may not vary much over the operating pressure and temperature range.
- Many processes depend more on changes in values rather than the actual value. REL can be used to establish a baseline and variations can be tracked from that value.

If possible, find values for γ at or near the operating temperature and pressure for best accuracy. Operate at a stable temperature. The block heater is an easy way to achieve this.

Invalid Gas Message

Stored settings (in the BGA244) and stored configuration (on BGAMon) can reference User Gases. An error is created if a setting or configuration tries to recall a non-existent User Gas. If this occurs, the screen message “Invalid Gas” is displayed and an error is reported. To correct this first select a different gas, then re-add the User Gas and select it.

Using REL with User Gases

In some cases the BGA244 will operate at high concentrations of the User gas. In this case you can REL to the User gas as a Reference. Follow the procedure *REL to a Reference Gas* described earlier in this chapter.

Chapter 5: BGAMon

BGAMon is a Windows based program for controlling the BGA244. It can:

- Display and configure all instrument parameters
- Save and recall instrument configurations to disk
- Perform data acquisition
- Add User Gases
- Update the BGA244 firmware

BGAMon is extremely useful to configure units with accessory BGA-M installed (metal cover) or units installed in remote or inaccessible locations.

This chapter provides details on using BGAMon but does not describe operation of the BGA244 in detail. See Chapter 3: *Operation Guide* for specifics on operating the BGA244.

The following conventions are used to refer to controls and messages for BGAMon. In general **bold text** is used to indicate a control or reading on BGAMon

- **[Button]** refers to a button that can be clicked to execute a function.
- **<Entry>** indicates that a value needs to be entered where “Entry” is the name of the parameter. Use the keyboard to enter the value. If the entered value is out of range, it will be ignored.
- **<Entry ↓>** indicates that a value needs to be selected from a list where “Entry ↓” is the name of the list. Pressing the key will open a drop down list of possible selections. [Press] the desired selections to choose it.
- **Choice** indicates a “radio button” that is used to select one of a number of possible choices. Press the circle to select that choice and de-select all other choices.
- **Choice** indicates a check box that is used to enable a choice. Unlike Radio Buttons, you can select any or all of check boxes as needed.
- **(Here/There)** refers to a location accessed from the Menu bar. For example to reach (Page/Pressure) you would first click the Page menu and select Pressure.
- **{Message}** refers to a popup Windows dialog box. This box will contain buttons and possible other controls. This box must be closed to return to the normal operation of BGAMon.

Installation

This section describes how to install and begin using BGAMon. BGAMon communicates with the BGA244 over USB. It requires a USB driver be installed to operate.

Drivers

BGAMon requires that the FTDI D2XX USB driver has been installed on the computer for proper operation.

Note: Windows may check for a new driver each time a different BGA244 is connected to the computer.

The first time the BGA244 is connected to the USB port of a computer running Microsoft Windows, you will likely be prompted with a “New Hardware Found” message and an invitation to search for the USB Driver. There are two USB drivers for the device (VCP and D2XX drivers). Depending on the version and configuration, Windows may either automatically install the drivers or prompt you to search for them. Allow Windows to install the drivers. Occasionally only a single driver will install. In some cases disconnecting and reconnecting the BGA244 will cause the second driver to load correctly.

If the drivers do not load successfully see Appendix E for details on manually installing the drivers.

Note: In some circumstances it may take a long time to load the drivers (up to 30 minutes or more). This is primarily caused by high demand on the Microsoft Update website. If possible, try again later where there will (hopefully) be less traffic. Otherwise use the manual installation procedure described in Appendix E (page 255).

Downloading BGAMon

The BGAMon Installer can be found on the thinksrs.com website. There is also a readme.txt file that contains information on the latest version of BGAMon. Use the following steps to download, extract and install BGAMon to your computer.

- Go www.thinksrs.com. Navigate to the Binary Gas Analyzer page.
- Click on the BGAMon Software link under “Downloads”. Save the BGAMon zipped file to your computer.
- Go to the folder where the file was downloaded. Double click on the file to extract the BGAMon Installer.
- Double click the BGAMon Installer icon to launch the BGAMon Installer. Follow the installer instructions.

After installation the BGAMon icon should appear on your desktop. Double click it to start BGAMon.

BGAMon Quick Start

Use the following procedure to begin using BGAMon to control the BGA244. If you encounter problems with any of the following steps see *Troubleshooting BGAMon* (page 144) for further instructions.

Connect the BGA244 to the computer's USB port using a USB Type A to USB Type B cable. If you plan power the BGA244 using a 24 VDC power supply, connect it prior to connecting the USB port to the computer. Be sure the USB power supply and cable are sufficient to operate the BGA244 if using USB for power. See *Power* (page 22) for details on appropriate power supply and cabling.

Double click on the BGAMon icon to start the program. Wait a few seconds after the BGA244 is plugged in so Windows can detect that it's connected. Then click [**Connect to Device**] on the BGAMon System area. The message adjacent the button should change from ***NOT CONNECTED*** to ***CONNECTED***.




FIGURE 19: BGAMON SNAPSHOT SCREEN

BGAMon starts on the Snapshot page. Here you can see the primary measurements and settings for the BGA244. Assuming the BGA244 has come from the factory and the gas port caps haven't been removed, it should display < 0.1% $\Delta W/W$ for Argon.

Using BGAMon

The operating state of BGAMon is based on the connected BGA244. Changes made to either the BGA244 or BGAMon are automatically transferred to the other. Any changes made to BGAMon when not connected are lost as soon as the BGA244 is connected.

The BGAMon screen consists of four main areas: the Menu bar, System area, Instrument page and Status bar.




FIGURE 20: BGAMON LAYOUT

Menu Bar

The Menu bar contains the navigation functions to access the different parts of BGAMon. These include File, Device Page and Utilities.

File

The File menu accesses the Load and Save Configuration pages. See *Loading and Saving Configurations* (page 142) for details.

Device

Detect Device can be used to see if the BGA244 USB driver is properly installed and the BGA244 has been detected by Windows. See *Troubleshooting BGAMon* (page 144) for details.

Page

The BGAMon instrument pages are accessed from this Menu. These include:

- Snapshot: Overview of most of the instrument settings and measurements
- Analog Input/Output: Configuration for the analog inputs and outputs
- Pressure: Configuration for pressure sensing and entry
- Limits: Configuration for all Limit values
- Events: Configuration and control of the Event Relays
- Strip Chart: Access and display for the BGAMon data acquisition
- Gas Selection: Gas Table and gas selection controls
- Faults/Alerts: Monitor and control for all active Fault and Alert conditions
- Computer I/O: Configuration for RS-232 and RS-422 interfaces
- Store/Recall: Store and Recall BGA244 settings to flash memory

Utilities

Utilities are functions that aren't part of the normal operation of BGAMon.

- See *Updating Firmware* (page 143) for details on updating the BGA244 firmware.
- See *Utility Mode* for information on the Utility Mode.

System Area

This area contains controls and indicators that are visible for all the instrument pages.

Fault

[FAULT] is a dual purpose control. First it indicates that a fault has occurred by turning red. Clicking on it when active opens the Faults/Alerts page for additional information about the active faults.

Alert

[ALERT] is also a dual purpose control. It indicates that an alert has occurred and clicking on it when active also opens the Faults/Alerts page for additional information about the active alerts.

Running

The **[RUNNING|STOPPED]** key acts the same as the [RUN|STOP] key on the BGA244 setup page. When [STOPPED] is selected most measurements on the BGA244 are halted. See Run|Stop (page 58) for details.

Screen Messages

Screen Messages may appear under the Fault and Alert indicator buttons. These are the same messages that appear on the BGA244 Home pages. See *Screen Messages* (page 50) for more information.

Connect or Disconnect

The **[Connect]** or **[Disconnect]** buttons connects or disconnects the BGAMon software with the BGA244. The current state is indicated by the ***CONNECTED*** or ***DISCONNECTED*** message that appears next to the button.

See *Troubleshooting BGAMon* (page 144) for further instructions if a **{Device Not Found}** or **{Unable to open BGA Device}** message box appears.

Autosave

Click **[AUTOSAVE NOW]** to immediately save the BGA244 configuration to flash. Autosave ordinarily will not occur until about 30 seconds after the last change to the BGA244 configuration occurs. Recent changes may not be saved if power is removed prior to Autosave occurring.

Instrument Page

The Instrument Page is selected from the Page Menu Bar. This section displays the BGA244 measurements and configuration. See the Instrument Pages in the following section for details on the different pages.

Status Bar

The Status Bar displays the operating state of BGAMon.

Instrument Pages

The Instrument pages show the operation of the BGA244. Most of the settings and measurements in BGAMon are the same as the BGA244 GUI. They are organized a bit differently to take advantage of the larger screen. The BGAMon functions are described in terms of their counterparts in the BGA244 GUI. See Chapter 3: *Operations Guide* for details on the use and meaning of the different controls. One feature that is unique to BGAMon is the Strip Chart. This can acquire, graph and store BGA244 measurements over intervals ranging from seconds to days.

Snapshot

Snapshot is an overview of the key measurements and settings of the BGA244. It combines the functions of a number of different pages including the Home page, Setup page and a number of Control Panel pages.

Analysis Results		Unit Families		Status	
Instrument Mode	Binary Gas Analyzer	Ratio	%	Speed of Sound	325.288 m/s
<input type="button" value="Gas Selection..."/>		Speed	m/s	Speed of Sound (NTP)	318.894 m/s
Binary Gas Analyzer		Temperature	C	Gas Temperature	31.836 °C
Primary	Argon	Pressure	psi	Analysis Pressure	0.00 psi (g)
	100.052 %	Heater Control		Block Temperature	33.196 °C
Secondary	Hydrogen	Heater	<input checked="" type="radio"/> Off <input type="radio"/> On	PCB Temperature	32.443 °C
	-0.052 %	Heater Temp	70 °C	Heater Power	0.00 W
Accuracy	+/- 0.049 %	Max Current	2.00 A	USB Voltage	4.9500 V
Rel		Power	0.00 W	External Voltage	24.4210 V
<input checked="" type="radio"/> Off <input type="radio"/> On		Block Temp	33.196 °C	Analog In 1	-0.000 mA
<input type="button" value="Rel to 100%"/>		Regulation	N/A	Analog In 2	0.0001 V
<input type="button" value="Rel to 0%"/>		Degas Heater	<input checked="" type="radio"/> Off <input type="radio"/> On	Meas Out	20.000 mA
				Analog Out 1	1.0000 V
Gas Temperature	31.836 °C			Analog Out 2	0.0000 V
Pressure	0.00 psi (g)			Measurement	
				Average Mode	On
				Number of Averages	10
				Amplitude	Auto
				Binary Gas Measurement	Mole Fraction
				Relaxation Correction	On

FIGURE 21: SNAPSHOT PAGE

Analysis Results

This section selects the instrument mode, the measured gas and REL. It reports the measured value, estimated accuracy, gas temperature and pressure. Most of the reported values and controls depend on the selected Instrument Mode. See the Home page for each Instrument Mode in Chapter 3 for further details.

Select Binary Gas Analyzer, Gas Purity Analyzer or Physical Measurements using **<Instrument Mode ↓>**

Click **[Gas Selection]** to jump to the Gas Selection page to select gases.

The Measured Value and Estimated Accuracy are determined by the Instrument Mode, selected gases and measured value. The Gas Temperature and Analysis Pressure are displayed at the bottom of this section.

User Speed of Sound and REL are specific to particular Instrument Modes. See the Home page for each Instrument Mode in Chapter 3 for details.

Unit Families

These are the settings from the Units page in the Control Panel. See *Units* (page 74) for details.

Heater Control

These are the settings and measurements from the Heaters page in the Control Panel. See *Heater* (page 78) for details.

Status

These are the measured values from the Status page in the Control Panel. See *Status* (page 77) for details.

Measurement

These are the settings from the Measurement page in the Control Panel. See *Measurement* (page 75) for details.

Analog I/O

The Analog I/O page combines the separate Analog Input and Output pages from the BGA244 Control Panel. See *Analog I/O* (page 81) for details.




FIGURE 22: ANALOG I/O PAGE

Note: The Analog Inputs can be configured from either the Pressure page or Analog I/O page. Changes made in one location will be reflected in the other location.

Pressure

The Pressure page combines all of the separate Pressure pages from BGA244 Setup menu. See *Pressure* (page 64) for details.

Use for Analysis: User Entered (dropdown) Analysis Pressure: 0.00 psi (g)

Ambient Pressure: 14.70 psi

Analog Input 1

Disabled Enabled Enable as Pressure Gauge

Input Type: Voltage (dropdown) Units: Absolute Gauge

Loop Power Voltage: 12.0 V Min (0 V): 0.00 psi

Input Reading: -0.0003 V Max (10 V): 150.00 psi

Reading: 0.00 psi

Analog Input 2

Disabled Enabled Enable as Pressure Gauge

Input Type: Voltage (dropdown) Units: Absolute Gauge

Loop Power Voltage: 12.0 V Min (0 V): 0.00 psi

Input Reading: -0.0001 V Max (10 V): 150.00 psi

Reading: 0.00 psi

User Defined Pressure

Units: Absolute Gauge

User Pressure: 0.00 psi

FIGURE 23: PRESSURE PAGE

Note: The Analog Inputs can be configured from either the Pressure page or Analog I/O page. Changes made in one location will be reflected in the other location.

Limits

The Limits page combines the limit settings from all of the separate graphs that have limits. This includes limits for the Binary Gas Analyzer, Gas Purity Analyzer and Physical Measurements (all *Limits*, page 54), plus the Pressure Meters (*Pressure* page 64) and Temperature Meter (*Temperature* page 67).

<p>BGA Limits</p> <p>Active</p> <p><input checked="" type="checkbox"/> Upper 102.000 %</p> <p><input checked="" type="checkbox"/> Lower 98.000 %</p> <p>Hysteresis 1.000 %</p> <p>Upper Meter Scale 105.000 %</p> <p>Lower Meter Scale 95.000 %</p>	<p>Purity Analyzer Limits</p> <p>Active</p> <p><input checked="" type="checkbox"/> Upper 2.000 %</p> <p><input checked="" type="checkbox"/> Lower -2.000 %</p> <p>Hysteresis 0.100 %</p> <p>Upper Meter Scale 5.000 %</p> <p>Lower Meter Scale -5.000 %</p>	<p>Physical Measurements Limits</p> <p>Active</p> <p><input type="checkbox"/> Upper 1000.000 m/s</p> <p><input type="checkbox"/> Lower 0.000 m/s</p> <p>Hysteresis 1.000 m/s</p>
<p>Pressure Meter 1 Limits</p> <p>Active</p> <p><input type="checkbox"/> Upper 100.00 psi</p> <p><input type="checkbox"/> Lower 10.00 psi</p> <p>Hysteresis 1.00 psi</p> <p>Upper Meter Scale 150.00 psi</p> <p>Lower Meter Scale 0.00 psi</p>	<p>Pressure Meter 2 Limits</p> <p>Active</p> <p><input type="checkbox"/> Upper 100.00 psi</p> <p><input type="checkbox"/> Lower 10.00 psi</p> <p>Hysteresis 1.00 psi</p> <p>Upper Meter Scale 150.00 psi</p> <p>Lower Meter Scale 0.00 psi</p>	<p>Temperature Limits</p> <p>Active</p> <p><input checked="" type="checkbox"/> Upper 40.000 °C</p> <p><input checked="" type="checkbox"/> Lower 20.000 °C</p> <p>Hysteresis 1.000 °C</p> <p>Upper Meter Scale 50.000 °C</p> <p>Lower Meter Scale 10.000 °C</p>

FIGURE 24: LIMITS PAGE

Events

The Events page combines the BGA244 Event 1 and Event 2 Relay Configuration pages. See *Event Relay 1 & 2* (page 55) for details.

<p>Event Relay 1</p> <p>Event Relay 1 Configuration</p> <p><input type="checkbox"/> Above Limit</p> <p><input type="checkbox"/> Below Limit</p> <p><input type="checkbox"/> Pressure Meter 1 Above Limit</p> <p><input type="checkbox"/> Pressure Meter 1 Below Limit</p> <p><input type="checkbox"/> Pressure Meter 2 Above Limit</p> <p><input type="checkbox"/> Pressure Meter 2 Below Limit</p> <p><input type="checkbox"/> Temperature Above Limit</p> <p><input type="checkbox"/> Temperature Below Limit</p> <p><input type="checkbox"/> No Signal</p> <p><input type="checkbox"/> System Fault</p> <p>Force On Force Off</p>	<p>Event Relay 2</p> <p>Event Relay 2 Configuration</p> <p><input type="checkbox"/> Above Limit</p> <p><input type="checkbox"/> Below Limit</p> <p><input type="checkbox"/> Pressure Meter 1 Above Limit</p> <p><input type="checkbox"/> Pressure Meter 1 Below Limit</p> <p><input type="checkbox"/> Pressure Meter 2 Above Limit</p> <p><input type="checkbox"/> Pressure Meter 2 Below Limit</p> <p><input type="checkbox"/> Temperature Above Limit</p> <p><input type="checkbox"/> Temperature Below Limit</p> <p><input type="checkbox"/> No Signal</p> <p><input type="checkbox"/> System Fault</p> <p>Force On Force Off</p>
--	--

FIGURE 25: EVENTS PAGE

Strip Chart

The BGAMon Strip Chart can acquire, plot, print and export measurements from the BGA244 as a function of time. The following measurements can be captured, printed and exported to a file.

- Primary and Secondary Gas Concentrations
- Gas Purity
- Measured Speed of Sound and Speed of Sound normalized to NPT
- Gas Temperature
- Heater Power
- Pressures (User, Pressure Gauge 1 & 2, Analysis Pressure)
- Analog Input 1 and 2




FIGURE 26: STRIP CHART PAGE

Measurements

BGAMon continuously records 13 different measurements when connected to the BGA244. All of the data taken is stored regardless of what is displayed on the chart. Measurements are started when BGAMon is connected to the BGA244. They are stopped if BGAMon is disconnected from the BGA244. Measurements are made at the BGA244's measurement rate (~4 Hz) and are time stamped by the BGA244 to eliminate any time uncertainty from Windows.



Note: Data is not saved if program is closed. Either print the graph or export the data to a file to save it after the program is closed.

Click **[Clear Data]** to clear the existing strip chart data and start collecting new data. Note that all previous data is permanently lost when this button is clicked.

Measurements for the Binary Gas Analyzer, Gas Purity Analyzer and Physical Measurements are always being captured, regardless of which instrument mode is selected.

Tip: Ordinarily BGA and GPA measurements aren't displayed at the same time. One case where it may be useful to display both is when dealing with nearly pure gases.

Certain measurements can be disabled in the BGA244. These include Heater Power, User Entered Pressure, Pressure Meters 1 & 2 and Analog Inputs 1 & 2. Disabled measurements will hold their last valid reading. That valid reading will be zero (0) if they haven't been enabled since the BGA244 was connected.


Units for the measurements are fixed to the values listed below. They do not reflect the Global units set on the BGA244.

- BGA Primary & Secondary Concentration (%): Binary Gas Analyzer Primary & Secondary Gas concentrations reported in %.
- GPA Delta Speed (frac): Gas Purity Analyzer $\Delta W/W$ reported in fraction.
- PHYS Speed of Sound (Meas) (m/s): Physical Measurements Measured Speed of Sound reported in m/s.
- PHYS Speed of Sound (NTP) (m/s): Physical Measurements NTP Speed of Sound reported in m/s.
- Gas Temp (C): Gas Temperature reported in °C.
- Heater Power (W): Heater Power reported in Watts. Value is 0 if disabled.
- User Entered Press. (psi): User Entered Pressure reported in psi. This will read in either absolute or gauge units depending on the BGA244 User Pressure units.
- Press. Gauge 1|2 (psi): Value from Pressure Gauge 1|2 reported in psi. This value depends on the BGA244 Gauge 1|2 configuration parameters. Value is 0 if disabled.
- Analysis Press. (psi): Value of the selected Analysis Pressure. This value depends on the BGA244 analysis pressure setting and the parameters associated with it.
- Analog In 1|2 (V or mA): Value of Analog Input 1 reported in either V or mA. The units depend on whether voltage or current input is selected on the BGA244. Value is 0 if disabled.

Graph

Click the desired Measurement check box to display it on the Strip Chart. Traces are assigned colors in the order they are added to the chart. Note that measurements for File Export are selected on the Export page independent of the graph.

Every point acquired is displayed on the strip chart for time scales of 30 seconds per division or less. For time scales of 1 minute or greater there are more data points than pixels available on the strip chart. In this case either the mean or the minimum, mean and maximum values can be displayed for each time point. The mean value is plotted in a darker color than the minimum and maximum values. Right click on the trace to open the **{Show Min/Max}** dialog box. Check or uncheck the box to enable or disable the Min/Max feature.




The **[Chart Title]** can be entered to help identify printed graphs.

Time Axis

The Time Axis operates in two different modes. Use **[Display Mode ↓]** to select between the Scroll and Fixed modes.

The **Scroll** mode displays the last “n” points of data taken. The Time Axis scrolls left as needed to show these points. Use the Time per Division display to show more or less data. This mode is commonly used while collecting data as it lets you view trends over either long or short time intervals.




The Time Axis is static for the **Fixed** mode. Use the Time per Division control to zoom about the cursor position. The left and right keyboard arrows can be used to pan the time axis data when zoomed in.

The Time Axis can be set between 5 sec to 50 days per division using the **slider control**. This enables BGAMon to display data over a wide range of time scales.

Y Axis

Each Measurement has a separate Y Axis that can be scaled independently. Each scale can be set using the **<Minimum>** and **<Maximum>** entry boxes, the up and down keyboard arrow keys or Autoscale.



[Autoscale Now] scales the minimum and maximum values at the moment it was pressed. **Always Autoscale Y Axis** will automatically rescale the display for each value acquired.

While many traces can be simultaneously displayed, only a single Y axis can be active at a given time. The active Y axis is indicated by the selected measurement and by the Y axis unit's label.

Cursor

The cursor reports of the value of the active trace (the selected Y axis) at the cursor Time axis value. Use the (computer) mouse to drag the cursor along the time axis to the desired location. If Show Min/Max is active the cursor reports the Mean value.

Cursor	
Time	1.266667 min
Primary Gas	100.0482 %

Printing

Click **[Print]** to open the Strip Chart Print Preview dialog box. From here you can select and setup installed printers and adjust the appearance of the strip chart prior to printing.

Select and configure the desired printer from Print Preview. Set the **Orientation** and **Margins** to scale the strip chart to the desired size. Use the **Detail** slider to adjust the font size and grid spacing. Click **[Print]** to send the preview graph to the printer when you are satisfied with the print preview.

Note: The print preview dialog box will not automatically close after **[Print]** is clicked. Click **[Close]** after the graph has been printed to close the window.




FIGURE 27: PRINT PREVIEW WINDOW

Export

Click **[Export]** to open the Strip Chart Export dialog box. From here you can select the measurements and time scale of the exported data. Data can be exported in either CSV (comma separated value) or TXT (text) format.

Select the measurements to be exported under **Measurements** or check **Select All**. Check **Save Raw Data** if you want to save every point of the acquired. Otherwise select the time interval between saved points using the **Measurement Time Interval** slider. Set the Start and End range using the either the **Range** sliders or **<Start>** and **<End>**.

Tip: Raw Data export files can be really big if you are saving anything longer than about an hours' worth of data. It normally works better to use a longer measurement time interval if you want to save more than 30 minutes of data.

Include Instrument Information prepends the time and date the file was saved, the unit serial number, firmware build number/date and calibration date. **Additional Comments** can also be added to the instrument information to document the data.

After the data to be saved has been properly configured, click **[OK]** to open the Windows **Save As** dialog box and save the file.




FIGURE 28: EXPORT WINDOW

Gas Selection

The measured gases used by the BGA244 are chosen on the Gas Selection page. The page includes a searchable copy of the Factory Gas table, an editable User Gas Table and controls to select the measured gases. See *Selecting Gases* (page 58) for specific details on setting gases.

Gas Selection

Primary Argon
Secondary USER/Heliox 80:20

Set Selected as Primary
Set Selected as Secondary

Selected Gas	
CAS #	7440-37-1
Formula	Ar
Names	Argon
Molar Mass (g)	39.948
Heat Capacity gamma	1.67
Speed of sound (m/s) @ NTP	318.889

Gas Table

Number of Entries: 921 Filter

Name	CAS #	Formula	Molar Mass	Heat Capacity	Speed of Sound
Methyl alcohol	67-56-1	CH4O	32.042	1.235	306.484
Hydrazine	302-01-2	H4N2	32.045	1.208	303.183
Fluoromethane	593-53-3	CH3F	34.033	1.288	303.693
Methyl fluoride	593-53-3	CH3F	34.033	1.288	303.693
R-41	593-53-3	CH3F	34.033	1.288	303.693
Hydrogen sulfide	7783-06-4	H2S	34.081	1.323	307.605
Hydrogen chloride	7647-01-0	HCl	36.461	1.399	305.834
Fluorine	7782-41-4	F2	37.997	1.363	295.703
Argon	7440-37-1	Ar	39.948	1.667	318.888
1-Propyne	74-99-7	C3H4	40.064	1.161	265.743
Methyl acetylene	74-99-7	C3H4	40.064	1.161	265.743
Prop-1-yne	74-99-7	C3H4	40.064	1.161	265.743
1,2-Propadiene	463-49-0	C3H5	40.064	1.167	266.405
Allene	463-49-0	C3H5	40.064	1.167	266.405
Propadiene	463-49-0	C3H5	40.064	1.167	266.405
Acetonitrile	75-05-8	C2H3N	41.052	1.192	266.016
1-Propene	115-07-1	C3H6	42.080	1.150	258.128
Prop-1-ene	115-07-1	C3H6	42.080	1.150	258.128

User Gas Table

Refresh User Gas Table

Name	Molar Mass (g/mol)	Heat Capacity
USER/Heliox 80:20	9.620	1.580
USER/air	28.900	1.400

Add Gas
Remove Gas

FIGURE 29: GAS SELECTION PAGE

Factory Gas Table

The Factory Gas Table is automatically loaded when BGAMon connects to the BGA244. Gases can be sorted by name, CAS #, chemical formula, molar mass, heat capacity (γ) or speed of sound (NTP).

To scroll through the gas table use the slider bar or up and down arrow keys. Use **[Filter]** to search the entire gas table for matching entries.

User Gas Table

The User Gas Table is also loaded when the BGA244 connects to BGAMon. The User Gas Table is empty unless gases have been previously added by the user. See *Adding and Removing User Gases* below for details.

Set Gases

First select the desired gas from either the Factory or User Gas Tables by clicking on it. Data for the selected gas is displayed to the left on the Gas Table.


Click **[Set Selected as Primary]** or **[Set Selected as Secondary]** to send the selected gas to the BGA244 to be measured. The selected Primary and Secondary gases are highlighted in green in the gas tables.

Adding and Removing User Gases

Gases can be added or removed from the User Gas Table. See *User Gases* (page 117) in Chapter 4 for instructions on determining the different parameters required for User Gases.

To add a User gas click **[Add Gas]** to open the Add User Defined Gas dialog box. Enter the **<Name>**, molar **<Mass>** and **<Heat Capacity Ratio>** of the gas. (Heat Capacity Ratio = γ)

Click **[OK]** to save the User Gas table to the BGA244.



To remove a User Gas, first select it and then click **[Remove Gas]** to remove it from the User Gas Table. Note that gases that are selected as the Primary and Secondary gases (highlighted in green) cannot be deleted. To delete one of these gases you must first select another primary or secondary gas and then delete it.



Note: If a User Gas is deleted it cannot be retrieved.

[Refresh User Gas Table] will reload the User Gas Table from the BGA244. The User Gas Table is loaded when BGAMon is first started. However if a User Gas is added over the computer interface while BGAMon is running the Gas Table would need to be refreshed to display the new gas.

Invalid Gas Message

Stored settings (in the BGA244) and stored configuration (on BGAMon) can reference a nonexistent User Gas. This is indicated by the screen message "Invalid Gas" and a reported error. If this occurs, select a different gas. If the User gas had been inadvertently deleted, re-enter the data and then select it.


Faults/Alerts

The Faults/Alerts page combines the Faults page and the Alerts page from the BGA244 Control panel. See Faults (page 85) and Alerts (page 87) in Chapter 3 for details.



FIGURE 30: FAULT/ALERT PAGE

Active Alerts are displayed as buttons. Click **[Alert]** to open an Alert Action dialog box. This allows the Alert to be cleared or disabled. See Alerts Action (page 89) for information on this and the **[Enable All Alerts]** button.



Computer I/O

The Computer I/O page combines the RS-232 page and RS-422 page from the BGA244 Control Panel. See *Computer I/O* (page 79) for details.

Note that the USB page and the RS-232 and RS-422 Transmit and Receive buffer pages are not accessible from BGAMon.




FIGURE 31: COMPUTER I/O PAGE

Store / Recall

The Store / Recall page combines the all of the Store, Recall and Erase pages from BGA244 Setup page. See *Store/Recall* (page 67) for details.

Store / Recall Instrument Settings

0	DEFAULT
1	HELIUM
2	ARGON
3	NITROGEN
4	CO2 H2
5	*EMPTY*
6	*EMPTY*
7	*EMPTY*
8	*EMPTY*
9	*EMPTY*
10	*EMPTY*
11	*EMPTY*
12	*EMPTY*
13	*EMPTY*
14	*EMPTY*
15	*EMPTY*
16	*EMPTY*
17	*EMPTY*
18	*EMPTY*
19	*EMPTY*
20	*EMPTY*

Store...
Recall
Erase

FIGURE 32: STORE/RECALL PAGE

Loading and Saving Configurations

Entire BGA244 configurations can be saved to or loaded from disk. This can be useful when configuring multiple BGA244s. Every setting of the BGA244 is saved or loaded EXCEPT the following:

- Stored settings (See the Store/Recall page for information on that function)
- User Gas Table



Note: If a configuration is loaded to the BGA244, all of the previous settings are erased and cannot be retrieved.

Click **(File/Save)** to open the File Save window. It's a good idea to use a specific directory to save your configuration files. Use descriptive file names to minimize confusion if different files are to be saved. For example: "BGA Argon and Helium.cfg"

Click **(File/Load)** to open the File Load window. Navigate to the appropriate directory and select the desired file.

Updating Firmware

The BGA244 firmware can be updated by customers using BGAMon. New versions of the firmware are not frequently released, but occasionally new features or bug fixes are added. If necessary, the new version of the BGA244 firmware will be provided by SRS. Save the file provided by SRS so it can be read by the computer running BGAMon.



Do not disconnect the BGA244 during firmware update as this may render the unit inoperable and require factory service. See *Problems during Update* below if an error occurs during an update.

Connect the BGA244 to your computer and run BGAMon. Click [**Connect to Device**] then select (**Utilities/Update Firmware**). Navigate to the location of the unzipped firmware file and select it. A {**Firmware Update**} message box will appear in BGAMon listing the firmware size. Click [OK] to begin the update.

The BGA244 will report the progress of the update on the LCD display. After the update is complete two buttons become active on the BGA244. Press [**RESTART**] to reset the BGA244 and run the new firmware. Press [**HOME**] to continue running the old firmware until the next time the unit is reset.

Problems during Update

A fatal error can occur if a file is corrupted while being stored to flash during a firmware update. If this occurs, each time the unit is reset it will enter the Factory Safe Mode. Reinstall the Firmware to correct this error. If this problem occurs repeatedly contact SRS for assistance.


Troubleshooting BGAMon

Connecting

If **{Device Not Found}** appears when starting BGAMon, you may have tried to connect the BGA244 before Windows has detected it. Click **[OK]** here and on **{Unable to open BGA Device}** if it appears. Then try **[Connect to Device]** again.

If the BGA244 doesn't power up successfully, confirm that the power source is able to properly power the BGA244. See *Power* (page 22) for details on troubleshooting power supply problems.

If BGAMon repeatedly displays the **{Device Not Found}** dialog box it indicates that either the drivers aren't installed or the BGA244 isn't properly connected to the USB port. Retry the Driver installation procedure described at the beginning of this chapter. If the drivers appear to be installed correctly view the Device List dialog box at (Device/Detect Device).



Device	Flags	Type	ID	LocID	Serial #	Description	Handle
0	1	9	67330069	536	DA40KI4	FT220X 4-BIT FT1248	2632888

Unplug the BGA244 from the USB port and then re-connect it. Check if Windows displays the "New hardware found" message. If so follow the driver installation procedure described above. If this problem occurs repeatedly contact SRS for assistance.

Running Slowly

Measurements for the BGA244 are always being logged to the Windows Temp directory whenever the BGA244 is connected to BGAMon. These files can become large if the BGA244 has been connected to BGAMon for more than a few days. In some cases this can slow your systems response time. If this is a problem either click **[Clear Data]** on the strip Chart page or disconnect the BGA244 from BGAMon. Note that **[Clear Data]** will delete all previously acquired data.

When data is cleared or the BGA244 is reconnected to BGAMon the logged files will be cleared and overwritten, freeing up that memory.

Chapter 6: Remote Programming

Introduction

The BGA244 can be remotely programmed over the USB interface, the RS-232 serial interface, or the RS-422 interface. Any host computer interfaced to the instrument can easily control and monitor its operation.

Note that the RS-422 interface is only available if a 24 V_{DC} supply is connected to the BGA244.

Interface Configuration

The interface configuration parameters can be controlled from the Computer I/O pages located at (Home/Setup/Control Panel/Computer I/O). See *Computer I/O* (page 79) for details.

The only two configuration parameters are the baud rates for the RS-232 and RS-422 interfaces. The default baud rate for both interfaces is 9600 baud. All interfaces are enabled by default, but the RS-232 and RS-422 interfaces may be disabled individually if desired. Any modification made to an interface takes effect immediately.

Responses to commands are returned to the interface that sent the command. For example if you query “*ESE?” over RS-232, the reply will only be sent back over the RS-232 interface.

Front-Panel Indicators

Two front panel LED indicators located to the right of the RS-232 connector help assist with programming: COMM and ERR. The COMM LED is an activity indicator that flashes every time a character is received or transmitted over one of the remote interfaces. The ERR LED will flash when a remote command fails to execute due to illegal syntax or invalid parameters.

Transmit and Receive Buffers

Each interface has its own set of transmit and receive buffers that show the most recent commands and responses over that interface. These can be helpful for remote interface debugging. The buffers are located on the specific computer I/O pages (Home/Setup/Control Panel/Computer IO). See *Computer I/O* (page 79) for details.

Break Signal

The RS-232 and RS-422 interface supports a “break condition”. This is a non-command based, low level abort signal used to reset the interface if communications are locked. When a “break” is received, the input and output queues and buffers are all flushed. Any currently in-process “set” commands are completed, but no responses are returned. Status bits are not modified when a break occurs. The break signal is a framing error of all zeros.

USB

A USB Type B communication port is included on the front panel of the instrument. The BGA244 complies with the High Speed USB 2.0 standard. Refer to *Computer Interfaces* (page 24) for details on cabling requirements when connecting the BGA244 to a host computer.

The appropriate software driver must be installed on the host computer to properly communicate with BGA244 over the USB interface. There are two separate drivers available for the BGA244: a Virtual Com Port (VCP) and a Direct Driver (USB driver + DLL). The VCP allows communication in the same fashion as any other COM port. The Direct Driver allows function calls to be embedded directly into a controlling program. Refer to *Using the USB Drivers* (page 197) for more information.

USB Configuration

There is no status information available for the USB interface. The transmit/receive buffers can be viewed at (Home/Setup/Control Panel/Computer IO/USB/Buffer).

RS-232

An RS-232 communication port is included on the front panel of the instrument. This is a standard 9 pin DB, female connector configured as a DCE (transmit on pin 2, receive on pin 3). See *Computer Interfaces* (page 24) details on cabling requirements when connecting the BGA244 to a host computer.

RS-232 Configuration

The RS-232 interface can be enabled or disabled. Status information and the transmit/receive buffers can be viewed at (Home/Setup/Control Panel/Computer IO/RS-232).

In order to communicate properly over RS-232, both the BGA244 and the host computer must be set to the same configuration. The following baud rates are supported: 2400, 4800, 9600 (default), 19.2k, 38.4k, 57.6k and 115.2k. The remaining communication parameters are set as follows: no parity, 8 data bits, 1 stop bit and CTS/RTS hardware flow control.

RS-422

An RS-422 communication port is included on the front panel of the instrument. It is configured as a 4 wire point-to-point, non-multidrop that is connected via a terminal strip. It supports single transmitter and receiver pair only (not multi-drop). See *Computer Interfaces* (page 25) details on cabling requirements when connecting the BGA244 to a host computer.

Note that the RS-422 interface requires that a +24 V_{DC} power supply is connected to the BGA244.

RS-422 Configuration

The RS-422 interface can be enabled/disabled. Status information and the transmit/receive buffers can be viewed at (Home/Setup/Control Panel/Computer IO/RS-422).

In order to communicate properly over RS-422, both the BGA244 and the host computer must be set to the same configuration. The following baud rates are supported: 2400, 4800, 9600 (default), 19.2k, 38.4k, 57.6k and 115.2k. The remaining communication parameters are set as follows: no parity, 8 data bits, 1 stop bit and no hardware flow control.

Command Syntax

Although the BGA244 doesn't have a GPIB interface, its commands follow IEEE-488.2 Standard. All commands use ASCII characters, are 4-characters long, and are case-insensitive. Standard IEEE-488.2 defined commands begin with the '*' character followed by 3 letters. Instrument specific commands are composed of 4 letters.

The four letter mnemonic (shown in capital letters) in each command sequence specifies the command. The rest of the sequence consists of parameters.

Commands may take either *set* or *query* form, depending on whether the '?' character follows the mnemonic. *Set only* commands are listed without the '?', *query only* commands show the '?' after the mnemonic, and *query optional* commands are marked with a '(?)'.

Parameters shown in { } and [] are not always required. Parameters in { } are required to set a value and are omitted for queries. Parameters in [] are optional in both set and query commands. Parameters listed without any surrounding characters are always required.

Do NOT send () or {} or [] as part of the command.

The command buffer is limited to 64 kbytes, with 25 byte buffers allocated to each of up to 3 parameters per command. If the command buffer overflows, both the input and output buffers will be flushed and reset. If a parameter buffer overflows, a command error will be generated and the offending command discarded.

Commands are terminated by a semicolon, a <CR> (ASCII 13), or a <LF> (ASCII 10). Execution of the command does not begin until a command terminator is received. The response terminator is set by the XTRM command (default <CR> <LF>).

Errors

Aside from communication errors, commands may fail due to either syntax or execution errors. Syntax errors can be detected by looking at bit 5 (CME) of the event status register (*ESR?). Execution errors can be detected by looking at bit 4 (EXE) of the event status register. In both cases, when an error occurs the red ERROR LED will flash and an error code indicating the specific cause of the error is placed in the error buffer. The error buffer may be queried with the LERR? command. Descriptions of all error codes can be found in the *Error Codes* section of this chapter, starting on page 187.

Parameter Conventions

The command descriptions use parameters, such as i, d, and v. These parameters represent integers or floating point values expected by the command. The parameters follow the conventions summarized in Table 22.

TABLE 22: COMMAND PARAMETER CONVENTIONS

Parameter	Meaning										
i, j, k	An integer value										
d	A floating point value										
p	A floating point value representing a pressure in global units										
v	A floating point value representing a speed in global units										
r	A floating point value representing a ratio in global units										
t	A floating point value representing temperature in global units										
s	An ASCII string										
u	An identifier of units. Allowed units depend on the type as identified below: <table border="1" data-bbox="699 596 1395 783"> <thead> <tr> <th>Type</th> <th>Allowed Units</th> </tr> </thead> <tbody> <tr> <td>Ratio</td> <td>'%', 'ppm' or 'Frac'</td> </tr> <tr> <td>Speed</td> <td>'m/s', 'kph' or 'mph'</td> </tr> <tr> <td>Temperature</td> <td>'C', 'K', or 'F'</td> </tr> <tr> <td>Pressure</td> <td>'psi', 'atm', 'bar', 'Pa', 'mmHg' or 'torr'</td> </tr> </tbody> </table>	Type	Allowed Units	Ratio	'%', 'ppm' or 'Frac'	Speed	'm/s', 'kph' or 'mph'	Temperature	'C', 'K', or 'F'	Pressure	'psi', 'atm', 'bar', 'Pa', 'mmHg' or 'torr'
Type	Allowed Units										
Ratio	'%', 'ppm' or 'Frac'										
Speed	'm/s', 'kph' or 'mph'										
Temperature	'C', 'K', or 'F'										
Pressure	'psi', 'atm', 'bar', 'Pa', 'mmHg' or 'torr'										

Numeric Conventions

Floating point values may be decimal ('123.45') or scientific ('1.2345e2'). Integer values may be decimal ('12345') or hexadecimal ('0x3039').

Measurement Errors

If an error in a measurement command occurs, the unit will return an overload value (9.9E37) to inform the user that there was something wrong with the measurement. The controlling program should test for this value and check the appropriate Status bytes to determine what the problem is.

Missing Power Supplies

Commands that require a power supply that is not present will generate an execution error. If a command cannot be successfully executed because of a missing 24 V_{DC} power supply, an Error 17 (No 24V Available) will be generated. See the Chapter 3: *Operations Guide* for details on which functions require the 24 V_{DC} power supply.

Configuration commands that depend on the 24 V_{DC} can be set if the supply isn't present, although they may not be active until it is. However some query commands will generate errors if it isn't present.

If a command requires 24 V_{DC} power supply, it is usually a good practice to query if the 24 V_{DC} is available using the EXPA command prior to sending the command. These can be done at the beginning of the controlling program during initialization.

Abridged Index of Commands

Common IEEE-488.2 Commands

*CAL?	Page 154	Run auto calibration routine
*CLS	Page 154	Clear Status
*ESE(?) <i>{i}</i>	Page 154	Standard Event Status Enable
*ESR?	Page 154	Standard Event Status Register
*IDN?	Page 154	Identification String
*OPC(?)	Page 154	Operation Complete
*PSC(?) <i>{i}</i>	Page 155	Power-on Status Clear
*RCL <i>i</i>	Page 155	Recall Instrument Settings
*RST	Page 155	Reset the Instrument
*SAV <i>i</i>	Page 155	Save Instrument Settings
*SRE(?) <i>{i}</i>	Page 155	Service Request Enable
*STB?	Page 155	Status Byte
*TRG	Page 156	Trigger
*TST?	Page 156	Self Test
*WAI	Page 156	Wait for Command Execution

Instrument Status Commands

ALRD	Page 157	Disable all Alerts
ALRE	Page 157	Enable all Alerts
ANAE(?) <i>{i}</i>	Page 157	Analog Enable Register
ANAI? <i>{i}</i>	Page 157	Analog Status Register (Immediate)
ANAR? <i>{i}</i>	Page 158	Analog Status Register (Latched)
BGOE(?) <i>{i}</i>	Page 158	BGA 0 Enable Register
BGOI? <i>{i}</i>	Page 159	BGA 0 Status Register (Immediate)
BGOR? <i>{i}</i>	Page 159	BGA 0 Status Register (Latched)
BG1E(?) <i>{i}</i>	Page 159	BGA 1 Enable Register
BG1I? <i>{i}</i>	Page 160	BGA 1 Status Register (Immediate)
BG1R? <i>{i}</i>	Page 160	BGA 1 Status Register (Latched)
FALE(?) <i>{i}</i>	Page 160	Fault Enable Register
FALI? <i>{i}</i>	Page 161	Fault Status Register (Immediate)
FALR? <i>{i}</i>	Page 161	Fault Status Register (Latched)

Event Register Commands

EVNC(?) <i>i,j</i>	Page 162	Event Configuration Register
EVNE(?) <i>{i}</i>	Page 162	Event Enable Register
EVNI? <i>{i}</i>	Page 162	Event Status Register (Immediate)
EVNR? <i>{i}</i>	Page 162	Event Status Register (Latched)
RLYF(?) <i>i,j</i>	Page 163	Relay Force
RLYU(?) <i>i</i>	Page 163	Relay Position

Interface Commands

BAUD(?)i{j}	Page 164	Set Baud Rate
LERR?	Page 164	Inspect Error Buffer
UARE(?)i{j}	Page 164	Enable Computer Interface
XTRM i{j,k}	Page 164	Interface Terminator

Measurement Commands

NSOS?[u]	Page 165	Normalized Speed of Sound
PRES?[u]	Page 177	Analysis Pressure
PUDL?[u]	Page 165	Purity Mode Measurement
RATO?i[u]	Page 165	Binary Gas Ratio Measurement
RAT2?i[u]	Page 166	Binary Gas Ratio 2 nd Measurement
SSOS?[u]	Page 166	Measured Speed of Sound
TCEL?[u]	Page 166	Cell Temperature
XALL?	Page 166	Query All

Measurement Related Commands

AVGE(?)i	Page 167	Enable Averaging
AVGN(?)i	Page 167	Number of Averages
BCTP(?)i	Page 167	Binary Concentration Type
MSMD(?)i	Page 167	Instrument Mode
RELH	Page 167	REL to 100%
RELM(?)	Page 168	REL Mode
RELV(?)r[u]	Page 168	REL Value
RELZ	Page 168	REL to 0%
RELX(?)i	Page 168	Relaxation Correction
RTBA r,r[u]	Page 168	Add Multipoint REL Table Entry
RTBC	Page 169	Clear Multipoint REL Table
RTBD i	Page 169	Delete i th Raw/REL Multipoint data pair
RTBE i,r,r [u]	Page 169	Update i th Raw/REL Multipoint data pair
RTBN?	Page 169	Query # of Entries in Multipoint REL Table
RTBR?i,[u]	Page 169	Query Multipoint REL Table Entry
RTW1(?)r,r[u]	Page 169	2 Point REL Point 1 Values
RTW2(?)r,r[u]	Page 169	2 Point REL Point 2 Values
RUNM(?)i	Page 168	Run Mode
SETT?	Page 170	Measurement Settled
UNCT?[u]	Page 170	Measurement Uncertainty

Limit Commands

LIME(?)i,j,k}	Page 170	Limit Enable
LIMH(?)i,d}{u}	Page 171	Limit Hysteresis
LIMM(?)i,j,d}{u}	Page 171	Meter Scale
LIMS?i	Page 171	Limit State
LIMT(?)i,j,d}{u}	Page 171	Limit Trip Value

Gas Selection Commands

GASP(?)s}	Page 172	Set Primary Gas
GASS(?)s}	Page 173	Set Secondary Gas
SWAP	Page 173	Swap Binary Gases

Heater Commands

BLTM?{u}	Page 185	Endplate Temperature
HEDG(?)i}	Page 174	Degas Heater Enable
HEEN(?)i}	Page 174	Block Heater Enable
HEIL(?)d}	Page 174	Max Heater Current
HEPW?	Page 174	Heater Power
HEST?	Page 174	Heater Power Status
HETM(?)d}	Page 175	Block Heater Temperature

Pressure Commands

PMAX(?)i,p}{u}	Page 176	Pressure Gauge Scale Maximum
PMIN(?)i,p}{u}	Page 176	Pressure Gauge Scale Minimum
PRAC(?)i}	Page 177	Select Analysis Pressure
PRAM(?)p}{u}	Page 177	Set Atmospheric Pressure
PREN(?)i,j}	Page 177	Pressure Gauge Enable
PRES?{u}	Page 177	Analysis Pressure
PRRD?i{u}	Page 177	Pressure Gauge Reading
PRSU(?)i,j}	Page 178	Pressure Measurement Scale Units
PUSR(?)p}{u}	Page 178	User Cell Pressure

Measure Analog Output Commands

MOCN(?)i}	Page 179	Measure Out Concentration
MOEN(?)i}	Page 179	Measure Out Enable
MOMN(?)i,d}{u}	Page 180	Measure Out Scale Min
MOMX(?)i,d}{u}	Page 180	Measure Out Scale Max
MOTY(?)i}	Page 180	Measure Out Type
MOVA?	Page 180	Measure Out Value

Analog Output Commands

AOEN(?)i{j}	Page 181	Analog Out Enable
AOMN(?)i,j{,d}[u]	Page 181	Analog Out Scale Min
AOMX(?)i,j{,d}[u]	Page 181	Analog Out Scale Max
AOSE(?)i{j}	Page 182	Analog Out Measure Setting
AOTY(?)i{j}	Page 182	Analog Out Type
AOUS(?)i{,d}	Page 182	Analog Out User Setting
AOVA?i	Page 182	Analog Out Value

Analog Input Commands

AILP(?)i{,d}	Page 183	Analog Loop Power Voltage
AINE(?)i{j}	Page 183	Analog Input Enable
AIRE?i	Page 183	Read Analog Input
AITY(?)i{j}	Page 184	Analog Input Type

Miscellaneous Commands

ATSV	Page 185	Autosave the current configuration
AUSRi,j,s	Page 185	Add User Gas
BAKL(?)i}	Page 185	Backlight
BLTM?[u]	Page 185	Endplate Temperature
DSPO(?)i}	Page 186	Display Orientation
DUSRs	Page 186	Delete User Gas
HOME	Page 186	Go to Home Page
ERASi	Page 186	Erase Stored Setup
EXPA?	Page 186	External Power Available?
MUTE(?)i}	Page 186	Key click Enable
PASE(?)i}	Page 187	Password Enable
PASL?	Page 187	Password Locked Status
PASS(?)i}	Page 187	Set Password
PCTM?[u]	Page 187	PCB Temperature
SGAM(?)i}	Page 187	Speaker Amplitude Mode
UNFA(?)i{,s}	Page 188	Global Units
VOLT?i	Page 188	External Power Supply Voltage

Detailed Command List

Common IEEE-488.2 Commands

*CAL?	Auto calibration This command does nothing and returns 0.																		
*CLS	Clear Status Clear Status immediately clears the ESR, BGOR, BG1R, FALR, ANAR, and EVNR registers as well as the LERR error buffer.																		
*ESE(?)<i>{i}</i>	Standard Event Status Enable Set (query) the Standard Event Status Enable register <i>{to i}</i> . Bits set in this register cause ESB (in STB) to be set when the corresponding bit is set in the ESR register.																		
*ESR?	Standard Event Status Register Query the Standard Event Status Register. Upon executing a *ESR? query, the returned bits of the *ESR register are cleared. The bits in the ESR register have the following meaning. <table border="1"> <thead> <tr> <th>Bit</th> <th>Meaning</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>OPC – operation complete</td> </tr> <tr> <td>1</td> <td>Reserved</td> </tr> <tr> <td>2</td> <td>QYE – query error</td> </tr> <tr> <td>3</td> <td>DDE – device dependent error</td> </tr> <tr> <td>4</td> <td>EXE – execution error</td> </tr> <tr> <td>5</td> <td>CME – command error</td> </tr> <tr> <td>6</td> <td>Reserved</td> </tr> <tr> <td>7</td> <td>PON – power-on</td> </tr> </tbody> </table> <p>Example *ESR? A return of '176' would indicate that PON, CME, and EXE are set.</p>	Bit	Meaning	0	OPC – operation complete	1	Reserved	2	QYE – query error	3	DDE – device dependent error	4	EXE – execution error	5	CME – command error	6	Reserved	7	PON – power-on
Bit	Meaning																		
0	OPC – operation complete																		
1	Reserved																		
2	QYE – query error																		
3	DDE – device dependent error																		
4	EXE – execution error																		
5	CME – command error																		
6	Reserved																		
7	PON – power-on																		
*IDN?	Identification String Query the instrument identification string. Example *IDN? Returns a string similar to 'Stanford Research Systems,BGA244,s/n000000,build1234'																		
*OPC(?)	Operation Complete The set form sets the OPC flag in the ESR register when all prior commands have completed. The query form returns '1' when all prior commands have completed, but does not affect the ESR register.																		

PSC(?) {i}*Power-on Status Clear**

This command does nothing and returns 1.

RCL i*Recall Instrument Settings**

Recall instrument settings from location i. The parameter i may range from 0 to 20. Location 0 is reserved for the recall of factory default instrument settings. Locations 1 to 20 are user configurable.

Example

*RCL 3 Recall instruments settings from location 3.

RST*Reset the Instrument**

Reset the instrument to default settings. This is equivalent to *RCL 0. It is also equivalent to recalling Setup 0 (Default) at (Home/Setup/Store-Recall/Recall). See *Default Setup* (page 69) for a list of the Default settings.

If this command is received while the unit is updating firmware, an Error 21 (Time Out Error) may be generated.

Example

*RST Resets the instrument to default settings

SAV i*Save Instrument Settings**

Save instrument settings to location i. The parameter i may range from 1 to 20.

Example

*SAV 3 Save current settings to location 3.

SRE(?) {i}*Service Request Enable**

Set (query) the Service Request Enable register {to i}.

STB?*Status Byte**

Query the standard IEEE 488.2 serial poll status byte. The bits in the STB register have the following meaning:

<u>Bit</u>	<u>Meaning</u>
0	BG0B - BG0R summary bit
1	BG1B - BG1R summary bit
2	FALB - FALR summary bit
3	ANAB - ANAR summary bit
4	MAV – message available
5	ESB – ESR summary bit
6	MSS – master summary bit
7	EVNB – Event summary bit

Example***STB?**

A return of '113' would indicate that BGOB, MAV, ESB, and MSS are set. BGOB indicates that an enabled bit in BGOR is set. MAV indicates that a message is available in the output queue. ESB indicates that an enabled bit in ESR is set. MSS reflects the fact that at least one of the summary enable bits is set and the instrument is requesting service.

TRG*Trigger**

This command does nothing.

TST?*Self Test**

Runs the instrument self test and returns 0 if successful. Otherwise it returns the encoded errors per the following table. Note that bits for tests that are not executed are set to 0. Refer to *Self Test* (page 89) for more details.

Bit	Meaning
0	Reserved
1	Power Supplies
2	Reserved
3	CODEC
4	ADC1, ADC2
5	USB
6	Thermistor limits or imbalance
7	Membrane Resistance
8	Magnets
9	Analog Output DACs
10	Analog Input ADCs
11	Thermistor shorted
12	Self Test Timeout
13-15	Reserved

WAI*Wait for Command Execution**

The instrument returns '0' when the OPC flag in the ESR register is set.

Example***WAI**

Wait until OPC flag is set.

Instrument Status Commands

Instrument Status Registers

ALRD **Disable All Alerts**

Disable all Alerts on the GUI.

This command disables all Alerts from appearing on the GUI. It has no effect on any of the status registers or the actual Alert conditions.

Example

ALRD This would stop any Alert messages from appearing on the BGA244 display.

ALRE **Enable All Alerts**

Enable all Alerts on the GUI.

This command enables all Alerts from appearing on the GUI. It has no effect on any of the status registers. It is the equivalent to the [ENABLE ALL ALERTS] key on the Alerts page.

ANAE(?)*{i}* **Analog Enable Register**

Set (query) the ANAE enable register *{to i}*. Bits set in this register cause ANA (in STB) to be set when the corresponding bit is set in the ANAR register.

ANAI?*{i}* **Analog Status Register (Immediate)**

Query the ANAI register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is not latched and reflects the value of the register the instant it was read. Refer to the *Instrument Status Register Model* (page 185)

for a description of the different conditions reported. The bits in the ANAI register have the following meaning:

Bit	Meaning
0	Measure Out Current Alert
1	Measure Out Temperature Alert
2	Output 1 Current Alert
3	Output 1 Temperature Alert
4	Output 2 Current Alert
5	Output 2 Temperature Alert
6	Input 1 Over Current Alert
7	Input 1 Under Current Alert
8	Input 1 Over Voltage
9	Input 1 Under Voltage
10	Input 2 Over Current
11	Input 2 Under Current
12	Input 2 Over Voltage
13	Input 2 Under Voltage
14	Loop Power 1 Alert
15	Loop Power 2 Alert

Example

ALAI? A return of '2048' indicates that Input 2 was under current.

ANAR?[i]

Analog Status Register (Latched)

Query the ANAR status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is a latched version of the ANAI register. Upon executing an ANAR? query, the register is cleared. The meaning of the bits in the ANAR register is the same as the ANAI register.

Example

ANAI? A return of '1' would indicate that Measure Out had a current fault since the last time this register has been read or cleared.

BGOE(?) {i}

BGA Enable Status Register 0

Set (query) the BGOE status enable register {to *i*}. Bits set in this register cause BGOB (in STB).

Example

BGOE 65535 This would enable all of the BGA0 register bits to set the BGOB bit of the STB register.

BG0I?[i]**BGA Status Register 0 (Immediate)**

Query the BG0I status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is not latched and reflects the value of the register the instant it was read. Refer to the *Instrument Status Register Model* (page 185) for a description of the different conditions reported. The bits in the BG0I register have the following meaning:

Bit	Meaning
0	2 Concentration Values (BGA Mode only)
1	Degas heater on
2	Measurement is stopped
3	Measurement is below allowable range
4	Measurement is above allowable range
5	No Signal
6	> 2 °C mismatch on temperature measurements
7	Cell Temperature < 0 °C
8	Cell Temperature > 70 °C
9	Bad Analysis Pressure Reading
10	Condensation Warning
11	Safe Mode
12-14	reserved
15	Summary bit for Fault Register (FALR)

Example

BG0I? A return of '2' would indicate that the Membrane Heater is on.

BG0R?[i]**BGA Status Register 0 (Latched)**

Query the BG0R status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is a latched version of the BG0I register. Upon executing a BG0R? query, the register is cleared. The meaning of the bits in the BG0R register is the same as the BG0I register.

Example

BG0R? A return of '192' would indicate that there had been a mismatch between the temperature sensors and that the temperature has dropped below < 0 °C since last read or cleared.

BG1E(?) {i}**BGA Status Enable Register 1**

Set (query) the BG1E status enable register {to *i*}. Bits set in this register cause BG1B (in STB) to be set when the corresponding bit is set in the BG1R register.

Example

BG1E? A return of '2' would indicate that bit 1 of the BGA1 register is enabled set the BG1B bit of the STB register.

BG1I?[i]**BGA Status Register 1 (Immediate)**

Query the BG1I status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is not latched and reflects the value of the register the instant it was read. Refer to the *Instrument Status Register Model* (page 185) for a description of the different conditions reported. The bits in the BG1I register have the following meaning:

Bit	Meaning
0	reserved
1	Heater is not Regulating
2	Measurement Meter Limits exceeded
3	Pressure Meter 1 Limits exceeded
4	Pressure Meter 2 Limits exceeded
5	Temperature Meter Limits exceeded
6	USB Voltage is over range (USB Power only)
7	USB Voltage is under range (USB Power only)
8	+24 V is over range (24 V Power only)
9	+24 V is under range (24 V Power only)
10	REL mode active
11	Key Press Detected
12	Password Lock
13	Average Settled (0 = settled, 1 = not settled)
14-15	reserved

Example

BG1I? A return of '1026' would indicate that REL mode was active and the Heaters were not regulating.

BG1R?[i]**BGA Status Register 1 (Latched)**

Query the BG1R status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is a latched version of the BG1I register. Upon executing a BG1R? query, the register is cleared. The meaning of the bits in the BG1R register is the same as the BG1I register.

Example

BG1R? A return of '4' would indicate the measurement limit has been exceeded since last read or cleared.

FALE(?)*{i}***Fault Enable Register**

Set (query) the FALE enable register *{to i}*. Bits set in this register cause FALT (in STB) to be set when the corresponding bit is set in the FALR register.

FALI?[i]**Fault Status Register (Immediate)**

Query the FALI register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is not latched and reflects the value of the register the instant it was read. Refer to the *Instrument Status Register Model* (page 185) for a description of the different conditions reported. The bits in the FALI register have the following meaning:

Bit	Meaning
0	Cell Under Temperature Fault
1	Cell Over Temperature Fault
2	Fault in Flash Memory
3	Firmware update Fault
4	Heater Fault (>90 °C)
5	1.4 V Supply Out of Range Fault
6	3.3 V Supply Out of Range Fault
7	4.5 V Supply Out of Range Fault
8	5V_I Supply Out of Range Fault
9	USB Under Voltage Fault
10	USB Over Voltage Fault
11	unused
12	24 V Over Voltage Fault
13	>5°C Mismatch on Temperature Measurements
14	unused
15	Self Test Failure

Example

FALI? A return of '128' would indicate that the 3.3V power supply was under voltage.

FALR?[i]**Fault Status Register (Latched)**

Query the FALR status register. If *i* is included the query returns the bit indicated by *i* ($0 \leq i \leq 15$). This register is a latched version of the FALI register. Upon executing a FALR? Query, the register is cleared. The meaning of the bits in the FALR register is the same as the FALI register.

Example

FALR? A return of '32768' would indicate that Self Test Failed since last time this register has been read or cleared.

Event Registers

EVNC{?}i{j}

Event Configuration Register

Set (query) Event Relay Configuration Register i to (j) . $i = 1|2$ for Event Relay $1|2$. The meaning of the bits in the EVNC registers is the same as the EVNI register. This command sets the Event Relay Configuration page as described in *Events* (page 55). Bits set in this register do not affect behavior of the Event Status registers.

Example

EVNC1,3 This would configure Event Relay 1 to switch to the ON position for either System Fault or No Signal.

EVNE(?)i

Event Enable Register

Set (query) the EVNE enable register $\{to\ i\}$. Bits set in this register cause EVENT (in STB) to be set when the corresponding bit is set in the EVNR register. The meaning of the bits in the EVNE register is the same as the EVNI register.

EVNI?[i]

Event Status Register (Immediate)

Query the EVNI register. If i is included the query returns the bit indicated by i ($0 \leq i \leq 15$). This register is not latched and reflects the value of the register the instant it was read. See Events in the *Operations Guide* for details on the Event registers. The bits in the EVNI register have the following meaning:

Bit	Meaning
0	System Fault
1	No Measurement
2	+Measurement Limit Exceeded
3	-Measurement Limit Exceeded
4	+Pressure 1 Limit Exceeded
5	-Pressure 1 Limit Exceeded
6	+Pressure 2 Limit Exceeded
7	-Pressure 2 Limit Exceeded
8	+Temperature Limit Exceeded
9	- Temperature Limit Exceeded
10-15	reserved

Example

EVNI? A return of '2' would indicate that the BGA244 cannot recover the acoustic signal.

EVNR?[i]

Event Status Register (Latched)

Query the EVNR status register. If i is included the query returns the bit indicated by i ($0 \leq i \leq 15$). This register is a latched version of the EVNI register. Upon executing an EVNR? Query, the register is cleared. The meaning of the bits in the EVNR register is the same as the EVNI register.

Example

EVNR? A return of '4' indicates that the signal exceeded the Measurement Limit since the last read or cleared.

RLYF{?}{i,j}**Relay Force**

Set (query) the Event Relay (i) Force value {to j}. i = 1 for Event Relay 1 and 2 for Event Relay 2. j can be one of the values in the table. This command is the same as the Event Relay Force functions described in *Events* (page 55).

If 24 V is not present, you can set the relay position, but the relay is inoperable until 24V is applied. Use the EXPA command to check the 24V power supply.

<u>J</u>	<u>meaning</u>
0	No Force
1	Force Off
2	Force ON

Example

RLYF 1,1 Force Event Relay 1 ON

RLYU?[i]**Relay Position**

Query the position of Event Relay (i). i = 1 for Event Relay 1 and 2 for Event Relay 2. The return value is 0 for OFF and 1 for ON.

This command will generate an Error 18 (No 24 V Available) if 24 V is not present.

Example

RLYU?2 A return value of 1 indicates that Event Relay 2 is in the ON position.

Interface Commands

BAUD(?)i{j}
Set Baud Rate

Set (query) the Baud Rate of interface i {to j}. Parameter i = 0 for RS-232; i = 1 for RS-422. Parameter j (baud rate) is set per the following table:

j	Baud Rate
0	2400
1	4800
2	9600
3	19200
4	38400
5	57600
6	115200

Example

BAUD 0,3 Set RS-232 baud rate to 19200

LERR?
Inspect Error Buffer

Query the last error in the error buffer. Upon executing a LERR? the returned error is removed from the error buffer. See Error Codes on Page 194 for a description. The Error Buffer can store up to 20 errors. If more than 19 errors occur without being queried, the 20th error will be 254 (too many errors), indicating that errors may have been dropped.

Example

LERR? A return of 26 indicates that the last error was Invalid Gas.

UARE(?)i{j}
Enable Computer Interface

Set (query) computer interface i enable mode {to j}. Parameter i = 0 for RS-232; i = 1 for RS-422. Parameter j = 1 for enabled, otherwise 0.

Example

UARE 0,0 Disable RS-232

XTRM i{j,k}
Interface Terminator

Set the interface terminator that is appended to each response to i, j, k. Parameters i, j and k are ASCII characters.

The default terminator is 13, 10, which is a carriage return followed by a line feed.

Measurement and Related Commands

Measurement commands report results from the BGA244. Certain commands require that the BGA244 is in the proper Instrument Mode. If not, an Error 11 (Illegal Mode) will be generated.

Measurements

If an error in a measurement occurs during a query, the unit will return an overload value (9.9E37) to inform the user that there was something wrong with the measurement. The controlling program should identify this and check the BGA0 status register if it occurs. A return value of '0' in the BGA0 status register confirms that a measurement is valid. Any other value indicates there is some problem with the measurement. See *BGA0 Status Register* (page 185) for more information.

Each measurement has its own unit family associated with it. If units are omitted, the selected global unit is used. If units are included with the command, they must be in the correct unit family or an Error 127 (Illegal Units) will be generated.

NSOS?[u]

Normalized Speed of Sound

Query the Normalized Speed of Sound for the current Instrument Mode. If omitted, units default to the global speed units.

Example

NSOS? Kph Query the normalized speed of sound in kilometers per hour
NSOS? Query the normalized speed of sound in global units

PRES?[u]

Analysis Pressure

Query the Cell Analysis Pressure. If omitted, units default to the global temperature units.

See *Pressure Commands* (page 176) for a list of possible errors that occur with the PRES command.

Example

PRES? Query the Analysis Pressure in global pressure units

PUDL?[u]

Purity Mode Measurement

Query the Gas Purity Measurement. If omitted, units default to the global ratio units.

This command is valid for Gas Purity Mode. If the command is received when unit is in the Binary Gas or Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

RATO? i[u]

Binary Gas Ratio Measurement

Query the Binary Gas Ratio of Gas i. i = 1 for Primary Gas; i = 2 for Secondary Gas. If omitted, units default to the global ratio units. The answer is returned in either mole or mass fraction, depending on the concentration type (BCTP).

This command is valid for Binary Gas Analyzer Mode. If the command is received when unit is in the Gas Purity or Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

Example

RATO? 1% Query the Gas Ratio of Gas 1 in %.
RATO? 2 Query the Gas Ratio of Gas 2 in the global ratio units.

RAT2? i[u]**Binary Gas Ratio 2nd Measurement**

Query the 2nd result¹ of the Binary Gas Ratio of Gas i. i = 1 for Primary Gas; i = 2 for Secondary Gas. If omitted, units default to the global ratio units. The answer is returned in either mole or mass fraction, depending on the concentration type (BCTP).

This command is valid for Binary Gas Analyzer Mode. If the command is received when unit is in the Gas Purity or Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

¹This is the second concentration (larger value) of a dual concentration measurement. If there isn't a dual concentration it will return the same value as the RATO? command. Refer to *Binary Gas Analyzer* (page 40) for more information.

SSOS?[u]**Measured Speed of Sound**

Query the Measured Speed of Sound for the current Instrument Mode. If omitted, units default to the global speed units.

Example

SSOS? m/s Query the Speed of Sound in m/s.

TCEL?[u]**Cell Temperature**

Query the Cell Temperature. If omitted, units default to the global temperature units.

If an error in a temperature measurement occurs during a query, the unit will return an overload value (9.9E37) to inform the user that there was something wrong with the measurement. The controlling program should identify this and check bits 6, 7 & 8 of the BGA0 status register if it occurs.

Example

TCEL? C Query the Cell Temperature in °C.

XALL?**Query All**

Query Ratio 1, Ratio 2, Gas Temperature, Analysis Pressure, Normalized Speed of Sound and Block Temperature in global units. The returned values are comma delimited.

See the individual commands for details on their specific behavior.

Related Commands

These commands configure measurements and set the screen appearance of the BGA244. Certain commands require that the BGA244 is in the proper Instrument Mode. If not, an execution error will be generated (Error 11, Illegal Mode).

AVGE(?)*i*

Enable Averaging

Set (query) Average Enable {to *i*}. *i* = 0 for disabled; *i* = 1 for enabled.

Example

AVGE1 Enable Averaging

AVGN(?)*i*

Number of Averages

Set (query) the number of spectrums to be averaged {to *i*}. *i* = 2 – 1000

Example

AVGN20 Set the number of averages to 20

BCTP{?}*i*

Binary Concentration Type

Set (query) the Binary Concentration Type {to *i*}. *i* = 1 for mole fraction; *i* = 2 for mass fraction.

MSMD(?)*i*

Instrument Mode

Set (query) the Instrument Mode to *i*. The value of *i* is determined from the following table. Note that some measurements depend on the Instrument Mode and are only valid for that particular mode.

<i>i</i>	meaning
1	Binary Gas Analyzer
2	Gas Purity Analyzer
3	Physical Measurements

Example

MSMD3 Set the Instrument Mode to Physical Measurements.
 MSMD? A return value of 1 indicated the BGA244 is in the Binary Gas mode.

RELH

REL to 100%

Set the 1 Point REL value to so the measurement reads 100% (or 1,000,000 ppm or 1.0 fraction).

This command is valid for Binary Gas 1 Point REL Mode. If the command is received when unit is in the Gas Purity or Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

RELM(?) {i}**REL Mode**

For GPA Mode:

Set (query) the REL Mode {to i}. i = 0 for no REL; i = 1 for REL.

For BGA Mode:

Set (query) the REL Mode {to i}. i = 0 for no REL; i = 1 for 1 Point REL; i = 2 for 2 Point REL; i = 3 for Multi Point REL.

This command is valid for Binary Gas and Gas Purity Modes. If the command is received when unit is in the Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

Example

RELM1 Set BGA to 1 Point REL Mode.

RELV(?) {r} [u]**REL Value**

Set (query) the REL value {to r}. If omitted, units default to the global ratio units.

This command is valid for Binary Gas 1 Point REL mode and Gas Purity Modes. If the command is received when unit is in the Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

Example

RELV? Query the present REL value in global ratio units.

RELZ**REL to Zero**

Set the REL value to so the measured value reads 0% (or 0ppm, 0 fraction).

This command is valid for Binary Gas 1 Point REL and Gas Purity Modes. If the command is received when unit is in the Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

Example

RELZ REL the present measurement to 0% (or 0ppm, 0 fraction).

RELX(?) {i}**Relaxation Correction**

Set (query) the Relaxation Correction {to i}. i = 0 | 1 for OFF | ON.

Example

RELX1 Set the Relaxation Correction ON.

RTBA r1,r2 [u]**Add Multi Point REL Entry**

Add a new Raw/REL value pair r1, r2 to the Multipoint REL Table. If omitted, units default to global ratio units.

Example

RTBA1.0,1.1% Add a Raw/REL pair of 1%/1.1% to the REL Table.

RTBC	Clear Multi Point REL Table
	Clear all entries in the Multi Point REL Table.

RTBD i	Delete Multi Point REL Table Entry
	Delete the i^{th} entry in the Multipoint Table. $0 \leq i \leq (\text{max table entry})$.
	Example
RTBD3	Delete the 3 rd entry in the Multi Point REL Table.

RTBE i,r1,r2 [u]	Edit Multi Point REL Entry
	Edit Multi Point Table Entry i to Raw/REL value pair $r1, r2$. If omitted, units default to global ratio units. $0 \leq i \leq (\text{max table entry})$.
	Example
RTBE4,1.0,1.1%	Set REL Table entry 4 to 1%/1.1%.

RTBN?	Query num points in Multi Point REL Table
	Query the number of points in the Multi Point REL Table.

RTBR?i[u]	Query Multi Point REL Table entry
	Query the i^{th} point in the Multi Point REL Table. If omitted, units default to global ratio units. $0 \leq i \leq (\text{max table entry})$. The returned values (Raw/REL) are comma delimited.
	Example
RTBR?4	Query the 4 th entry in the Multi Point REL Table

RTW1(?) {r1,r2} [u]	Set (Query) Point 1 Raw/REL value
	Set (query) the Point 1 Raw/REL value pair {to $r1, r2$ } for 2 Point REL Mode. If omitted, units default to global ratio units.
	Example
RTW1 1.0,1.1%	Set 2 Point REL Point 1 to 1%/1.1%.

RTW2(?) {r1,r2} [u]	Set (Query) Point 2 Raw/REL value
	Set (query) the Point 2 Raw/REL value pair {to $r1, r2$ } for 2 Point REL Mode. If omitted, units default to global ratio units.
	Example
RTW2?	Query 2 Point REL Point 2.

RUNM(?) {i}	Run Mode
	Set (query) the Run Mode {to i }. $i = 0$ for STOP; $i = 1$ for RUN.
	Example
RUNM0	Set the BGA244 to STOP.

SETT?**Averaged Measurement Settled**

Query the Measurement Settled status. Returns 0 for no averaging or average not settled. Returns 1 for average settled.

UNCT?[u]**Measurement Uncertainty**

Query the Measurement Uncertainty of the currently active Instrument Mode. If omitted, units default to the global ratio units.

This command is valid for Binary Gas and Gas Purity Modes. If the command is received when unit is in the Physical Measurements mode, an Error 11 (Illegal Mode) will be generated.

Example

UNCT?% Query the measurement uncertainty in %.

Limits

Limit commands configure the upper and lower limits for the six measurements that can have limits. See the *Limits* section for each measurement in the Chapter 3: *Operations Guide* for more details on their specific Limits.

Each measurement has its own unit family associated with it. If units are omitted, the selected global unit is used. If units are included with the command, they must be in the correct unit family or an Error 127 (Illegal Units) will be generated.

The table below lists the measurement ID, types and their associated units.

ID	Measurement	Unit Family
1	Binary Measurement	'r' (ratio)
2	Gas Purity Measurement	'r' (ratio)
3	Physical Measurement	'v' (speed)
4	Pressure Meter 1	'p' (pressure)
5	Pressure Meter 2	'p' (pressure)
6	Temperature	't' (temperature)

LIME(?)i,j{k}**Limit Enable**

Set (query) the enable state of the selected limit type {to k}. Parameter i is the Measurement ID. Parameter j = 1|2 for Upper|Lower limits. Parameter k = 0 for disabled; k = 1 for enabled.

Example

LIME 1,2,1 Enable the Binary Gas Analyzer lower limit

LIME?4,1 Query the enable mode for the Pressure Meter 1 upper limit

LIMH(?)i,{d}[u]**Limit Hysteresis**

Set (query) the hysteresis value of the selected limit type {to d}. Parameter i is the Measurement ID. Parameter d is a floating point value in units of the associated measurement. If omitted, units default to the Measurement global units.

Example

LIMH 1,3% Set the Binary Gas Analyzer hysteresis to 3%
 LIMH 6,5 Set the Temperature hysteresis to 5 using the global temp units
 LIMH?4,Pa Query the Pressure Meter 1 hysteresis value in Pascal

LIMM(?)i,j,{d}[u]**Meter Scale**

Set (query) the meter scale value of the selected measurement {to d}. Parameter i is the Measurement ID. Parameter j = 1|2 for Upper|Lower scale. Parameter d is a floating point value in units of the associated measurement. If omitted, units default to the Measurement global units.

The LIMM command is not valid for the Physical Measurement Mode. An Error 11 (Illegal Mode) will be generated for i = 3.

Example

LIMM 2,1,5% Set the Gas Purity Analyzer meter scale upper value to +5%
 LIMM 4,2,25 Set the Press Meter 1 lower scale to 25 global pressure units
 LIMM?1,2ppm Query the Binary Gas Analyzer lower scale value in ppm

LIMS?i**Limit State**

Query the state of the selected limit type. Parameter i is the Measurement ID. Returns the following:

0	within limits
1	below limit
2	above limit

Example

LIMS? 1 Query the limit state for the Binary Gas Analyzer mode

LIMIT(?)i,j,{d}[u]**Limit Trip Value**

Set (query) the limit trip value of the selected measurement {to d}. Parameter i is the Measurement ID. Parameter j = 1|2 for Upper|Lower limits. Parameter d is a floating point value in units of the associated measurement. If omitted, units default to the Measurement global units.

Example

LIMIT 1,1,60% Set the Binary Gas Analyzer upper limit value to 60%
 LIMIT 6,2,10C Set the Temperature lower limit upper value to 10°C
 LIMIT?4,2 Query Pressure Meter 1 lower limit value in global units.

Configuration Commands

Configuration commands set the BGA244 to make measurements. Some commands may set parameters that are not currently relevant depending on the unit's operating configuration. If so the newly set parameters will be applied when the configuration is changed to make them relevant.

Example

The instrument is in Binary Gas mode and the limits are changed for Gas Purity. The limits for the Gas Purity Analyzer have no effect on Binary Gas measurements. However if the Instrument Mode is changed to Gas Purity, the new selection will be used for those measurements.

Gas Selection Commands

Gases from either the Factory Gas table or User Gas table can be set over the remote interface.

Gases from the Factory Gas Table are selected by their CAS numbers (CAS#). The CAS# is a string which can include or omit the dashes. The CAS# of a gas can be found online using Google, Wikipedia or Appendix A, which lists all supported gases in the BGA244 Factory Gas Table and their CAS#'s. The preferred gas name will be displayed on the Home page when the gas is set over the remote interface. If an invalid CAS# is entered, an Error 26 (Invalid Gas) will be generated.

Gases from the User Gas table are selected by the string "USER/s" where s is the User Gas name. See *User Gases* (page 117) for details on entering User gases. The gas name will be displayed on the Home page as "*User/gas name*". If an invalid User Gas value is set, an Error 26 (Invalid Gas) will be generated.

GASP(?) {s}

Set Primary Gas

Set (query) the Primary Gas {to s}.

's' is a string that identifies the gas. This can be the CAS# of any gas in the Factory Gas Table or the name of a User Gas in the User Gas Table.

Example

GASP 7440-37-1

Set the Primary Gas to 7440-37-1 (argon)

GASP?

A return of 7727-37-9 indicates that the Primary Gas is set to nitrogen.

GASS(?) {s}**Set Secondary Gas**

Set (query) the Secondary Gas {to s}.

's' is a string that identifies the gas. This can be the CAS# of any gas in the Factory Gas Table or the name of a User Gas in the User Gas Table.

Example

GASS USER/Blend1 Set the Secondary Gas to User Gas "Blend1".

SWAP**Swap Binary Gases**

Swap the Primary Gas and Secondary Gas. No return. An Error 26 (Invalid Gas) will be generated in either the Primary Gas or the Secondary Gas are not specified.

Heater Commands

Note that even if the 24 V_{DC} power supply is not present, all Heater parameters can be set even though the heaters will not operate. Use the HEST command to see if the heater is operational or the EXPA command to confirm that the 24 V power supply is present.

Bit 1 of the BGA1 status register will report if the Block Heater is actively regulating the temperature.

BLTM?[u]	Endplate Temperature Query the End Plate Temperature. Returns the endplate temperature. If omitted, units default to the global temperature units.
-----------------	--

HEDG(?) {i}	Degas Heater Enable Set (query) the Degas Heater Enable Mode {to i}. i = 0 for disabled; i = 1 for enabled. Example HEDG 1 Enable the Degas Heater.
--------------------	---

HEEN(?) {i}	Block Heater Enable Set (query) the Block Heater Enable Mode {to i}. i = 0 for disabled; i = 1 for enabled. Note that enabling the Block Heater will turn the heater only if 24 V is present. Use the HEST command to confirm that the heater is actually on. Example HEEN 1 Enable the Block Heater
--------------------	--

HEIL(?) {d}	Maximum Heater Current Set (query) the Maximum Heater Current {to d} in amps. Example HEIL? Query the maximum heater current in amps
--------------------	--

HEPW?	Heater Power Query the Block Heater Power. Return value is in watts (float). Example HEPW? A return value of 10 indicates the heater is delivering 10 watts of power to the BGA244 block
--------------	--

HEST?	Block Heater Status Query the Block Heater Status. Returns 1 for heater on, otherwise 0
--------------	---

HETM(?) {d}**Block Heater Temperature**

Set (query) the Block Heater Temperature {to d} in °C.

Example

HETM 50 Set the Block Heater Temperature to 50 °C

Pressure Commands

There are two different methods used to enter pressure information: the value can be directly entered or a pressure transducer can be interfaced using one of the Analog Inputs. Direct entry can always be used regardless of external power supplies. The analog inputs can be used to interface to a pressure transducer if a 24 V_{DC} power supply is connected.

Note that even if the 24 V_{DC} power supply is not present, all Pressure Meter parameters can be set, even though they will not operate. Use the EXPA command to confirm that the 24 V_{DC} power supply is present. If you attempt to query the pressure (PRRD or PRES¹) when the 24 V_{DC} supply is not available, an Error 18 (No 24V Available) will be generated. An Error 25 (Input Error) will be generated if the analog input is disabled.

If an analog over or under range occurs during a pressure query (PRRD or PRES¹), an overload value (9.9E37) will be returned to inform the user that there was something wrong with the measurement. The controlling program should identify this and check the BGA0 and Analog status registers if it occurs. See *Instrument Status Registers* (page 185) for more information.

¹ if Analog In 1 or 2 are configured as pressure gauges and/or selected for Analysis Pressure

PMAX(?)i{,p}[u]

Pressure Gauge Scale Maximum

Set (query) Pressure Gauge scale maximum {to p}. Parameter i = 1|2 for Pressure Gauge 1|2. If omitted, units default to the global pressure unit.

Example

PMAX 1,100psi Set the Scale Maximum for Pressure Gauge 1 to 100 psi

PMIN(?)i{,p}[u]

Pressure Gauge Scale Minimum

Set (query) Pressure Gauge scale minimum {to p}. Parameter i = 1|2 for Pressure Gauge 1|2. If omitted, units default to the global pressure unit.

Example

PMIN 1,0psi Set the Scale Minimum for Pressure Gauge 1 to 0 psi

PMIN?2 Query the Scale Minimum for Pressure Gauge 2 in global units

PRAC(?)i}**Select Analysis Pressure Source**

Set (query) the pressure analysis source {to i}. i is set per the following table. If the selected source is not configured as a pressure gauge, an Error 23 (Illegal Gauge) will be generated.

<u>Value</u>	<u>Pressure Analysis Source</u>
1	Analog Input 1
2	Analog Input 2
3	User Entered

Example

PRAC 1 Set the Pressure Analysis Source to Analog Input 1
 PRAC? A return value of 3 indicates User Entered pressure is selected.

PRAM(?)p][u]**Set Atmospheric Pressure**

Set (query) the Ambient Atmospheric Pressure {to p}. If omitted, units default to the global pressure unit.

Example

PRAM 14.7 psi Set the Atmospheric Pressure to 14.7 psi
 PRAM? Torr Query the Atmospheric Pressure in torr

PREN(?)i,j}**Pressure Gauge Enable**

Set (query) the Pressure Gauge enable mode. i = 1|2 for Pressure Gauge 1|2. Parameter j = 0 for disabled; j = 1 for enabled.

Example

PREN 1,1 Enable Pressure Gauge 1
 PREN? 2 Query if Pressure Gauge 2 is enabled

PRES?[u]**Analysis Pressure**

Query the Cell Analysis Pressure. If omitted, units default to the global pressure units.

Example

PRES? Query the Analysis Pressure in global pressure units

PRRD?i][u]**Pressure Gauge Reading**

Query the pressure of analog pressure gauge i. i is 1|2 for Pressure Gauge 1|2. If omitted, units default to the global pressure units. If the selected source is not configured as a pressure gauge, an Error 23 (Illegal Gauge) will be generated.

Example

PRRD?1 psi Query the value of Pressure Gauge 1 in psi

PRSU(?)i{,j}**Pressure Measurement Scale Units**

Set (query) pressure measurement I scale units {to j}. Parameter j = 0 for absolute units; j = 1 for gauge units. Parameter i is set per the following table.

<u>i</u>	<u>Pressure Measurement</u>
1	Analog Input 1
2	Analog Input 2
3	User Entered

Example

PRSU 1,0 Set the Pressure Gauge 1 to absolute pressure
PRSU? 0 Query User Pressure scale units

PUSR(?)p{u}**User Cell Pressure**

Set (query) the User Pressure. If omitted, units default to the global pressure units.

Example

PUSR 20psi Set the User Pressure to 20 psi

Analog I/O Commands

Some Analog I/O commands may set parameters that are not currently relevant depending on the unit's configuration. The newly set parameters will be applied when the configuration is changed to make them relevant.

Even if the 24 V_{DC} power supply is not present, all Analog I/O parameters can be set although they will not operate. Use the EXPA command to confirm that the 24 V_{DC} power supply is present. The following command will generate an Error 18 (No 24V Available) if the 24 V_{DC} power supply is not present: AIRE, MOMA and AOMA.

Measure Analog Output Commands

The Measure Output is linked to the active Instrument Mode. See *Measure Output* (page 83) for more details.

Note that although the Measure Out is linked to the current Instrument Mode, any of the Measure Out parameters can be set at any time.

A parameter ID is used to specify the Instrument Mode for some commands. Each measurement has its own unit family associated with it. If units are omitted, the current global unit is used. If units are included with the command, they must be in the correct unit family or an Error 127 (Illegal Units) will be generated. The table below lists the measurement ID, types and their associated units.

ID	Measurement	Unit Family
1	Binary Measurement	'r' (ratio)
2	Gas Purity Measurement	'r' (ratio)
3	Physical Measurement	'v' (speed)

MOCN(?)*{i}*

Measure Out Concentration

Set (query) the Measure Out Binary Gas concentration type *{to i}*. Parameter *i* = 1 for 1st solution; *i* = 2 for 2nd solution.

This command is only active for Binary Gas mode. Refer to *Details* (page 42) for more information description.

MOEN(?)*{i}*

Measure Out Enable

Set (query) the Measure Out enable mode. Parameter *i* = 0 for disabled; *i* = 1 for enabled.

Example

MOEN 1 Enable Measure Out

MOMN(?)i{,d}[u]**Measure Out Scale Min**

Set (query) the Measure Out Scale Min of Instrument Mode i {to d}. Parameter i is the Measurement ID shown at the beginning of this section.

MOMX(?)i{,d}[u]**Measure Out Scale Max**

Set (query) the Measure Out Scale Max of Instrument Mode i {to d}. Parameter i is the Measurement ID shown at the beginning of this section.

Example

MOMX 1,0% Set the Measure Out Scale Max of the BGA to 0%
 MOMX 2,+10 Set the Measure Out Scale Max of the Purity Analyzer to +10%
 MOMX?2 Query the Measure Out Scale Max of Physical Measurements in global units

MOTY(?)i}**Measure Out Type**

Set (query) the Measure Out type {to i}. Parameter i selects one of the following output types.

i	Output Type
1	0 – 5 V
2	0 – 10 V
3	4 – 20 mA

Example

MOTY 2 Set Measure Out to 0 – 10V

MOVA?**Measure Out Value**

Query the present Measure Out value. The return value is in volts or amps determined by the Measure Output Type.

An Error 18 (No 24V Available) will be generated if the 24 V_{DC} power supply is not available. An Error 24 (Output Error) will be generated if the output is disabled.

If an over temperature or current alert is active the command will return an overload value (9.9E37). The controlling program should identify this and check the Analog Status register if it occurs. Refer to *Instrument Status Registers* (page 185) for more information.

Analog Output Commands

The Analog Outputs can be linked to user values or one of several different measurements. See Analog Outputs (page 82) for more details.

Note that although the Analog Out is only linked to the selected measurement, any of the Analog Out parameters can be set at any time.

A parameter ID is used to specify the Measurement for some commands. Each measurement has its own unit family associated with it. If units are omitted, the selected global unit is used. If units are included with the command, they must be in the correct unit family or an Error 127 (Illegal Units) will be generated. The table below lists the measurement ID, types and their associated units.

ID	Measurement	Unit Family
1	Speed of Sound	'v' (ratio)
2	Normalized Speed of Sound	'v' (ratio)
3	Cell Temperature	't' (speed)
4	Pressure Meter 1	'p' (pressure)
5	Pressure Meter 2	'p' (pressure)
6	User Setting	none

AOEN(?)i{,j}

Analog Out Enable

Set (query) the Measure Out enable mode. Parameter i = 1|2 for Analog Out 1|2. Parameter j is 0 for disabled; j is 1 for enabled.

Example

AOEN 1,1 Enable Analog Out 1

AOMN(?)i,j{,d}[u]

Analog Out Scale Min

Set (query) Analog Out i, Measurement j, Scale Min {to d}. Parameter i = 1|2 for Analog Out 1|2. Parameter j is the Measurement ID shown at the beginning of this section.

AOMX(?)i,j{,d}[u]

Analog Out Scale Max

Set (query) Analog Out i, Measurement j, Scale Max {to d}. Parameter i = 1|2 for Analog Out 1|2. Parameter j is the Measurement ID shown at the beginning of this section.

Example

AOMX 1,4,10Pa Set Analog Out 1, Pressure Meter 1, Scale Max to 0 Pascal

AOMX 2,10C Set Analog Out 1, Cell Temperature, Scale Max to 10 °C

AOMX?2 Query the Measure Out Scale Max of Physical Measurements in global units

AOSE(?)i{j}**Analog Out Measure Setting**

Set (query) the Analog Out i type {to j}. Parameter i = 1|2 for Analog Out 1|2. Parameter j is the Measurement ID shown at the beginning of this section.

Example

AOSE 1,3 Set Analog Out 1 to Cell Temperature

AOTY(?)i{j}**Analog Out Type**

Set (query) the Analog Out i type {to j}. Parameter i = 1|2 for Analog Out 1|2. Parameter j selects one of the following output types:

j	Output Type
1	0 – 5 V
2	0 – 10 V
3	4 – 20 mA

Example

AOTY1,3 Set Analog Out 1 to 4 – 20 mA

AOUS(?)i{,d}**Analog Out User Setting**

Set (query) Analog Out i User setting {to d} in volts or amps. Parameter i = 1|2 for Analog Out 1|2.

Example

AOUS 2,9.00 Set Analog Out 2 to 9.00 volts (for Analog Out 2 set to volts)
AOUS 1,20e-3 Set Analog Out 1 to 20 mA (for Analog Out 1 set to current)

AOVA?i**Analog Out Value**

Query the present Analog Out i. The returned value is in volts or amps determined by the Analog Out Type. Parameter i = 1|2 for Analog Out 1|2.

An Error 18 (No 24V Available) will be generated if the 24 V_{DC} power supply is not available. An Error 24 (Output Error) will be generated if the output is disabled.

If an over temperature or current alert is active the command will return an overload value (9.9E37). The controlling program should identify this and check the Analog Status Register if it occurs. Refer to *Instrument Status Registers* (page 185) for more information

Example

AOVA?1 A return value of 1.000000 indicates that Analog Output 1 is outputting 1.000 V.

Analog Inputs Commands

Analog Inputs can be read directly or configured as a pressure meter. Refer to *Pressure Commands* (page 176) and *Analog Inputs* (page 81) for more details.

Some Analog In commands set parameters that are not currently relevant depending on the unit's configuration. These parameters will be applied when the configuration is changed to make them relevant.

AILP(?)i{,d}

Analog Loop Power Voltage

Set (query) the Analog Input i Loop Power Voltage {to d} in volts. i = 1|2 for Analog Input 1|2.

Example

AILP 1,9.0	Set the Loop Power Voltage of Analog Input 1 to 9.0 V.
AILP? 2	Query the Loop Power Voltage of Analog Input 2.

AINE(?)i{,j}

Analog Input Enable

Set (query) the Analog Input i Enable Mode {to j}. Parameter i = 1|2 for Analog Input 1|2. Parameter j = 0 for disabled; j = 1 for enabled.

Example

AINE 1,1	Enable Analog Input 1
----------	-----------------------

AIRE?i

Read Analog Input

Query Analog Input i. The returned value is in volts or amps determined by the Analog In Type. i = 1|2 for Analog Input 1|2.

An Error 18 (No 24V Available) will be generated if the 24 V_{DC} power supply is not available. An Error 25 (Input Error) will be generated if the input is disabled.

The controlling program should check the Analog status register (ANAI or ANAR) to ensure that no voltage or current alerts are active when reading the Analog Input to determine that the reading is valid. See *Instrument Status Registers* (page 185) for more information.

Example

AIRE? 2	Query Analog Input 2
ANAI?	A return value of 0 indicates the measurement is valid.

AITY(?)i{,j}**Analog Input Type**

Set (query) the Analog Input i type {to j}. Parameter i = 1|2 for Analog Input 1|2. Parameter j selects one of the following input configurations.

Note that there are two input configurations that are query only. These indicate that the current input is in its overload protection state. See Analog Inputs (page 81) for details.

<u>j</u>	<u>Input Configuration</u>
1	Voltage
2	Current
3	Current w/ Loop Power
4 ¹	Current (overload protection)
5 ¹	Current w/ Loop Power (overload protection)
6 ¹	Current w/ Loop Power (Loop Power protection)
7 ¹	Current w/ Loop Power (overload and Loop Power protection)

¹Query only values

Example

AITY 1,1

Set Analog Input 1 to Current

AITY? 2

A return of 4 would indicate that Analog Input 2 was set to current and is in the overload protection state.

Miscellaneous Commands

These commands control functions that aren't described elsewhere.

ATSV

Autosave

Immediately Autosave the current configuration.

If this command is received while the flash memory is currently being accessed, an Error 27 (Not Allowed) may be generated. If there is an error while accessing or writing to flash memory, an Error 14 (Flash Write Error) may be generated.

Example

ATSV Autosave the current configuration.

AUSRi,j,s

Add User Gas

Add a User Gas specified by i (molar mass), j (gamma) and s (name). See *User Gases* (page 117) for details on adding User Gases.

If an invalid value is set for a User Gas, an Error 26 (Invalid Gas) will be generated.

Example

AUSR 9.62,1.58,HELIOX 80:20 Add Heliox 80:20 to the User Gas Table.

BAKL(?) {i}

Backlight

Set (query) the Display Backlight Intensity {to i}. Parameter i varies from 1 – 10, where 1 is the dimmest and 10 the brightest.

Example

BAKL 5 Set the Backlight Intensity to 50%.

BLTM?[u]

Endplate Temperature

Query the End Plate Temperature. If omitted, units default to the global temperature units.

Example

BLTM? C Returns the temperature of the End Plates in °C

DSPO(?) {i}**Display Orientation**

Set (query) the Display Orientation {to i}. Parameter i is defined per the following table. The orientation is viewed relative to the front panel name and logo.

<u>i</u>	<u>Display Orientation</u>
0	0°
1	+90°
2	180°
3	270°

Example

DSPO 1 Rotate the Display to 0°

DUSRs**Delete User Gas**

Delete User Gas specified by s (name).

If the User Gas with the name 's' is not in the User Gas Table, an Error 26 (Invalid Gas) will be generated.

Example

DUSR HELIOX 80:20 Delete Heliox 80:20 from the User Gas Table.

HOME**Go to Home Page**

Return the GUI to the Home page of the selected Instrumented Mode.

ERAS i**Erase Stored Setting**

Erase instrument setting at location i. The parameter i may range from 1 to 20.

Example

ERAS 3 Erase the setting at location 3.

EXPA?**External Power Available?**

Query if External 24 V power is available. Returns 1 if available, otherwise 0.

Example

EXPA? A Return of 1 indicates +24V is present.

MUTE(?) {i}**Keyclick Mute**

Set (query) the Key Click Mute {to i}. Parameter i = 0 for Mute off and i = 1 for Mute on (no sound).

Example

MUTE 0 Turn the Key Click off.

PASE(?) {i}**Password Enable**

Set (query) the Password Lock mode {to i}. Parameter i = 1 for enabled, i = 0 for disabled.

Example

PASE 1 Set Password lock

PASL?**Password Locked Status**

Query the Password Locked status. Returns 0 for unlocked, 1 for locked.

PASS(?) {i}**Password**

Set (query) the Password {to i}. Parameter i must be between 0 and 9999.

Example

PASS 1234 Set Password to '1234'

PCTM?[u]**PCB Temperature**

Query the PCB Temperature. If omitted, units default to the global temperature units.

Example

PCTM? Returns the temperature of the PCB in global temperature units

SGAM(?) {i}**Set Speaker Amplitude Mode**

Set (query) the Speaker Amplitude Mode {to i}. The value of i is determined from the following table.

<u>i</u>	<u>meaning</u>
1	Auto
2	Full Scale
3	½ Full Scale
4	¼ Full Scale

Example

SGAM1 Set the Speaker amplitude to Auto

UNFA(?)i{s}**Global Units**

Set (query) Global Unit i {to s}. Parameter i and the allowable strings for 's' are listed below.

i	unit type	allowed unit strings
1	Ratio	'%', 'ppm', 'frac'
2	Speed	'm/s', 'kph', 'mph'
3	Temperature	'C', 'K', 'F'
4	pressure	'psi', 'atm', 'bar', 'Pa', 'mmHg', 'torr'

Example

UNFA 1,% Set the Global Ratio Units to %

UNFA?4 A return of 'Pa' indicates the Global Pressure units are Pascal.

VOLT? i**External Power Voltage**

Query Power Supply Voltage i. Returns voltage in volts (float). Parameter i = 1 for USB supply, i = 2 for 24 V_{DC} Supply.

Example

VOLT?2 Returns the value of the 24 V_{DC} power supply.

Status Byte Definitions

The BGA244 reports on its status by means of several status registers. These include the serial poll status byte, the standard event status register (*ESR) and several additional instrument status registers (INSRs). These read-only registers record the occurrence of defined events inside the unit. If the event occurs, the corresponding bit is set to one. Bits in the status registers are latched¹. Once an event bit is set, subsequent state changes do not clear the bit. All bits are cleared when the registers are queried, with a *ESR?, for example. The bits are also cleared with the clear status command, *CLS. The bits are not cleared, however, with an instrument reset (*RST) or a device clear (break signal).

Each of the unit's event status registers has an associated enable register. The enable registers control the reporting of events in the serial poll status byte (*STB). If a bit in the event status register is set and its corresponding bit in the enable register is set, then the summary bit in the serial poll status byte (*STB) will be set. The enable registers are readable and writable. Reading the enable registers or clearing the status registers does not clear the enable registers. Bits in the enable registers must be set or cleared explicitly. To set bits in the enable registers, write an integer value equal to the binary weighted sum of the bits you wish to set.

The serial poll status byte (*STB) also has an associated enable register called the service request enable register (*SRE). This register functions in a similar manner to the other enable registers, except that it controls the setting of the master summary bit (bit 6) of the serial poll status byte.

¹ Besides the latched INSRs, there are unlatched (or immediate) copies of some of the status registers that reflect the value of the bits the instant the register was read. Having both the latched and unlatched copies of the registers can make it easier to determine the state of the BGA244.

Power On Behavior of Status Registers

All Enable Registers are cleared on power on. These include the following registers: *SRE, *ESE, ALAE, BGOE, BG1E, FALE and EVNE.

Serial Poll Status Byte

<u>Bit</u>	<u>Name</u>	<u>Meaning</u>
0	BGOB	An unmasked bit in the BGA0 status register (BGOR) has been set.
1	BG1B	An unmasked bit in the BGA1 status register (BG1R) has been set.
2	FALB	An unmasked bit in the FAULT status register (FALR) has been set.
3	ANAB	An unmasked bit in the ANALOG status register (ANAR) has been set.
4	MAV	The interface output buffer is non-empty
5	ESB	An unmasked bit in the standard event status register (*ESR) has been set.
6	MSS	Master summary bit. Indicates that the instrument is requesting service because an unmasked bit in this register has been set.
7	EVNB	An unmasked bit in the EVENT status register (EVNR) has been set.

The serial poll status byte may be queried with the *STB? Command.

Standard Event Status Register

<u>Bit</u>	<u>Name</u>	<u>Meaning</u>
0	OPC	Operation complete. All previous commands have completed. See command *OPC.
1	Reserved	
2	QYE	Query error occurred.
3	DDE	Unused.
4	EXE	Execution error. A command failed to execute correctly because a parameter was invalid.
5	CME	Command error. The parser detected a syntax error.
6	Reserved	
7	PON	Power on. The unit has been power cycled.

The standard event status register may be queried with the *ESR? command. The standard event status enable register (*ESE) may be used to control the setting of the ESB summary bit in the serial poll status byte.

Instrument Status Register Model

The BGA244 has five groups of Instrument Status Registers (INSRs). Each group contains three separate 16 bit registers. The first register contains a latched copy of each bit. The second register is an enable register for the first register and is used as a mask for the summary bits in the Status Byte register. The third register contains the unlatched version of each bit. The unlatched (or immediate) register shows the value of the bits at the instant it was read. Each Status Register is described below, together with the page in the *Operations Guide* that refers to specific meaning of each bit.

A non-zero value in the BGA0 register indicates there may be something that can make the measurement invalid. It is a good practice to test this register for each measurement query.

BGA Status Register 0

Bit	Meaning	See
0	Dual Concentration (BGA Mode only)	Page 42
1	Degas Heater on	Page 79
2	Measurement is stopped	Page 58
3	Measurement is below allowable range	Page 50
4	Measurement is above allowable range	Page 50
5	No Signal	Page 50
6	> 2 °C mismatch on temperature measurements	Page 87
7	Cell Temperature < 0 °C	Page 87
8	Cell Temperature > 70 °C	Page 87
9	Bad Analysis Pressure Reading	Page 66
10	Condensation Warning	Page 51
11	Safe Mode	Page 205
12-14	reserved	
15	Summary bit for Fault Register (FALR)	Page 192

BGA Status Register 1

Bit	Meaning	See
0	reserved	
1	Heater is not Regulating	Page 78
2	Measurement Meter Limits exceeded	Page 54
3	Pressure Meter 1 Limits exceeded	Page 66
4	Pressure Meter 2 Limits exceeded	Page 66
5	Temperature Meter Limits exceeded	Page 67
6	USB Voltage is over range (USB Power only)	Page 88
7	USB Voltage is under range (USB Power only)	Page 88
8	+24 V is over range (24 V Power only)	Page 88
9	+24 V is under range (24 V Power only)	Page 88
10	REL mode active	Page 61
11	Key Press Detected	
12	Password Lock	Page 91
13	Average Not Settled (0=settled, 1= not settled)	Page 75
14-15	reserved	

Fault Status Register

Bit	Meaning	See
0	Cell Under Temperature fault	Page 86
1	Cell Over Temperature fault	Page 86
2	Fault in Flash Memory	Page 86
3	Firmware update fault	Page 86
4	Heater Fault (>90 °C)	Page 86
5	1.4 V Supply Out of Range Fault	Page 86
6	3.3 V Supply Out of Range Fault	Page 86
7	4.5 V Supply Out of Range Fault	Page 86
8	5V_I Supply Out of Range Fault	Page 86
9	USB Under Voltage Fault	Page 86
10	USB Over Voltage Fault	Page 86
11	24 V Under Voltage Fault	Page 86
12	24 V Over Voltage Fault	Page 86
13	>5°C Mismatch on Temperature Measurements	Page 86
14	unused	
15	Self Test Failure	Page 89

Analog Status Register

Bit	Meaning	See
0	Measure Out Current Alert	Page 82
1	Measure Out Temperature Alert	Page 82
2	Output 1 Current Alert	Page 82
3	Output 1 Temperature Alert	Page 82
4	Output 2 Current Alert	Page 82
5	Output 2 Temperature Alert	Page 82
6	Input 1 Over Current Alert	Page 81
7	Input 1 Under Current Alert	Page 81
8	Input 1 Over Voltage	Page 81
9	Input 1 Under Voltage	Page 81
10	Input 2 Over Current	Page 81
11	Input 2 Under Current	Page 81
12	Input 2 Over Voltage	Page 81
13	Input 2 Under Voltage	Page 81
14	Loop Power 1 Alert	Page 81
15	Loop Power 2 Alert	Page 81

Event Status Register

The Event Registers operate independently from the Event Relay configuration in the unit. A single register group (latched, enable and immediate) contains all of the event condition. These registers have no effect on the Event Relay configuration. User code should be used to determine the appropriate actions based on the event conditions.

Use the EVNC command to set the Event Relay Configuration. Refer to *Events* (page 55) for details on the specific conditions for items in the Event register. Use the RLYF command to force the relays on or off.

Event Status Register

Bit	Meaning	See
0	System Fault	Page 56
1	No Measurement	Page 56
2	+Measurement Limit Exceeded	Page 56
3	-Measurement Limit Exceeded	Page 56
4	+Pressure 1 Limit Exceeded	Page 56
5	-Pressure 1 Limit Exceeded	Page 56
6	+Pressure 2 Limit Exceeded	Page 56
7	-Pressure 2 Limit Exceeded	Page 56
8	+Temperature Limit Exceeded	Page 56
9	- Temperature Limit Exceeded	Page 56
10-15	reserved	

Error Codes

The instrument contains an error buffer that may store up to 20 error codes associated with errors encountered during power-on self tests, command parsing, or command execution. The ERR LED will flash when a remote command fails for any reason. The errors in the buffer may be read one by one by executing successive LERR? commands. The meaning of each of the error codes is described below.

Execution Errors

- 0 No Error**
No more errors left in the queue.
- 10 Illegal Value**
A parameter was out of range.
- 11 Illegal Mode**
The action is illegal in the current mode. This would happen if the user tries to RELZ when the Instrument Mode is set to Physical Measurements.
- 12 Flash Read Error**
There was an error when attempting to read flash memory.
- 13 Recall Failed**
The recall of instrument settings from nonvolatile storage failed because its instrument settings were invalid.
- 14 Flash Write Error**
There was an error when attempting to write flash memory.
- 15 Flash Erase Error**
There was an error when attempting to erase flash memory.
- 18 No 24V Available**
The requested action failed because +24V power supply was not available.
- 19 Cal Error**
The requested action failed because of an error in the calibration data.
- 20 Firmware Error**
The requested action failed because of an error in the Firmware Update.
- 21 Time Out Error**
The requested action failed because of a Time Out error.
- 22 Binary Stream Error**
The requested action failed because of an error in the binary stream.

- 23 Illegal Gauge**
The requested action failed because the Analog Input is not defined as a pressure gauge.
- 24 Output Error**
The requested action failed because the Analog Output is disabled.
- 25 Input Error**
The requested action failed because the Analog Input is disabled.
- 26 Invalid Gas**
The requested action failed because of an improperly specified gas.
- 27 Not allowed**
The requested action failed because another operation blocked it.

Query Errors

- 30 Lost Data**
Data in the output buffer was lost. This occurs if the output buffer overflows or if a communications error occurs and data in output buffer is discarded.

Parsing or Command Errors

- 110 Illegal Command**
The command syntax used was illegal. A command is normally a sequence of four letters, or a '*' followed by three letters.
- 111 Undefined Command**
The specified command does not exist.
- 112 Illegal Query**
The specified command does not permit queries
- 113 Illegal Set**
The specified command can only be queried.
- 114 Null Parameter**
The parser detected an empty parameter.
- 115 Extra Parameters**
The parser detected more parameters than allowed by the command.
- 116 Missing Parameters**
The parser detected missing parameters required by the command.

- 117 Parameter Overflow**
The buffer for storing parameter values overflowed. This probably indicates a syntax error.
- 118 Invalid Floating Point Number**
The parser expected a floating point number, but was unable to parse it.
- 120 Invalid Integer**
The parser expected an integer, but was unable to parse it.
- 121 Integer Overflow**
A parsed integer was too large to store correctly.
- 122 Invalid Hexadecimal**
The parser expected hexadecimal characters but was unable to parse them.
- 126 Syntax Error**
The parser detected a syntax error in the command.
- 127 Illegal Units**
The units supplied with the command are not allowed.
- 128 Missing Units**
The units required to execute the command were missing.

Communication Errors

- 170 Communication Error**
A communication error was detected. This is reported if the hardware detects a framing, or parity error in the data stream.
- 171 Over run**
The input buffer of the remote interface overflowed. All data in both the input and output buffers will be flushed.

Other Errors

- 254 Too Many Errors**
The error buffer is full. Subsequent errors have been dropped.

Using the USB Drivers

The BGA244 uses an FTDI FT220X USB interface chip. FTDI provides two different types of software drivers to communicate with the interface chip: VCP, a Virtual COM Port and D2XX, a dynamic linked library (DLL). Drivers are available for various versions of Microsoft Windows, Mac OS-X, Linux 2.6 & greater and Android. Some operating systems may not have both drivers available. Detailed information of the drivers and their installation is provided on the FTDI web site: www.ftdichip.com/Drivers.

VCP or Virtual COM Port makes the BGA244 look like an additional COM port available to the computer. Application software can access the BGA244 in the same way as it accesses any standard COM port. This is usually the simplest way to control the BGA244 using the USB port.

The D2XX driver gives access the BGA244 through a DLL. Application software can access the BGA244 through a series of DLL function calls. See the *D2XX Programmer's Guide* on the FTDI web site for a list of available functions. The D2XX driver is a bit more complicated to use, but is faster than the VCP.

Both methods use the remote commands previously described in this chapter to communicate with the BGA244. Be sure that the appropriate driver is installed on any computer that communicates with the BGA244 over USB.

Windows Drivers for the BGA244

The first time the BGA244 is connected to the USB port of a computer running Microsoft Windows, you will likely be prompted with a "New Hardware Found" message and an invitation to search for the USB Driver. There are two USB drivers for the device (VCP and D2XX drivers). Depending on the version and configuration, Windows may either automatically install the drivers or prompt you to search for them. Allow Windows to install the drivers. Occasionally only a single driver will install. In some cases disconnecting and reconnecting the BGA244 will cause the second driver to load correctly.

If the drivers do not load successfully see Appendix E for details on manually installing the drivers.

Drivers for Other Operating Systems

FTDI has USB drivers for several different operating systems besides Microsoft Windows. These include Linux, Android, MAC OSX, OS8 & OS9. See Appendix E for details on manually installing the drivers.

Chapter 7: Service

Troubleshooting

The following section describes different problems that may arise while using the BGA244 and possible solutions. If you don't see a solution to a problem you experience please contact SRS for additional information. If you are contacting SRS about a problem with your BGA244, please have the following information readily available.

- The unit serial number
- The firmware version number
- A description of the problem
- Any error messages, faults or alerts

Power On

Ordinarily the LEDs will flash briefly when power is first applied to the BGA244, followed by the LCD screen lighting up and then displaying the boot up screen. Check the power source if the LEDs and the LCD screen do not light when power is applied. Testing with a second power source or cable may help determine if the BGA244 or power source is defective.

Certain problems during the initialization process are indicated by blink codes of the POWER and ERROR LEDs. There must be a minimum operational voltage of for the BGA244 to operate.

- USB $\geq 3.5V$
- +24 $\geq 18 V$

POWER LED

The power LED will flash at about 4 Hz for a fault on the active power supply.

- USB Fault: Voltage $< 4.5V$ or $> 5.3V$
- +24 V Fault: Voltage $< 20V$ or $> 26V$

In case of a fault check the external power supply voltage for the BGA244 is within specification, as well as the power supply cable. If operating on USB power make sure that the USB power supply or computer can supply enough current and that the USB cable is a heavy enough gauge. See *Power* (page 22) for details.

Testing with a second power source or cable may help determine if the BGA244 or power source is defective.

ERROR LED

If the ERROR LED is continuously lit after power on it indicates a System Fault occurred during Self Test.

Power cycle the unit if this occurs. If the error occurs repeatedly see the following section for more information.

LCD Display not Active

If the LCD display does not light up shortly after power is applied there is either a problem with the BGA244 or the backlight brightness has been reduced too much to be viewed. Try a Hard Reset to force the backlight brightness to full on if the POWER and ERROR LEDs indicate that the unit is operating correctly. See *Hard Reset* (page 92) for details.

Self Test Fail

The unit will initially display the splash screen then go to the About page during initialization and while running Self Test. Normally the display will change to the Home page as soon as Self Test is complete. However if Self Test fails, the BGA244 will remain on the About page and display “Self Test Fail” at the bottom of the page.

Power cycle the unit if this occurs. If the error occurs repeatedly see the following section for more information.


Self Tests

Self Test automatically runs during power on or reset or can be started from the front panel or over the computer interface. It verifies that the majority of the mechanical and electrical parts of the BGA244 are operating in their nominal state. After a successful Self Test the unit will briefly display “Self Test Pass” and then jump to the Home page when run during power on.

If Self Test fails, the BGA244 will remain on the About page and display “Self Test Fail” at the bottom of the page. If this occurs only the [SELF TEST] key will appear. Press this to see which of the Self Tests failed.

The BGA244 will not operate properly after a failed Self Test. Occasionally a test will fail due to some transient condition, so the first step is to re-run it. Either power cycle the unit or start the Self Test from the front panel or over the computer interface. See *Self Tests* (page 89) for a detailed explanation of each the Self Tests.

In some unusual cases it may be necessary to completely power off the unit for at least 10 seconds to completely reset the BGA244.



The only recurring faults in Self Test that can be corrected by the user relate to the power supplies. Refer to *Power* page 22 more information on resolving these.

TABLE 23: POWER SUPPLY FAULT DESCRIPTION

USB Supply	External USB Power Supply is within range Not performed if +24V is present
24 V Supply	External 24 V Power Supply is within range Performed if +24V is present

If the Self Test failure persists the unit will have to be returned to SRS for service. Please note which self tests fail when contacting SRS to assist in repairing the unit.

System Faults

The BGA244 constantly monitors a number of different parameters to ensure it is operating correctly. If it detects a problem it reports it as a System Fault. A few of them may be self clearing or be able to be corrected by the user. But most indicate that something is broken in the BGA244. Power cycle or reset the unit to see if the fault corrects itself. If the error occurs repeatedly see the following table for directions on how to proceed.

TABLE 24: SYSTEM FAULT DESCRIPTION

Fault Name	Meaning	Action Required
USB Undervoltage	External USB Supply <4.45V	Check USB power source and cable
USB Overvoltage	External USB Supply > 5.50V	Check USB Power source. Do not operate at this voltage.
24V Overvoltage	External 24 Supply > 28V	Check 24 V Power source. Do not operate at this voltage.
Under Temperature	Cell Temperature reads < -20 °C	Verify temperature. If valid turn on heaters to increase temperature.
Over Temperature	Cell Temperature reads > 80 °C	Verify temperature. Turn off heaters if they are on.
Flash Fault	Fatal Error in Flash Memory	See <i>Safe Mode</i> later in this chapter.
Firmware	Error in Firmware Update	See <i>Safe Mode</i> later in this chapter.
Heater Fault	Block Temperature reads > 90 °C	Turn off heater
1.4V Under Voltage	Internal 1.4V Supply too low	Fatal if it persists; Return to SRS
1.4V Over Voltage	Internal 1.4V Supply too high	Fatal if it persists; Return to SRS
3.3V Under Voltage	Internal 3.3V Supply too low	Fatal if it persists; Return to SRS
3.3V Over Voltage	Internal 3.3V Supply too high	Fatal if it persists; Return to SRS
5V_I Under Voltage	Internal 5V Supply too low	Fatal if it persists; Return to SRS
5V_I Over Voltage	Internal 5V Supply too high	Fatal if it persists; Return to SRS
Therm Imbalance	Thermistors > 5 °C imbalance for longer than 1 minute	Possibly caused by condensation. May be able to reverse by putting the cell under vacuum.
Self Test	One of the Self Tests failed	See Self Tests Fatal if it persists; Return to SRS

Alerts

An Alert indicates that non-critical problem has occurred, usually with an external device, power supply or wiring. Ordinarily the BGA244 is able to make accurate measurements with an active alert. There are several groups of Alerts: Temperature, Power Supply, General and Analog I/O.

Temperature Alerts indicate that the gas in the cell is outside the normal operating range.

TABLE 25: TEMPERATURE ALERT DESCRIPTION

Alert Name	Action Required
Under Temperature (< 0 °C)	Use the Block Heater to increase temperature
Over Temperature (> 70 °C)	Turn off the Block Heaters; cool the cell if necessary
Temperature Mismatch (> 2 °C)	This may occur if the gas temperature changes dramatically over a short period of time for some orientations of the BGA244. Change mounting orientation if it is a problem.

Power Supply Alerts indicate that external power is approaching its minimum or maximum values. See *Power* (page 22) for details on power supplies and cabling.

TABLE 26: POWER SUPPLY ALERT DESCRIPTION

Alert Name	Meaning
USB Undervoltage	External USB Supply < 4.6 V
USB Overvoltage	External USB Supply > 5.3 V
24V Undervoltage	External 24 Supply < 22 V
24V Overvoltage	External 24 Supply > 26 V

Heater Not Regulating indicates that the heater power is pinned either full on or full off. This will usually occur for a few minutes when the heaters are first turned on or the temperature is changed by a large amount.

Factory Safe Mode indicates a fatal error occurred and the unit re-booted into the Factory Safe Mode. See that section below for more details.

Analog I/O Alerts are usually caused by disconnected or shorted wires or improper configuration of the BGA244 or the external devices.

TABLE 27: ANALOG I/O ALERT DESCRIPTION

Alert Name	Probably Cause
Input Undervoltage (V In only)	Input polarity reversed
Input Overvoltage (V In only)	4 – 20 mA device connected with voltage input selected
Input Undercurrent (I in only)	Current loop wire disconnected
Input Overcurrent (I in only)	Current loop wire shorted
Input Loop Overcurrent (I in only)	Improper device connected
Output Current (I Out)	Current loop wire disconnected or load resistor > 840 Ω
Output Temp	Output shorted

Screen Messages

Screen Messages indicate that there is an issue with the measurement or the BGA244. They can range from the unit being stopped or a configuration interfering with a measurement to a fault within the BGA244 or a problem with external wiring. If a message only appears for a few seconds it can ordinarily be ignored. However if it repeatedly occurs or remains on for more than a few seconds it may require further attention.

Stopped

This indicates that the BGA244 is in STOP mode and no measurements are being made. Pressing the [STOP] key will take you to the Setup page where the Run mode can be selected. This can also be accessed at (Home/Setup).

Degas On

This indicated that the Degas heaters are on and measurements can't be made. Go to (Home/Setup/Control Panel/Heaters) to turn the Degas heaters off and return to the normal operating state.

>102%, <-2%

>102% or <-2% (or their equivalent in ppm or fraction) indicate that the binary gas measurement is out of range for the selected gas combination (Binary Gas Analyzer only).

Confirm that the correct gases have been selected. Check that the operating pressure is correct if using gases with similar speeds of sound. Confirm that there are no contaminating gases in the system. Remember that any third gas can cause problems with binary (two) gas measurements.

No Signal

This indicates that the BGA244 cannot recover the acoustic signal or identify a valid speed of sound for the gas. It can occur for any of several different reasons.

A rapid change in gas species, temperature or pressure with a large number of averages set can smear the acoustic spectrum making it impossible to identify the lines. Try reducing the number of averages or waiting until successful measurements have resumed. In some cases it may be necessary to minimize rapid changes to the gas.

Confirm that the operating pressure is adequate for gases being used. See *Minimum Operating Pressures* (page 99) for information on the minimum recommended operating pressure for different gas species.

Run Self Test (either power cycle the unit or go to Home/Setup/Control Panel/Self Test) to confirm that a hardware problem has not developed since the instrument was last started.

Check for vibrations or stray AC electric or magnetic fields that may interfere with the transducers. See *Interference* (page 103) for more information.

Invalid

Invalid indicates that the BGA244 cannot report a valid result based on the measured speed of sound. This can briefly occur after a major disruption of the signal. Other causes include the following.

Extremely dense gases operating at high pressures can cause a couple of different problems. First check that the output amplitude isn't set too high. Go to Home/Setup/Control Panel/Measurement. The amplitude is normally set to Auto. Try setting it to $\frac{1}{4}$ to see if it resolves the problem. Dense gases also may also have a speed of sound below the BGA244s operating range (<100 m/s).

If the BGA244 cannot calculate a valid solution for the binary gas concentration from the measured speed of sound it will report invalid (Binary Gas Analyzer only). This can occur with problem gases (see *Two Solutions* below) if the measured speed of sound is at or below the minimum. Confirm that the correct gases have been selected and that the operating pressure is correct. Confirm that there are no contaminating gases in the system.

Bad Pressure

This indicates that the measured pressure is invalid. It's usually caused by a problem with the pressure transducer configuration or a missing 24 V_{DC} power supply. Confirm the 24 V_{DC} power supply is connected and make sure the Analog Input is properly configured for the pressure transducer. Be sure to set the ambient pressure if you are using gauge units. See *Pressure* (page 65) for more information.

If the configuration is OK and power is connected, check the wiring to the pressure transducer. The Analog Input page (Home/Setup/Control Panel/Analog IO/Input 1|2) will indicate an Alert if the transducer is improperly configured and connected.

Condensation

This indicates that at least one of the gases in the cell is at or near condensation. The BGA244 cannot measure gases that are at or near their condensation point. See *Condensation* (page 95) for more information.

The Block Heaters may be able to increase the cell temperature above the condensation point. Otherwise you must reduce the concentration of the gas that is in danger of condensing.

Invalid Gas Message

Stored settings (in the BGA244) and stored configuration (on BGAMon) can reference User Gases. An error is created if a setting or configuration tries to recall a non-existent User Gas. If this occurs, the screen message “Invalid Gas” is displayed and an error is reported. To correct this first select a different gas, then re-add the User Gas and select it.

Two Sol

Two Sol (Two Solutions) indicates there are two valid molar ratios for a given speed of sound (Binary Gas Analyzer only). This only occurs for a few problem mixtures. Refer to *Problem Mixtures* (page 112) for more information.

System Fault

System Fault indicates that a serious problem has occurred with the BGA244. Press [SYSTEM FAULT] to go to the Faults page. See *System Faults* earlier in this chapter for diagnostic and troubleshooting information.

Alert

An Alert indicates that non-critical problem has occurred, usually with an external device, power supply or wiring. Press [ALERT] to view the active Alert. See *Alerts* earlier in this section for diagnostic and troubleshooting information.

Factory Safe Mode

If the BGA244 firmware encounters a fatal error, the BGA244 will time out after about 5 seconds and restart in the Factory Safe Mode. While the BGA244 can be temporarily operated in this mode, this should be corrected as soon as possible to ensure full functionality. The Factory Safe Mode is indicated by the following:

- The About screen displays the message: “Factory Default Setup” in large yellow text at the bottom of the screen. This screen can be viewed during initialization, as well as from (Home/Setup/Control Panel/About).
- The Factory Safe Mode Alert is active. See *General Alerts* page 88 for information on alerts.
- Bit 11 of the BGA0 status register is set.



There are three possible conditions that can cause the firmware to re-boot to the Factory Safe Mode. Use the following techniques to correct the conditions.

Firmware Update

A fatal error can be caused if a file is corrupted while being stored to flash during a firmware update. If this occurs, each time the unit is powered on it will enter the Factory Safe Mode.

Reinstall the Firmware to correct this error. Refer to *Updating Firmware*, page 143 for details. If this problem occurs repeatedly contact SRS for assistance.

Stored Setup

In rare cases, a corrupted stored setup can cause a fatal error. As part of the Factory Safe Mode, the Default setup will be loaded to enable this to be corrected. To determine if a corrupted stored setup caused the fatal error, reload it when the unit is in the Factory Safe Mode. If this causes a restart then that setup is corrupted.

The corrupted setup must be deleted to correct this error. From the Factory Safe Mode, delete the setup and then restart the unit.

Firmware Error

In the unlikely event that the BGA244 gets a fatal error that isn't related to updating the code or a corrupted stored setup, the installed version of the firmware may contain a fatal error.

Confirm that this error is not caused by a stored setup before proceeding. Reinstall the firmware as described in *Updating Firmware*, page 143. Contact SRS for assistance if this problem occurs with the newly installed version.

Locked

The BGA244 keypad can be locked out with a password enable to keep unauthorized personnel from changing settings on the BGA244. This is indicated by the "Locked" message at the top left of the screen. Only the Home page can be viewed when the keypad is locked.

If you forget the password, the unit can be reset using the Hard Reset function described below. This will disable password locking and reset the password to the default value of 0000.

Store and Recall Failure

Setups are stored into Flash memory with a typical life time of about 100,000 erase cycles. In rare cases a large number of erase cycles can permanently "wear out" a particular memory location, rendering it unusable. Ordinarily a different memory location can be used. But if your application requires this particular location, the unit will have to be returned to SRS to replace and reprogram the Flash memory.

Note: Don't continuously store settings over the computer interfaces to avoid wearing out the memory. Practically this is the only way to wear out a location.

Touchscreen Problems

The LCD touchscreen can occasionally drift so that the displayed keys and keypress areas don't align. Running Screen Calibration (page 74) should fix this.

Use the following procedure in case the display has drifted so far that you cannot activate screen calibration.

- Remove power to the unit.
- Press and hold a finger on the display while applying power. Keep it there until after the unit has fully booted and the About screen has vanished.
- At this point the screen will display "Release screen to start screen calibration". Remove your finger and follow the instruction to calibrate the screen.

Hardware Reset

The hardware reset button can be used to reset the BGA244. There are two different types of resets with slightly different behaviors. Use an unbent paper clip or a small screw driver to press the reset button. See *Hardware Reset* (page 92) for more details.

Normal Reset

Press the reset button once. Releasing the button forces a Normal Reset. This acts the same as power cycling the BGA244. This is ordinarily used if the BGA244 is in an unknown state and it is difficult to remove power to restart the unit.

Hard Reset

Press the reset button once. Release it, then immediately press and hold the button down for at least 10 seconds to execute a Hard Reset. This acts the same as power cycling the BGA244 and Recalling Setup 0 (Default Setup). Hard Reset is ordinarily used if the BGA244 is in an inoperable state where loading the Default Setup is impossible, or to reset a forgotten password.

Installing BGA-M

This section describes how to replace the standard cover (with LCD display) with accessory BGA-M, a rugged solid metal cover. This is typically done if the unit is operating in an environment where the LCD display may be damaged or broken.

Remember that units must either be configured prior to installation or over one of the computer interfaces since the LCD display and touch screen will be removed. The simplest way to configure the BGA244 over a computer interface is using the BGAMon software provided by SRS. See Chapter 5: *BGAMon* for more information. Otherwise it can be configured by user written code. See Chapter 6: *Remote Programming* for more information.

Replacement of the standard cover with BGA-M requires a Philips #1 screwdriver. Perform these steps with the unit on a flat surface clear of any obstructions.

Remove the eight 4-40 x 1/4" black screws holding the top cover to the rest of the BGA244.

Lift the cover off the BGA244. You may need to wiggle the cover a bit to remove it.

Place the BGA-M cover on the unit. Make sure the ground lug is on the same side as the terminal strip headers.

Loosely install the eight 4-40 x 1/4" black screws holding the top cover to the rest of the BGA244. Tighten the screws.

Maintenance

The BGA244 does not require any regular maintenance or replacement of sensors, reference gases or filaments. The only user serviceable parts are the gas fittings (BGA244 only).

Replacing Gas Fittings

Ordinarily the NPT gas fittings of the BGA244 should never require service. However if they become damaged or worn due to improper or repeated threading they can be replaced. They can also be replaced by one of the accessory gas fittings or other male $\frac{1}{8}$ "-27 NPT fittings for applications that require minimum insertion length. It is not recommended that NPT pipes be directly connected to the BGA244.

Removal and installation of the standard gas fittings requires a 9/16th inch wrench. Disconnect any gas lines and electrical connections prior to removing a fitting. Perform these steps with the unit on a flat surface clear of any obstructions.

Removal Guidelines

1. Grip the unit firmly across the top cover. Put the wrench on the desired gas fitting and turn the wrench counter-clockwise to initially loosen the fitting. Note that it may initially take a large amount of force to break the fitting loose.
2. Continue unscrewing the fitting counter-clockwise until it comes out of the cylinder.
3. Remove any loose thread sealant from the cylinder threads. Try to keep any contaminants or debris out of the BGA244 cylinder.

Use thread sealant when replacing the gas fittings. SRS recommends Loctite 565 thread sealant to seal leaks and prevent galling/seizing of the threads. A 0.2-oz tube of the Loctite is available from SRS (accessory 45855K12).

Installation Guidelines

1. Apply a strip of thread sealant around the male pipe threads, leaving the first 2 turns uncovered.
2. Finger tighten the adapter into the fitting.
3. Grip the unit firmly across the top cover. Wrench tighten the adapter 2 $\frac{1}{4}$ turns past finger tight.
4. Allow the Loctite 565 at least 24 hours to cure before applying pressure or vacuum to the system.

Replacing Gas Fittings for the BGA244E

As with the standard BGA244, the NPT gas fittings of the BGA244E should never require service. However if they become damaged or worn due to improper or repeated threading they can be replaced.

Removal and installation of the standard gas fittings requires a 9/16th inch wrench and either a pipe wrench or pliers. Disconnect any gas lines and electrical connections prior to removing a fitting. Be sure to have sufficient working space before attempting to replace the fitting.

Removal Guidelines

1. Grip the tubing just below the enclosure seal of the BGA244E using a pipe wrench or pliers. Put the 9/16th inch wrench on the gas fitting and turn it counter-clockwise to initially loosen the fitting. Note that it may initially take a large amount of force to break the fitting loose.
2. Continue unscrewing the fitting counter-clockwise until it comes out of the tubing.
3. Remove any loose thread sealant from the cylinder threads. Try to keep any contaminants or debris out of the BGA244E tubing.

Use thread sealant when replacing the gas fittings. SRS recommends Loctite 565 thread sealant to seal leaks and prevent galling/seizing of the threads. A 0.2-oz tube of the Loctite is available from SRS (accessory 45855K12).

Installation Guidelines

1. Apply a strip of thread sealant around the male pipe threads, leaving the first 2 turns uncovered.
2. Finger tighten the adapter into the fitting.
3. Grip the tubing just below the enclosure seal of the BGA244 using a pipe wrench or pliers. Wrench tighten the adapter 2 ¼ turns past finger tight.
4. Allow the Loctite 565 at least 24 hours to cure before applying pressure or vacuum to the system.

Terminal Strips

Replacement terminal blocks are available from Digikey or other distributors.

Connector	Phoenix Contact p/n	Digikey p/n
C4, C6 (5 pin)	1803604	277-1164-ND
C7 (8 pin)	1803633	277-1167-ND
C8 (6 pin)	1803617	277-1165-ND

Cleaning

Clean the LCD display with either a spray glass cleaner or optical cloth.

Clean the casing with a mild detergent or spray cleaner. Take care to prevent any cleaner from entering the unit through the ventilation holes.

Calibration

The items that determine the BGA244's accuracy are the physical dimensions of the steel cavity, the crystal oscillator frequency and the temperature sensing thermistors. The physical dimension of the cavity doesn't change over time. The crystal oscillator and thermistors age at a rate several orders of magnitude lower than the BGA244's accuracy specification. Because of the lack of drift for any of the critical components in the BGA244 there is no specified periodic calibration interval.

The only item that suffers from any appreciable drift is the display. This can be calibrated using the Screen Calibration procedure (page 74) in Chapter 3.

The Performance Verification procedure described in the following section can be used to confirm that the BGA244 is operating properly and meeting its specification. Contact SRS for repair and calibration information if the unit fails any the following procedures.

Performance Verification

There are several parts to the performance verification procedure. Self Test confirms that the primary functions of the BGA244 are fully functional. The Analog I/O Tests verifies that the Analog Inputs and Outputs are operational and meet specification. The Measurement Accuracy Tests confirm that the gas measurements meet the specified accuracy.

Self Test

Self Test checks the functional operation of many important internal components. It is automatically run upon power on. Self Test reports the status of the different tests on the Self Test page. It can also be started from the Self Test page (Home/Setup/Control Panel/Self Test). See *Self Test* (page 89) in Chapter 3 for details.

Self Test can also be initiated using the *TST? command. See *TST? (page 156) for a description of the command and definition of the return codes.

See *Troubleshooting* (page 199) for possible solutions to any Self Test failures.

Analog I/O Accuracy

These tests confirm that the BGA244s Analog Inputs and Outputs meet specification. Voltage and current outputs are tested using a DMM. Voltage and current inputs are tested comparing the BGA244 measurements with a DMM.

Required Equipment

- Keysight 33461A 6½ digit DMM or equivalent. The key specification is 0.0035% DC Accuracy.
- 8 and 5 pin headers for the BGA244 terminal strips. See *Terminal Strips* (page 211) for replacements if needed.
- Wires and leads to connect the BGA244 to the DMM.

Analog Measure Output

The Measurement Output value cannot be set directly. Instead the output is set by adjusting the Output Scale values and comparing the expected output to the measured output.




FIGURE 33: SETUP FOR MEASURE OUT VOLTAGE TEST




FIGURE 34: SETUP FOR MEASURE OUT CURRENT TEST

1. Connect the Measurement Output (C6) to the DMM as shown in Figure 33. Set the DMM to measure DC Volts at maximum resolution. Select auto range.
2. Recall the Default Setup. Go to (Home/Setup/Select Gas) and select whatever gas is currently in the gas chamber. This sets the BGA244 to measure the purity of the gas that is in the chamber. Confirm that the measurement is $\pm 20\%$ or less.
3. Go to (Home/Setup/Control Panel/Analog IO/Measurement Output). Set Scale Max to 100%. Verify that the BGA244 Output Value¹ and the DMM agree to within the limits listed in Table 28. Set Scale Max back to 10%.
4. Set Scale Min to -100%. Record the BGA244 Output Value¹ and the DMM measured value. Set Scale Min back to -10%. Verify that the BGA244 Output Value¹ and the DMM agree to within the limits listed in Table 28.
5. Change the Output Type to 0 - 5V. Repeat steps 3 and 4.
6. Connect the Measurement Output (C6) to the DMM as shown in Figure 34. Set the DMM measure DC Current at maximum resolution. Select auto range.
7. Repeat steps 3 and 4.

¹ The BGA244 Output Value is displayed just below the Scale Max value.

TABLE 28: MEASURE OUT TEST LIMITS

	BGA244 Output Value	DMM Reading	Limits
0 – 10V Low			± 1.22 mV
0 – 10V High			± 3.25 mV
0 – 5V Low			± 1.11 mV
0 – 5V High			± 1.12 mV
4 – 20 mA Low			± 15.45 μ A
4 – 20 mA High			± 28.54 μ A

Analog Inputs and Outputs 1 & 2

The Analog Outputs can be set directly from the User setting. The outputs and inputs are connected together in loopback fashion and monitored using the DMM. Note that the Analog In limits are calculated relative to the DMM measurement, not the Analog Output values.




FIGURE 35: SETUP FOR ANALOG I/O VOLTAGE TEST




FIGURE 36: SETUP FOR ANALOG I/O CURRENT TEST

1. Connect Analog Output 1, Analog Input 1 (C7) and the DMM HI input as shown in Figure 35. Set the DMM to measure DC Volts at maximum resolution. Select auto range.
2. Recall the Default Setup. Go to (Home/Setup/Control Panel/Analog IO).
3. Set Analog Output 1 User Setting to 1 V. Verify that the DMM, BGA244 Output and Input values all agree to within the limits listed in Table 29.
4. Set Analog Output 1 User Setting to 10 V. Verify that the DMM, BGA244 Output and Input values all agree to within the limits listed in Table 29.
5. Change the Output Type to 0 - 5V. Repeat steps 4 and 5 for 0.5V and 5V.
6. Connect Analog Output 1, Analog Input 1 and the DMM I input as shown in Figure 36. Set the DMM to measure DC Current at maximum resolution. Select auto range.
7. Change the Output 1 Type to 4 – 20mA.
8. Set Analog Output 1 User Setting to 4 mA. Verify that the DMM, BGA244 Output and Input values all agree to within the limits listed in Table 29.

9. Set Analog Output 1 User Setting to 20 mA. Verify that the DMM, BGA244 Output and Input values all agree to within the limits listed in Table 29.
10. Repeat steps 1 to 10 for Analog Output 2 and Analog Input 2 (C7).

TABLE 29: ANALOG I/O TEST LIMITS

	Set Value	DMM	V Out Limit	Analog In	V In Limit ¹
0 – 10V	1 V		± 1.25 mV		± 1.25 mV
0 – 10V	10.0 V		± 3.50 mV		± 3.50 mV
0 – 5V	0.5 V		± 1.12 mV		± 1.12 mV
0 – 5V	5.0 V		± 2.25 mV		± 2.25 mV
4 – 20 mA	4 mA		14.0 µA		14.0 µA
4 – 20 mA	20 mA		30.0 µA		30.0 µA

¹Note that the Analog In limits are calculated relative to the DMM measurement, not the Analog Output values.

Measurement Accuracy

BGA244 calculates concentration or normalized purity from two measurements: gas temperature and speed of sound.

Temperature

This test confirms that the BGA244's gas temperature sensing thermistors are operating properly. It compares the gas temperature measurement with the block temperature and a temperature sensor on the printed circuit board.

The most important requirement is that all parts of the BGA244 are at thermal equilibrium. This requires no gas flow, heaters off and a stable temperature environment.

Required Equipment

Either USB or +24 V_{DC} to power the BGA244.

1. Connect power to the BGA244 and allow the BGA244 to reach thermal equilibrium. This may take several hours.
2. Go to Home/Setup/Control Panel/Status
3. Observe the Gas Temperature, Block Temperature and PCB Temperature.
 - The Gas Temperature should be within ± 2 °C from the Block Temperature.
 - The Gas Temperature should be within ± 2.5 °C from the PCB Temperature

Gas Purity

This test confirms that the BGA244 meets its accuracy specification for gas measurements. This is done by flowing a pure reference gas (Argon or Nitrogen) through the cell and confirming that the measured speed of sound is within tolerance of the ideal speed of sound.

The exact configuration of the gas systems can vary and may be incorporated within the user's installation. The most important requirements are to:

- Ensure that pure reference gas is flowing through the BGA244
- Enter the correct analysis pressure

Required Equipment

Tank of at least 99.99% pure Argon or Nitrogen

The BGA244 should vent to atmosphere with minimal pressure drop. A short length of tubing or other apparatus can be connected from the BGA244 outlet to atmosphere if necessary. Be certain that the pressure drop from the BGA244 outlet to atmosphere is <1 psi (6.9 kPa).

1. Power on the BGA244. Recall the Default Instrument setup. (Home/Setup/Store|Recall/Recall/Default). If using Nitrogen, change the primary gas to Nitrogen (Goto Home/Setup/Select Gases/Select Primary & select N2).
2. Go to (Home/Setup/Pressure/Enter_User_Pressure/Ambient). Enter the atmospheric pressure of your location as the Ambient pressure value. Ordinarily this value will depend primarily on your elevation. Use the closest value from the Table 30.

TABLE 30: PRESSURE VS ALTITUDE

Elevation	Nominal Atmospheric Pressure
Sea level (0' or m)	14.7 psi (101 kPa)
1000' (305 m)	14.2 psi (98 kPa)
2000' (610 m)	13.8 psi (95 kPa)
3000' (914 m)	13.2 psi (91 kPa)
4000' (1219 m)	12.8 psi (88 kPa)
5000' (1524 m)	12.3 psi (85 kPa)
6000' (1829 m)	11.9 psi (82 kPa)
7000' (2134 m)	11.5 psi (79 kPa)
8000' (2438 m)	11.2 psi (77 kPa)
9000' (2743 m)	10.7 psi (74 kPa)
10,000' (3048 m)	10.3 psi (71 kPa)

3. Flow at least 2 liters of the reference gas) at 200 - 1000 sccm to fully purge the BGA244.
4. Check that the Gas Purity measurement has fully settled. If not, wait until it settled.
5. Confirm that the Gas Purity measurement is $<\pm 0.05\%$ (Argon) or $<\pm 0.08\%$ (Nitrogen).

Appendix A: Gas Table

Factory Gas Table

The Factory Gas Table contains data for about 500 different gases that are supported by the BGA244. Each entry includes the formula, common name, CAS number, up to two alternate names and physical data about the gas. The gas(es) to be analyzed are chosen on the Setup page (page 58). The individual gas can be selected using any of the names, the chemical formula or the CAS number.

Additional gases can be added to the BGA244's User Gas Table. This can be done using BGAMon or over the remote interface. See User Gases (page 117) for details on adding new gases to the User Gas Table.

The table below lists gases in the Factory Gas Table, sorted by molecular weight. Each entry contains the preferred name, up to two alternate names, the formula (in Hill notation), the molecular weight and the CAS number.

Note: New gases are periodically added to the Factory Gas Table. Go to [Gases Measured by the BGA244](#) for the latest list.

TABLE 31: FACTORY GAS TABLE

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Hydrogen, normal	Normal hydrogen		H2	2.01588	1333-74-0n
Hydrogen, ortho	Ortho hydrogen		H2	2.01588	1333-74-0o
Hydrogen, para	Para hydrogen		H2	2.01588	1333-74-0p
Hydrogen			H2	2.01588	1333-74-0
Helium-3	He-3		He-3	3.01603	14762-55-1
Deuterium hydride			DH	3.022	13983-20-5
Helium	Helium-4		He	4.002602	7440-59-7
Deuterium			D2	4.0282	7782-39-0
Methane	R-50	Carbon tetrahydride	CH4	16.0428	74-82-8
Ammonia	Nitrogen trihydride	NH3	H3N	17.03026	7664-41-7
Water	Dihydrogen Monoxide		H2O	18.01527	7732-18-5
Hydrogen fluoride			HF	20.006	7664-39-3
Heavy Water	Deuterated Water		D2O	20.02751	7789-20-0
Neon			Ne	20.179	7440-01-9
Ethyne	Acetylene		C2H2	26.037	74-86-2
Hydrogen cyanide			CHN	27.025	74-90-8
Diborane	Boroethane	B2H6	H6B2	27.67	19287-45-7
Carbon monoxide	Carbon oxide		CO	28.0101	630-08-0
Nitrogen			N2	28.01348	7727-37-9
Ethylene	Dicarbene	Ethene	C2H4	28.05376	74-85-1
Air, dry			N2-O2-Ar	28.96	132259-10-0
Nitrogen oxide	Nitrogen monoxide	Nitric oxide	NO	30.006	10102-43-9
Formaldehyde			CH2O	30.026	50-00-0
Ethane			C2H6	30.06904	74-84-0

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Methanamine	Aminomethane	Methylamine	CH5N	31.057	74-89-5
Oxygen			O2	31.9988	7782-44-7
Methanol	Methyl alcohol		CH4O	32.04216	67-56-1
Hydrazine			H4N2	32.045	302-01-2
Fluoromethane	R-41	Methyl fluoride	CH3F	34.033	593-53-3
Hydrogen sulfide			H2S	34.08088	7783-06-4
Hydrogen chloride			HCl	36.461	7647-01-0
Fluorine			F2	37.99681	7782-41-4
Argon			Ar	39.948	7440-37-1
1,2-Propadiene	Allene	Propadiene	C3H5	40.064	463-49-0
1-Propyne	Prop-1-yne	Methyl acetylene	C3H4	40.0639	74-99-7
Acetonitrile			C2H3N	41.052	75-05-8
1-Propene	Prop-1-ene	Propylene	C3H6	42.07974	115-07-1
Cyclopropane	Trimethylene		C3H6	42.081	75-19-4
Ethyleneimine			C2H5N	43.068	151-56-4
Carbon dioxide	R-744	Carbonyl oxide	CO2	44.0098	124-38-9
Dinitrogen monoxide	R-744A	Nitrous oxide	N2O	44.0128	10024-97-2
Acetaldehyde			C2H4O	44.053	75-07-0
Ethylene oxide			C2H4O	44.053	75-21-8
Propane	Dimethylmethane		C3H8	44.09562	74-98-6
Formamide			CH3NO	45.041	75-12-7
N-Methylmethanamine	N,N-Dimethylamine	Dimethylamine	C2H7N	45.084	124-40-3
Ethanamine	Aminoethane	Ethylamine	C2H7N	45.084	75-04-7
Formic acid	Methanoic acid		CH2O2	46.026	64-18-6
Methoxymethane	1,1'-Oxybismethane	Dimethyl ether	C2H6O	46.06844	115-10-6
Ethanol	Ethyl alcohol		C2H6O	46.06844	64-17-5
Methylsilane			CH6Si	46.144	992-94-9
Ozone	Triatomic oxygen		O3	47.998	10028-15-6
Fluoroethane	R-161	Ethyl fluoride	C2H5F	48.06	353-36-6
Methanethiol	Methylsulfide	Methyl mercaptan	CH4S	48.107	74-93-1
Chloromethane	R-40	Methyl chloride	CH3Cl	50.488	74-87-3
Difluoromethane	R-32	methylene fluoride	CH2F2	52.024	75-10-5
Cyanogen			C2N2	52.035	460-19-5
Vinyl acetylene			C4H4	52.075	689-97-4
Acrylonitrile			C3H3N	53.063	107-13-1
1,3-Butadiene	Butadiene		C4H6	54.09	106-99-0
1-Butyne	Ethylacetylene	But-1-yne	C4H6	54.09	107-00-6
Dimethyl acetylene			C4H6	54.09	503-17-3
1,2-Butadiene			C4H6	54.09	590-19-2
1,2-Oxazole	1-oxa-2-azacyclopentadiene	Isoxazole	C3H3NO	55.058	288-14-2
Propionitrile			C3H5N	55.079	107-12-0
Acrolein			C3H4O	56.063	107-02-8
1-Butene	Butylene		C4H8	56.10632	106-98-9
2-Methyl-1-propene	2-Methylpropene	Isobutene	C4H8	56.10632	115-11-7
Cyclobutane	Tetramethylene		C4H8	56.106	287-23-0
(Z)-2-Butene	cis-2-Butene	Z-2-Butene	C4H8	56.10632	590-18-1
(E)-2-Butene	E-2-butene	trans-2-Butene	C4H8	56.10632	624-64-6
Methyl isocyanate			C2H3NO	57.051	624-83-9
Methyl vinyl ether			C3H6O	58.079	107-25-5
Propionaldehyde			C3H6O	58.079	123-38-6

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Acetone	dimethyl ketone	2-Propanone	C3H6O	58.07914	67-64-1
Butane	n-Butane		C4H10	58.1222	106-97-8
Isobutane	i-Butane	2-Methylpropane	C4H10	58.1222	75-28-5
Acetamide			C2H5NO	59.067	60-35-5
1-Propanamine	1-Aminopropane	n-Propyl Amine	C3H9N	59.11	107-10-8
2-Propaneamine	2-Aminopropane	Isopropyl amine	C3H9N	59.11	75-31-0
N,N-Dimethylmethanamine	Trimethylamine		C3H9N	59.11	75-50-3
Methyl formate	Methyl methanoate		C2H4O2	60.052	107-31-3
Acetic acid			C2H4O2	60.052	64-19-7
carbonyl sulfide	carbon oxide sulfide		COS	60.0751	463-58-1
Methoxyethane	Methyl ethyl ether	Methylethyl ether	C3H8O	60.095	540-67-0
2-Propanol	Isopropyl alcohol	Isopropanol	C3H8O	60.095	67-63-0
1-Propanol	Propyl alcohol	Propan-1-ol	C3H8O	60.095	71-23-8
Ethylenediamine			C2H8N2	60.098	107-15-3
Dimethylsilane			C2H8Si	60.17	1111-74-6
Nitromethane	nitrocarbol		CH3NO2	61.04	75-52-5
Ethylene glycol			C2H6O2	62.068	107-21-1
Ethanethiol	Thioethanol	Ethyl mercaptan	C2H6S	62.134	75-08-1
Methylthiomethane	Thiobismethane	Dimethyl sulfide	C2H6S	62.134	75-18-3
Vinyl chloride			C2H3Cl	62.498	75-01-4
1,1-Difluoroethene	R-1132A	1,1-Difluoroethylene	C2H2F2	64.035	75-38-7
Sulfur dioxide		SO2	O2S	64.0638	7446-09-5
Chloroethane	1-Chloroethane	Ethyl chloride	C2H5Cl	64.514	75-00-3
1,2-Difluoroethane			C2H4F2	66.05	624-72-6
1,1-Difluoroethane	R-152a	Algofrene 67	C2H4F2	66.051	75-37-6
2-Methyl -1-butene-3-yne			C5H6	66.101	78-80-8
1H-Pyrrole	Azole	Pyrrole	C4H5N	67.09	109-97-7
Furan	1,4-Epoxy-1,3-butadiene		C4H4O	68.074	110-00-9
Cyclopentene			C5H8	68.117	142-29-0
3-Methyl-1-butyne			C5H8	68.117	598-23-2
3-Methyl-1,2-butadiene			C5H8	68.117	598-25-4
1-Pentyne	Ethylmethylacetylene	Pent-1-yne	C5H8	68.117	627-19-0
2-Pentyne			C5H8	68.117	627-21-4
Butyronitrile			C4H7N	69.105	109-74-0
Fluoroform	R-23	Trifluoromethane	CHF3	70.01385	75-46-7
1-Pentene	cis-Pentene	Pent-1-ene	C5H10	70.133	109-67-1
Cyclopentane	Pentamethylene		C5H10	70.133	287-92-3
2-Methyl-2-butene	2-Methylbut-2-ene	Amylene	C5H10	70.133	513-35-9
2-Methyl-1-butene			C5H10	70.133	563-46-2
3-Methyl-1-butene	Methylethylethylene	3-Methylbut-1-ene	C5H10	70.134	563-45-1
(2Z)-2-Pentene	cis-2-Pentene	Z-2-Pentene	C5H10	70.134	627-20-3
Chlorine			Cl2	70.906	7782-50-5
Nitrogen trifluoride			F3N	71.019	7783-54-2
Acrylic acid			C3H4O2	72.063	79-10-7
Trimethylaluminium	TMA	Aluminum trimethanide	C3H9Al	72.085	75-24-1
Tetrahydrofuran	Butylene oxide	Diethylene oxide	C4H8O	72.106	109-99-9
Butyraldehyde			C4H8O	72.106	123-72-8
Methylethyl ketone	Butanone	MEK	C4H8O	72.106	78-93-3
Neopentane	Dimethylpropane	2,2-Dimethylpropane	C5H12	72.14878	463-82-1
Pentane	n-Pentane		C5H12	72.14878	109-66-0

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Isopentane	2-Methyl butane		C5H12	72.14878	78-78-4
N,N-Dimethyl formamide			C3H7NO	73.094	68-12-2
N-Methyl acetamide			C3H7NO	73.094	79-16-3
N-Ethylethanamine	N,N-Diethylamine	Diethylamine	C4H11N	73.137	109-89-7
1-Butanamine			C4H11N	73.138	109-73-9
2-Methyl-1-propanamine	Isopropylmethylamine	Isobutylamine	C4H11N	73.138	78-81-9
Ethyl formate	Ethyl methanoate	Ethyl ester of formic acid	C3H6O2	74.079	109-94-4
Propionic acid			C3H6O2	74.079	79-09-4
Methyl acetate	Methyl ethanoate		C3H6O2	74.079	79-20-9
Methylpropyl ether			C4H10O	74.122	557-17-5
Methylisopropyl ether			C4H10O	74.122	598-53-8
Ethoxyethane	1,1'-Oxybisethane	Diethyl ether	C4H10O	74.122	60-29-7
1-Butanol	Propylmethanol	1-Butyl alcohol	C4H10O	74.122	71-36-3
2-Methyl-2-propanol	ter-Butyl alcohol	tert-Butanol	C4H10O	74.122	75-65-0
2-Butanol	Methyl-2-propanol	sec-Butanol	C4H10O	74.122	78-92-2
2-Methyl-1-propanol	2-Methylpropanol	Isobutanol	C4H10O	74.123	78-83-1
Nitroethane			C2H5NO2	75.067	79-24-3
1,2-Propylene glycol			C3H8O2	76.094	57-55-6
Carbon disulfide			CS2	76.141	75-15-0
Propyl mercaptan			C3H8S	76.161	107-03-9
Methylethyl sulfide			C3H8S	76.161	624-89-5
2-Propyl mercaptan			C3H8S	76.161	75-33-2
Germane			GeH4	76.642	7782-65-2
Arsine	AsH3		H3As	77.945	7784-42-1
Benzene	1,3,5-Cyclohexatriene	Cyclohex-1,3,5-triene	C6H6	78.11184	71-43-2
Dimethyl sulfoxide			C2H6OS	78.133	67-68-5
1-Chloropropane	Propyl chloride		C3H7Cl	78.541	540-54-5
2-Chloropropane			C3H7Cl	78.541	75-29-6
Pyridine	Azabenzene		C5H5N	79.101	110-86-1
Sulfur trioxide			O3S	80.063	7446-11-9
Methylchlorosilane			CH5ClSi	80.589	993-00-0
Hydrogen bromide	Bromane	HBr	BrH	80.912	10035-10-6
2-Methylfuran	Methylfuran	Sylvan	C5H6O	82.102	534-22-5
Cyclohexene			C6H10	82.144	110-83-8
3-Methylcyclopentene			C6H10	82.144	1120-62-3
1-Hexyne			C6H10	82.144	693-02-7
1-Methylcyclopentene			C6H10	82.144	693-89-0
2-Hexyne			C6H10	82.144	764-35-2
3-Hexyne			C6H10	82.144	928-49-4
Krypton			Kr	83.798	7439-90-9
1,1,1-Trifluoroethane	R-143a	Freon 143a	C2H3F3	84.041	420-46-2
1,1,2-Trifluoroethane	R-143		C2H3F3	84.041	430-66-0
Cyclopentanone			C5H8O	84.118	120-92-3
Thiophene	Thiofuran		C4H4S	84.14	110-02-1
Cyclohexane			C6H12	84.15948	110-82-7
1-Hexene	Hex-1-ene	Hexene	C6H12	84.159	592-41-6
Methylcyclopentane			C6H12	84.159	96-37-7
4-Methyl-1-pentene	4-Methylpent-1-ene		C6H12	84.15948	691-37-2
Dichloromethane	Methane dichloride	Methylene chloride	CH2Cl2	84.933	75-09-2
Vinyl acetate			C4H6O2	86.089	108-05-4

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Methacrylic acid			C4H6O2	86.089	79-41-4
Methyl acrylate			C4H6O2	86.089	96-33-3
2-Pentanone	Pentan-2-one	Methyl propyl ketone	C5H10O	86.132	107-87-9
Pentanal			C5H10O	86.132	110-62-3
3-Methyl-2-butanone	3-Methyl butanone	Methyl isopropyl ketone	C5H10O	86.132	563-80-4
3-Pentanone	Diethyl ketone	Metacetone	C5H10O	86.132	96-22-0
2-Methylpentane	1,1-Dimethylbutane	Isohexane	C6H14	86.17536	107-83-5
Hexane	n-Hexane		C6H14	86.17536	110-54-3
2,3-Dimethylbutane	Diisopropyl	Biisopropyl	C6H14	86.175	79-29-8
2,2-Dimethylbutane	Neohexane		C6H14	86.17536	75-83-2
3-Methylpentane			C6H14	86.177	96-14-0
Chloro(difluoro)methane	R-22	Chlorodifluoromethane	CHClF2	86.468	75-45-6
Tetrafluoromethane	R-14	carbon tetrafluoride	CF4	88.01	75-73-0
Butyric acid			C4H8O2	88.105	107-92-6
Propyl formate	Propyl methanoate	n-Propyl methanoate	C4H8O2	88.105	110-74-7
1,4-Dioxane			C4H8O2	88.105	123-91-1
Ethyl acetate	Acetic ester	Ethyl ethanoate	C4H8O2	88.105	141-78-6
Methyl propionate			C4H8O2	88.105	554-12-1
Isobutyric acid			C4H8O2	88.105	79-31-2
1,3-Dioxane			C4H8O2	88.106	505-22-6
3-Methyl-1-butanol			C5H12O	88.148	123-51-3
Methyl tert-butyl ether	MTBE	tert-Butyl methyl ether	C5H12O	88.148	1634-04-4
2-Pentanol			C5H12O	88.148	6032-29-7
Methylisobutyl ether			C5H12O	88.148	625-44-5
Ethylisopropyl ether			C5H12O	88.148	625-54-7
Methylbutyl ether			C5H12O	88.148	628-28-4
Ethylpropyl ether			C5H12O	88.148	628-32-0
1-Pentanol	n-Amyl alcohol	Pentanol	C5H12O	88.148	71-41-0
2-Methyl-2-butanol	tert-Amyl Alcohol	2-Methylbutan-2-ol	C5H12O	88.15	75-85-4
Tetrahydrothiophene			C4H8S	88.171	110-01-0
Oxalic acid			C2H2O4	90.035	144-62-7
1,3-Butanediol			C4H10O2	90.121	107-88-0
1,1-Dimethoxyethane			C4H10O2	90.121	534-15-6
1,2-Butanediol			C4H10O2	90.121	584-03-2
Butyl mercaptan			C4H10S	90.187	109-79-5
Methylisopropyl sulfide			C4H10S	90.187	1551-21-9
Diethyl sulfide	ethyl sulfide	1,1-Thiobisethane	C4H10S	90.187	352-93-2
Methylpropyl sulfide			C4H10S	90.187	3877-15-4
sec-Butyl mercaptan			C4H10S	90.187	513-53-1
Methylbenzene	Toluene		C7H8	92.13842	108-88-3
2-Chlorobutane	1-Methylpropyl chloride	sec-Butyl chloride	C4H9Cl	92.568	78-86-4
4-Methylpyridine	4-Methylazine	4-Picoline	C6H7N	93.128	108-89-4
3-Methylpyridine	3-Mepy	3-Picoline	C6H7N	93.128	108-99-6
2-Methylpyridine	2-Picoline		C6H7N	93.128	109-06-8
Aniline	1-Aminobenzene	Benzenamine	C6H7N	93.128	62-53-3
Phenol	Benzenol	Hydroxybenzene	C6H6O	94.111	108-95-2
Dimethyl disulfide			C2H6S2	94.199	624-92-0
Bromomethane			CH3Br	94.939	74-83-9
Dimethylzinc	Zinc dimethanide	DMZn	C2H6Zn	95.478	544-97-8
Fluorobenzene			C6H5F	96.102	462-06-6

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
1-Heptyne			C7H12	96.17	628-71-7
Cyclohexanone			C6H10O	98.143	108-94-1
Methylcyclohexane			C7H14	98.186	108-87-2
Ethylcyclopentane			C7H14	98.186	1640-89-7
1-Heptene	Heptylene	Hept-1-ene	C7H14	98.186	592-76-7
(1R,3R)-1,3-Dimethylcyclopentane	1,trans-3-Dimethylcyclopentane	trans-1,3-Dimethylcyclopentane	C7H14	98.188	1759-58-6
(1R,3S)-1,3-Dimethylcyclopentane	cis-1,3-Dimethylcyclopentane		C7H14	98.188	2532-58-3
1,2-Dichloroethane	1,2-DCE	1,2-Ethylene dichloride	C2H4Cl2	98.959	107-06-2
1,1-Dichloroethane	1,1-DCE		C2H4Cl2	98.959	75-34-3
Tetrafluoroethene	1,1,2,2-Tetrafluoroethylene	Perfluoroethylene	C2F4	100.016	116-14-3
Methyl methacrylate			C5H8O2	100.116	80-62-6
Methylisobutyl ketone			C6H12O	100.159	108-10-1
Cyclohexanol	Cyclohexyl Alcohol	1-Cyclohexanol	C6H12O	100.159	108-93-0
Ethylisopropyl ketone			C6H12O	100.159	565-69-5
3-Hexanone	Hexan-3-one		C6H12O	100.159	589-38-8
2-Hexanone	Butyl methyl ketone	Hexan-2-one	C6H12O	100.159	591-78-6
Hexanal			C6H12O	100.159	66-25-1
2,4-Dimethylpentane	Diisopropylmethane		C7H16	100.202	108-08-7
Heptane	n-Heptane		C7H16	100.202	142-82-5
2,2,3-Trimethylbutane	Triptane	Pentamethylethane	C7H16	100.202	464-06-2
3,3-Dimethylpentane			C7H16	100.202	562-49-2
2,3-Dimethylpentane	Ethylisopropylmethyl methane		C7H16	100.202	565-59-3
3-Methylhexane	2-Ethylpentane		C7H16	100.202	589-34-4
2,2-Dimethylpentane			C7H16	100.202	590-35-2
2-Methylhexane	Isoheptane		C7H16	100.202	591-76-4
3-Ethylpentane			C7H16	100.202	617-78-7
1-Chloro-1,1-difluoroethane	R-142b	1,1-Difluoro-1-chloroethane	C2H3ClF2	100.495	75-68-3
Di-isopropyl amine			C6H15N	101.19	108-18-9
N,N-Diethylethanamine	TEA	Triethylamine	C6H15N	101.19	121-44-8
Dipropyl amine			C6H15N	101.19	142-84-7
1,1,2,2-Tetrafluoroethane	R-134	a,w-Dihydroperfluoroethane	C2H2F4	102.032	359-35-3
1,1,1,2-Tetrafluoroethane	R-134a	Tetrafluoroethane	C2H2F4	102.032	811-97-2
Acetic anhydride			C4H6O3	102.089	108-24-7
Ethyl propionate			C5H10O2	102.132	105-37-3
Pentanoic acid			C5H10O2	102.132	109-52-4
Propyl acetate	Propyl ethanoate	n-Propyl ethanoate	C5H10O2	102.132	109-60-4
2-Methylbutanoic acid			C5H10O2	102.132	116-53-0
Methyl butyrate			C5H10O2	102.132	623-42-7
Di-isopropyl ether			C6H14O	102.175	108-20-3
1-Hexanol	Caproic alcohol	Hexyl alcohol	C6H14O	102.175	111-27-3
2-Hexanol			C6H14O	102.175	626-93-7
Methyl pentyl ether			C6H14O	102.175	628-80-8
Dichloro(fluoro)methane	R-21	Dichlorofluoromethane	CHCl2F	102.92	75-43-4
Benzonitrile			C7H5N	103.121	100-47-0
Malonic acid			C3H4O4	104.061	141-82-2
Silicon tetrafluoride		SiF4	F4Si	104.079	7783-61-1
1,2-Dimethoxypropane			C5H12O2	104.148	7778-85-0

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Styrene			C8H8	104.149	100-42-5
Pentyl mercaptan			C5H12S	104.214	110-66-7
2-Pentyl mercaptan			C5H12S	104.214	2084-19-7
Methylbutyl sulfide			C5H12S	104.214	628-29-5
Chloro(trifluoro)methane	R-13	Trifluorochloromethane	CClF3	104.459	75-72-9
Diethanol amine			C4H11NO2	105.136	111-42-2
Ethylbenzene			C8H10	106.165	100-41-4
p-Xylene	1,4-Xylene	1,4-Dimethylbenzene	C8H10	106.165	106-42-3
m-Xylene	1,3-Xylene	1,3-Dimethylbenzene	C8H10	106.165	108-38-3
o-Xylene	ortho-Xylene	1,2-Dimethylbenzene	C8H10	106.165	95-47-6
2,4-Dimethylpyridine	2,4-Lutidine		C7H9N	107.155	108-47-4
2,6-Dimethylpyridine	2,6-Lutidine		C7H9N	107.155	108-48-5
3,4-Dimethylpyridine	3,4-Lutidine		C7H9N	107.155	583-58-4
2,3-Dimethylpyridine	2,3-Lutidine		C7H9N	107.155	583-61-9
2,5-Dimethylpyridine	2,5-Lutidine	5-methyl-2-methylpyridine	C7H9N	107.155	589-93-5
3,5-Dimethylpyridine	3,5-Lutidine		C7H9N	107.155	591-22-0
Sulfur tetrafluoride	Tetrafluoro-λ4-sulfane	SF4	F4S	108.06	7783-60-0
Quinone			C6H4O2	108.095	106-51-4
Benzyl alcohol			C7H8O	108.138	100-51-6
Anisole			C7H8O	108.138	100-66-3
p-Cresol	4-Methylphenol	4-Hydroxytoluene	C7H8O	108.138	106-44-5
m-Cresol	3-Methylphenol	1-Hydroxy-3-methylbenzene	C7H8O	108.138	108-39-4
o-Cresol	2-Methylphenol	2-Hydroxytoluene	C7H8O	108.138	95-48-7
Bromoethane			C2H5Br	108.965	74-96-4
Benzenethiol			C6H6S	110.177	108-98-5
1-Octyne			C8H14	110.197	629-05-0
cis-1,2-Dimethylcyclohexane			C8H16	112.213	2207-01-4
1-Octene	Octene		C8H16	112.213	111-66-0
Ethylcyclohexane			C8H16	112.213	1678-91-7
1,1-Dimethylcyclohexane			C8H16	112.213	590-66-9
trans-1,2-Dimethylcyclohexane			C8H16	112.213	6876-23-9
Cyclooctane			C8H16	112.215	292-64-8
Chlorobenzene	Phenyl chloride	Benzyl chloride	C6H5Cl	112.557	108-90-7
1,2-Dichloropropane	Propylene dichloride		C3H6Cl2	112.986	78-87-5
1,1-Dichloropropane			C3H6Cl2	112.986	78-99-9
3-Heptanone			C7H14O	114.185	106-35-4
2-Heptanone			C7H14O	114.185	110-43-0
Heptanal			C7H14O	114.185	111-71-7
Di-isopropyl ketone			C7H14O	114.185	565-80-0
1-Methylcyclohexanol			C7H14O	114.185	590-67-0
trans-2-Methylcyclohexanol			C7H14O	114.185	7443-52-9
cis-2-Methylcyclohexanol			C7H14O	114.185	7443-70-1
Octane	n-Octane		C8H18	114.229	111-65-9
2,2,4-Trimethylpentane	Isooctane (trimethylpentane)		C8H18	114.229	540-84-1
2,3,3-Trimethylpentane			C8H18	114.229	560-21-4
3,3-Dimethylhexane			C8H18	114.229	563-16-6
2,2,3-Trimethylpentane			C8H18	114.229	564-02-3

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
3,4-Dimethylhexane			C8H18	114.229	583-48-2
2,3-Dimethylhexane			C8H18	114.229	584-94-1
2,4-Dimethylhexane			C8H18	114.229	589-43-5
4-Methylheptane			C8H18	114.229	589-53-7
2,2-Dimethylhexane			C8H18	114.229	590-73-8
2,5-Dimethylhexane	Diisobutane		C8H18	114.229	592-13-2
2-Methylheptane	i-Octane	Isooctane (methylheptane)	C8H18	114.229	592-27-8
2,2,3,3-Tetramethylbutane	Hexamethylethane		C8H18	114.229	594-82-1
3-Ethyl-2-methylpentane	2-Methyl-3-ethylpentane		C8H18	114.229	609-26-7
3-Ethylhexane			C8H18	114.229	619-99-8
Trimethylgallium	TMGa	Me3Ga	C3H9Ga	114.827	1445-79-0
Methyldichlorosilane			CH4Cl2Si	115.034	75-54-7
Ethyl butyrate			C6H12O2	116.158	105-54-4
Butyl acetate			C6H12O2	116.158	123-86-4
Hexanoic acid			C6H12O2	116.158	142-62-1
2-Ethyl butanoic acid			C6H12O2	116.158	88-09-5
1-Heptanol	1-Hydroxyheptane	Heptan-1-ol	C7H16O	116.201	111-70-6
2-Heptanol			C7H16O	116.201	543-49-7
Cyclohexyl mercaptan			C6H12S	116.224	1569-69-3
1,1-Dichloro-1-fluoroethane	R-141b	Dichlorofluoroethane	C2H3Cl2F	116.9496	1717-00-6
Succinic acid			C4H6O4	118.088	110-15-6
Ethyl (2S)-2-hydroxypropanoate	Ethyl lactate		C5H10O3	118.133	687-47-8
alpha-Methyl styrene			C9H10	118.176	98-83-9
Hexyl mercaptan			C6H14S	118.24	111-31-9
Phenyl isocyanate			C7H5NO	119.121	103-71-9
Chloroform	R-20	Trichloromethane	CHCl3	119.378	67-66-3
1,1,1,2,2-Pentafluoroethane	R-125	Pentafluoroethane	C2HF5	120.0214	354-33-6
Propylbenzene	n-Propylbenzene		C9H12	120.192	103-65-1
1,2,3-Trimethylbenzene	Trimethylbenzene		C9H12	120.192	526-73-8
1,2,4-Trimethylbenzene	pseudocumene	Pseudocumene	C9H12	120.192	95-63-6
1-methylethyl benzene	2-Phenylpropane	Cumene	C9H12	120.192	98-82-8
Mesitylene	Sym-Trimethylbenzene	1,3,5-Trimethylbenzene	C9H12	120.194	108-67-8
1-Ethyl-4-methylbenzene	1,4-Methylethylbenzene	4-Ethyltoluene	C9H12	120.194	622-96-8
Dichloro(difluoro)methane	R-12	dichlorodifluoromethane	CCl2F2	120.913	75-71-8
Benzamide			C7H7NO	121.137	55-21-0
Benzoic acid			C7H6O2	122.121	65-85-0
2,4-Dimethylphenol		2,4-Xylenol	C8H10O	122.167	105-67-9
3,5-Dimethylphenol	1,3,5-Xylenol	3,5-Xylenol	C8H10O	122.167	108-68-9
4-Ethylphenol	p-Ethylphenol	4-Hydroxyethylbenzene	C8H10O	122.167	123-07-9
2,3-Dimethylphenol	1-Hydroxy-2,3-dimethylbenzene	Xylenol	C8H10O	122.167	526-75-0
2,6-Dimethylphenol	1-Hydroxy-2,6-dimethylbenzene	2,6-Xylenol	C8H10O	122.167	576-26-1
3-Ethylphenol	3-Ethylbenzanol	1-Hydroxy-3-ethylbenzene	C8H10O	122.167	620-17-7
2-Ethylphenol	2-Ethylphenyl alcohol	1-Hydroxy-2-ethylbenzene	C8H10O	122.167	90-00-6
3,4-Dimethylphenol	1,2-Dimethyl-4-hydroxybenzene	3,4-Xylenol	C8H10O	122.167	95-65-8
2,5-Dimethylphenol	2,5-Dimethylbenzanol	1,2,5-Xylenol	C8H10O	122.167	95-87-4
Propenylcyclohexene			C9H14	122.207	13511-13-2
Benzyl mercaptan			C7H8S	124.203	100-53-8

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
1-Nonyne			C9H16	124.223	3452-09-3
1-Nonene	Non-1-ene		C9H18	126.239	124-11-8
Hydrogen iodide	Iodane	Hydroiodic acid	HI	127.912	10034-85-2
Naphthalene			C10H8	128.171	91-20-3
3-Octanone			C8H16O	128.212	106-68-3
2-Octanone			C8H16O	128.212	111-13-7
Octanal			C8H16O	128.212	124-13-0
Nonane	n-Nonane		C9H20	128.2551	111-84-2
2,2,4,4-Tetramethylpentane	Di-tert-butylmethane		C9H20	128.2551	1070-87-7
2,2,3,4-Tetramethylpentane			C9H20	128.2551	1186-53-4
2,3,3,4-Tetramethylpentane			C9H20	128.2551	16747-38-9
2-Methyloctane	Isononane		C9H20	128.2551	3221-61-2
Heptanoic acid			C7H14O2	130.185	111-14-8
1-Octanol	Capryl alcohol	1-Octyl alcohol	C8H18O	130.228	111-87-5
2-Octanol			C8H18O	130.228	123-96-6
Dibutyl ether			C8H18O	130.228	142-96-1
Ethylhexyl ether			C8H18O	130.228	5756-43-4
Bromo(difluoro)methane	bromodifluoromethane	Difluorobromomethane	CHBrF2	130.92	1511-62-2
Xenon			Xe	131.293	7440-63-3
1,2,3,4-Tetrahydronaphthalene			C10H12	132.202	119-64-2
Heptyl mercaptan			C7H16S	132.267	1639-09-4
1,1,2-Trichloroethane			C2H3Cl3	133.404	79-00-5
1,1,2,2,3-pentafluoropropane	R-245ca		C3H3F5	134.0479	679-86-7
1,1,1,3,3-pentafluoropropane	R-245fa		C3H3F5	134.0479	460-73-1
Butylbenzene	n-Butylbenzene	Butyl benzene	C10H14	134.218	104-51-8
1,4-Diethylbenzene	para-Diethylbenzene	p-Diethylbenzene	C10H14	134.221	105-05-5
1,2,4,5-Tetramethylbenzene	Durene		C10H14	134.221	95-93-2
Trichlorosilane	SiHCl3		HCl3Si	135.452	10025-78-2
Methyl benzoate			C8H8O2	136.148	93-58-3
Benzyl ethyl ether			C9H12O	136.191	539-30-0
2-Chloro-1,1,1,2-tetrafluoroethane	R-124	1,1,1,2-tetrafluoro-2-chloroethane	C2HClF4	136.475	2837-89-0
1-Chloro-1,1,2,2-tetrafluoroethane	R-124a	1,1,2,2-Tetrafluorochloroethane	C2HClF4	136.476	354-25-6
Trichloro(fluoro)methane	R-11	Trichlorofluoromethane	CCl3F	137.368	75-69-4
Hexafluoroethane	R-116	Perfluoroethane	C2F6	138.0118	76-16-4
1-Decyne			C10H18	138.25	764-93-2
cis-Decahydronaphthalene	(4as,8as)-Decahydronaphthalene	cis-Decalin	C10H18	138.253	493-01-6
trans-Decahydronaphthalene	(4ar,8ar)-Decahydronaphthalene	trans-Decalin	C10H18	138.253	493-02-7
1-Decene	Dec-1-ene	Decylene	C10H20	140.266	872-05-9
2,2,5-Trimethylheptane			C10H22	142	20291-95-6
1-Methylnaphthalene	Methyl-1-naphthalene		C11H10	142.2	90-12-0
Nonanal			C9H18O	142.239	124-19-6
2,2,3,3-Tetramethylhexane			C10H22	142.2817	13475-81-5
3,3,5-Trimethylheptane			C10H22	142.2817	7154-80-5
Decane	n-decane		C10H22	142.2817	124-18-5

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Dihydrogen hexafluorosilicate	Hexafluorosilicic acid	SiF ₆	H ₂ F ₆ Si	144.09	16961-83-4
Octanoic acid			C ₈ H ₁₆ O ₂	144.211	124-07-2
2-Ethyl hexanoic acid			C ₈ H ₁₆ O ₂	144.211	149-57-5
1-Nonanol	Nonylol	Nonyl alcohol	C ₉ H ₂₀ O	144.255	143-08-8
2-Nonanol			C ₉ H ₂₀ O	144.255	628-99-9
Sulfur hexafluoride	Hexafluoro-λ ₆ -sulfane	SF ₆	F ₆ S	146.0554	2551-62-4
Octyl mercaptan			C ₈ H ₁₈ S	146.294	111-88-6
p-Dichlorobenzene			C ₆ H ₄ Cl ₂	147.002	106-46-7
m-Dichlorobenzene			C ₆ H ₄ Cl ₂	147.002	541-73-1
o-Dichlorobenzene			C ₆ H ₄ Cl ₂	147.002	95-50-1
1,1,1,3,3-pentafluorobutane	R-365mfc		C ₄ H ₅ F ₅	148.0745	406-58-6
Phthalic anhydride			C ₈ H ₄ O ₃	148.116	85-44-9
Bromo(trifluoro)methane	R-13B1	CFC 13B1	CBrF ₃	148.91	75-63-8
Ethyl benzoate			C ₉ H ₁₀ O ₂	150.175	93-89-0
1,1,1,2,3,3-hexafluoropropane	R-236ea		C ₃ H ₂ F ₆	152.0384	431-63-0
1,1,1,3,3,3-hexafluoropropane	R-236fa		C ₃ H ₂ F ₆	152.0384	690-39-1
2,2-Dichloro-1,1,1-trifluoroethane	R-123	Dichlorotrifluoromethyl methane	C ₂ HCl ₂ F ₃	152.931	306-83-2
Phosphoric trichloride	Phosphoryl chloride	Phosphorus oxychloride	Cl ₃ OP	153.332	10025-87-3
Tetrachloromethane	R-10	Carbon tetrachloride	CCl ₄	153.823	56-23-5
Biphenyl	1,1'-Biphenyl	Phenylbenzene	C ₁₂ H ₁₀	154.208	92-52-4
1-Chloro-1,1,2,2,2-pentafluoroethane	R-115	1-Chloropentafluoroethane	C ₂ ClF ₅	154.4664	76-15-3
Di-1,3-cyclopentadien-1-yl magnesium	Bis(cyclopentadienyl) magnesium	Cp ₂ Mg	C ₁₀ H ₁₀ Mg	154.491	1284-72-6
Decanal			C ₁₀ H ₂₀ O	156.265	112-31-2
Undecane			C ₁₁ H ₂₄	156.308	1120-21-4
Triethylgallium	TeGa		C ₆ H ₁₅ Ga	156.906	1115-99-7
Bromobenzene			C ₆ H ₅ Br	157.008	108-86-1
Nonanoic acid			C ₉ H ₁₈ O ₂	158.238	112-05-0
2-Methyloctanoic acid			C ₉ H ₁₈ O ₂	158.238	3004-93-1
1-Decanol	1-Hydroxydecane	Decan-1-ol	C ₁₀ H ₂₂ O	158.281	112-30-1
Bromine			Br ₂	159.808	7726-95-6
Trimethylindium	TMI		C ₃ H ₉ In	159.922	3385-78-2
Nonyl mercaptan			C ₉ H ₂₀ S	160.32	1455-21-6
Vinyl trichlorosilane			C ₂ H ₃ Cl ₃ Si	161.49	75-94-5
1,3,5-Triethylbenzene			C ₁₂ H ₁₈	162.276	102-25-0
Ethyltrichlorosilane			C ₂ H ₅ Cl ₃ Si	163.506	115-21-9
Bromo(chloro)difluoro methane	R-12B1	Chlorobromodifluoro methane	CBrClF ₂	163.365	353-59-3
Terephthalic acid			C ₈ H ₆ O ₄	166.131	100-21-0
1,2,3,4,5-Pentafluorobenzene	Pentafluorobenzene		C ₆ HF ₅	168.066	363-72-4
Tetrachlorosilane	Silicon tetrachloride	SiCl ₄	Cl ₄ Si	169.896	10026-04-7
1,1,1,2,3,3,3-heptafluoropropane	R-227ea		C ₃ HF ₇	170.0289	431-89-0
Diphenyl ether			C ₁₂ H ₁₀ O	170.207	101-84-8
Dodecane	n-Dodecane		C ₁₂ H ₂₆	170.3348	112-40-3

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
1,1-Dichloro-1,2,2,2-tetrafluoroethane	R-114a	Dichlorotetrafluoroethane	C2Cl2F4	170.921	374-07-2
1,2-Dichloro-1,1,2,2-tetrafluoroethane	R-114	Cryofluorane	C2Cl2F4	170.921	76-14-2
Decanoic acid			C10H20O2	172.265	334-48-5
1-Undecanol			C11H24O	172.308	112-42-5
Dibromomethane			CH2Br2	173.835	74-95-3
Decyl mercaptan			C10H22S	174.347	143-10-2
Phenanthrene			C14H10	178.229	85-01-8
Anthracene			C14H10	178.233	120-12-7
Benzophenone			C13H10O	182.218	119-61-9
Tridecane		n-Tridecane	C13H28	184.361	629-50-5
Hexafluorobenzene	Perfluorobenzene		C6F6	186.056	392-56-3
1-Dodecanol	1-Hydroxydodecane	Dodecan-1-ol	C12H26O	186.338	112-53-8
1,1,2-Trichloro-1,2,2-trifluoroethane	R-113	Daiflon 113	C2Cl3F3	187.375	76-13-1
1,2-Dibromoethane	Ethylene dibromide		C2H4Br2	187.861	106-93-4
1,1-Dibromoethane			C2H4Br2	187.861	557-91-5
Octafluoropropane	R-218	2,2-Dihydroperfluoropropane	C3F8	188.0193	76-19-7
Dimethyl terephthalate			C10H10O4	194.184	120-61-6
Dimethyl phthalate			C10H10O4	194.184	131-11-3
trifluoroiodomethane	trifluoroiodomethane		CF3I	195.9104	2314-97-8
Tetradecane	n-Tetradecane		C14H30	198.388	629-59-4
Octafluorocyclobutane			C4F8	200.04	115-25-3
1-Chloro-2,3,4,5,6-pentafluorobenzene	Perfluorochlorobenzene	Chloropentafluorobenzene	C6ClF5	202.511	344-07-0
Tetraethyl orthosilicate	TEOS	Silicic acid	C8H20O4Si	208.327	78-10-4
Dibromo(difluoro)methane	R-12B2	Dibromodifluoromethane	CBr2F2	209.816	75-61-6
Pentadecane	n-Pentadecane		C15H32	212.415	629-62-9
1,3,5-Trinitrobenzene			C6H3N3O6	213.105	99-35-4
Tetrachlorogermane	Germanium chloride	GeCl4	Cl4Ge	214.421	10038-98-9
Radon			Rn	222.018	10043-92-2
Tetrakis(dimethylamido) titanium(IV)	TDMAT	[(CH3)2N]4Ti	C8H24N4Ti	224.17	3275-24-9
Hexadecane	n-Hexadecane	Cetane	C16H34	226.441	544-76-3
2,4,6-Trinitrotoluene			C7H5N3O6	227.131	118-96-7
o-Terphenyl			C18H14	230.304	84-15-1
perfluorobutane	decafluorobutane		C4F10	238.03	355-25-9
Heptadecane			C17H36	240.468	629-78-7
Octadecane	n-Octadecane		C18H38	254.494	593-45-3
Nonadecane			C19H40	268.521	629-92-5
Eicosane			C20H42	282.547	112-95-8
Dodecafluoropentane	Perfluoro-n-pentane	Perflenapent	C5F12	288.03	678-26-2
Tungsten hexafluoride	Tungsten(VI) fluoride	WF6	F6W	297.83	7783-82-6
Tetrabromomethane	Carbon tetrabromide		CBr4	331.627	558-13-4
Tetrakis(diethylamido) titanium(IV)	TDEAT	[(C2H5)2N]4Ti	C16H40N4Ti	336.383	4419-47-0
Tetradecafluorohexane	n-Tetradecafluorohexane	Perfluorohexane	C6F14	338.044	355-42-0
Tetrabromosilane	Silicon tetrabromide		Br4Si	347.702	7789-66-4
Hexadecafluoroheptane	Perfluoroheptane		C7F16	388.052	335-57-9
Octadecafluorooctane	Perfluorooctane		C8F18	438.059	307-34-6

Preferred Name	Alternate Name 1	Alternate Name 2	Formula	Weight	CAS#
Icosafluorononane	Perfluorononane	Eicosafluorononane	C9F20	488.067	375-96-2
Docosafluorodecane	Perfluorodecane		C10F22	538.075	307-45-9
Titanium tetrachloride	Tetrachlorotitanium	Tickle-4	Cl4Ti	189.679	7550-45-0

Gas Table Properties

The BGA244 binary gas analyzer determines gas purity and mixture compositions by measuring the temperature and the speed of sound in the gas. The speed of sound in an ideal gas is given by the follow equation.

$$W_o = \sqrt{\frac{\gamma_0 R T}{M}} \quad \text{Eq. 1}$$

Where W_o is the speed of sound in the ideal gas (low pressure) limit, γ_0 the ratio of heat capacities, M the molar mass, T the absolute temperature, and R the ideal gas constant. For argon, $\gamma_0 = C_p/C_v = C_p/(C_p - R) = 5/3$ and $M = 0.039948 \text{ kg/mol}$, and so the ideal speed of sound at 20°C (293.15 K) is about 318.89 m/s.

For the BGA244 to determine the purity of an *ideal* gas it needs to know the isobaric heat capacity (C_p) and M . The measurement of *real* gases will be improved if the other thermodynamic and transport properties, and the gas pressure, are also known.

The Gas Table contains information on about 500 gases listed by common names, formulas, and registration numbers. In addition to the molar mass, the Gas Table provides parameters to estimate each of the following (all of which can modify the speed of sound in the gas): The temperature dependence of the heat capacity, the second and third virial coefficients, the viscosity and thermal conductivity as functions of temperature, the translational/vibrational relaxation of the heat capacity, and the saturation vapor pressure.

In addition, the Gas Table provides information on the critical parameters (P_c, V_c, T_c, Z_c), the Pitzer acentric factor, the electric dipole moment, and chemical family allowing the computation of cross virial coefficients by the BGA244 firmware.

Finally, to allow for gas specific corrections caused by uncertainties in all the other parameters, there are three offset parameters to repair a fixed offset or offsets which change linearly with pressure or temperature.

As shown in *Table 32* there are 46 data fields (columns) associated with each gas. The table shows the column number, names, units, value for nitrogen, and a brief description for each of the column entries. A more complete description of each column follows.

TABLE 32: GAS TABLE PROPERTIES

Column	Name	Units	N2 Value	Description
1	CAS#	Alpha-numeric	7727-37-9	Chemical Abstract Service number
2	Preferred name	Alpha-numeric	Nitrogen	Preferred name. Often the systematic name if not arcane.
3	Alt 1 name	Alpha-numeric		Alternate name or ASKRAE designation. Example: R-134A
4	Alt 2 name	Alpha-numeric		Alternate name or non-Hill formula. Example: SF6 rather than F6S
5	Formula	Hill form	N2	Chemical formula, carbon(s) first, hydrogen(s) next, alpha thereafter
6	Molar mass	g/mol	28.01348	Molecular weight in grams per mole
7	a0(Cp)	dimensionless	3.4379	Scaled coefficient of polynomial for isobaric heat capacity in units of R
8	a1(Cp)×1E+03	1/K	0.7884	Scaled coefficient of polynomial for isobaric heat capacity in units of R
9	a2(Cp)×1E+05	1/K^2	-0.3505	Scaled coefficient of polynomial for isobaric heat capacity in units of R
10	a3(Cp)×1e+08	1/K^3	0.6090	Scaled coefficient of polynomial for isobaric heat capacity in units of R
11	a4(Cp)×1e+11	1/K^4	-0.2508	Scaled coefficient of polynomial for isobaric heat capacity in units of R
12	Cp(Tnorm)/R	dimensionless	3.5026997	Isobaric heat capacity (in units of R) calculated from polynomial at T=T(norm)
13	Wo(Tnorm, P=0)	m/s	348.963	Ideal gas speed of sound at T_norm. Wo=sqrt(γRT/M)
14	Pc	bar	33.958	Critical pressure (bar)
15	Vc	cc/mol	89.414	Critical volume (cc/mol)
16	Tc	K	126.192	Critical temperature (K)
17	Zc	dimensionless	0.289387	Critical compressibility
18	Acentric factor	dimensionless	0.0	Pitzer acentric factor
19	Dipole moment	D (debye)	0.00	Electric dipole moment
20	Family	Integer	1	Integer (1-6) to describe chemical "family" per table, below. (For CSP methods)
21	a_Tsono	dimensionless	0.00	Tsonopoulos correlation a-parameter for 2nd virial (see PG&L Pg 4.15)
22	b_Tsono	dimensionless	0.00	Tsonopoulos correlation b-parameter for 2nd virial (see PG&L Pg 4.15)
23	av	cc/mol	144.14	Zuckerwar parameter for exponential form of 2nd virial. (See Eq 4.30, pg 94)
24	bv	cc/mol	101.08544	Zuckerwar parameter for exponential form of 2nd virial. (See Eq 4.30, pg 94)
25	cv	K	115.778	Zuckerwar parameter for exponential form of 2nd virial. (See Eq 4.30, pg 94)
26	dv	[cc/mol]^2	16689.36	Zuckerwar parameter for exponential form of 3rd virial. (See Eq 4.35, pg 98)
27	ev	[cc/mol]^2	49.618	Zuckerwar parameter for exponential form of 3rd virial. (See Eq 4.35, pg 98)
28	fv	K	552.72650	Zuckerwar parameter for exponential form of 3rd virial. (See Eq 4.35, pg 98)
29	gv	1/K	0.015	Zuckerwar parameter for exponential form of 3rd virial. (See Eq 4.35, pg 98)
30	C_asymptote	[cc/mol]^2	1188.1	Zuckerwar parameter for exponential form of 3rd virial. (See Eq 4.35, pg 98)
31	a0(η)	μPa-s	1.66702397	Polynomial coefficient for viscosity, η(T)
32	a1(η)×100	μPa-s/K	6.22979799	Polynomial coefficient for viscosity, η(T)
33	a2(η)×10,000	μPa-s/K^2	-0.28410365	Polynomial coefficient for viscosity, η(T)
34	a0(κ)	mW/m.K	1.904	Polynomial coefficient for thermal conductivity κ(T)
35	a1(κ)×100	mW/m.K^2	8.706	Polynomial coefficient for thermal conductivity κ(T)
36	a2(κ)×10,000	mW/m.K^3	-0.267	Polynomial coefficient for thermal conductivity κ(T)
37	Vib Temp	K	3352.00	Characteristic temperature for excitation of vibrational mode
38	a1_r	dimensionless	-3.65	Coefficient for relaxation-time x pressure product (us-atm)
39	a2_r	K^1/3	71.63	Coefficient for relaxation-time x pressure product (us-atm)
40	a3_r	K	0.00	Coefficient for relaxation-time x pressure product (us-atm)
41	A_vp	Result in Pa	0.00	Coefficient for the Antoine Equation for vapor pressure. (PG&L Eq 7-3.1, pg 7.4)
42	B_vp	Result in Pa	0.00000	Coefficient for the Antoine Equation for vapor pressure. (PG&L Eq 7-3.1, pg 7.4)
43	C_vp	Result in Pa	0.00000	Coefficient for the Antoine Equation for vapor pressure. (PG&L Eq 7-3.1, pg 7.4)
44	W_offset	ppm	0.00	Empirical offset for the computed speed of sound
45	W_Pslope	ppm/psi	0.00	Empirical pressure slope for the computed speed of sound
46	W_Tslope	ppm/T	0.00	Empirical temperature slope for the computed speed of sound

CAS Registry number (Column 1)

The CAS registry number is a unique identifier assigned by the Chemical Abstract Service (CAS) to every chemical substance described in reviewed scientific literature. The CAS numbers are used by chemical suppliers, governmental regulators, safety data sheets (MSDS), chemical and thermodynamic databases. Using the CAS number (instead of a chemical name) will assure that correct gas has been selected. The CAS Registry Number is a Registered Trademark of the American Chemical Society.

Preferred name (Column 2)

Except for a few arcane cases (for example, “carbon dioxide” is used instead of “methanedione” and “ozone” is used instead of “2-trioxiden-2-ium-1-ide”) the preferred name is usually the IUPAC name (International Union of Pure and Applied Chemistry).

First and second alternate names (Column 3 & 4)

Many compounds are commonly called by more than one name. For example, difluoromethane (CH_2F_2) is also called methylene fluoride and is an ASHRAE registered refrigerant designated as R-32. The alternate name columns accommodate these other names and industry designations.

Formula (Columns 5)

Formulas for compounds may be written in several ways. The BGA244 uses a simple, non-structural form called Hill notation. The rules are simple: For molecules containing carbon, the carbons are listed first, followed by the hydrogens, followed by everything else in alphabetical order of the element symbols. If the compound contains no carbons everything is listed in alphabetical order of the element symbols, including hydrogen. This system yields sensible results in most cases, for example, methane is CH_4 . There are a few exceptions however: SF_6 is written F6S in Hill notation. In such cases the formula SF_6 is included as one of the alternate names so that a user looking for SF_6 will indeed find sulfur hexafluoride.

Source of data

Many references (listed at the end of this section) were used to compile (or calculate) the data used in the Gas Table. Data from those references was critically evaluated, primarily by comparison between the various references but also taking into consideration the sources used by those references.

Not all references had all of the data required for the Gas Table. Different references also used different correlation functions and applied those functions to different temperature ranges. As an example, consider the parameterization of the isobaric heat capacity vs. temperature. The BGA244 uses a 4th order polynomial for this task. This polynomial is well behaved and offers more than enough flexibility to accurately follow the heat capacity over the BGA244's operating temperature range. This is the same polynomial as used in one of the references, however, the accuracy of the coefficients presented in that reference were found to be less accurate than could be found by fitting a polynomial over the operating temperature range to data extracted from the correlation functions of other references, or to data extracted from NIST's REFPROP.

The same methodology was applied to other gas parameters and parameters for correlation functions to estimate the second and third virial coefficients, the viscosity and thermal conductivity as functions of temperature, the translational/vibrational relaxation of the heat capacity, and the saturation vapor pressure. The distillation of all of these estimates and the parameters for new correlation functions are presented in the Gas Table.

Corresponding States Principle (CSP)

CSP methods allow the estimation of fluid properties which depend on intermolecular forces by correlation functions scaled with the critical constants of the fluid. For example, the departure from ideal gas compressibility can be described with a virial equation of state (EOS). The first two coefficients of that EOS, called the 2nd and 3rd virial coefficients, can be measured directly or estimated using CSP methods.

CSP methods were used to estimate the gases' 2nd and 3rd virial coefficients, thermal conductivity and viscosity. These results would be used in the Gas Table if experimental results (or correlation functions derived from those results) were not available.

Molar mass (Column 6)

The molar mass is in units of grams per mole. Note that these are not SI units, for which the molar mass is in kg/mole. An accurate molar mass is critically important as a 200 ppm error in the molar mass results in a 100 ppm error in the speed of sound. The molar mass is often rounded to two or three digits in published tables which is insufficient for the lighter gases.

Isobaric heat capacity, C_p/R (Columns 7-12)

The isobaric heat capacity in the ideal gas limit, in units of the ideal gas constant, is computed from the scaled coefficients to the 4th order polynomial given in Columns 7-11. (The coefficients are scaled to keep their magnitude on order 1.) C_p/R is computed from Eq. 2 (where T is the temperature in K):

$$\frac{C_p}{R} = a_0 + \frac{a_1 \cdot T}{1000} + \frac{a_2 \cdot T^2}{100,000} + \frac{a_3 \cdot T^3}{100,000,000} + \frac{a_4 \cdot T^4}{100,000,000,000} \quad \text{Eq. 2}$$

And the ratio of heat capacities, γ_0 , is computed from Eq. 3:

$$\gamma_0 = \frac{C_p}{C_v} = \frac{C_p}{C_p - R} = \frac{C_p/R}{C_p/R - 1} \quad \text{Eq. 3}$$

Column 12 of the Gas Table holds the value of $C_p(T_{norm})/R$ where $T_{norm} = 293.15 \text{ K}$. This value serves as a computational check point, allowing the BGA244 firmware to validate the table contents and its method for computing heat capacity.

Speed of sound at normal temperature (Column 13)

The speed of sound at normal temperature (293.15 K), and in the zero pressure limit, is given in Column 13 in units of meters/second. This value can serve as a computational check of other table values and firmware algorithms, and is useful for sorting the Gas Table by the speed of sound. Such a sort provides prospective on the ability of the BGA244 to measure the composition of a gas mixture: It is difficult to measure the composition of a mixture of two gases which have nearly the same speed of sound.

Tsonopoulos correlation

The Tsonopoulos correlation is a CSP method which is used to estimate the 2nd virial. Absent experimental measurements, this method was used to estimate the 2nd and 3rd virial coefficients of pure gases in the Gas Table.

The method is also used by the BGA244 firmware to estimate the 2nd cross-virials of a gas mixture. This computation is done in firmware (instead of listed in the Gas Table) due to the large number (125,000) possible mixtures for 500 different gases.

To compute the 2nd cross-virial via the Tsonopoulos correlation the firmware will require the critical constants (P_c , V_c , T_c , and Z_c), acentric factor, dipole moment, and categorization into one of six chemical families. The Gas Table details for those values are described below.

Critical constants (Columns 13-17)

The critical constants for each substance, P_c , V_c , T_c , and Z_c , are listed in Columns 13-17. The critical pressure is in units of bar. (1 bar = 100,000 Pa). The critical volume is in units of cc/mole. The critical temperature is in units of kelvin and the critical compressibility is dimensionless.

Acentric Factor (Column 18)

The Pitzer acentric factor is used to characterize the departure of thermodynamic properties of a substance from those predicted by two-parameter corresponding states principles (CSP). As CSP applies strictly only to spherical molecules, the acentric factor is considered to be a measure of the acentricity of the molecule. The acentric factor is dimensionless, and is determined from the departure of vapor pressures from those predicted for spherical molecules (such as noble gases.)

Dipole moment (Column 19)

The electric dipole moment of the substance in the gas phase is given in Column 19 in units of D (debye). Debye are not SI units, but provide a dipole value which is on order 1 for molecules. $1 D \approx 3.33564 \times 10^{-30} C \cdot m = 0.208194 e \cdot \text{\AA}$.

Chemical Family (Column 19)

The Tsonopoulos correlation uses different correlation terms depending on the nature of the species, which are categorized into one of six chemical “families” as shown below.

Family	Description
1	Simple, Normal gases
2	Keytones, Aldehydes, Alkyl Nitriles, Ethers, Carboxylic Acid Esters
3	1-Alkanols except methanol
4	Methanol
5	Water

Tsonopoulos a,b parameters (Column 20-21)

These parameters, which are detailed in Table 4-5 on page 4.15 in PG&L are used in the Tsonopoulos correlation to compute the 2nd virial coefficients and 2nd cross-virial coefficients via a CSP method. (Note there is an error in the table in PG&L for the exponent for the a-parameter for Alkyl Halides, Mercaptans, Sulfides, and Disulfides: -2.188×10^{-4} should read -2.188×10^{-11} per original published papers.)

Second virial coefficient parameters (Column 23-25)

The second virial coefficient, $B(T)$, is parameterized as a function of temperature as shown in Eq. 4. (See Zuckerwar 4.30-4.32b):

$$B(T) = a_v - b_v e^{(c_v/T)} \quad \text{Eq. 4}$$

The coefficients, a_v , b_v and c_v (with units of cc/mol, cc/mol and K) are found in columns 23-25 of the Gas Table. Several sources were used for these parameters including: Fitting the above equation to values for the 2nd virial computed over the operating temperature range from REFPROP, or by using coefficients directly from Zuckerwar or from Kaye & Laby, or (lastly) fitting the above equation to values for the 2nd virial computed from the Tsonopoulos correlation.

The 2nd virial and its first and second temperature derivatives alter the speed of sound in a manner which increases linearly with gas density. For many gases the virial correction for the speed of sound will be a small correction. For example, in methane at 1.6 atm and 300K, virial effects reduce the speed of sound by only 0.12%. However, even this small factor may be important in determining relative concentrations of species which are close in molecular weight.

Third virial coefficient parameters (Column 26-30)

The 3rd virial as a function of temperature is represented Eq. 5 (See Zuckerwar's Eq 4.35).

$$C(T) = \left[d_v - e_v \cdot \exp\left(\frac{f_v}{T}\right) \right] \cdot \exp(-g_v \cdot T) + C_{asymptote} \quad \text{Eq. 5}$$

The parameters of this equation, d_v , e_v , f_v , g_v and $C_{asymptote}$ are given in columns 26-30 of the Gas Table. The units for these parameters are cc/mol^2 , cc/mol^2 , K , K^{-1} and cc/mol^2 .

Coefficients for the 3rd virial were sourced from Zuckerwar, computed from a CSP method by Orbey, or extracted from NIST's REFPROP program. All three methods provided reasonable agreement for common gases, but very disparate results for less common gases. There is a scarcity of reliable experimental results for the 3rd virial, and the method of Orbey does not address polar compounds, hence coefficients for the 3rd virial are missing for many substances.

The 3rd virial and its first and second temperature derivatives alter the speed of sound in a manner which increases quadratically with gas density. The impact is vanishingly small below a few atmospheres for most gases, but impacts the speed of sound by about 0.3% in SF₆ at ten atmospheres at 25C.

Finally, we suspect that literature reports of 2nd virial measurements have often conjoined 3rd virial effects into their fits. (We have seen cases where explicitly adding in the 3rd virial overcorrects.)

Viscosity (Columns 31-33)

The viscous boundary layer (typically a few 0.001") impacts the cavity resonance frequency by a fraction of a percent. The effect can be compensated for if we know the gas viscosity. The viscosity of a gas is approximated by the polynomial as shown in Eq. 6.

$$\eta = A_{\eta} + B_{\eta} \cdot T + C_{\eta} \cdot T^2 \quad \text{Eq. 6}$$

Here, T is the absolute temperature in K. Coefficients for the polynomial, A_{η} , B_{η} and C_{η} , are tabulated for each gas and scaled so that the result has units of $\mu\text{Pa}\cdot\text{s}$.

Polynomial coefficients were determined by least squares fitting to viscosity data points over the operating temperature range. Data points were extracted from NIST's REFPROP, computed from correlation functions provided by Perry or PG&L, or computed from the CSP method of Lucas (see PG&L 9.9).

Thermal conductivity (Columns 34-36)

The thermal boundary layer (typically a few 0.001") impacts the cavity resonance frequency by a fraction of a percent. The effect can be compensated for if we know the thermal conductivity of the gas. The thermal conductivity of a gas is approximated by the polynomial in Eq. 7.

$$\eta = A_{\kappa} + B_{\kappa} \cdot T + C_{\kappa} \cdot T^2 \quad \text{Eq. 7}$$

Here, T is the absolute temperature in K. Coefficients for the polynomial, A_{κ} , B_{κ} and C_{κ} , are tabulated for each gas and scaled so that the result has units of $\text{mW}/\text{m}\cdot\text{K}$ (1000x larger than SI value with units of $\text{W}/\text{m}\cdot\text{K}$).

Polynomial coefficients were determined by least squares fitting to thermal conductivity data points over the operating temperature range. Data points were extracted from NIST's REFPROP, computed from correlation functions provided by Perry or PPL&G, or computed from the CSP method of Chung (1984, 1988) following PG&L Eq. 10-3.14

Heat capacity relaxation parameters (Columns 37-40)

There are a few gases for which vibrational energy levels (which contribute to the heat capacity) do not have time to equilibrate with the translational energy during the period of an acoustic cycle. The relaxation correction may be thought of as a correction to the heat capacity of the gas: If the relaxation time, τ_d , is longer than the acoustic cycle then vibrational degrees of freedom are partially frozen out, the heat capacity is lower, and so γ and the speed of sound will be higher. Without a correction for this effect, the computed speeds of sound (as is reported by NIST's REFPROP program, for example, which computes the speed of sound at zero frequency) are lower than the actual speed of sound.

We follow Zuckerwar's approach for the relaxation correction to the square of the speed of sound W^2 as shown in Eq. 8:

$$W^2 = W_\theta^2 \left[1 + \frac{\varepsilon}{1 - \varepsilon} \cdot \frac{(\omega\tau_d)^2}{1 + (\omega\tau_d)^2} \right] \quad \text{Eq. 8}$$

Here W_θ^2 is the speed of sound (squared) corrected for the static heat capacity and virial effects (but not relaxation effects), ε is the relaxation strength, τ_d is the relaxation time and $\omega = 2\pi \cdot f_{101}$. The vibrational temperature (in K) in column 37 allows the firmware to compute ε and the coefficients in columns 38-40 allow the computation of τ_d . Computationally, we cast Zuckerwar's correction as a frequency dependence of the heat capacity.

Vapor pressure (Columns 41-43)

The saturation vapor pressure of the gas is computed to warn that the system may be close to condensation (which can impact the speed of sound or fill the chamber with liquid) and to limit the range of composition computational results.

The Antoine equation, Eq. 9, is used to model the saturation vapor pressure over the operating temperature range.

$$P_{vp}^{sat}(\text{bar}) = 10 \left[A - \frac{B}{T+C} \right] \quad \text{Eq. 9}$$

Values for A , B and C are given in columns 41-43 of the Gas Table. These values were found by fitting the Antoine equation to vapor pressure data points over the operating temperature range. Vapor pressure data points were generated from multiple sources, including, NIST's REFPROP and correlation functions and parameters found in Perry, PG&L, PPL&G.

Vapor pressure data is omitted for fluids with critical temperatures below 265 K (as these fluids cannot be liquids at the BGA244 operating temperatures). In cases where no vapor pressure data is available, the BGA244 will not provide condensation warnings and will not limit composition results to below the saturation vapor pressure.

Speed of sound offsets (Columns 44-46)

The BGA244 computes the theoretical speed of sound using accurate molar masses, temperature corrected heat capacities, virial corrections, and translational-vibrational relaxation corrections. The speed of sound is measured using the thermo-viscous corrected resonance frequency and a cavity factor which has been corrected for perturbations and transducer resonances. Uncertainties in all of these corrections can lead to discrepancies between the measured and the computed speed of sound on order of +/- 100 ppm. An empirical speed of sound offset $W_{offset}(ppm)$, pressure correction $W_{P-slope}(ppm/psi)$ and temperature correction $W_{T-slope}(ppm/K)$ for each gas are used to repair these discrepancies. Values for those corrections are given in columns 44-46 of the Gas Table (and are initially set to zero).

References

1. REFPROP. Reference Fluid Thermodynamic and Transport Properties. *NIST Standard Reference Database 23, Version 9.1*. E.W.Lemmon, M.L. Huber, and M.O. McLinden. U.S. Dept of Commerce, (2013)
2. Second virial coefficients of normal alkanes, linear 1-alkanols and their binaries. C. Tsonopoulos, J.H. Dymond, and A.M Szafranski. *Pure and Appl. Chem.*, Vol 61, No.8, pp. 1387-1394, (1989)
3. *Handbook of the Speed of Sound in Real Gases*, Allan J. Zuckerwar, Academic Press, (2002)
4. *Tables of Physical and Chemical Constants, Section 3.5, Kaye & Laby, (Critical constants and second virial coefficients)*, http://www.kayelaby.npl.co.uk/chemistry/3_5/3_5.html
5. Correlation for the third virial coefficient using T_c , P_c and ω as parameters. H. Orbey and J. H. Vera, *AIChE J.* 29:107 (1983).
6. *Perry's Chemical Engineers Handbook*, 8th Ed., B.E. Poling et. al., Section 8. McGraw-Hill (2008)
7. *The Properties of Gases and Liquids*, 5th Ed., B.E. Poling, J.M. Prausnitz, J.P. O'Connell, McGraw-Hill (2001)
8. Chung, T.-H., M. Ajlan, L. L. Lee, and K. E. Starling: *Ind. Eng. Chem. Res.* **27**: 671 (1988).
9. Chung, T.-H., L. L. Lee, and K. E. Starling: *Ind. Eng. Chem. Fundam.*, **23**: 8 (1984).
10. *Physical Properties of Liquids and Gases*, Appendix C, http://booksite.elsevier.com/9780750683661/Appendix_C.pdf

Appendix B: BGA244E

The BGA244E Enclosure Version allows the BGA244 to be mounted in exposed locations. When properly installed it should meet the following NEMA Type specifications.

Type 6 (with cover latched)

This is intended for indoor or outdoor use primarily to provide a degree of protection against solid foreign objects (falling dirt), hose-directed water, the entry of water during occasional temporary submersion at a limited depth and damage from external ice formation.

Type 6P (with cover screws installed)

This is intended for indoor or outdoor use primarily to provide a degree of protection against solid foreign objects (falling dirt), hose-directed water, the entry of water during prolonged submersion at a limited depth and damage from external ice formation.

Operation

The BGA244E operates the same manner as the regular BGA244. Configuration and control can be performed over either the front panel or one of the computer interfaces. Results are reported over the front panel, the computer interfaces or the analog outputs.

Because the BGA244E is inside of an enclosure it is less convenient to access the front panel. This is especially true if the enclosure is configured to Type 6P, with the cover screws installed. If regular access to the front panel is necessary, be sure to allow adequate clearance to open the cover. Often measurements are monitored remotely over one of the computer interfaces or analog outputs after the unit is configured.

Environment

The enclosure permits a harsher ambient environment, protecting the unit from dirt, water and ice. The same temperature, pressure, flow and gas requirements apply for the BGA244E as for the standard BGA244. No extra ventilation is required. Refer to *Operating Environment* (page 13) for details.

If the unit will not be exposed to prolonged submersion, insert the plugs into the cover screw holes to keep debris out. In this case the latch alone can adequately seal the cover to the box. Be sure to align the gasket in the cover with the sealing bead of the box when closing the cover. Fasten the latch to seal the box.

However if the unit will be exposed to prolonged submersion, you must install the four 10-32 x 0.75" cover screws to prevent leakage. Be sure to align the gasket in the cover with the sealing bead of the box when closing and fastening the cover latch. Insert and loosely tighten the four screws. Tighten screws in opposite corners to 10 inch-lbs (1.13 N-m) to seal the box.

Do not use PVC cement to connect fittings to the enclosure. Keep solvents that may react with polycarbonate away from the enclosure. This includes the following: acetone, ammonium hydroxide, ammonium nitrate, phenol, carbon tetrachloride, hydraulic brake fluid, lacquer thinner, Loctite, methylene chloride, perchlorethylene, potassium hydroxide, sodium hydroxide, toluene and xylene. Use light soapy water to clean the enclosure if necessary.

Access

Leave adequate clearance around the door if you will need to access the unit after installation. See the mechanical drawings Figure 38 and Figure 39 in Appendix C for dimensions. The door requires a minimum of 1.625" (41mm) clearance past the side of the box to fully open (to 90°). Be sure to leave adequate clearance to release the cover latch.

The gas fittings require a wrench to tighten or loosen. Be sure there is sufficient clearance, both in length and rotation. See the *Installation* section below for details.

The conduit fittings are tightened or loosened by hand. Be sure there is sufficient clearance, both in length and rotation. See the following *Installation* section for details.

Leave adequate clearance for the mounting flanges if used. If the flanges are to be fastened with bolts be sure to have sufficient wrench clearance, both in length and rotation.

Unpacking

Before You Open the Box

Do not remove the gas port caps until just prior to connecting the unit to your gas manifold to reduce the chance of contamination of the cell.

Read the *Installation* section on the following page prior to installing the BGA244E into your system.

Read the Chapter 3: *Operations Guide* and Chapter 4: *Applications Guide* in the main manual prior to operating the BGA244E.

Inspect all components of the SRS BGA244E upon unpacking. Report any damage to Stanford Research Systems immediately. Compare the contents of the shipping container to the list below and report any discrepancies.

See www.thinksrs.com/downloads/PDFs/Manuals/BGA244m.pdf for the complete BGA244 User's Manual.

What is included with the BGA244E

1. One BGA244E
2. One 6' (1.8 m) USB cable
3. One Quick Start Guide

Accessories (if ordered)

- | | |
|-------------|------------------------------------|
| 1. BGA-24 | +24 V _{DC} Power Supply |
| 2. BGA-S | Acrylic shield for the LCD display |
| 3. BGA-M | Metal cover to replace display |
| 4. Adapters | Gas fitting adapters |

Quick Start

Use the *Quick Start Guide* (page 9) in Chapter 1 to get started making measurements with the BGA244E.

Installation

Installation of mounting the BGA244E consists of the following steps. First mount the units in its operating location. Next connect the gas lines, conduit and cables. Plan the mounting of the BGA244E, gas pipes and electrical conduit prior to beginning installation. Be sure to clean out gas lines prior to connecting them to the BGA244E to avoid contamination.

Mounting

Whenever possible mount the BGA244E in a location that is free of large vibrations or impacts. If there are strong mechanical vibrations it may be necessary to mount the BGA244E in a different location.

Never rely on the inlet and outlet tubing or the conduit to support the weight of the BGA244E as it weighs about 9 lbs (4 kg) and is likely to damage either the tubing or itself.

Plan the location of the BGA244E, gas pipes and electrical conduit prior to mounting the unit. See the mechanical drawings (Appendix C, Figure 38 and Figure 39) for the location of the gas fittings, electrical connectors and mounting points. There is no preferred gas flow direction; either gas port can be input or output.

The enclosure is a modified Integra Enclosure H8084HCF-6P. It includes hardware to mount it to a fixed location. The BGA244E can be installed in two different ways. The enclosure can be directly fastened to a mounting plate if the back of the mounting plate is accessible. Or mounting flanges can be installed on the BGA244E if the back of the mounting plate is not accessible.

Direct Mount

Mount the BGA244E from the rear of the mounting plate using four ¼"-20 UNC-B2 screws per the mechanical drawing (Figure 38 Appendix C). The maximum penetration depth of the screws is 0.25" (6.35mm). Tighten the screws to 20 in-lbs (2.26 N-m).

Flange Mounting

Place the BGA244E face down on a flat, soft surface to avoid marring the cover. Place the mounting flanges over the square bosses along the top and bottom edges. Fasten the two flanges using the ¼"-20 flat head Philips head screws. Tighten the screws to 20 in-lbs (2.26 N-m).

Mount the BGA244E from either the front or rear using four ¼" screws or bolts per the mechanical drawing (Figure 39, Appendix C). Mount to the ½" long slots only; the longer slots are not reinforced to support the weight of the BGA244E.

It may be necessary to remove the left side conduit nut to access screws in the lower flange. Alternatively bolts can be used to fasten the flanges. Do not over tighten screws or bolts to the flanges so as to not damage them.

Gas Fittings

The gas fittings are Swage 1/8"-27 NPT female-female couplings (SS-2-HGC). If damaged, the adapters can be replaced by the customer. See the Chapter 7: *Service* for details on the proper procedure to replace the gas fittings.

Use a 9/16th inch wrench and either a pipe wrench or pliers to tighten the tubing so as to not damage the unit. Use thread sealant when connecting tubing to the BGA244E. SRS recommends Loctite 565 thread sealant to seal leaks and prevent galling/seizing of the threads. A 0.2-oz tube of the Loctite is available from SRS (accessory 45855K12). Avoid getting thread sealant on the enclosure as it may react with polycarbonate. Be sure to clean out gas lines prior to connecting them to the BGA244E to avoid contamination.

Installation Guidelines

1. Apply a strip of thread sealant around the male pipe threads, leaving the first 2 turns uncovered.
2. Finger tighten the pipe into the fitting.
3. Hold the fitting with the 9/16th inch wrench. Wrench tighten the pipe 1.5 – 3.0 turns past finger tight using a pipe wrench or pliers.
4. Allow the Loctite 565 at least 24 hours to cure before applying pressure or vacuum to the system.

Gas Fitting Adapters

The BGA244E can accept the same gas fitting adapters as the standard BGA244. See *Gas Fitting Adapters* (page 17) for instructions on their installation.

Removal and installation of the standard gas fittings requires a 9/16th inch wrench and either a pipe wrench or pliers. Disconnect any gas lines and electrical connections prior to removing a fitting. Be sure to have sufficient working space before attempting to replace the fitting.

Conduit

The conduit connectors on the BGA244E are Arlington Snap²IT Non-metallic Push-On Connectors (NMLT7). They are designed to connect to 3/4" LFNC-B conduit (Liquid-tight, Flexible, Non-metallic Conduit, type B). Use the following procedures to install or remove the conduit.

- To install, simply push the liquid tight conduit over the ferrule end of the fitting. While pushing the connector, turn the conduit slightly clockwise to seat it.
- To remove, unscrew the nut and disassemble the connector. Re-assemble the connector for re-use.

Electrical



Make sure that the BGA244 is properly grounded. Depending on the installation, the power supply wiring and gas tubing may not be properly connected to earth ground. If necessary, connect the BGA244 chassis ground lug (C1) to a suitable earth ground.



Be sure that your installation confirms to all safety and electrical code requirements.



For CE compliance it is recommended that all power and I/O cables are shielded and grounded.



The BGA244 has no line voltages connections. Applying line voltage to any pin of any connector on the BGA244 will cause severe damage to the instrument and is a fire and smoke hazard.

The BGA244E has the same electrical connections that are in the standard BGA244. See the *Electrical Connections* (page 19) for more details on the electrical connections. Pay special attention to the following sections that may be more critical for BGA244E installations.

Electrical Noise Precautions

Installations may require long cable runs. Pay special attention to the grouping of wires in the conduits. Separate the power and computer interface wires from the analog I/O wires if possible.

Electrical Connections

It may be more difficult to identify wires that have been pulled through conduits. Be sure to properly identify each wire before connecting it to the terminal strips. Be especially careful to properly identify the +24 V_{DC} wire, since connecting it to an incorrect pin can damage the unit.



Be sure to use the correct pinout for each connector. Failure to do so can result in damage to the BGA244. Pay special attention to C4 and C6, as they are adjacent and have the same number of pins.

Tip: It's usually easier to remove the terminal strip from the BGA244 before connecting wires. Unplug the terminal strip by pulling straight back from the unit. Loosen the screws before inserting the wires. Make sure to observe the correct pinouts!

USB

The USB interface is not particularly well suited to power and control the BGA244E, although it can work in some circumstances. This would typically involve a USB adapter located near the BGA244E. A USB type B connector will fit through the 3/4" conduit. It is more common to power the BGA244E using a +24 V_{DC} power supply and control it using either RS-232 or RS-422.

The maximum cable length for USB is about 5m (16.4'). A 5m cable should perform adequately for data communications, but powering the BGA244E over long cables is challenging. Voltage drops in the power wires may cause the voltage to drop below the minimum required voltage (4.75 V). Always use USB cables with 20 AWG power wires to minimize voltage drops. The BGA244E has been verified to operate with 3m USB cables with the appropriate power wires.

RS-232

Since a standard 9 pin RS-232 connector will not fit through the 3/4" conduit, a cable must first be pulled through the conduit, before the RS-232 connector is fastened.

Connect the cable to a male DB9 connector following the pin out listed in following table. The minimum RS-232 implementation is 3 wires: Rx, Tx and Ground. The CTS and RTS wires can be added to provide handshaking. The CD and DSR wires are pulled high to the asserted state and can be connected if required.

TABLE 33: C9: RS-232 CONNECTOR FOR BGA244E

Pin	Signal
1	CD
2	Rx
3	Tx
4	DTR
5	GND
6	DSR
7	RTS
8	CTS
9	RI

It is recommended that the cable shield be terminated to the connector. Provide strain relief if the cable may be pulled on.

Follow the manufacturer's instructions for connecting the cable wires to the male DB9 connector. The following parts (available from Digikey) should be suitable.

TE Connectivity p/n	Digkey p/n
1571650-4	1571650-4-ND
749805-7	A105099-ND

Appendix C: Mechanical Drawings

Figure 37 is a mechanical drawing of the BGA244 and BGA244HP that shows mounting locations and gas fittings.

Figure 38 is a mechanical drawing for the BGA244E without the Mounting Flanges that shows mounting locations and gas fittings.

Figure 39 is a mechanical drawing for the BGA244E with the Mounting Flanges that shows mounting locations and gas fittings.




FIGURE 37: BGA244 MECHANICAL DRAWING




FIGURE 38: BGA244E MECHANICAL DRAWING (NO MOUNTING FLANGES)




FIGURE 39: BGA244E MECHANICAL DRAWING (WITH MOUNTING FLANGES)

Appendix D: Declaration of Contamination

The repair and/or service of the BGA244, BGA244HP or BGA244E can only be carried out if the declaration of contamination portion of the RMA process has been completed. SRS reserves the right to refuse acceptance of any equipment submitted for repair or service where the declaration has been omitted or has not been fully or correctly completed. SRS also reserves the right to refuse servicing any equipment which could be potentially harmful to the personnel carrying out the repair and service of the equipment. SRS will not accept any equipment which has been radioactively or explosively contaminated without written evidence that such equipment has been properly decontaminated.

Contact SRS if you have any questions regarding the contents of this declaration. The declaration of contamination portion of the RMA requires the following information.

- Description of the operating environment the instrument was exposed to.
- Whether the instrument was exposed to potentially harmful substances.
- If exposed to potentially harmful substances, list the following for each substance.
 - Chemical Name
 - Chemical Formula
 - Precautions
 - First Aid measures in the event of accident
- If exposed to potentially harmful substances, were any of the substances:
 - Radioactive
 - Toxic
 - Corrosive
 - Explosive
- If the unit will be decontaminated before shipping to SRS.
- Confirmation that the information supplied with this form is complete and accurate and that it is shipped in accordance with the appropriate regulations covering Package, Transportation and Labeling of Dangerous Substances.

Appendix E: Installing USB Drivers

In most cases, the USB drivers for the BGA244 should automatically install when the unit is first connected to a Windows computer. However, in some cases this may not work due to a lack of internet connection, Windows version or other incompatibility. In these cases the drivers can be manually installed using the following procedure.

The BGA244 uses a FTDI FT220X USB interface chip. FTDI provides USB drivers for several different operating systems. Drivers and their documentation are available on the FTDI website: www.ftdichip.com/FTDrivers.htm

Driver Types

There are two different drivers that can be used to communicate with the BGA244 over USB: VCP and D2XX. Depending on the application only one of the drivers may be required.

VCP or Virtual COM Port makes the BGA244 look like an additional COM port available to the computer. Application software can access the BGA244 in the same way as it accesses any standard COM port. This is commonly used for communications between a user created program and the BGA244.

The D2XX driver gives access the BGA244 through a DLL. Application software can access the BGA244 through a series of DLL function calls. BGAMon uses this driver.

Driver Version

Be sure to install the appropriate driver for your operating system. The FTDI website contains drivers for Windows versions from the present back to Windows 98. Drivers for Linux, MAC and Android are also available.

Driver Installation

Click on the D2XX Drivers tab. Click on the link to download both the D2XX & VCP drivers (preferred) or the VCP only link, then download the zip file for the drivers to the computer. If the computer is not connected to the internet, download the drivers on a different computer and copy them to a USB thumb drive. Move the thumb drive to the computer that needs the drivers installed.

Go to <http://www.ftdichip.com/Support/Documents/InstallGuides.htm> and download the appropriate installation guide for your computer. Follow the Install Guide instructions to install the drivers.

In some cases, Windows may complain that the driver is not certified or verified. If you are certain of the source of the driver, the warning can be ignored and you can continue the installation.

Connect the BGA244 to the computer. The “Found New Hardware” message may appear. Say “No” to Windows Update and instead select “Install Software Automatically”.

Contact FTDI if you are unable to successfully install the FTDI USB drivers. Contact information is available at www.ftdichip.com/FTContact.htm

Appendix H: Revisions

Version 0.61	Final Development Version
Version 1.00	Initial Release Version
Version 1.10	Updated functionality & GUI
Version 1.20	Added advanced REL feature, updated functionality & GUI; fixed errors. Removed schematics, parts list & circuit description.